



Calibration Certificate

Ref No Cal-00006

Device Re Dev-00007

Recepit R Rec-00003

Date 2025-11-27

Customer شركة الاصايل للمقاولات

Instrument معايرة جهاز GPS

Manufacturer Trimble

Model R2

Serial No 45454

High-precision		Real-Time Kinematic Surveying	
Horizontal	3 mm+1 ppm RMS	Horizontal	8 mm+1 ppm RMS
Vertical	5 mm+1 ppm RMS	Vertical	15 mm+1 ppm RMS
Satellites Tracking			
<input checked="" type="checkbox"/>	GPS	L1CA,L2P(Y),L2C,LS	
<input checked="" type="checkbox"/>	GLONASS	L1,L2	
<input checked="" type="checkbox"/>	Galileo	E1,E5a,E5b,E6	
<input checked="" type="checkbox"/>	BDS	B1I,B2I,B3I,B1C,B2a,B2b	
<input checked="" type="checkbox"/>	QZSS	L1,L2,L5,L6	
<input checked="" type="checkbox"/>	L-Band	B2b PPP	

(Only for the Asian pacific region)

We certify that the following survey instrument has been tested and calibrated in accordance to the instructions of the manufacturing company, and it complies with specification in accordance with user's manual supplied on delivery.

This certificate is valid for 6 months.

Date of calibration 2025-11-27

Expiry Date 2026-05-26

Certified By Rawad Almasaha Company.

Prepared by Mohammed al-obaid