BD Care Work Order Service Report (synthetic test document)

Institution: Harbor Clinical Lab
Service date: 2026-03-14
Installed Product 647794E6 - LSRFortessa SO
Serial Number: R647794E6092
Work Order Task Code T113 Repair / Troubleshooting Visit
Subject: No events on the sample channel
Description: Fluidics pressure low during the run