Service Report - Pick Fields Sample

Laser delay 29.40

Sheath pressure 4.9 psi

Technician: Dana Levi