

# Certificate of Analysis

Becton Dickinson and Company  
BD Diagnostic Systems  
PO Box 999  
Sparks MD 21152-0999 US

Page: 1 of 2

**Product Name** : Bottle Tryptose Agar 500G  
**Catalog Number** : 264300 **Manufacture Date**: 2019/03/26  
**Batch Number** : 9142595  
**Expiration Date** : 2024/02/29

01. Dehydrated Medium Appearance: Light beige, free-flowing, homogeneous
02. Solubility: 4.1% solution, soluble in distilled or deionized water on boiling
03. Solution Appearance: Light amber, slightly opalescent
04. Cultural Response: Medium was prepared per label instructions. Plates were inoculated with the test organisms and incubated at 35 ± 2°C for 40-48 hours under CO<sub>2</sub>.

TEST ORGANISMS		RECOVERY
Brucella canis	ATCC® 23365	good
Brucella abortus RB51	BEI NR-2552	good

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C :		7.4	7.0	7.4
Bulk Lot Number:	-	9084871		

Animal source	Country of Origin	Tissue Category		
		BIC	SIC	ABC
Bovine	New Zealand	IV	IV	MLK
Porcine	USA	II	II	IB
Porcine	USA	III	III	IB

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 13485:2016 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at [www.bd.com/regdocs](http://www.bd.com/regdocs). For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual

Creation Date: 2019/06/10 16:12:27

# Certificate of Analysis

Becton Dickinson and Company  
BD Diagnostic Systems  
PO Box 999  
Sparks MD 21152-0999 US

Page: 2 of 2

**Product Name** : Bottle Tryptose Agar 500G  
**Catalog Number** : 264300 **Manufacture Date**: 2019/03/26  
**Batch Number** : 9142595  
**Expiration Date** : 2024/02/29

lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at [www.bd.com/regdocs](http://www.bd.com/regdocs) under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Diagnostics - Diagnostic Systems, Animal Origin Position Statement", at [www.bd.com/regdocs](http://www.bd.com/regdocs) under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see

[www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information](http://www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information).



Megan Reeher  
Associate Director, Quality Management