

TRIUMF nEXO SiPM Studies



Fabrice Retière
Carl Rethmeier
Lloyd James
Paul Lopes Gomes



August 5, 2015




Figure: Crosstalk rate against overvoltage for different devices.

Remote control of the power supply / picoammeter was implemented and interfaced with the scope DAQ software so that voltage scans can now be done automatically.




Figure: Delta-time distribution for VUV3 device.

To determine the source of the inconsistency in PE measurements, the following steps were taken:

- Installing temperature sensors in the setup to keep track of temperature at different places in the box.
- Other stuff?




Figure: 10 thermocouples were installed in various locations around the box so that temperature could be recorded in each location over time.




Figure: PE variation over 200 minutes in different cooling conditions.




Figure: Variation of ratio of PE to initial PE over 200 minutes in different cooling conditions.