



OPEN

Calibration of spectra in presence of non-stationary background using unsupervised physics-informed deep learning

Alessandro Puleio, Riccardo Rossi[✉] & Pasqualino Gaudio

Calibration is a key part of the development of a diagnostic. Standard approaches require the setting up of dedicated experiments under controlled conditions in order to find the calibration function that allows one to evaluate the desired information from the raw measurements. Sometimes, such controlled experiments are not possible to perform, and alternative approaches are required. Most of them aim at extracting information by looking at the theoretical expectations, requiring a lot of dedicated work and usually involving that the outputs are extremely dependent on some external factors, such as the scientist experience. This work presents a possible methodology to calibrate data or, more generally, to extract the information from the raw measurements by using a new unsupervised physics-informed deep learning methodology. The algorithm allows to automatically process the data and evaluate the searched information without the need for a supervised training by looking at the theoretical expectations. The method is examined in synthetic cases with increasing difficulties to test its potentialities, and it has been found that such an approach can also be used in very complex behaviours, where human-drive results may have huge uncertainties. Moreover, also an experimental test has been performed to validate its capabilities, but also highlight the limits of this method, which, of course, requires particular attention and a good knowledge of the analysed phenomena. The results are extremely interesting, and this methodology is believed to be applied to several cases where classic calibration and supervised approaches are not accessible.

Spectroscopy is among the most relevant techniques to extract information from the interaction of electromagnetic waves with matter^{1–3}. Absorption spectroscopy, laser-induced fluorescence, Raman spectroscopy, and laser-induced breakdown spectroscopy are only a few of the vast number of techniques that, by analysing a spectrum, allow one to probe the medium, providing information about its atoms, molecules, crystals, etc.^{4–7}. Their properties have made spectroscopy a leading technology not only in physical science but also in industries, medicine, art, etc. For example, Raman spectroscopy finds applications in everything: from medicine^{8,9}, by allowing even cancer diagnosis¹⁰, to art and cultural heritage¹¹, from food science to environmental and explosive monitoring^{12–15}. Laser-Induced Breakdown Spectroscopy (LIBS) is also widely used to analyse the atomic composition of matter and recently, also due to the development of machine learning algorithms, it has been applied in several fields of science and industries^{16–18}, such as food¹⁹, chemical²⁰, pharmaceutical²⁰, and medicine²⁰. The same applies for absorption spectroscopy, that changing the electromagnetic radiation wavelength goes from radio wave to X-ray, leading to a huge number of techniques (nuclear magnetic resonance spectroscopy, IR absorption spectroscopy, UV–vis absorption spectroscopy, etc.) and applications^{21–23}. Being most of them non-contact measurements, in extreme environments, the use of light and its interaction with matter is the only way to make diagnostics. An example is the plasma in nuclear fusion reactors that reaching temperature of the order of 150 million degrees it can be analysed only with external measurements (for example, magnetic coils), while the core of the plasma is analysed by studying the interaction of electromagnetic waves with the plasma^{24,25}. It involves that in a nuclear fusion reactor there is a huge number of spectroscopic diagnostics²⁶, used, for example, to monitor plasma ions concentration, the presence of impurities, observe some edge instabilities, etc.^{25,27,28}. A similar argument applies, of course, to space and atmosphere science^{29–31}.

Optical spectroscopy techniques rely on studies of the resulting spectra from light-matter interactions. In addition to the accurate work that scientists need to do for hardware selection, alignment, control of the

Department of Industrial Engineering, University of Rome “Tor Vergata”, Via del Politecnico 1, 00133 Rome, Italy.
✉email: r.rossi@ing.uniroma2.it

environmental conditions to ensure specific properties of the matter, etc. There is also a huge work required to develop efficient algorithms able to extract the desired pieces of information from a measured spectrum^{32,33}. Some spectroscopy problems can be conceptualised as follows. Suppose to have a spectroscopic technique that allows one to measure the spectrum I_d and we want to evaluate a quantity y that is linked with the spectrum property such that $y = f(I_d)$. In ideal cases, $f(I_d)$ is usually known (imagine, for example, the concentration measurement of a gas by using absorption spectroscopy, where you know the absorption cross-section spectrum of chemical³⁴). Unfortunately, in real cases $f(I_d)$ is not known a priori and is difficult to estimate, even with numerical simulations, for several reasons. At first, components are not ideal (for example, the optics have properties that must be calibrated). Second, even small misalignment can change the optics response. Moreover, spectra are usually the results of not only the spectrum I_d relative to the physical effect that we are considering, but it contains spurious effects due to other minor (hopefully) phenomena, such as background light, secondary emission or absorption effects (for example, Raman spectroscopy spectra are usually accompanied by a strong background due to fluorescence and elastic scattering^{6,35}), and much more. Last, measurements are also affected by noise that sometimes play a relevant role in the quality of the measurement (typical examples are some Raman spectroscopy applications where the Raman light is of the order of the noise and differential absorption lidar where the noise defines the maximum range and the sensitivity of the technique³⁶). Thus, the previous equation may be written as $y = g(I_d, I_b, I_n)$, where I_b is what will be named in the article as “background noise”, which represents all the effects due to other phenomena, I_n is the noise and the function g is of course different from the physical known function f .

In general, some countermeasures can be taken to address such problems. For example, noise is usually addressed by reducing the noise due to the instruments and by performing statistics, when possible, by averaging several spectrum measurements. The background is also usually faced instrumentally, for example by filtering out the wavelengths that contain only the background radiation.

However, even reducing such effect, it is usually required to find the function g that allows to evaluate our desired information y and the only way to find this function is very often based on data-driven approaches that are typically known as calibration^{37,38}. A calibration can be done in different ways. The most simple and widespread method is based on performing controlled measurements of I (where I is the measured spectrum due to I_d , I_b , and I_n) where y is known and by a fitting approach one can find the function that allows to evaluate y from I ³⁹. When the problem is quite complex, it may be hard to find a simple equation (polynomial, power law, etc.) that fits the data, and therefore more complex approaches are used. Today, one of the most utilised methodologies is supervised machine and/or deep learning. The rationale is very similar to function fitting, since the machine learning algorithm is a very general function (considering as machine learning tool a sufficiently deep method, such as neural networks or decision trees), that it is “trained” with labelled data, i.e. measurements where both the input (in our case I) and the output (y) are known. These artificial intelligence methods are very powerful and usually they far exceed by far the classical calibration method^{40–45}.

While a ‘supervised’ method is the best approach to perform a calibration^{46,47}, it may sometimes be impossible to be applied. For example, sometimes it is not viable to reproduce controlled experiments where y is a priori known and, therefore, the training phase (or the fitting) is impractical to perform. In such cases, only theoretical assumptions based on the physics of the phenomenon can help to find an acceptable calibration function.

In this paper, a new unsupervised deep learning calibration methodology based on Physics-Informed Neural Network (PINN) is introduced. Physics-Informed Neural Network^{48,49}, or PINN, is an emerging machine learning technology that is gaining attention in all fields of science (and recently also industries). The main difference from classic supervised machine learning is the way they are trained, since they do not require the knowledge of the output to be trained but use physical expectation. Contrary to standard machine learning, that uses expected vs actual output based loss functions, the physics-informed machine learning aims at minimising a loss function that looks at the physics. Typically, PINNs have been used to simulate flows (Navier–Stokes equations)^{50,51}, and solve specific PDEs⁴⁸, but are now finding applications also in various other sectors such as medicine⁵² and energy⁵³.

In this work, we present a new way to use PINN to develop a methodology that can calibrate or evaluate specific information from spectra without the need to use a supervised method. This new approach is of course interesting for all those cases where it is impossible to perform a typical calibration or train a supervised machine learning algorithm, since a “training set” cannot be produced.

The paper is divided into sections as follows. Section “[Unsupervised physics-informed calibration](#)” aims at showing the methodology used, by introducing the general calibration problem and proposing a new methodology to evaluate the information that can be extracted from the spectra. Section “[Results](#)” shows the results of the new method by analysing the performances in numerical (or synthetic) cases and one experimental case reproduced in our laboratory. Section “[Discussion and conclusions](#)” is dedicated to discussion, conclusion, and future developments, while in Section “[Materials and methods](#)” it is possible to find all the details regarding the materials and methods to replicate this work.

Unsupervised physics-informed calibration

Consider a light-matter interaction that produces a spectrum $I_d(\lambda)$ as a function of the concentration c of a specific element such that:

$$I_d(\lambda) = g(c, \lambda), \quad (1)$$

where $g(c, \lambda)$ is a general function that describes the intensity of the spectrum generated at each wavelength λ once the concentration c is defined. For the sake of simplicity, suppose that there is linear functionality between $I_d(\lambda)$ and c , so the previous function can be written as:

$$I_d(\lambda) = I_{d,0}(\lambda)c, \quad (2)$$

where $I_{d,0}(\lambda)$ is the spectrum when c is equal to one, which will be named “reference spectrum” from now on. Suppose also that during the measurement other effects occur such that the recorded spectrum $I(\lambda)$ is the sum of $I_d(\lambda)$, $I_b(\lambda)$, and $I_n(\lambda)$, where $I_b(\lambda)$ is a variable spectrum due to other effects (the “background spectrum”) and $I_n(\lambda)$ is a random spectrum due to the noise (electrical, optical, etc.). The only way to measure c (without using supervised learning) is to estimate $I_b(\lambda)$, so that it can be removed from $I(\lambda)$ and Eq. (1) can be applied ($I_{d,0}(\lambda)$ is considered known). In this work, we present an innovative method to estimate $I_b(\lambda)$, and thus calculate $I_d(\lambda)$ and c , by using a physics-informed neural network.

The PINN architecture, described in detail in Section “Materials and methods”, is shown in Fig. 1. A typical autoencoder structure⁵⁴ is applied to the input, the measured spectrum $I(\lambda)$, with the aim of evaluating the background noise $I_{p,b}(\lambda)$. Thus, the predicted background is removed from the measurement and a new convolutional layer (with the associated MaxPooling and ReLU layers) is applied. Then, a fully connected layer is used to predict the concentration c_p .

The neural network is based on minimising a loss function composed of two terms. The first is the reconstruction error. In fact, we expect that the sum of the predicted background spectrum and the predicted $I_{p,d}$, calculated as the reference spectrum ($I_{d,0}(\lambda)$) multiplied by the predicted concentration (c_p) is equal to the measured one $I(\lambda)$ (excluding the error associated with the noise, assumed to be averagely zero. Please note that all the systematic errors will be automatically computed in the background noise):

$$L_{rec} = \sum (I(\lambda) - I_{p,d}(\lambda) - I_{p,b}(\lambda))^2 = \sum (I(\lambda) - c_p I_{d,0}(\lambda) - I_{p,b}(\lambda))^2. \quad (3)$$

However, this solution is not sufficient to find a unique solution (the PINN may always choose to predict $I_{p,b}(\lambda) = I(\lambda)$ and $c_p = 0$). Therefore, a regularisation term has been added:

$$L_{reg} = \sum \left(\frac{dI_{p,b}}{d\lambda} \right)^2. \quad (4)$$

This term asks the PINN to avoid background spectra when possible, *i.e.* when the shape of the measured spectrum fits $c_p I_{d,0}(\lambda)$. Thus, the loss term is written as follows:




Figure 1. Physic-Informed Neural Network (PINN) architecture used in this work. The top shows the autoencoder that aims to predict the background spectrum, while the bottom shows the net that aims to predict the concentration.

$$L_{tot} = L_{rec} + \alpha L_{reg} = \sum (I(\lambda) - c_p I_{d,0}(\lambda) - I_{p,b}(\lambda))^2 + \alpha \sum \left(\frac{dI_{p,b}}{d\lambda} \right)^2, \quad (5)$$

where α weights the importance of the regularisation term. As usual, the regularisation term plays a relevant role in the quality of the algorithm. In Section “[Deep learning architecture](#)”, we provide a methodology to estimate a good value of α for this neural network. However, it is worth to highlight that is a good practice to train the algorithm with different regularisation values and by inspecting the results critically.

The previous methodology, which has been shown for one interesting emitting agent (or class), can be easily extended to a multi-agent (multiclass) reconstruction. In that case, the architecture of the net is the same, except the output layer that predicts two or more concentrations. The loss function is modified as follows:

$$L_{tot, multi\ class} = L_{rec} + \alpha L_{reg} = \sum \left(I(\lambda) - \sum_{j=1}^N c_{pj} I_{0,j}(\lambda) - I_{p,b}(\lambda) \right)^2 + \alpha \sum \left(\frac{dI_{p,b}}{d\lambda} \right)^2, \quad (6)$$

where c_{pj} is the concentration and $I_{0,j}$ is the reference spectrum of the j -th agent respectively and N is the number of agents.

Another consideration regards the linearity assumption. In fact, at the beginning of this section, we assumed that the process is working in a linear regime, a hypothesis that is not always satisfied. In such a case, the linear function should be substituted with the non-linear one and a more general loss function becomes:

$$L_{tot, multi\ class} = L_{rec} + \alpha L_{reg} = \sum \left(I(\lambda) - \sum_{j=1}^N g_j(c_{pj}, \lambda) - I_{p,b}(\lambda) \right)^2 + \alpha \sum \left(\frac{dI_{p,b}}{d\lambda} \right)^2, \quad (7)$$

where $g_j(c_{pj}, \lambda)$ is the function to estimate the expected spectrum $I_{d,j}$ from the concentration c_{pj} see Eq. (1).

Results

The proposed methodology has been tested with both synthetic and experimental datasets. At first, we show the synthetic cases that allowed us to better investigate the advantages and the limits of the new methodology. Then, an experimental case with measurements performed in our laboratory is analysed.

Synthetic datasets. Synthetic tests have the advantage of controlling all the variables and ensure that the hypotheses at the basis of the algorithm are fulfilled. Six datasets have been generated to test the algorithm at different difficulties:

1. N01: Peaked reference spectrum with a steady constant background;
2. N02: Peaked reference spectrum with a unsteady constant background;
3. N03: Peaked reference spectrum with a steady shaped background;
4. N04: Peaked reference spectrum with a unsteady shaped background;
5. N05: Not peaked reference spectrum with a unsteady shaped background;
6. N06: Two classes peaked reference spectra with an unsteady shaped background.

In the case of N01 and N02, the use of a deep learning is not justifiable since a simple offset allows to ensure the highest performances. However, for the sake of completeness and to show that the method works also for these simple cases, these datasets have been analysed and the performances have been compared with a simple offset removal approach. Also, in the case of N03 the deep learning method proposed may be not needed since after a statistical analysis one may be able to automatically remove the background and analyse the polished spectrum. Things worsen in the case of N04, where the background strongly varies from measurement to measurement, involving that the statistics becomes complex, and it is not easy to maximise its efficiency. N05 analyses the case of a reference spectrum that does not reach a peak, making the analysis more challenging since the shapes of the background and the reference are quite similar. N06 validates the multiclass approach under the same conditions as N04. A detailed description of the functions used to generate the synthetic spectra is provided in Section “[Materials and methods](#)”. Just to give an idea of the complexity of the task, the “measured spectra” with concentrations ranging between 0.1 and 0.15 are provided in Fig. 2. In this figure, the part of the spectrum related to our agent is represented by the peaks at 656 nm and 486 nm. The graph clearly shows that the variability of the background is very large compared to the intensity of the two peaks, involving the automatic separation of the spectrum due to the agent from the background spectrum being not an easy task.

For each case, the PINN has been trained on a set of data (“training set”) and has been evaluated on new data (“test set”). Despite the fact that the PINN cannot overfit the data (it does not use supervised training), we decided to test them also with new data (test set) to demonstrate that, once the net is trained, it can also be easily used for real-time applications. The results of these six cases (from N01 to N06) are shown in Fig. 3. Both the training and test target vs predicted data are shown and the R-squared is displayed. Each graph shows the target concentration (i.e. the one used to generate the data) vs the predicted concentration by the PINN. For very simple situations (N01, N02), the prediction is practically perfect. This good result is also observed for more challenging cases (N03, N04), while a slight decrease in performances is observed in the most challenging cases (N05, N06). However, it has to be highlighted that also in these difficult cases, the relative prediction errors were below 1%.




Figure 2. Example of spectra in the N04 dataset, showing spectra calculated with agent concentrations ranging from 0.1 and 0.15. The image aims to show in what extreme conditions the algorithm has been tested.




Figure 3. Results obtained in the six synthetic cases. Each figure shows the target concentration vs the predicted concentration for the six synthetic cases for both training and test datasets.

These results clearly demonstrate that the application of PINN allows us to obtain very accurate results when the physical phenomena is known.

Experimental datasets. Experimental tests have been performed on Laser-Induced Fluorescence spectra. The authors set up an experiment in which samples of riboflavin (B2 vitamin) were irradiated with a 280 nm power LED. The emitted fluorescence at larger wavelengths (typically from 450 to 700 nm) is recorded at 90° through a spectrometer. To simulate variable background radiation, an intensity-variable tungsten lamp has been used. By varying the voltage of the lamp, the intensity (and slightly the shape) of the background radiation has been varied from measurement to measurement. A detailed description of the experimental apparatus, procedure, and typical measured spectra is reported in the “Materials and methods” section.

The results are shown in Fig. 4. Also, in this case, the target concentration (estimated by measuring the mass of riboflavin used to make the sample) is compared with the predicted one. It can be seen that, in this case, the reconstruction quality is worse compared to the synthetic cases. In this case, the concentration prediction error is on average about 10%. This larger error is mainly due to three reasons:

1. The fluorescence and background spectra are continuous and similar. This involves a larger confusion. It is clear that in the case where the background and reference spectra are equal, inversion is not possible (this is also true for the supervised calibration).
2. The target concentrations are affected by uncertainties (even if they should be smaller than 5%).
3. For this experiment, the linear reconstruction loss function was used Eq. (3). However, our spectrometer shows a clear non-linear behaviour (the counts do not increase linearly with the input intensity). This involves that the measured spectrum is distorted with respect to the reference one, and concentrations are systematically overestimated or underestimated by a value proportional to the non-linearity.

Discussion and conclusions

A new physics-informed deep learning method has been developed to calibrate spectra without the need for dedicated calibration experiments. The method consists of finding the concentrations of the desired species by reconstructing the measured spectra. Contrary to classic calibrations, that require the set of measurements where the output is known, this approach does not require a dedicated set but the knowledge of the physics of the phenomenon. This unsupervised method can play a relevant role in all situations where standard calibration methodologies and supervised machine learning algorithms are not deployable (while in these other cases, it is better to use the standard methodologies, as supervised methods are more accurate than unsupervised).

A second relevant feature of this method is that it allows the decoupling of the effects composing the spectra. In fact, the first part of the neural network aims at finding the background, which in some cases can contain useful information about the physical phenomenon. For example, in laser-induced fluorescence when a fourth harmonic Nd:YAG wavelength is used to induce the fluorescence, the second harmonic scattering is usually present in a spectral region where also fluorescence is present (532 nm). By using our method, we may clean the fluorescence spectrum from the scattering one, without the necessity to filter the signal through optical filters.

Generally speaking, the methodology that we presented in this paper can have several applications in optics, from calibration of diagnostics (that is the case study used to introduce the methodology) to cleaning of spectra and the solution of inverse problems.

As with all the supervised machine learning techniques, once the net is trained, it can be easily run in real-time. By analysing the average time required by our neural network on new data (statistics performed on 6000 samples), it has been found that the average prediction time is 10 ms, with a minimum value of 7 ms and a




Figure 4. Results obtained on experimental Laser-Induced Fluorescence spectra. The figure shows the target vs predicted concentrations for both training and test datasets.

maximum one of 15 ms. Even if this prediction time is compatible with most real-time applications, it is worth mentioning that the tests have been performed on a simple laptop with a code that it has not been optimised for real-time prediction. It is expected that the prediction speed can be easily reduced by, at least, an order of magnitude.

It is also worth mentioning that this methodology can be integrated with a sort of adaptive training. In fact, while standard calibration requires the knowledge of the output, and therefore it is trained on ad hoc experiments, the physics informed can be trained on all experiments. Therefore, new experiments mean new data for PINN training, allowing the extension of the ‘training set’, the “evolution” of the neural network, and therefore a network with a wide range of applicability.

A fairly general regularisation loss function for similar work has been developed. However, readers are strongly encouraged to perform parametric analysis to evaluate the best regularisation term. It is also worth mentioning that this methodology may be implemented with other physics-guided ad hoc loss functions, developed for specific cases to ensure higher performance. In fact, as this methodology based on the knowledge of the physics of the phenomenon, it is clear that it is impossible for the authors to develop loss functions that work in all areas of optics. Thus, the methodology shown in this paper should be considered as a recipe to develop ad hoc neural network architectures and loss functions to optimise specific problems.

Of course, such a powerful methodology has its limits. First, it is crucial to know the physics of the phenomenon (the ‘reference spectra’ or, in general, the $g(c, \lambda)$). Moreover, the responses of the instrumentation must be known. Non-linearities, cut-offs, etc. must be taken into account and written in the loss functions; otherwise, systematic errors would arise. In our experimental case, for example, the non-linearity of our spectrometer leads to large errors (see “Experimental results” section).

It is worth highlighting that the methodology shown in this work has been used without any type of preprocessing tool, method, or analysis. However, it is clear that the use of some preprocessing algorithms to better select the input data before the training may help in ensuring a better training set, resulting in more performant algorithms. Typical procedures may be smoothing techniques or the averaging of more measurements, but also the use of some advanced processing tools, such as those based on fuzzy logic, that allow data association without supervision^{55–57}.

In conclusion, a new methodology combining physics knowledge and artificial intelligence has been developed to automatically evaluate some quantities from spectra without the need to supervise the machine learning tool. It is obvious that this approach may lead to huge improvements in all the scientific and industrial cases where ad hoc calibration measurements cannot be performed or the actual values are not accurately defined.

Moreover, even if this methodology has been shown for optical spectroscopic measurements, it has to be highlighted that the approach may be transferred not only other spectroscopic measurements, but it may be deeply generalised to be potentially applied, by changing the neural network architecture and the loss function, to any type of measurements that require a calibration procedure. The future developments of this work go exactly in this direction, since we will try to generalise the methodology for other physics and engineering cases.

Materials and methods

In this section, the details concerning the neural network, the synthetic cases generated for the test, and the experimental setup to collect the LIF spectra are described.

Deep learning architecture. The architecture of our physics-informed deep learning algorithm is shown in Fig. 1. The input is a spectrum, and, by an autoencoder-like structure, the background is predicted. Autoencoders are usually divided into Encoder and Decoder. The encoder is composed of four layers: two convolutional layers followed by two MaxPool layers⁵⁸. After each convolutional layer, a ReLU (Rectified Linear Unit) layer is used. The output of the code is then deconvolved through the decoder, which has two deconvolutional and MaxUnPool layers. Deconvolutional layers are followed by a ReLU layer each. The output of the decoder is the predicted background spectrum.

Then, the predicted background spectrum is removed from the measured spectrum, and a convolutional layer, followed by a ReLU and MaxPool are applied. Therefore, a fully connected layer is applied and the output layer returns the net output. The transfer function of the fully connected layer is a ReLU, while the output layer transfer function is a linear one. Several attempts have been made to “tune” the hyperparameters of the neural network architecture (convolutional size, fully connect size, etc.) and the chosen hyperparameter is resumed showed in Table 1.

The loss function is described in the main text, and no additional information must be provided.

For what concerns the test performed on the datasets from N01 to N05, the number of batches is 100 while the number of epochs is $5 \cdot 10^3$. For the N06 dataset, the number of epochs equal is $5 \cdot 10^4$. Instead, to test the algorithm on the real case, the batch number has been set equal to 60 with a number of epochs equal to $25 \cdot 10^4$. The training algorithm is ADAM⁵⁹.

For our cases, no significant differences have been observed for most hyperparameters. Of course, an important role is played by the code size (maximum compression of the data). It has been found that, for our synthetic cases, a 5×7 size is usually enough (increasing the size no improvements have been observed), for both one-class and double-class tests, while the experimental cases we used a 2×20 code.

Another hyperparameter is the regularisation term (α in Eq. (5)). As usual, there is no magic number, and it is strongly suggested to try several tries before accepting a result. In our case, we found that a good solution is:

	Neural network architecture tests on datasets from N01 to N06				
	FC neurons	Kernel	Number of filters	Pool size	Stride
C-R-1	–	[20 1]	20	–	–
M1	–	–	–	10	10
C-R-2	–	[20 1]	7	–	–
M2	–	–	–	14	14
T-R-1	–	[20 1]	7	–	–
T-R-2	–	[20 1]	20	–	–
C-R	–	[20 1]	5	–	–
M	–	–	–	50	100
FC-R	10	–	–	–	–
FC-output	1, 2 (As a function of output concentrations)	–	–	–	–

Table 1. Neural Network Hyperparameters. Please refer to Fig. 1 for layer labels.

$$\alpha = \frac{1}{\Delta\lambda} \frac{\sum I_0}{\sum \left| \frac{\partial I_0}{\partial \lambda} \right|} \quad (8)$$

This number has been derived just by imposing two conditions, the first is that the relative reconstruction error should be small, and the second is that the background spectrum should contain smoother features in respect to reference spectra (these conditions are satisfied in several conditions, since the background is the sum of several effects that tend to smooth the resulting total spectrum). However, we would like to highlight again that even if this number provided to be a very good choice for our cases, a parametric analysis should always be performed to select the best regularisation term.

Synthetic approach: datasets and tests. Six synthetic tests have been performed:

1. N01: Peaked reference spectrum with a steady constant background;
2. N02: Peaked reference spectrum with a unsteady constant background;
3. N03: Peaked reference spectrum with a steady shaped background;
4. N04: Peaked reference spectrum with a unsteady shaped background;
5. N05: Not peaked reference spectrum with a unsteady shaped background;
6. N06: Two classes peaked reference spectra with an unsteady shaped background.

The “measured” spectra have been calculated by generating a reference spectrum, random concentrations ranging from 0 to 2, background and noise spectra, and summing up all of them.

In the case of the reference spectrum for one species (N01, N02, N03, N04, N05), I_0 has been calculated as follows (parameter values are shown in Table 2):

$$I_0 = I_1 e^{-\left(\frac{(\lambda-\lambda_1)^2}{(2\sigma_1)^2}\right)} + I_2 e^{-\left(\frac{(\lambda-\lambda_2)^2}{(2\sigma_2)^2}\right)} \quad (9)$$

While in the case of two species:

		N01		N02		N03		N04		N05		
I ₁		180		180		180		180		180		
I ₂		60		60		60		60		60		
λ ₁		656.3		656.3		656.3		656.3		656.3		
λ ₂		486.1		486.1		486.1		486.1		486.1		
σ ₁		2		2		2		2		45		
σ ₂		2		2		2		2		40		
N06	I _{1,1}	I _{1,2}	I _{2,1}	I _{2,2}	λ _{1,1}	λ _{1,2}	λ _{2,1}	λ _{2,2}	σ _{1,1}	σ _{1,2}	σ _{2,1}	σ _{2,2}
	180	60	240	100	656.3	486.1	587.5	480.0	2	2	2	2

Table 2. Values used to generate the reference spectra.

$$\begin{aligned}
 I_{0,species\ 1} &= I_{1,1}e^{-\left(\frac{(\lambda-\lambda_{1,1})^2}{(2\sigma_{1,1})^2}\right)} + I_{1,2}e^{-\left(\frac{(\lambda-\lambda_{1,2})^2}{(2\sigma_{1,2})^2}\right)} \\
 I_{0,species\ 2} &= I_{2,1}e^{-\left(\frac{(\lambda-\lambda_{2,1})^2}{(2\sigma_{2,1})^2}\right)} + I_{2,2}e^{-\left(\frac{(\lambda-\lambda_{2,2})^2}{(2\sigma_{2,2})^2}\right)}
 \end{aligned}
 \tag{10}$$

The background spectra are generated as follows:

$$\begin{aligned}
 \text{N01 : } I_{back}(\lambda) &= 10 \\
 \text{N02 : } I_{back}(\lambda) &= 10\varepsilon \text{ where } \varepsilon \text{ is a random number from 1 to 6} \\
 \text{N03 : } I_{back}(\lambda) &= 50e^{-(\lambda-500)^2/(2*75^2)} \\
 \text{N04, N05, N06 : } I_{back}(\lambda) &= 50\varepsilon_1 e^{-(\lambda-500-200\varepsilon_2)^2/(2*75^2)} \\
 &\text{where } \varepsilon_1 \text{ is a random number from 1 to 6 and } \varepsilon_2 \text{ is a random number from 0 to 1}
 \end{aligned}
 \tag{11}$$

Random noise has been added by generating a random number normally distributed with zero mean and standard deviation equal to five.

Real-case potential used tests: datasets and tests. Experimental tests have been performed by irradiating samples of riboflavin with ultraviolet (UV) light. The fluorescence spectra have been recorded by a spectrometer. The light source is a 280 nm centred wavelength diode. A laser line filter with a central wavelength equal to 280 nm and an aperture of 20 nm (± 10 nm) has been used to filter out the long radiation tail of the diode. The spectrometer is a Flame from the Ocean Insight company. It has 2048 pixels, with a working wavelength range from 180 to 874 nm. The average resolution is 0.33 nm. The exposure time for all measurements was 2 s. In order to simulate undesired background radiation, an intensity-variable tungsten lamp has been used. By varying the voltage, the intensity and the shape of the spectrum of the background vary. During the various measurements, the voltage (and therefore the intensity) has been varied to ensure a high variability of the background. However, since the fluorescence is observed from about 450 nm, the spectra have been cut-off for $\lambda < 400$ nm.

Several riboflavin samples have been prepared at different concentrations, by diluting riboflavin in distilled water:

- Sample 1 = $5 \cdot 10^{-4}$ g/l
- Sample 2 = $2 \cdot 10^{-4}$ g/l
- Sample 3 = $1 \cdot 10^{-4}$ g/l
- Sample 4 = $5 \cdot 10^{-5}$ g/l
- Sample 5 = $2 \cdot 10^{-5}$ g/l
- Sample 6 = 0 g/l

In total, 240 spectra have been recorded with variable background. Half of them have been used for the training, while the other 120 have been tested (see Fig. 4). The reference spectrum has been evaluated by measuring the LIF spectra of riboflavin in the absence of background radiation (calculate as the average of 10 spectra to reduce the presence of the noise in the reference one). The riboflavin concentration for the reference spectrum was 10^{-4} g/l.

Data availability

The experimental measurements and the codes to generate the synthetic spectra and train the PINN will be provided by the author upon request (contact: r.rossi@ing.uniroma2.it).

Received: 8 November 2022; Accepted: 3 February 2023

Published online: 07 February 2023

References

1. Broadbent, A. D. Colorimetry, Methods. In *Encyclopedia of Spectroscopy and Spectrometry* 3rd edn (ed. John, C.) 321–327 (Elsevier, 2017).
2. Hollas, J. M. (John M. Modern spectroscopy. 452.
3. Gauglitz, G. & Moore, D. S. *Handbook of Spectroscopy* 2nd edn, Vol. 1–4, 1–1878 (Wiley, 2014).
4. Parson, W. W. Modern optical spectroscopy: With exercises and examples from biophysics and biochemistry, second edition. *Modern Optical Spectroscopy: With Exercises and Examples from Biophysics and Biochemistry, Second Edition* 1–572 (2015) doi:<https://doi.org/10.1007/978-3-662-46777-0/COVER>.
5. Lakowicz, J. R. *Principles of Fluorescence Spectroscopy* 1–954 (Springer, 2006). <https://doi.org/10.1007/978-0-387-46312-4>.
6. Smith, E. & Dent, G. *Modern Raman Spectroscopy: A Practical Approach* (Wiley, 2019).
7. Li, J. & Zhang, M. Physics and applications of Raman distributed optical fiber sensing. *Light* **11**, 1–29 (2022).
8. Staritzbichler, R. *et al.* Raman spectroscopy on blood serum samples of patients with end-stage liver disease. *PLoS ONE* **16**, e0256045 (2021).
9. Choo-Smith, L. P. *et al.* Medical applications of Raman spectroscopy: From proof of principle to clinical implementation. *Biopolymers* **67**, 1–9 (2002).
10. Auner, G. W. *et al.* Applications of Raman spectroscopy in cancer diagnosis. *Cancer Metastasis Rev.* **37**, 691–717 (2018).

11. Bellot-Gurlet, L., Pagès-Camagna, S. & Coupry, C. Raman spectroscopy in art and archaeology. *J. Raman Spectrosc.* **37**, 962–965 (2006).
12. Zhang, W., Ma, J. & Sun, D. W. Raman spectroscopic techniques for detecting structure and quality of frozen foods: Principles and applications. *Crit. Rev. Food Sci. Nutr.* **61**, 2623–2639. <https://doi.org/10.1080/10408398.2020.1828814> (2020).
13. Kuhar, N., Sil, S., Verma, T. & Umapathy, S. Challenges in application of Raman spectroscopy to biology and materials. *RSC Adv.* **8**, 25888–25908 (2018).
14. Gaft, M. & Nagli, L. UV gated Raman spectroscopy for standoff detection of explosives. *Opt. Mater. (Amst)* **30**, 1739–1746 (2008).
15. López-López, M. & García-Ruiz, C. Infrared and Raman spectroscopy techniques applied to identification of explosives. *TrAC Trends Anal. Chem.* **54**, 36–44 (2014).
16. Singh, J. P. & Thakur, S. N. *Laser-induced breakdown spectroscopy*.
17. Khan, Z. H. *et al.* Laser-Induced Breakdown Spectroscopy (LIBS) for Trace Element Detection: A Review. *J. Spectrosc.* **2022**, 3887038. <https://doi.org/10.1155/2022/3887038> (2022).
18. Legnaioli, S. *et al.* Industrial applications of laser-induced breakdown spectroscopy: A review. *Anal. Methods* **12**, 1014–1029 (2020).
19. Markiewicz-Keszczka, M. *et al.* Laser-induced breakdown spectroscopy (LIBS) for food analysis: A review. *Trends Food Sci. Technol.* **65**, 80–93 (2017).
20. Anabitarte, F., Cobo, A., Lopez-Higuera, J. M., Byrne, H. J. & Louarn, G. Laser-induced breakdown spectroscopy: Fundamentals, applications, and challenges. *Int. Sch. Res. Netw. ISRN Spectrosc.* **2012**, 12 (2012).
21. Yano, J. & Yachandra, V. K. X-ray absorption spectroscopy. *Photosynth. Res.* **102**, 241–254 (2009).
22. Lagalante, A. F. Atomic absorption spectroscopy: A tutorial review. *Appl. Spectrosc. Rev.* **34**, 173–189. <https://doi.org/10.1081/ASR-100100844> (2007).
23. Liu, C. & Xu, L. Laser absorption spectroscopy for combustion diagnosis in reactive flows: A review. *Appl. Spectrosc. Rev.* **54**, 1–44. <https://doi.org/10.1080/05704928.2018.1448854> (2018).
24. Wesson, J. *Tokamaks*. (2011).
25. Hutchinson, I. H. *Principles of plasma diagnostics*. (1987).
26. Orsitto, F. P. *et al.* Diagnostics and control for the steady state and pulsed tokamak DEMO. *Nucl. Fusion* **56**, 026009 (2016).
27. Biel, W. *et al.* Diagnostics for plasma control—From ITER to DEMO. *Fusion Eng. Des.* **146**, 465–472 (2019).
28. Iguchine, V. *Active Control of Magneto-hydrodynamic Instabilities in Hot Plasmas* Vol. 83 (Springer, 2015).
29. Chen, W., Venables, D. S. & Sigrist, M. W. *Advances in Spectroscopic Monitoring of the Atmosphere* (Elsevier, 2021).
30. Winnewisser, G. & Kramer, C. Spectroscopy between the stars. *Space Sci. Rev.* **90**, 181–202 (1999).
31. Clark, R. N., Swayze, G. A., Carlson, R., Grundy, W. & Noll, K. Spectroscopy from space. *Rev. Mineral Geochem.* **78**, 399–446 (2014).
32. Slavov, C., Hartmann, H. & Wachtveitl, J. Implementation and evaluation of data analysis strategies for time-resolved optical spectroscopy. *Anal. Chem.* **87**, 2328–2336 (2015).
33. Nenninger, G. G., Piliarik, M. & Homola, J. Data analysis for optical sensors based on spectroscopy of surface plasmons. *Meas. Sci. Technol.* **13**, 2038 (2002).
34. Rothman, L. S. *et al.* The HITRAN2012 molecular spectroscopic database. *J. Quant. Spectrosc. Radiat. Transf.* **130**, 4–50 (2013).
35. Wei, D., Chen, S. & Liu, Q. Review of fluorescence suppression techniques in Raman spectroscopy. *Appl. Spectrosc. Rev.* <https://doi.org/10.1080/05704928.2014.99993650.387-406> (2015).
36. Kovalev, V. A. & Eichinger, W. E. *Elastic Lidar: Theory, Practice, and Analysis Method* (Wiley Interscience, 2004).
37. Howard, M. *Principles and Practice of Spectroscopy Calibration* (Wiley Interscience, 1991).
38. Mark, H. Quantitative Spectroscopic Calibration Update based on the original article by H. Mark, *Encyclopedia of Analytical Chemistry*, © 2000, John Wiley & Sons, Ltd. *Encyclopedia of Analytical Chemistry* (2012) doi:<https://doi.org/10.1002/9780470027318.A8108.PUB2>.
39. Fahrmeir, L., Kneib, T., Lang, S. & Marx, B. *Regression: Models, Methods and Applications* Vol. 9783642343339 (Springer-Verlag, 2013).
40. Ralbovsky, N. M. & Lednev, I. K. Towards development of a novel universal medical diagnostic method: Raman spectroscopy and machine learning. *Chem. Soc. Rev.* **49**, 7428–7453 (2020).
41. MezaRamirez, C. A., Greenop, M., Ashton, L. & Rehman, I. U. R. Applications of machine learning in spectroscopy. *Appl. Spectrosc. Rev.* **56**, 733–763. <https://doi.org/10.1080/05704928.2020.1859525> (2020).
42. Fu, W. & Hopkins, W. S. Applying machine learning to vibrational spectroscopy. *J. Phys. Chem. A* **122**, 167–171 (2018).
43. Fellner, L. *et al.* Determination of composition of mixed biological samples using laser-induced fluorescence and combined classification/regression models. *Eur. Phys. J. Plus* **136**, 1–6 (2021).
44. Wang, Z. *et al.* VECTOR: Very deep convolutional autoencoders for non-resonant background removal in broadband coherent anti-Stokes Raman scattering. *J. Raman Spectrosc.* <https://doi.org/10.1002/jrs.6335> (2022).
45. Huang, L., Luo, R., Liu, X. & Hao, X. Spectral imaging with deep learning. *Light* **11**, 1–19 (2022).
46. Zuo, C. *et al.* Deep learning in optical metrology: A review. *Light* **11**, 1–54 (2022).
47. Pan, B. Optical metrology embraces deep learning: Keeping an open mind. *Light* **11**, 1–3 (2022).
48. Raissi, M., Perdikaris, P. & Karniadakis, G. E. Physics-informed neural networks: A deep learning framework for solving forward and inverse problems involving nonlinear partial differential equations. *J. Comput. Phys.* **378**, 686–707 (2019).
49. Karniadakis, G. E. *et al.* Physics-informed machine learning. *Nat. Rev. Phys.* **3**, 422–440 (2021).
50. Cai, S., Mao, Z., Wang, Z., Yin, M. & Karniadakis, G. E. Physics-informed neural networks (PINNs) for fluid mechanics: A review. *Acta Mechanica Sinica/Lixue Xuebao* **37**, 1727–1738 (2021).
51. Arthurs, C. J. & King, A. P. Active training of physics-informed neural networks to aggregate and interpolate parametric solutions to the Navier-Stokes equations. *J. Comput. Phys.* **438**, 110364 (2021).
52. Sahli Costabal, F., Yang, Y., Perdikaris, P., Hurtado, D. E. & Kuhl, E. Physics-informed neural networks for cardiac activation mapping. *Front. Phys.* **8**, 42 (2020).
53. Misyris, G. S., Venzke, A. & Chatzivasileiadis, S. Physics-Informed Neural Networks for Power Systems. *IEEE Power and Energy Society General Meeting 2020-August*, (2019).
54. Ian Goodfellow, Yoshua Bengio, and A. C. Deep I Learning. *Genet Program Evolvable Mach* **19**, 305–307 (2017).
55. Ali, Z., Mahmood, T. & Yang, M. S. Complex T-spherical fuzzy aggregation operators with application to multi-attribute decision making. *Symmetry* **12**, 1311 (2020).
56. Versaci, M. *et al.* A Fuzzy similarity-based approach to classify numerically simulated and experimentally detected carbon fiber-reinforced polymer plate defects. *Sensors* **22**, 4232 (2022).
57. Radhakrishna, V., Aljawarneh, S. A., Kumar, P. V. & Choo, K. K. R. A novel fuzzy gaussian-based dissimilarity measure for discovering similarity temporal association patterns. *Soft Comput.* **22**, 1903–1919 (2018).
58. Nagi, J. *et al.* Max-pooling convolutional neural networks for vision-based hand gesture recognition. *2011 IEEE International Conference on Signal and Image Processing Applications, ICSIPA 2011* 342–347 (2011) doi:<https://doi.org/10.1109/ICSIPA.2011.6144164>.
59. Kingma, D. P. & Ba, J. L. Adam: A Method for Stochastic Optimization. *3rd International Conference on Learning Representations, ICLR 2015 - Conference Track Proceedings* (2014) doi:<https://doi.org/10.48550/arxiv.1412.6980>.

Author contributions

R.R. conceived of the presented idea and developed the theory. A.P. developed the code and performed the numerical and experimental tests. P.G. supervised the work. A.P. wrote the manuscript, while R.R. and P.G. reviewed it.

Competing interests

The authors declare no competing interests.

Additional information

Correspondence and requests for materials should be addressed to R.R.

Reprints and permissions information is available at www.nature.com/reprints.

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2023

Terms and Conditions

Springer Nature journal content, brought to you courtesy of Springer Nature Customer Service Center GmbH (“Springer Nature”).

Springer Nature supports a reasonable amount of sharing of research papers by authors, subscribers and authorised users (“Users”), for small-scale personal, non-commercial use provided that all copyright, trade and service marks and other proprietary notices are maintained. By accessing, sharing, receiving or otherwise using the Springer Nature journal content you agree to these terms of use (“Terms”). For these purposes, Springer Nature considers academic use (by researchers and students) to be non-commercial.

These Terms are supplementary and will apply in addition to any applicable website terms and conditions, a relevant site licence or a personal subscription. These Terms will prevail over any conflict or ambiguity with regards to the relevant terms, a site licence or a personal subscription (to the extent of the conflict or ambiguity only). For Creative Commons-licensed articles, the terms of the Creative Commons license used will apply.

We collect and use personal data to provide access to the Springer Nature journal content. We may also use these personal data internally within ResearchGate and Springer Nature and as agreed share it, in an anonymised way, for purposes of tracking, analysis and reporting. We will not otherwise disclose your personal data outside the ResearchGate or the Springer Nature group of companies unless we have your permission as detailed in the Privacy Policy.

While Users may use the Springer Nature journal content for small scale, personal non-commercial use, it is important to note that Users may not:

1. use such content for the purpose of providing other users with access on a regular or large scale basis or as a means to circumvent access control;
2. use such content where to do so would be considered a criminal or statutory offence in any jurisdiction, or gives rise to civil liability, or is otherwise unlawful;
3. falsely or misleadingly imply or suggest endorsement, approval, sponsorship, or association unless explicitly agreed to by Springer Nature in writing;
4. use bots or other automated methods to access the content or redirect messages
5. override any security feature or exclusionary protocol; or
6. share the content in order to create substitute for Springer Nature products or services or a systematic database of Springer Nature journal content.

In line with the restriction against commercial use, Springer Nature does not permit the creation of a product or service that creates revenue, royalties, rent or income from our content or its inclusion as part of a paid for service or for other commercial gain. Springer Nature journal content cannot be used for inter-library loans and librarians may not upload Springer Nature journal content on a large scale into their, or any other, institutional repository.

These terms of use are reviewed regularly and may be amended at any time. Springer Nature is not obligated to publish any information or content on this website and may remove it or features or functionality at our sole discretion, at any time with or without notice. Springer Nature may revoke this licence to you at any time and remove access to any copies of the Springer Nature journal content which have been saved.

To the fullest extent permitted by law, Springer Nature makes no warranties, representations or guarantees to Users, either express or implied with respect to the Springer nature journal content and all parties disclaim and waive any implied warranties or warranties imposed by law, including merchantability or fitness for any particular purpose.

Please note that these rights do not automatically extend to content, data or other material published by Springer Nature that may be licensed from third parties.

If you would like to use or distribute our Springer Nature journal content to a wider audience or on a regular basis or in any other manner not expressly permitted by these Terms, please contact Springer Nature at

onlineservice@springernature.com