Acoustic Shadow Detection: Study and Statistics of B-Mode and Radiofrequency Data

Ricky Hu[a,*], Rohit Singla[a], Farah Deeba[a], Robert N. Rohling[a,b]

[a]*Department of Electrical and Computer Engineering, University of British Columbia, Vancouver, Canada*
[b]*Department of Mechanical Engineering, University of British Columbia, Vancouver, Canada*

Abstract

An acoustic shadow is an ultrasound artifact occurring at boundaries between significantly tissue impedances, resulting in signal loss and a dark appearance. Shadow detection is important as shadows can identify anatomical features or obscure regions of interest. A study was performed to scan human subjects (N=35) specifically to explore the statistical characteristics of various shadows from different anatomy and with different transducers. Differences in shadow statistics were observed and used for shadow detection algorithms with a fitted Nakagami distribution on radiofrequency speckle (RF) or cumulative entropy on brightness-mode (B-mode) data. The fitted Nakagami parameter and Entropy values in shadows were consistent across different transducers and anatomy. Both algorithms utilized adaptive thresholding, needing only the transducer pulse width as an input parameter for easy utilization by different operators or equipment. Mean Dice coefficients

*Corresponding Author: Ricky Hu, Robotics and Control Laboratory, University of British Columbia, Room 3090, 2332 Main Mall, Vancouver, BC, Canada V6T 1Z4. Email: rhu@ece.ubc.ca

(± standard deviation) of 0.90±0.07 and 0.87±0.08 were obtained for the RF and B-mode algorithms, which is within the range of manual annotators. The high accuracy in different imaging scenarios indicate that the shadows can be detected with high versatility and without expert configuration. The understanding of shadow statistics can be used for more specialized techniques to be developed for specific applications in the future, including pre-processing for machine learning and automatic interpretation.
Keywords: Acoustic Shadow, Ultrasound, Speckle, Radiofrequency, Segmentation

---

Introduction

Ultrasound devices have become increasingly affordable and portable, encouraging applications such as point-of-care ultrasound (Bouhemad et al., 2011), novice usage (Sippel et al., 2011), and analysis by machine learning (Ghose et al., 2013). However, ultrasound is susceptible to unique artifacts that increase the difficult of interpretation and processing of images. One artifact is an acoustic shadow, which occurs when an ultrasound wave crosses a boundary of two materials with high impedance differences (Kremkau and Taylor, 1986). The wave is almost completely reflected and depicted beyond the boundary is a continuous dark region and a loss of anatomical features. Shadows occur in air-tissue, tissue-bone, and tissue-lesion interfaces. Shadows can aid interpretation, such as identifying gall stones (Good et al., 1979) or spinal levels (Galiano et al., 2005). However, shadows, such as from poor transducer contact, can lead to misinterpretation of anatomy, particularly by novice users and automated processing algorithms. Thus, the identification of shadows is an important preprocessing step in many applications.

Several methods have been used in literature to detect shadows and illustrative examples are discussed. Geometric techniques model the path of an ultrasound signal for an expected image along the scanline using a random walk (Karamalis et al., 2012). Pixels are then flagged as a shadow if it is below a confidence threshold. However, geometric techniques require knowledge of ultrasound transducer properties to parameterize random walk weights, such as the focal length, radius of curvature, and thickness. The technique is therefore challenging to implement across different ultrasound equipment. This also reduces applicability for machine learning applications

as accurate transducer parameter labels are required for each image.

Pixel intensity methods ignore the transducer properties and analyze only the graphical properties of an image (Hellier et al., 2010). Shadows have been detected on brain images by analyzing the entropy along a scanline to flag pixels of sudden low entropy as a potential shadow. The technique achieved a comparable Dice similarity coefficient as geometric methods but require specific thresholding, window sizing, filtering, and image mask parameterization for different anatomy and transducers. The drawback is again the need for parameterization and tuning, which requires image processing expertise and prior knowledge of specific applications.

Machine learning methods have gained significant interest in medical imaging analysis. To our knowledge, no machine learning method has demonstrated capability of general shadow detection from multiple anatomy. Deep learning methods have identified features in a specific image sets that contain shadows, such as neuroanatomical regions in cranial scan (Milletari et al., 2017) or spinal levels in a posterior scan (Hetherington et al., 2017). Although machine learning has the potential of providing automated feature recognition in multiple applications, a large data set is required for an algorithm to recognize certain features. Ultrasound imaging is highly variable due to unique artifacts, operator technique, and equipment. In addition, shadows are a common feature that occur in various imaging scenarios. Previous techniques focus on a single anatomical region and training data was from a consistent imaging scenario. However, it is difficult to construct a training data set with the generality required to recognize shadows in different scenarios usable for a variety of ultrasound applications.

There are two objectives to this paper. First, to address the need for understanding general characteristics of shadows, a study was conducted to scan multiple anatomy and transducers specifically to analyze the statistics of different types of shadows. Second, to address existing needs for versatile detection with minimal parameterization, previous methods were then extended utilizing statistical thresholding of radiofrequency (RF) or brightness-mode (B-mode) data to detect shadows from various imaging scenarios.

Materials and Methods

Data Collection

Ultrasound RF and B-mode data were acquired by scanning 37 adult participants with informed written consent, approved by the University of British Columbia Research Ethics Board (Study ID: H18-01199). The scans included a forearm scan near the distal end of the pronator quadratus, an elbow scan near the cubital fossa, and a rib scan on the anterior surface of right ribs 11-12. Each scan was taken with both a curvilinear (Model C5-2/60, Ultrasonix Medical Corporation, Richmond, BC, Canada) and linear (Model L14-5/38, Ultrasonix Medical Corporation, Richmond, BC, Canada) transducer. Different transducer settings were used for each anatomical region and transducer, summarized in Table 1. Shadows were expected to occur due to superficial and deep bones and from an air gap created by the lateral edges of the transducer not being in flush contact with the skin. The experiment was designed to generate a dataset from various imaging scenarios to explore general shadow characteristics and to validate the versatility of the two simple shadow detection methods.

Radiofrequency Speckle Analysis

To analyze shadows, windows of speckle were analyzed on the RF signal. Speckle occurs due to multiplicative scattering of acoustic waves in a material, resulting in a granular appearance on the image. The benefit of RF analysis is that B-mode image processing commonly attempts to remove speckle, but speckle contains information of the acoustic interactions in tissue (Burckhardt, 1978). Speckle can then characterize different regions, such as a region of tissue or a region of signal loss in a shadow. In addition, B-mode image formation can be manipulated by an operator to visually enhance an image, such as adjusting time-gain compensation or dynamic range. Thus, the underlying speckle analysis can provide shadow detection usable across different machines and operators.

One of the first models for speckle is the one parameter Rayleigh distribution to model the probability density of a random walk (Burckhardt, 1978). The Rayleigh distribution is capable for modeling fully developed speckle, which does not occur in limited scattering (Tuthill et al., 1988). More generalized models have been applied such as the Rician, Homodyned-K, and Nakagami distributions to characterize speckle (Destrempes and Cloutier, 2010). The utility of speckle has been demonstrated in the literature to classify tumorigenicity of breast lesions (Byra et al., 2016) or levels of liver fibrosis (Ho et al., 2012) by categorizing image regions based on the speckle pattern. Shadow characterization presents a simpler problem as shadow and non-shadow regions contain significantly different speckle patterns. Thus, the Nakagami distribution expressed in Eq. 1 was chosen to model speckle. The Nakagami distribution provides greater generality than the Rayleigh

distribution while being more computationally efficient than the Rician or Homodyned K distributions (Destrempes and Cloutier, 2010):

$$\Phi(x, \mu, \omega) = 2(\frac{\mu}{\omega})^{\mu} \frac{1}{\Gamma(\mu)} x^{(2\mu-1)} e^{\frac{-\mu}{\omega}x^2} \tag{1}$$

where x is RF intensity, μ is the shape parameter or Nakagami "m" parameter, ω is a scale parameter and $\Gamma(\mu)$ is the gamma distribution.

To characterize shadows, the raw RF data was first processed by computing the echo envelope of each scanline with a Hilbert transform. This was performed on an averaged RF signal from three image frames. This creates a pre-scan converted image, visually similar to B-mode but without filtering to remove speckle. Next, the RF image was divided into overlapped windows with a width of a single RF data point and a length of three times the pulse width. This patch size was demonstrated in literature to be sufficiently large to capture multiple wavelengths and scattering events while being small enough to be useful in differentiating different regions on the millimeter scale (Byra et al., 2016). Next, each window was fit to a Nakagami distribution using a maximum likelihood estimate to compute a map of Nakagami parameters μ and ω, as shown in Fig. 1.

To detect shadows, Otsu's method was applied on the entire image to automatically compute a threshold for the ω parameter. This was sufficient as the ω parameter is significantly different for shadow regions with abundant speckle and non-shadow regions with minimal speckle, Then, for each scanline, the axially deepest data point that is above the threshold is labeled as the shadow boundary and all data points below are labeled as a shadow.

The Nakagami shape parameter, m, was also investigated, though there

was not sufficient delineation between parameter values in shadow and non-shadow regions for this parameter to be effective in thresholding. The distributions of the two paraemters are displayed for shadow and non-shadow regions in Figure [XX].

B-mode Scanline Analysis

Many ultrasound machines do not provide access to RF data for speckle analysis. Thus, a previous pixel-intensity shadow detection method on B-mode images was modified and extended. Scanline entropy was investigated on B-mode images to characterize different types of shadows, but with the addition of adaptive thresholding of entropy to address the need for usability with minimum configuration. B-mode analysis was performed on an averaged image from 3 image frames, similar to RF analysis. First, the cumulative scanline entropy is computed for each pixel, similar to the "Rupture Criterion" (Hellier et al., 2010), with the window size fixed as three times the pulse width, η, as defined in Eq. 2. This is the same window size as the RF analysis.

$$S_{i,j} = \sum_{k=1}^{3\eta} I(i-k,j) log_2 \frac{I(i-k,j)}{I(i+k,j)} + I(i+k,j) log_2 \frac{I(i+k,j)}{I(i-k,j)} \qquad (2)$$

where $S_{i,j}$ is the cumulative entropy at pixel i on scanline j, η is the pulse width, and $I(i)$ is the gray level, or intensity, of pixel $(i, j$. For the case of curvilinear images, radial scanlines were linearly interpolated between the two symmetric lateral edges of the image.

Next, Otsu's method is applied similarly to compute a threshold entropy value. The intuition of the threshold is different than in RF analysis. In

RF analysis, the threshold separates patches of intense and minimal speckle. In B-mode analysis, the threshold separates pixels of a shadow boundary, which has high entropy, and pixels away from shadow boundary, which include shadow and non-shadow regions. Thus, shadows can be identified by finding the last pixel on a scanline with an entropy higher than the threshold, representing a bright shadow boundary.

Validation

A trained annotator (RH) manually outlined the boundary of the shadow regions on B-mode images. The manual regions were used as a gold standard, as manual identification is common in clinical practice and has been used in previous literature for comparison (Hellier et al., 2010). A Dice coefficient was computed to compare similarity of manual and automated shadow detection. The manual outline was used to define four regions for classification of statistical parameters: a non-shadow region above the boundary, a shadow region below the boundary, a "transition region", which is a window defined as three pulse widths long axially below the boundary, and a "deep shadow region", which is the data below the transition region. The validation was repeated with the RF and entropy window increased and decreased by 50%. The Ljung-Box Q-test was use to measure residual autocorrelation of the Dice coefficients.

Results

Examples of detected shadows from both methods are highlighted in gray in Fig. 2 in different shadow detection scenarios. The Dice coefficients for both methods for different anatomy and transducers are shown in Table

2. The mean Dice coefficients (± standard deviation) were 0.90±0.07 and 0.87±0.08 for RF and B-mode methods. Manual annotation was repeated five times with a mean Dice coefficient of 0.92±0.02 for all images and transducers. The Dice coefficient did not change by more than 0.03 when the window size was varied by 50%.

With the benefit of a varied dataset, general statistics of shadows can be analyzed, as summarized in Table 3 and Table 4. For shadow detection, the parameters between a shadow and non-shadow are of particular interest. Shadows were observed to have a mean Nakagami ω parameter of 4.14 ± 0.40 and a mean entropy of 1.03 ± 0.29 whereas non-shadows were observed to have a mean ω of 6.24 ± 0.92 and 2.20 ± 0.81. The values of entropy and Nakagami ω are consistent across different transducers and anatomical regions. The variance of entropy and Nakagami ω in one imaging region and transducer setting is less than the variance across different regions and transducers for shadows and non-shadows.

Discussion

The RF and B-mode shadow detection developed achi-eved a comparable Dice similarity coefficient to manual detection for all anatomy and transducer types ($p < 0.025$). The previous studies using B-mode entropy reported a mean Dice coefficient of 0.91±0.07 between manual annotators (Hellier et al., 2010). An important feature of shadow detection is being able to differentiate between a shadow and simply high attenuation of the signal. Both scenarios result in an eventual loss of signal. Shadow detection, however, has a characteristic high intensity shadow boundary before a significant loss in signal,

compared to gradual signal losses in attenuation. The high Dice similarity coefficient indicates that both methods were capable of this distinction. This is also visualized in Fig. 2, where regions of low intensity without a bright shadow boundary were correctly labeled as non-shadow. The high accuracy supports the versatility of the detection method as both methods are able to identify shadows across different anatomy and transducers with minimum configuration.

For a general observation for shadows, the computed Nakagami ω parameter of all manually outlined shadows indicate that there is a statistically significant difference between shadow and non-shadow regions, regardless of anatomy and transducer and even with the error in the transition regions considered. The speckle and its statistics from shadows is thus distinct from the speckle created by tissue, muscle, or fat. This observation can be utilized in the future for further analysis of shadows.

In RF detection, both false positive and false negative errors most frequently occurred immediately below a shadow boundary as opposed to B-mode detection where errors were in various regions. To study the frequent areas of error further, the "transition region" immediately below a manually annotated shadow boundary and a "deep shadow region" below the transition region was investigated. The Nakagami ω parameter of transition regions of all anatomy and transducers were within a standard deviation of both shadow and non-shadow regions. The deeper shadow regions were observed to have a lower Nakagami ω parameter than shadow regions and with a lower standard deviation as summarized in Table 3. The spread of the speckle also significantly decreases after the transition region. This indicates

that the transition region cannot be fully distinguished from either a shadow or non-shadow and presents as it is statistically similar to the two. This is likely the cause of the errors, as the speckle distribution is much more consistent in the deep shadow regions compared to any other region. Physically, speckle interactions appear to gradually lessen after a brightest point on a scanline, possibly due to incomplete total reflection at a boundary. The boundary is thus is not an instantaneous division between non-shadow and shadow, rather, there is a transition region with statistics between a shadow and non-shadow before the speckle fully resembles a shadow.

In the transition region of B-mode images, the entropy values were similar but consistently higher than non-shadow values. This is expected as entropy is the highest when there is the greatest change in pixel intensity, which occurs at a shadow boundary, even with the a non-instantaneous non-shadow to shadow transition. However, the averaged entropy of all non-shadow regions have a greater spread than the Nakagami parameters, likely due to the differing operator settings used. Thus, B-mode detection may not be as consistent as RF detection.

There is a limitation with analysis with the Nakagami distribution in that the Nakagami parameters change depending on frequency. (**?**) observed that in the 36-58MHz transducer frequency range, the Nakagami parameter decreased near the theoretical lower limit compared to a higher parameter value at 10MHz signal. This was reported to be due to the spatial organization of the cells to be "on the order of a fraction of the wavelength" and a Nakagami distribution cannot model the scatterers of red blood cells at this frequency. From limitations of the equipment used in our study, we cannot

conclude that shadow detection with Nakagami analysis will be accurate in frequencies higher than 13.3MHz, especially with the changes in the Nakagami parameters and scatterers at high frequencies. Diagnostic ultrasound mainly uses a frequency range of 2-15MHz (**?**) and the shadow detection method is expected to not be applicable in this range without issues from the high frequency behaviour of the Nakagami distribution.

In previous literature, shadows were defined qualitatively (Kremkau and Taylor, 1986) as a sudden loss of signal and brightness. The observed transition region in this study suggests that the qualitative definition of a shadow may be insufficient for accurate detection. One algorithm may detect the shadow starting immediately after the brightest location, or another may use a convention such as a full width at half maximum to define where the signal has sufficiently low intensity to resemble the start of a shadow. There is a decision point required for a clear definition for where a shadow begins to improve shadow detection accuracy, both from a signaling perspective for image processing and a visual perspective for manual inspection.

The findings in this study result in several implications. First, the statistics of acoustic shadows have been investigated on a dataset with shadows occurring from multiple scenarios as opposed to specific cases where shadows are observed. This provided a more generalizable observation that shadows can be characterized by distinctive speckle distributions in different of anatomy and equipment and that there exists a transition region before the loss of speckle in a shadow. Second, the shadow detection methods demonstrated high accuracy, indicating that the same shadow detection method can be used with different transducer or imaging location. In future stud-

ies, the speckle statistics observed can be used to develop further models for anatomical features containing shadows. In machine learning algorithms, an initial network could be used with the shadow detection methods presented. Future studies would also have to take into consideration the most frequent source of error of shadow detection as the shadow boundary.

Conclusions

Acoustic shadows from different imaging scenarios were investigated. RF and B-mode methods were developed for acoustic shadow detection requiring only the transducer pulse width as the input parameter. When comparing to manual detection, the methods achieved a Dice similarity coefficient within range of manual observers. The work focused on applying shadow detection and statistical analysis to a varied dataset of three different anatomical locations and two different transducer to provide a representative understanding of general acoustic shadows. The statistics of acoustic shadow indicate that shadows contain a distinct speckle distribution compared to non-shadows and the speckle characteristics transition at the shadow boundary. The statistical findings of shadows can aid interpretation of ultrasound images in the future using speckle analysis. The versatility of the shadow detection method has the potential to improve the interpretation of ultrasound images with shadow artifacts or to serve as a pre-processing step for machine learning methods.

Acknowledgements

This work is supported by the National Sciences and Engineering Research Council of Canada (Grant Number: F09-05533). We acknowledge

Victoria Lessoway for training and assistance in manually identifying acoustic shadows.

References

Bouhemad B, Brisson H, Le-Guen M, Arbelot C, Lu Q, Rouby JJ. Bedside ultrasound assessment of positive end-expiratory pressure-induced lung recruitment. American Journal of Respiratory and Critical Care Medicine, 2011;183:341–347.

Burckhardt CB. Speckle in ultrasound B-mode scans, 1978.

Byra M, Nowicki A, Wróblewska-Piotrzkowska H, Dobruch-Sobczak K. Classification of breast lesions using segmented quantitative ultrasound maps of homodyned K distribution parameters. Med. Phys., 2016;43:5561–5569.

Destrempes F, Cloutier G. A critical review and uniformized representation of statistical distributions modeling the ultrasound echo envelope. Ultrasound Med. Biol., 2010;36:1037–1051.

Galiano K, Obwegeser AA, Bodner G, Freund M, Maurer H, Kamelger FS, Schatzer R, Ploner F. Ultrasound guidance for facet joint injections in the lumbar spine: A computed tomography-controlled feasibility study. Anesthesia and Analgesia, 2005;101:579–583.

Ghose S, Oliver A, Mitra J, Martí R, Lladó X, Freixenet J, Sidibé D, Vilanova JC, Comet J, Meriaudeau F. A supervised learning framework of statistical shape and probability priors for automatic prostate segmentation in ultrasound images. Medical Image Analysis, 2013;17:587–600.

Good LI, Edell SL, Soloway RD, Trotman BW, Mulhern C, Arger Pa. Ultrasonic properties of gallstones. Effect of stone size and composition. Gastroenterology, 1979;77:258–263.

316 Hellier P, Coupé P, Morandi X, Collins DL. An automatic geometrical and
317 statistical method to detect acoustic shadows in intraoperative ultrasound
318 brain images. Medical Image Analysis, 2010;14:195–204.

319 Hetherington J, Lessoway V, Gunka V, Abolmaesumi P, Rohling R. SLIDE:
320 automatic spine level identification system using a deep convolutional neu-
321 ral network. International Journal of Computer Assisted Radiology and
322 Surgery, 2017;12:1189–1198.

323 Ho MC, Lin JJ, Shu YC, Chen CN, Chang KJ, Chang CC, Tsui PH. Using
324 ultrasound Nakagami imaging to assess liver fibrosis in rats. Ultrasonics,
325 2012;52:215–222.

326 Karamalis A, Wein W, Klein T, Navab N. Ultrasound confidence maps using
327 random walks. Medical Image Analysis, 2012;16:1101–1112.

328 Kremkau FW, Taylor KJ. Artifacts in ultrasound imaging. Journal of Ultra-
329 sound in Medicine, 1986;5:227–237.

330 Milletari F, Ahmadi SA, Kroll C, Plate A, Rozanski V, Maiostre J, Levin
331 J, Dietrich O, Ertl-Wagner B, Bötzel K, Navab N. Hough-CNN: Deep
332 learning for segmentation of deep brain regions in MRI and ultrasound.
333 Computer Vision and Image Understanding, 2017;164:92–102.

334 Sippel S, Muruganandan K, Levine A, Shah S. Review article: Use of ultra-
335 sound in the developing world. Int. J. Emerg. Med., 2011;4:72.

336 Tuthill TA, Sperry RH, Parker KJ. Deviations from rayleigh statistics in
337 ultrasonic speckle. Ultrasonic Imaging, 1988;10:81–89.

@articleCloutier2004, author = Cloutier, Guy and Daronatand, Michel and Savé Ry, David and Garcia, Damien and Durand, Louis-Gilles and Foster, F Stuart, isbn = 00014966, issn = 00014966, journal = Acoust. Soc. Am., keywords = 4380Qf,4380Vj FD Pages,4380n,566 577,PACS numbers, mendeley-groups = Acoustic Shadowing, number = 1, pages = 566–577, pmid = 15296017, title = Non-Gaussian statistics and temporal variations of the ultrasound signal backscattered by blood at frequencies between 10 and 58 MHz, volume = 116, year = 2004

@articleJensen2007, author = Jensen, Jørgen Arendt, doi = 10.1016/j.pbiomolbio.2006.07.025, isbn = 0079-6107, issn = 00796107, journal = Prog. Biophys. Mol. Biol., keywords = Blood velocity estimation,Imaging,Medical ultrasound, mendeley-groups = Acoustic Shadowing, number = 1-3, pages = 153–165, pmid = 17092547, title = Medical ultrasound imaging, volume = 93, year = 2007

Figure Captions

Figure 1: A visualization of the B-mode and RF parameter maps. The b) Entropy Map was computed from processing of the a) original B-mode image and the d) Nakagami ω map was computed from the c) echo envelope. Note that the echo envelope contains noticeable speckle, which has been used to fit a Nakagami distribution to characterize shadow. The region at depth 2.50cm and scanlines 32-40 is attenuation and not a shadow. This is an important distinction in shadow detection and both maps show the region as below a threshold to flag a shadow boundary.

Figure 2: A comparison of the original B-mode images, the detected shadows manual detection, RF detection, and B-mode detection. Both detection methods perform similar to manual detection. Both methods perform slightly less accurately on curvilinear images, likely due to the reduced resolution from interpolating the scanlines. Most errors of RF detection occur near the shadow boundary, likely due to the transitioning speckle from non-shadow to shadow.

Tables

Table 1: Transducer properties for different imaging scenarios.

	Anatomy	**Frequency**	**Depth**	**Gain**
Linear Transducer (L14-5/38)	Forearm	11.0MHz	5.0cm	50%
	Elbow	11.0MHz	5.0cm	40%
	Ribcage	5.0MHz	10.0cm	30%
Curvilinear Transducer (C5-2/60)	Forearm	4.0MHz	5.0cm	50%
	Elbow	4.0MHz	5.0cm	40%
	Ribcage	3.3MHz	10.0cm	30%

Table 2: Mean Dice coefficients for different imaging scenarios ± standard deviation.

		RF	**B-Mode**
Linear (L14-5/38)	Forearm	0.91±0.05	0.89±0.06
	Elbow	0.94±0.06	0.90±0.07
	Ribcage	0.87±0.09	0.84±0.06
Curvilinear (C5-2/60)	Forearm	0.89±0.05	0.86±0.08
	Elbow	0.93±0.04	0.90±0.09
	Ribcage	0.83±0.08	0.83±0.10
Mean	**All Anatomy**	**0.90±0.07**	**0.87±0.08**

Table 3 : The mean Nakagami ω and Entropy values of different anatomy,

transducer, and shadowing region ± standard deviation. Values are consistent among different transducers and anatomical regions. The variance of entropy and Nakagami ω in one imaging region and transducer setting is less than the variance across different regions and transducers for shadows and non-shadows.

	Linear (L14-5/38)			**Curvilinear (C5-2/60)**		
	Forearm	Elbow	Ribcage	Forearm	Elbow	Ribcage
Nakagami ω (Log Scale)						
Shadow	4.15 ± 0.45	4.18 ± 0.45	4.04 ± 0.42	4.22 ± 0.32	4.19 ± 0.40	4.08 ± 0.37
Non-Shadow	6.19 ± 0.96	6.49 ± 0.97	6.29 ± 0.95	6.54 ± 0.88	6.29 ± 1.04	5.64 ± 0.71
Transition	4.94 ± 0.62	5.36 ± 0.62	4.96 ± 0.38	5.26 ± 1.02	5.37 ± 0.99	4.59 ± 0.92
Deep Shadow	4.13 ± 0.43	4.16 ± 0.43	4.03 ± 0.41	3.93 ± 0.20	4.09 ± 0.30	4.03 ± 0.26
Entropy (Log Scale)						
Shadow	0.92 ± 0.22	1.10 ± 0.36	1.04 ± 0.27	1.06 ± 0.28	0.96 ± 0.21	1.10 ± 0.37
Non-Shadow	2.34 ± 0.96	2.34 ± 0.80	2.14 ± 0.82	1.67 ± 0.82	1.75 ± 1.14	1.88 ± 0.42
Transition	2.45 ± 0.62	2.56 ± 0.53	2.15 ± 0.51	2.18 ± 1.21	1.93 ± 1.10	1.99 ± 1.10
Deep Shadow	0.71 ± 0.43	0.89 ± 0.26	0.92 ± 0.40	0.98 ± 0.21	0.82 ± 0.19	1.04 ± 0.26

Table 4 : The mean Nakagami ω and Entropy values of all anatomy and transducers for different shadowing regions ± standard deviation.

	Mean Nakagami ω (Log Scale)	**Mean Entropy (Log Scale)**
Shadow	4.14 ± 0.40	1.03 ± 0.29
Non-Shadow	6.24 ± 0.92	2.02 ± 0.81
Transition	5.08 ± 0.77	2.21 ± 0.84
Deep Shadow	4.06 ± 0.34	0.89 ± 0.27