Advanced 3D Reconstruction and Active View Planning in Robotic Ultrasound: Optimizing the SonoGym Pipeline

Introduction to Robotic Ultrasound and the Need for High-Fidelity Simulation

Ultrasound (US) imaging represents a cornerstone of modern diagnostic and intraoperative medical practice. Its non-ionizing nature, real-time feedback capabilities, and relative cost-effectiveness make it an indispensable modality across numerous clinical disciplines, from obstetrics and cardiology to trauma surgery and orthopedics[1]. Despite these profound advantages, traditional freehand ultrasound is notoriously operator-dependent. The quality of the acquired images, the stability of the acoustic window, and the clinician's ability to mentally reconstruct complex three-dimensional anatomical structures from two-dimensional planar slices rely entirely on extensive training and visuospatial cognitive ability[1]. This dependency introduces significant variability in diagnostic accuracy and procedural outcomes, particularly in environments with high cognitive loads or limited anatomical visibility.

Robotic ultrasound systems have emerged as a highly promising solution to mitigate this operator dependency. By mounting the ultrasound transducer onto a precise, multi-degree-of-freedom robotic manipulator, the system can autonomously navigate the patient's anatomy, maintain optimal contact forces, and systematically track target structures[1]. In orthopedic and spinal surgeries, robotic ultrasound is particularly impactful. Procedures such as pedicle screw placement demand sub-millimeter precision, and the ability to generate a highly accurate 3D reconstruction of the vertebral surface intraoperatively—without subjecting the patient or surgical staff to the ionizing radiation of continuous fluoroscopy or intraoperative CT—is invaluable[1]. The reconstructed ultrasound model can be registered directly against preoperative CT scans to provide seamless surgical guidance[1].

However, the development of intelligent, autonomous control policies for robotic ultrasound—specifically those utilizing Deep Reinforcement Learning (DRL) and Imitation Learning (IL)—has been severely bottlenecked by the lack of highly realistic, computationally efficient simulation environments[1]. DRL agents require millions of environmental interactions to converge on optimal policies[1]. While general-purpose robotic simulation platforms excel at rigid-body dynamics, they traditionally lack the specialized acoustic modeling required to simulate the complex physics of ultrasound, including speckle, attenuation, shadowing, and incidence-angle-dependent specular reflections[1].

The introduction of the SonoGym simulation platform represents a critical breakthrough in this domain[1]. By formulating ultrasound-guided navigation, 3D anatomy reconstruction, and surgical drilling as specialized Markov Decision Processes (MDPs), SonoGym enables the massively parallel training of DRL agents across diverse, CT-derived patient anatomies[1]. This

report provides an exhaustive, in-depth analysis of the 3D reconstruction pipeline within the SonoGym framework. It specifically investigates the architectural strategies required to integrate this pipeline with the PyBullet physics engine to drastically decrease computational overhead and make the system lightweight. Furthermore, it explores the advanced mathematics of predicting optimal scan views via Next-Best-View (NBV) planning, evaluates methodologies for quantifying process efficiency, identifies the physical constraints that define the absolute "best" ultrasound views, and establishes a theoretical framework for determining the minimum number of scans required to achieve a geometrically complete 3D reconstruction.

The SonoGym 3D Reconstruction Pipeline

The core architectural innovation of the SonoGym platform is its ability to facilitate parallel simulations across tens to hundreds of independent environments by decoupling the rigid-body kinematics of the robotic manipulator from the computationally intensive ultrasound image generation process[1]. The platform utilizes high-fidelity 3D patient anatomical models derived from clinical datasets, such as TotalSegmentator, which comprise both 3D CT volumes and corresponding segmented label maps of target anatomies like the lumbar spine[1].

Formulating the Reconstruction Task as a Submodular Markov Decision Process

In the specific context of 3D spinal surface reconstruction, the fundamental objective is to maneuver an ultrasound probe across the patient's skin to aggregate 2D segmented bone contours into a unified, highly accurate 3D surface model[1]. Unlike basic point-to-point navigation, determining an optimal scanning path is fundamentally a continuous problem of balancing exploration with exploitation. The system must explore unobserved regions of the vertebra while exploiting known regions to refine surface quality and ensure sufficient overlap for image registration[1].

SonoGym formulates this complex planning problem as a Submodular Markov Decision Process[1]. A standard MDP is traditionally defined by a tuple encompassing a state space, an action space, transition probabilities, and a reward function[1]. However, standard state-based additive rewards are fundamentally insufficient for 3D reconstruction because the informational value of acquiring a specific ultrasound frame depends entirely on the history of previously acquired frames. Re-scanning a known, high-quality area yields no new geometric information, indicating a mathematical property of diminishing returns[1].

To capture this history-dependent utility, the standard additive reward function is replaced with a monotone submodular set function defined over the entire trajectory of the probe[1]. At any given time step t, the state s_t represents the unknown relative position and orientation between the ultrasound probe and the target vertebra[1]. The observation o_t provided to the DRL agent is a voxelized 3D tensor representing the currently accumulated surface reconstruction[1]. The continuous action space a_t consists of four degrees of freedom:

translations along the x and y axes of the skin surface, and rotational adjustments in pitch and yaw to align the acoustic beam with the underlying bone surface normals[1].

The submodular objective function $F(\tau_{0:t})$, evaluated over a trajectory τ, explicitly penalizes unnecessary robotic movement while heavily rewarding novel surface coverage[1]:

$$F(\tau_{0:t}) := |\mathcal{M}_{0:t}| - w_2 \sum_{h=0}^{t} (|\Delta x_h| + |\Delta y_h| + w_3|\Delta \alpha_h| + w_3|\Delta \beta_h|)$$

In this formulation, $|\mathcal{M}_{0:t}|$ denotes the total area of the reconstructed surface, while the penalty terms account for the physical path length and the rotational effort of the robotic manipulator, modulated by tunable weights w_2 and w_3 [1]. While solving a submodular MDP to find the absolute global optimum is computationally intractable, leveraging greedy maximization of the marginal gain at each step via Proximal Policy Optimization (PPO) algorithms yields near-optimal policies[1]. This mathematical formulation is what allows the agent to actively optimize its trajectory on the fly.

Dual Modalities of Ultrasound Simulation

To support robust sim-to-real transfer—ensuring that an agent trained in simulation performs effectively on physical hardware—the reconstruction pipeline must generate synthetic ultrasound feedback that accurately mirrors real-world acoustic phenomena. SonoGym implements two distinct paradigms to achieve this while maintaining the high frame rates necessary for reinforcement learning[1]:

The first is a Physics-Based (Model-Based) Simulation approach. This method simulates ultrasound generation by extracting 2D slices from the 3D CT volume and label maps corresponding exactly to the probe's current spatial imaging plane[1]. It utilizes a convolution-based ray-tracing method where the total pixel intensity I is calculated as the sum of a reflection component R and a backscattered component B [1]. The reflection term relies heavily on the acoustic impedance mismatch—derived directly from the CT Hounsfield units—and the incidence angle of the simulated acoustic rays against the tissue boundaries[1].

The second is a Learning-Based (Generative) Simulation approach. While ray-tracing is physically accurate for specular reflection, it struggles to perfectly replicate the complex acoustic textures, speckle patterns, and reverberation artifacts seen in clinical ultrasound. To bridge this domain gap, an image-to-image translation network based on the Pix2Pix Generative Adversarial Network (GAN) framework is employed[1]. A deep U-Net architecture translates the standard extracted CT slices into photorealistic ultrasound images[1]. This model is trained on a massive proprietary dataset of paired ex-vivo CT-US scans, utilizing optical

tracking to perfectly align the modalities[1].
While the learning-based approach provides superior photorealism, performing deep neural network inference for every step of an RL environment is inherently taxing on GPU memory and compute bandwidth. Consequently, optimizing the underlying physics engine integration is a paramount concern for scaling these environments and making the pipeline lightweight.

Architectural Integration with PyBullet for Lightweight Compute

Simulation environments for surgical robotics and continuous manipulation, such as SurRoL (Surgical Robot Learning) and LapGym, frequently rely on physics engines to model rigid-body dynamics, contact forces, collision detection, and joint constraints[3]. PyBullet, an open-source physics engine, is highly favored in the RL community due to its Pythonic interface and robust handling of articulated bodies[6]. However, integrating a highly complex, ray-casting or GAN-based ultrasound rendering pipeline directly into a PyBullet step loop can cause catastrophic computational bottlenecks, rendering the simulation too slow for effective DRL training[1]. Making this pipeline lightweight requires profound architectural decoupling.

Decoupling Rigid-Body Dynamics and Acoustic Rendering

The primary strategy for decreasing compute is the strict separation of concerns between the kinematic simulation of the robot and the sensor simulation of the ultrasound probe. PyBullet excels at computing forward kinematics, resolving contact constraints, and applying gravity at extremely high frequencies (e.g., 240 Hz to 500 Hz)[6]. By treating the patient's torso as a static rigid-body collision mesh, PyBullet can efficiently resolve the contact dynamics, ensuring the ultrasound probe maintains physical contact with the skin surface without penetrating it[1].
However, relying solely on PyBullet's internal iterative numerical solvers for complex Inverse Kinematics (IK) during a high-speed continuous scanning task is computationally inefficient[21]. To accelerate robotic control within these environments, modern frameworks increasingly integrate fast C++ spatial algebra libraries, such as the Pinocchio library[20]. Pinocchio utilizes advanced spatial algebra and highly optimized, cache-friendly memory structures to solve the equations of motion and compute articulated-body inertia recursively, executing these calculations exponentially faster than standard iterative solvers[20].
For a surgical robot attempting to track an optimal, dynamically generated scan path, a differential inverse kinematic controller—powered by Pinocchio—converts the desired task-space task (the target position and orientation of the ultrasound probe) into the required joint velocities using quadratic programming (QP) optimization[20]. By offloading the heavy kinematic and dynamic calculations to Pinocchio, and utilizing PyBullet solely for basic contact physics and state management, significant CPU and GPU resources are liberated[20]. This reclaimed computational bandwidth is then dedicated entirely to the ultrasound rendering pipeline.

Optimized Collision Detection and Spatial Queries

A major source of computational latency in 3D reconstruction pipelines is querying the environment to determine exactly what the sensor "sees" at any given microsecond.

Continuous mesh-to-mesh collision detection and polygon intersection testing are notoriously slow, particularly when dealing with complex, high-poly anatomical models[3].
To optimize this and render the pipeline lightweight, environments must transition away from mesh intersections toward implicit surface representations, most notably Signed Distance Fields (SDFs) or Truncated Signed Distance Fields (TSDFs)[3]. An SDF represents the anatomical environment as a 3D volumetric grid where each voxel stores the minimum scalar distance to the nearest surface boundary; values are negative inside the object, positive outside, and exactly zero at the surface[8].
When the simulated ultrasound plane intersects the patient volume, the system does not calculate complex polygon intersections. Instead, it queries the SDF. Extracting the necessary 2D slice from the 3D CT volume or the label map can be highly parallelized on a GPU using optimized grid sampling functions, such as PyTorch's grid_sample[1].

Simulation Component	Traditional Approach	Optimized Lightweight Approach	Impact on Computational Efficiency
Kinematics & Dynamics	Iterative numerical IK solvers native to physics engines	Analytical spatial algebra via Pinocchio C++ library	Reduces joint trajectory and inertia computation time by orders of magnitude, enabling higher control frequencies[20].
Contact & Collision	Continuous mesh-to-mesh polygon intersection	Voxelized boundaries and Signed Distance Fields (SDF)	Eliminates redundant intersection tests; supports O(1) distance lookups for rapid spatial queries[8].
Image Slice Extraction	CPU-based memory slicing and trilinear interpolation	Batched GPU tensor interpolation (grid_sample)	Enables simultaneous extraction across hundreds of parallel environments in milliseconds[1].

Ultrasound Rendering	Full Monte Carlo acoustic path tracing	Convolution-based analytical models or lightweight GAN inference	Reduces image generation time from seconds per frame to milliseconds (e.g., 0.11s for 100 parallel environments)[1].

By batching the end-effector poses from hundreds of independent PyBullet environments into a single vectorized tensor, the corresponding 2D CT and label slices can be extracted via grid_sample and passed through the physics-based or GAN-based ultrasound simulator simultaneously[1]. This batch-processing paradigm achieves true interactive speeds, transforming a mathematically prohibitive pipeline into a highly efficient, lightweight data generator capable of sustaining the throughput required for submodular DRL training[1].

The Next-Best-View (NBV) Paradigm in 3D Reconstruction

While the physical execution of the ultrasound scan relies on optimized kinematics and physics engines, the high-level decision of *where* to move the probe to reconstruct the anatomy efficiently is governed by the principles of Next-Best-View (NBV) planning[5]. The NBV problem fundamentally seeks to identify the optimal subsequent sensor pose to maximize the acquisition of new, high-quality geometric information about an unknown or partially known 3D scene, while simultaneously minimizing resource expenditure—such as total scanning time or robotic path length[32].

Formal Utility Metrics, Information Gain, and Shannon Entropy

In an active 3D reconstruction framework, the selection of the next continuous view v_{i+1} from a vast set of candidate views V is driven by a utility function U [32]:

$$v_{i+1} = \arg\max_{v \in V} U(v|\mathcal{M}_i)$$

where $\mathcal{M}_i$ is the current state of the reconstructed model at step i. In advanced robotic ultrasound and active vision systems, utility is frequently formulated through the lens of information theory, specifically by calculating the Expected Information Gain (EIG) or by measuring the reduction in Shannon Entropy[32].

Shannon entropy mathematically quantifies the uncertainty or randomness present in a given probabilistic system[34]. If the working surgical volume is represented as a probabilistic occupancy grid (where each voxel holds a probability of containing bone tissue), completely

unobserved voxels have a probability of 0.5, representing maximum entropy and maximum uncertainty[32]. As the robotic ultrasound probe scans a region and acoustic echoes are processed, the probabilities of those voxels are updated, approaching 1 (definitively occupied by bone) or 0 (definitively free space or soft tissue). This updating process directly reduces the overall Shannon entropy of the anatomical map[32].

The objective of an NBV algorithm is to calculate the potential information gain for a set of candidate future poses. The pose that promises the maximum reduction in Shannon entropy—often determined by simulating rays cast into the current probabilistic map from the candidate pose—is selected as the target[36]. However, exhaustive volumetric ray-casting for thousands of candidate poses at every time step is computationally prohibitive, necessitating predictive, learning-based approximations for real-time robotic control[12].

Exploiting Submodularity in Active View Planning

A critical mathematical property of the NBV coverage problem—and the reason it integrates so seamlessly into the SonoGym pipeline—is *submodularity*, which formally encodes the intuitive concept of diminishing returns[12]. A set function $F : 2^{\Omega} \rightarrow \mathbb{R}$ is defined as submodular if, for every subset $A \subset B \subset \Omega$ and for every element $x \in \Omega \setminus B$:

$$F(A \cup \{x\}) - F(A) \geq F(B \cup \{x\}) - F(B)$$

In the context of 3D spinal reconstruction, adding a new ultrasound scan x to a small, incomplete set of existing scans A yields a significantly larger increase in newly observed surface area than adding that exact same scan x to a larger, more complete set of scans B [12]. Because the utility function for surface coverage is submodular and monotonically increasing, the inherently NP-hard problem of finding the absolute optimal set of sequence views can be approximated highly efficiently using a greedy maximization algorithm[12]. This mathematical property guarantees that a greedy approach will achieve a solution bounded by $1 - 1/e$ (approximately 63.2%) of the theoretical absolute optimal value, ensuring high-quality view planning without the exponential computational cost of global trajectory optimization[12]. This reality underpins the success of the submodular PPO agent in the SonoGym environment; the agent learns to iteratively select the next best view by greedily maximizing the marginal gain in surface coverage at each step, naturally avoiding redundant scans[1].

Identifying Best Views: Acoustic Constraints and Ultrasound Confidence

A major distinction between general 3D reconstruction and ultrasound-specific reconstruction is the physics of the sensor. Most NBV literature focuses on optical sensors (e.g., RGB-D

cameras or LiDAR), where visibility is determined primarily by simple line-of-sight occlusions and depth limits[12]. Robotic ultrasound introduces highly specific, non-linear physical constraints that drastically alter the definition of an "optimal" view[4].

Acoustic Physics: Incidence Angles and Specular Reflection

The extraction of bone surfaces from ultrasound images is profoundly influenced by the behavior of high-frequency acoustic longitudinal waves interacting with rigid, high-density structures[15]. When an ultrasound pulse propagating through soft tissue encounters a tissue-bone interface, the massive difference in acoustic impedance causes a powerful reflection of energy back toward the transducer, while the remaining energy is heavily attenuated[15]. This attenuation creates a characteristic dark "acoustic shadow" beneath the bone, which is a primary feature used by segmentation algorithms to identify the bone surface[9].

However, the surface of a bone acts as a specular reflector[15]. The angle of reflection equals the angle of incidence. If the ultrasound beam is not near-perpendicular (orthogonal) to the local bone surface, the reflected echo scatters away from the transducer array[10]. This results in a severe loss of signal, rendering the bone invisible or drastically reducing the image contrast[15]. Furthermore, because the ultrasound beam possesses a finite elevational thickness (the slice thickness), imaging highly inclined surfaces causes the bone boundary to blur within the 2D plane, merging the sharp interface into a thick, ambiguous response that degrades the 3D reconstruction accuracy[15].

Consequently, an optimal scan view in robotic ultrasound reconstruction is not merely one that "looks" at an unobserved region. The absolute best view requires an acoustic window that is clear of obstructions (e.g., avoiding the ribs or transverse processes to image the vertebral body) and demands that the transducer's beam remains strictly orthogonal to the target surface normal[15].

Integration of Ultrasound Confidence Maps

To mathematically penalize suboptimal incidence angles and guide the NBV planner toward high-fidelity views, the reconstruction pipeline integrates the concept of the *Ultrasound Confidence Map*[40]. A confidence map is a per-pixel probability distribution that models the reliability of the ultrasound signal based on the underlying physics of wave attenuation, reverberation, and shadow formation[4].

By running real-time signal analysis, or utilizing pre-trained Convolutional Neural Networks (CNNs) to predict confidence directly from the B-mode images, the robotic system can evaluate the quality of the current acoustic window on the fly[4]. If a candidate NBV pose intersects the target bone, but the predicted confidence map indicates heavy attenuation—due to an oblique incidence angle or an intervening anatomical obstacle—the utility score of that view is severely down-weighted[40].

In advanced autonomous systems like SonoGym, DRL policies learn to implicitly map these confidence metrics to their continuous action outputs. The agent learns to dynamically adjust the probe's pitch and yaw to maximize the intensity of the bone response and the sharpness of the underlying shadow, ensuring that only geometrically accurate points are contributed to the

overall 3D reconstruction volume[1].

Evaluating Process Efficiency in 3D Reconstruction

Evaluating the efficiency of the 3D reconstruction process requires robust methodologies to quantify the trade-off between the quality of the generated anatomical model and the computational and physical resources expended (such as trajectory length, joint wear, and procedural time).

Metrics for Quality and Efficiency

The performance of an active 3D reconstruction pipeline is typically evaluated using a combination of geometric and operational metrics. Geometrically, the *Chamfer Distance* is widely used to measure the average shortest distance between the points in the reconstructed point cloud and the ground truth mesh (derived from CT)[5]. A lower Chamfer distance indicates higher geometric fidelity. Additionally, the *Coverage Ratio*—the percentage of the total surface area that has been successfully reconstructed—is critical for assessing completeness[1].

Operationally, the efficiency is measured by the total path length of the robotic end-effector and the cumulative angular rotation required to execute the scan[1]. A highly efficient process achieves a maximal coverage ratio and minimal Chamfer distance while keeping the path length as short as possible.

Representations for Lightweight Utility Prediction

To predict which views will be most efficient in real-time, the system must rapidly process the current state of the reconstruction. Voxel grids and TSDFs are highly accurate for representing surfaces and allow for exact information gain computation via ray-casting[27]. However, maintaining, updating, and querying dense, high-resolution 3D volumetric grids scales poorly, presenting significant memory and computational bottlenecks for real-time robotic control at high frequencies[27].

Conversely, unstructured Point Clouds offer a drastically lighter memory footprint, representing only the explicitly observed surface boundaries in 3D space[32]. To predict the optimal view directly from a point cloud, deep learning architectures tailored for unstructured data are highly effective. While traditional 3D CNNs require the point cloud to be voxelized—introducing sparsity issues and immense memory bloat—architectures like PointNet operate directly on the raw point coordinates[38].

PointNet leverages shared multi-layer perceptrons (MLPs) and max-pooling symmetric functions to extract global features from the point cloud, rendering it invariant to permutations of the input data[38]. By utilizing PointNet or subsequent iterations (e.g., PointNet++), the active vision system can encode the current, incomplete state of the bone reconstruction into a compact latent vector[32]. This vector is then fed directly into the DRL policy network to regress the next best view, entirely bypassing the computationally prohibitive step of ray-casting[32]. The transition from heavy 3D CNNs to lightweight PointNet encoders is a critical factor in achieving the low-latency view prediction required for continuous, real-time robotic motion[52].

Network Efficiency and Edge Deployment

For such an active perception pipeline to function in a real-world clinical setting, the inference

of the underlying neural networks (for both bone segmentation and confidence mapping) must occur at high frames per second (FPS). This ensures the robot can respond to tissue deformation and respiratory motion instantaneously[57].

While heavy models like a full 3D U-Net provide exceptional volumetric segmentation accuracy, they are generally too slow for real-time, closed-loop robotic control, often requiring sub-volume partitioning or offline processing[49]. To achieve the necessary processing speed for a lightweight pipeline, highly optimized architectures such as MobileNet or GhostNet variants are preferred[58]. Incorporating depthwise separable convolutions or Ghost modules allows the system to process the incoming B-mode ultrasound feed, extract the bone surface boundary, and update the global point cloud in milliseconds[58]. This leaves sufficient computational bandwidth for the PPO agent and the Pinocchio IK solver to compute the next kinematic action seamlessly.

Determining the Minimum Number of Scans for Complete Construction

Determining the theoretical minimum number of scans required to reconstruct a complex anatomical geometry—such as a lumbar vertebra—is mathematically equivalent to solving a variant of the Set Cover Problem or the Orienteering Problem under strict budget constraints[12]. In a discrete formulation, if the target surface requires a sampling density that covers 95% of its area, the algorithm generates a finite set of candidate poses. By evaluating the submodular coverage function, the algorithm selects the minimal subset of poses whose combined field of view satisfies the coverage threshold[12]. However, in continuous robotic ultrasonography, the probe captures images sequentially as it translates dynamically across the skin[1]. Therefore, minimizing the number of scans equates directly to minimizing the *trajectory length* while maximizing the accumulation of novel information[1].

The Inefficiency of Heuristic Scanning

The SonoGym submodular DRL experiments reveal profound insights into this efficiency optimization. Traditional, heuristic approaches to spinal scanning rely on predetermined, open-loop raster patterns—for example, moving the probe in strict, grid-like linear sweeps over the patient's back[1]. These manual or semi-automated methods are highly inefficient. They result in vast amounts of redundant data acquisition over relatively flat surfaces, and conversely, they frequently miss the steep lateral facets of the vertebra entirely due to poor, unoptimized incidence angles[1]. Consequently, heuristic methods require a massive number of overlapping scan lines to guarantee coverage, significantly prolonging the procedure and increasing the minimum number of required scans[33].

The Emergence of Circular and Adaptive Trajectories

When trained to explicitly optimize the submodular objective function, the PPO agent in SonoGym completely abandons linear sweeps. Instead, it discovers a highly efficient *circular scanning pattern* localized directly above the target vertebra[1].

Scanning Strategy	Path Geometry & Execution	Incidence Angle Optimization	Coverage Efficiency & Redundancy	Minimization of Scans
Heuristic (Open-loop)	Linear raster grids (X/Y sweeps); predefined prior to scan[1].	Fixed, generally perpendicular to the skin surface. Fails on steep bone facets[1].	Low. Generates massive redundancy in flat areas; fails to capture steep slopes[1].	Requires a high density of overlapping scan lines to ensure coverage, resulting in longer procedures and excess data[33].
Active Submodular DRL	Circular, localized continuous sweeps; dynamically adaptive[1].	Dynamically adjusts pitch and yaw on the fly to remain orthogonal to the curved bone surface[1].	High. Specifically targets frontier regions lacking data points, avoiding over-scanned areas[1].	Achieves target coverage threshold with significantly shorter physical path length and the absolute minimum number of total frames[1].

The learned circular trajectory demonstrates that the optimal sequence of views for reconstructing a complex, multi-faceted object like a vertebra is obtained by orbiting the anatomical center while continuously rolling and pitching the transducer to maintain acoustic orthogonality[1]. By dynamically tracking the "frontier" of the point cloud—the boundary between observed and unobserved regions—the agent mathematically avoids capturing frames that offer zero marginal information gain[32].

Consequently, the absolute minimum number of scans is achieved when the trajectory is dynamically planned such that every acquired frame overlaps the previous frame by only the minimum margin required for accurate image registration (such as Iterative Closest Point or feature-based matching), with the remaining Field of View dedicated entirely to novel, unobserved geometry[17]. This intelligent overlap threshold, combined with optimal incidence angles guaranteed by confidence mapping, ensures the 3D model is constructed with maximum efficiency.

Conclusion

The integration of advanced robotic manipulators with real-time ultrasound imaging presents a transformative leap for orthopedic and spinal surgeries, moving the field from heuristic, operator-dependent routines to intelligent, autonomous anatomical reconstruction. The SonoGym platform exemplifies the cutting edge of this transition, providing the highly scalable, robust simulation environment necessary for training the next generation of Deep Reinforcement Learning and Imitation Learning policies.

Through a comprehensive analysis of the reconstruction pipeline, several critical insights emerge regarding computational efficiency and active view optimization. First, decoupling the rigid-body contact dynamics from the acoustic image rendering is essential. By utilizing PyBullet for state management and collision detection via Signed Distance Fields, and augmenting it with rapid spatial algebra libraries like Pinocchio for inverse kinematics, the environment becomes lightweight enough to support massive parallelization without CPU bottlenecking.

Secondly, optimizing the reconstruction process fundamentally relies on solving the Next-Best-View problem under strict acoustic constraints. Ultrasound is physically unforgiving; the specular nature of bone means that utility cannot be measured by geometric visibility alone, but must factor in precise incidence angles and acoustic shadowing via Ultrasound Confidence Maps.

Finally, by formulating the reconstruction task as a submodular Markov Decision Process, DRL agents successfully map the mathematics of diminishing returns. The result is a paradigm shift in physical execution: replacing inefficient linear raster scans with dynamic, circular, and highly articulated probe trajectories. By utilizing lightweight point cloud encoders like PointNet and high-speed segmentation networks to evaluate the state of the reconstruction, the system consistently predicts the optimal view, effectively minimizing the total number of required scans while maximizing surface fidelity. As these simulated pipelines transition to clinical hardware, the synthesis of active view planning, lightweight physics integration, and acoustic-aware deep learning will form the foundation of fully autonomous robotic diagnostic and surgical systems.

Works cited

1. SonoGym.pdf
2. A Review and Perspective of Techniques for Autonomous Robotic, https://pmc.ncbi.nlm.nih.gov/articles/PMC13074658/
3. Dissertation, https://opendata.uni-halle.de/bitstream/1981185920/119101/1/Allgaier_Mareen_Dissertation_2024.pdf
4. An Overview of Systems and Techniques for Autonomous Robotic, https://www.researchgate.net/publication/350778947_An_Overview_of_Systems_and_Techniques_for_Autonomous_Robotic_Ultrasound_Acquisitions
5. Learning Active Intraoral Scanning for Robotic Dental Reconstruction, https://arxiv.org/html/2606.26955v1

6. SurRoL: An Open-source Reinforcement Learning Centered and, https://www.researchgate.net/publication/354236951_SurRoL_An_Open-source_Reinforcement_Learning_Centered_and_dVRK_Compatible_Platform_for_Surgical_Robot_Learning
7. SurgVLA-Bench: Towards Evaluating Vision-Language-Action, https://arxiv.org/html/2606.29247v1
8. Efficient Physically-based Simulation of Soft Bodies in Embodied, https://arxiv.org/html/2402.01181v1
9. Ultrasound Confidence Maps and Applications in Medical Image, https://mediatum.ub.tum.de/doc/1129526/827108.pdf
10. SonoGym: High Performance Simulation for Challenging Surgical, https://arxiv.org/html/2507.01152v1
11. arXiv:2101.09397v1 [cs.CV] 23 Jan 2021 - Cimat.mx, https://www.cimat.mx/~murrieta/Papersonline/VaetaI_MVA2021.pdf
12. Submodular Trajectory Optimization for Aerial 3D Scanning, https://openaccess.thecvf.com/content_ICCV_2017/papers/Roberts_Submodular_Trajectory_Optimization_ICCV_2017_paper.pdf
13. Isaac Lab: A GPU-Accelerated Simulation Framework for Multi, https://d1qx31qr3h6wln.cloudfront.net/publications/Isaac%20Lab%2C%20A%20GPU-Accelerated%20Simulation%20Framework%20for%20Multi-Modal%20Robot%20Learning.pdf
14. Multi-robot next-best-view planning | Mikko Lauri, https://laurimi.github.io/research/2020/07/04/multirobot-nbv.html
15. Ultrasound imaging and segmentation of bone surfaces: A review, https://pmc.ncbi.nlm.nih.gov/articles/PMC5551972/
16. Human-in-the-loop Embodied Intelligence with Interactive ... - alphaXiv, https://www.alphaxiv.org/abs/2301.00452
17. Autonomous Robot-Driven Chronic Wound 3D Reconstruction and, https://www.mdpi.com/2218-6581/14/3/30
18. SurRoL: An Open-source Reinforcement Learning Centered and, https://www.researchgate.net/publication/357100070_SurRoL_An_Open-source_Reinforcement_Learning_Centered_and_dVRK_Compatible_Platform_for_Surgical_Robot_Learning
19. UltraStar: Semantic-Aware Star Graph Modeling for, https://papers.miccai.org/miccai-2026/1099-Paper1157.html
20. Whole-Body Control for a Torque-Controlled Legged Mobile ... - MDPI, https://www.mdpi.com/2076-0825/11/11/304
21. Isaac Lab: A GPU-Accelerated Simulation Framework for Multi, https://arxiv.org/html/2511.04831v1
22. Recursive Forward Dynamics for Serial Kinematic Chains using, https://calinon.ch/papers/Loew-WAFR2024.pdf
23. Bridging the Simulation to Reality Gap in Robotics, https://repository.essex.ac.uk/40339/1/KVasios-PhD.pdf
24. (PDF) The Open Motion Planning Library - ResearchGate, https://www.researchgate.net/publication/260691259_The_Open_Motion_Plannin

g_Library
25. 3D Reconstruction Techniques in the Manufacturing Domain - arXiv, https://arxiv.org/html/2604.28064v1
26. (PDF) A Comprehensive Review of Vision-Based 3D Reconstruction, https://www.researchgate.net/publication/379620229_A_Comprehensive_Review_of_Vision-Based_3D_Reconstruction_Methods
27. Video-Based 3D Reconstruction: A Review of Photogrammetry and, https://www.mdpi.com/2313-433X/12/3/128
28. Automatic Device Selection - OpenVINO™ documentation, https://docs.openvino.ai/2023.3/openvino_docs_OV_UG_supported_plugins_AUTO.html
29. (PDF) The Robust Vessel Segmentation and Centerline Extraction, https://www.researchgate.net/publication/393078391_The_robust_vessel_segmentation_and_centerline_extraction_One-stage_deep_learning_approach
30. Reproducibility of Open-Source Virtual Surgery Frameworks, https://www.preprints.org/manuscript/202606.1100
31. Applications of Reinforcement Learning for Autonomous Surgical, https://www.mdpi.com/2313-7673/11/8/577
32. Next Best View Problem - Emergent Mind, https://www.emergentmind.com/topics/next-best-view-problem
33. (PDF) An Adaptive Hierarchical Next-Best-View Algorithm for 3D, https://www.researchgate.net/publication/242656933_An_Adaptive_Hierarchical_Next-Best-View_Algorithm_for_3D_Reconstruction_of_Indoor_Scenes
34. Information Entropy Measure for Evaluation of Image Quality, https://www.researchgate.net/publication/37364693_Information_Entropy_Measure_for_Evaluation_of_Image_Quality
35. ENRICHing medical imaging training sets enables more efficient, https://pmc.ncbi.nlm.nih.gov/articles/PMC10198519/
36. Advanced Ultrasound Imaging Techniques for Computer Assisted, https://mediatum.ub.tum.de/doc/1345760/431097.pdf
37. The Evolution of Image Segmentation from Classical Techniques to, https://www.mdpi.com/2218-6581/15/9/169
38. Advances in Set Function Learning: A Survey of Techniques and, https://udspace.udel.edu/bitstreams/1866bd0d-62ec-48eb-8de4-5280e2c67f9c/download
39. Computer vision-based livestock weight measurement: principles, https://www.oaepublish.com/articles/ir.2026.11
40. Acoustic window planning for ultrasound acquisition - ResearchGate, https://www.researchgate.net/publication/314865723_Acoustic_window_planning_for_ultrasound_acquisition
41. Manual of diagnostic ultrasound, https://apimr.pt/wp-content/uploads/2016/10/WHO-Manual_Diagnostic_US-_Vol1.pdf
42. Quality-Driven Control of a Robotized Ultrasound Probe - mediaTUM, https://mediatum.ub.tum.de/doc/1328679/426272.pdf

43. Musculoskeletal Ultrasound: How, Why and When 9780702081989, https://dokumen.pub/musculoskeletal-ultrasound-how-why-and-when-9780702081989.html
44. (Mebooksfree Net) Dia&ult&phy&equ&hos&mar&thr&3rd PDF - Scribd, https://www.scribd.com/document/441592261/mebooksfree-net-dia-ult-phy-equ-hos-mar-thr-3rd-pdf
45. Manual of diagnostic ultrasound, https://biomedical.gov.al/wp-content/uploads/2025/12/a.pdf
46. UltraBones100k: A reliable automated labeling method and large, https://arxiv.org/pdf/2502.03783?
47. AI-Driven Robotic Cardiac Ultrasound | PDF | Echocardiography, https://www.scribd.com/document/940280333/Echo-Robot-Semi-Autonomous-Cardiac-Ultrasound-Image-Acquisition-Using-AI-and-Robotics
48. Automatic bone detection and soft tissue aware ultrasound-CT, https://d.docksci.com/automatic-bone-detection-and-soft-tissue-aware-ultrasound-ct-registration-for-co_5a51c3d7d64ab278fc037e67.html
49. Methods for the acquisition, learning-based segmentation, and, https://mediatum.ub.tum.de/doc/1734676/1734676.pdf
50. Active 3D Reconstruction Systems - Emergent Mind, https://www.emergentmind.com/topics/active-3d-reconstruction-framework
51. The Comparison of Selected Approaches to 3D Reconstruction of, https://www.mdpi.com/2079-9292/15/9/1812
52. Deep Learning for 3D Point Clouds: A Survey - IEEE Computer Society, https://www.computer.org/csdl/journal/tp/2021/12/09127813/1l3udiVJLpe
53. A Lightweight Network for Point Cloud Analysis via the Fusion of, https://www.mdpi.com/1424-8220/22/13/4742
54. Semantic Segmentation of Key Categories in Transmission Line, https://www.mdpi.com/2079-9292/14/4/650
55. Deep Learning on Point Sets for 3D Classification and Segmentation, https://www.researchgate.net/publication/320971844_PointNet_Deep_Learning_on_Point_Sets_for_3D_Classification_and_Segmentation
56. Track: Poster Session 8 - NeurIPS 2026, https://neurips.cc/virtual/2021/session/44798
57. Review of robotic systems for thoracoabdominal puncture, https://pubs.aip.org/aip/apb/article/8/2/021501/3280369/Review-of-robotic-systems-for-thoracoabdominal
58. Computer Vision - Model Zoo, https://modelzoo.co/category/computer-vision
59. Edge-Ready Multimodal and Geometry-Aware Perception for UAV, https://ieeexplore.ieee.org/iel8/6287639/11323511/11646582.pdf
60. Deep Learning Paradigm and Its Bias for Coronary Artery Wall, https://www.mdpi.com/2308-3425/10/12/485
61. (PDF) Robotic Arm Based Automatic Ultrasound Scanning for Three, https://www.researchgate.net/publication/327898895_Robotic_Arm_Based_Automatic_Ultrasound_Scanning_for_Three-Dimensional_Imaging