



Line Frequency X-ray Machine

Certified and Approved By:





S. No	Parameter	Specified value
1	Mains Voltage	230VAC, 50Hz, Single phase
2	Power output	8.0 kw
3	Output frequency	Two pulse, Full wave rectified
4	Kvp range	40 – 100 kvp in 15 steps
5	MA range	25 – 100 mA
6	MAs range	0.5 – 300 mAs
7	X-ray tube	Stationary anode
8	Focal Spot Nominal	2.6mm
9	Target angel	15°
10	Collimator	Manual
11	Exposure Release Switch	2 Steps
12	Digital Display	Kv, MA and MAs
13	Timer	Micro Controller Based
14	Protection	Over voltage Overload

Safety Standard : BIS IS7620: Part-1
Regulatory Requirement : AERB type approval



Innovation Meditech Pvt. Ltd.

H.O. and Works:

B-6, Sector-88, Phase 2, NOIDA-201305, UP, India
Phone: 0120-499-6085

Toll Free No - 18002026276

Email: sales@innovationmeditech.com

Web: www.innovationmeditech.com



**BRANCH OFFICE & RESIDENT: JAIPUR • LUCKNOW • PATNA • HYDERABAD
BHOPAL • HISAR • KOLKATA**