

Simoa blood biomarkers tested on Insight46 phase 1 plasma and serum samples from the LWBL/UKDRI biobank

Ashvini Keshavan | 11/07/2019

Contents

Document details	2
Methods	2
Variables in this document	3

Document details

Categories of variables	Simoa blood biomarkers tested on phase 1 plasma and serum samples from the LWBL/UKDRI biobank
Summary of work undertaken	Source data transferred to Stata file Missing data – labelled as -99. Methods section below
Names of source variables	SerumNFL, Tauplasma, Ab42plasma, Ab40plasma, Ab4240ratioplasma, plasma_available, serum_available
Output variables	serumnfl_i46p1, tauplasma_i46p1, ab42plasma_i46p1, ab40plasma_i46p1, ab4240ratioplasma_i46p1, plasma_available_i46p1, serum_available_i46p1
Papers using these variables	1. Blood and cerebrospinal fluid biomarkers in Alzheimer's disease – from clinical to preclinical cohorts. PhD thesis (Ashvini Keshavan). 2. Plasma amyloid, tau and serum neurofilament light chain in Insight 46 – associations with cognition and brain imaging. AAIC 2019 poster (Keshavan et al.) 3. The Insight 46 Cohort – Relationships with Cerebral PET Amyloid positivity Using Two Independent Methods for Plasma Amyloid Beta Measurement. AAIC 2019 Featured Research Session Oral presentation. (Schott et al.)

Methods

Taken from reference 1 above:

Serum was chosen for NFL analysis and plasma for analysis of the other three biomarkers (total tau, A β 40 and A β 42) as per the recommendations from our prior publication detailing plasma vs serum comparisons of quantifications of Simoa blood biomarkers in other samples (Keshavan et al. 2018, Alzheimer's and Dementia, Diagnosis, Assessment and Disease monitoring).

For serum NFL analysis: a single 500 μ L aliquot of serum for each individual was thawed directly to room temperature over 1 hour and vortexed for 2 seconds to ensure thorough mixing. 200 μ L was pipetted into a 1.5 ml polypropylene centrifuge tube and centrifuged at 13 000 g for 10 minutes, as per the kit manufacturer's recommendation; the remaining 300 μ L was replaced into -80 °C in the original cryovial. After the 200 μ L was centrifuged, 130 μ L of the supernatant was pipetted onto the plate for analysis in duplicate. If the coefficient of variation (CV) across the duplicates was >15% or no value was returned for either, the procedure above was repeated at a later date, employing one additional freeze-thaw cycle by starting with the 300 μ L volume that was in the original cryovial. At this point all 500 individuals who had blood sampling had a serum NFL value quantified with a CV <15%.

For plasma A β 40 and A β 42 analysis: a single 500 μ L aliquot of plasma for each individual was thawed directly to room temperature over 1 hour and vortexed for 2 seconds. 300 μ L was pipetted into a 1.5 ml polypropylene centrifuge tube and centrifuged at 13 000 g for 10 minutes as per the kit manufacturer's recommendation; the remaining 200 μ L was replaced into -80 °C in the original cryovial. After the 300 μ L was centrifuged, 100 μ L of the supernatant was pipetted onto each of two plates for analysis in duplicate, capitalising on the ability to load two different reagent kits at a time on the HD-1 analyser. When plates of samples for analysis of A β 40 and A β 42 were prepared in this way, the plate containing samples for A β 40 was always analysed first, and that containing samples for A β 42 was analysed second. The CV across the duplicates was <15% for all samples assayed for A β 40 but for some samples assayed for A β 42 the CV was >15% or no value was returned. In this case the procedure above was repeated at a later date, using a fresh 500 μ L aliquot of plasma and pipetting out and centrifuging 200 μ L then pipetting 100 μ L of the supernatant onto the plate for A β 42 analysis.

For plasma t-tau analysis: if only one aliquot of plasma had been used for A β 42, the second (fresh) 500 μ L aliquot of plasma for each individual was used for assaying total tau and the steps undertaken to prepare the sample thereafter were identical to those described above for serum NFL. However, if the A β 42 assay was

being repeated on the second aliquot, then on the same thaw of this sample, after vortexing for 2 seconds, 200 μ L was pipetted into a separate 1.5 ml polypropylene centrifuge tube and used for plasma t-tau analysis in parallel. In this situation, the plate of samples for analysis of A β 42 was analysed first and the plate for analysis of t-tau second. If the CV across duplicates was >15% on the first analysis of t-tau, the analysis was repeated at a later date, employing one additional freeze-thaw cycle by starting with the 300 μ L volume that was in the original cryovial.

By this method, all samples analysed for plasma A β 40 and A β 42 underwent one freeze-thaw cycle; 59 samples analysed for plasma t-tau and 72 samples analysed for serum NFL underwent two freeze-thaw cycles. This procedure was deemed to be acceptable based on the evidence that up to four freeze-thaw cycles do not affect measured plasma t-tau or serum NFL concentrations for these assays (Keshavan et al. 2018, Alzheimer's and Dementia, Diagnosis, Assessment and Disease monitoring).

For each blood biomarker, all aliquots were assayed using the same batch of reagents excepting the last plate of samples, which were those requiring repeat analysis. Every plate was analysed according to its own calibrators (made from the stock solution provided in the kit), and included two run validation controls also made from this stock. The measured values of the two controls used in each case had an inter-plate CV of <30% so all plates were deemed acceptable for inclusion in the statistical analysis. Details of the commercially available Simoa total tau 2.0, NF-light, A β 40 and A β 42 assays are given in the kit inserts and they were all used as per the manufacturer's instructions. All measurements were conducted by the same individual (Dr Ashvini Keshavan) on the same automated HD-1 analyser (Quanterix) at the UK DRI laboratory at UCL.

Where no aliquot of plasma or serum was available in the biobank for a given individual (due to failure of blood sampling or incomplete blood sampling) the variables plasma_available and serum_available indicate this. For the blood biomarker variables, the code -99 indicates:

- 1 in the case where an aliquot was available: the coefficient of variation across the duplicates was above 15% and therefore the quantification was unreliable.
- 2 in the case where an aliquot was not available: no quantification was performed.

It is recommended that prior to making a request for any of these variables, researchers discuss the plan with Dr Ashvini Keshavan.

Variables in this document

7 high-confidence matches

Variable	Label	How it was found
Sub-Studies [57] › Insight46 [699] › Biological samples [6037] › Blood biomarkers [60372]		
ab40plasma_i46p1	Simoa Plasma amyloid beta 40 (pg/ml) - Insight46 phase 1	topsheet field
ab4240ratioplasma_i46p1	Simoa Plasma amyloid beta 42 to amyloid beta 40 ratio - Insight46 phase 1	topsheet field
ab42plasma_i46p1	Simoa Plasma amyloid beta 42 (pg/ml) - Insight46 phase 1	topsheet field
plasma_available_i46p1	Plasma available for Simoa analyses in Phase 1 - Insight46 phase 1	topsheet field
serum_available_i46p1	Serum available for Simoa analyses in Phase 1 - Insight46 phase 1	topsheet field
serumnfl_i46p1	Simoa Serum NFL (pg/ml) - Insight46 phase 1	topsheet field
tauplasma_i46p1	Simoa Plasma total tau (pg/ml) - Insight46 phase 1	topsheet field

1 medium-confidence match

Variable	Label	How it was found
Questionnaire data [55] › Health and medical history [510] › Medical conditions [5101] › Other medical conditions [51011]		

Variable	Label	How it was found
blood	Have you had anaemia in the last 10 years - at age 43 years	prose scan

1 low-confidence match

Variable	Label	How it was found
Questionnaire data [55] › Sociodemographics [515] › Household [5152]		
own	Now I would like to get some general information about your household. Does your household own or rent this accommodation? - at age 53 years	prose scan