

Determination of Optimized Parameters for
the Separation of PAHs using Gas
Chromatography
CHEM 4303
Analytical Separations

Robby RENZ

October 15, 2018

Date Performed: September 11, 2018
Date Completed: October 2, 2018
Partner: Jaya Roe
Lab Instructor: Kevin Stroski

Abstract

The main objective of this experiment was to modify the parameters of a GC-FID in order to efficiently separate a mixture of PAHs. This was done with the help of a gas chromatography equipped with a flame ionization detector [1]. Also, check out the this citation [2].

This is working!!! I am the King!!! I will live forever!!!

1 Introduction

Separation is important and stuff.

The main objective of this experiment is to separate polynuclear aromatic hydrocarbons (PAHs) with the help of a high resolution gas chromatography (GC) that is equipped with a flame ionization detector. The initial part of the experiment involves configuring the GC in order to get it ready for PAH analysis. In addition, the resolution of the GC was also optimized in order to get the best resolution possible for the resulting chromatogram. This part involves altering the temperature of the detector, injector and oven. Furthermore, solute focusing methods, like for instance, solvent effects or cold trapping, can be taken advantage of in order to increase the resolution of the chromatogram. These methods involve the stationary phase "trapping" the solutes while the solvent evaporates through the column. And when the temperature is high enough, the solutes are released and are free to be carried down the column by the carrier gas.

In a typical gas chromatography (GC), the mobile phase, in which He is most commonly used as the carrier gas, transports the gaseous analyte throughout the column. Furthermore, due to partitioning, the individual components are separated because of its interaction with the stationary phase, which can be either a solid or a nonvolatile liquid which is bonded on the inside of the column itself.

One of the detectors that is available for the GC is called the flame ionization detector (FID). The type of detector works best when the carrier gas is N_2 as it greatly decreases its detection limit, compared to He. In addition, the FID is sensitive to most hydrocarbons, making it the ideal detector to use where PAHs are concerned.

PAHs are their own sub-group consisting of three or more fused and unsubstituted aromatic rings, although naphthalene is also considered a PAH, even though it has only two fused rings. Despite the fact that PAHs are present in plant matter and in petroleum, they are also considered contaminants. These class of compounds are also formed during the incomplete combustion of certain organic compounds, as well as in forest fires. Thus, there is a great importance of tracking, measuring and separating these concentrations of PAHs.

The last part of this experiment involves running simulations to get the best possible chromatogram. This is done with the help of a software called GC-SOS[®], and by altering the temperature conditions. The retention times, peak amplitudes and peak widths are monitored until a satisfying chromatogram is sim-

ulated, and the best overall program has been determined.

A useful equation to find the resolution with the help of widths at the bases of the peaks is given by:

$$R = \frac{21't_r}{w_{b_1} + w_{b_2}} \quad (1)$$

where t_r denotes the retention time of the corresponding peak.

$$N = 5.545 \left(\frac{t_r}{w_{\frac{1}{2}}} \right)^2 = 16 \left(\frac{t_r}{w_b} \right)^2 \quad (2)$$

In equation 5, N represents the number of plates, theoretically speaking; $w_{\frac{1}{2}}$ and w_b are the widths of the peak at half-height and at the base, respectively. As a result, equation 5 ultimately simplifies to the following equation:

$$w_b = 1.699w_{\frac{1}{2}} \quad (3)$$

Futhermore, equation 3 proves that there exists a direct relationship between $w_{\frac{1}{2}}$ and w_b . Appropriately, the resolution can also be calculated by the half-heights of the widths of the peaks like so:

$$R = \frac{21't_r}{1.699 \left(w_{\frac{1}{2}}^1 + w_{\frac{1}{2}}^2 \right)} \quad (4)$$

However, equation 4 can only be used for peaks of similar heights.

2 Objectives

First Objective

Optimize and assess vital GC parameters for the separation of PAHs with the help of high resolution gas chromatography.

Second Objective

Finding out the plate height as well as the number of theoretical plates for peak 7 in the chromatogram of PAH mixture.

Third Objective

Determining the resolution of peaks 5 and 6 in the chromatogram of PAH mixture

3 Methods and Instrumentation

The analysis of the PAH mixture was performed on an Agilent 7890A GC, which was equipped with an FID. Before each analysis, the apparatus was left on standby so that the flame was able to reach the desired temperature for that particular analysis. The separation was performed on an Agilent 122-5032 J&W DB-5 capillary column, 30 m length \times 0.25 mm i.d. \times 0.25 μ thickness, 7 inch cage. In addition, each analysis was conducted with the injection mode at splitless and the carrier gas was He?

For the first method, named ?, the oven temperature was initially set at 110°C for 0.5 min, increased by 7.5 °C all the way to 300 °C; it was held at this temperature for 40 minutes, and the total runtime for this method was 65.83 min. The results are shown in Figure ? and the integration of each peak is shown in Table ?

4 Results and Discussion

As shown in Figure ?

In conclusion, blah blah blah . . .

5 References

References

- (1) Smith, J. M.; Jones, A. B., *Chemistry*, 7th; Publisher: 2012.
- (2) Nam, J. J.; Sweetman, A. J.; Jones, K. C. *J. Environ. Monit.* **2009**, *11*, 45–48.

6 Appendix

6.1 Calculations

Mass required to make 10 mL of 1 $\frac{\text{mg}}{\text{mL}}$ (1000 ppm) naphthalene in toluene
Calculation is here.

Determining the plate height for peak 7

Peak 7 was identified as fluoranthene.

Determining the number of theoretical plates for peak 7

Peak 7 was identified as fluoranthene.

$$N = 5.545 \left(\frac{t_r}{w_{\frac{1}{2}}} \right)^2 = 16 \left(\frac{t_r}{w_b} \right)^2 \quad (5)$$

In equation 5, N represents the number of theoretical plates; $w_{\frac{1}{2}}$ and w_b are the widths of the peak at half-height and at the base, respectively.

Determining the resolution for peaks 5 and 6

Peaks 5 and 6 were identified as phenanthrene and anthracene.

$$R = \frac{21't_r}{w_{b_1} + w_{b_2}} \quad (6)$$

t_r in equation 6 denotes the retention time of the corresponding peak.

6.2 Chromatograms

There are ? pages of chromatograms that were collected during this experiment.