

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT
MRI-LINACs
Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

Perelman School of Medicine

University of Pennsylvania, Philadelphia, PA

Presented by **Lekan Molu** (Lay-con Moh-lu)

March 10, 2021

Funding Sources

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References



CANCER PREVENTION & RESEARCH
INSTITUTE OF TEXAS

Talk Outline

Automating
Treatment
Planning in
Radiation
Therapy

Olaelekan
Ogunmolu

BOO

MCTS
BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT
MRI-LINACs
Innovation

iDG

Robustness issues
Innovation
iDG Results

Futures

References

- Beam Orientation Optimization (BOO)
 - Monte Carlo Tree Search and Neuro-Dynamic Programming for BOO
 - Column Generation as Pretraining for MCTS for BOO
- Patient Head Motion Correction in External Beam Radiation Therapy (RT)
 - Intensity-Modulated RT (IMRT): Earlier PhD Work
 - Magnetic Resonance Imaging and Linear Accelerator Systems (MRI-LINACs)
- Robustness Margins and Robust Deep Policies for Nonlinear Control

Research Significance

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



IMRT Treatment Planning (Beam Delivery)

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO
MCTS
BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT
MRI-LINACs
Innovation


iDG
Robustness issues
Innovation
iDG Results

Futures

References



The Australian Synchrotron.



A Multi-leaf collimator, ©Varian.

Radiation Delivery Couch and Gantry

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References



Varian's TrueBeam Radiotherapy System.

Part I.A: Beam Orientation Optimization (BOO)

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

■ Beam Orientation Optimization (BOO)

- Monte Carlo Tree Search and Neuro-Dynamic Programming

Contributions

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO
MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs
Innovation

iDG

Robustness issues
Innovation
iDG Results

Futures

References

Relevant Publications

Ogunmolu, Olalekan, Michael Folkerts, Dan Nguyen, Nicholas Gans, and Steve Jiang. "Deep BOO! Automating Beam Orientation Optimization in Radiation Therapy." In *Algorithm Foundations of Robotics XIII*, Merida, Mexico. Published in *Springer's Proceedings in Advanced Robotics (SPAR) Book*, 2020.

- A sparse tree lookout strategy for games with large state spaces guides transition between beam angle sets
- Tree lookout strategy guided by a deep neural network policy

Prostate Cancer Example

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues
Innovation
iDG Results

Futures

References



Stage I Prostate Cancer




Found by: Needle biopsy

Grade Group: 1

PSA level: Less than 10

OR



Found by: Digital rectal exam

Grade Group: 1

PSA level: Less than 10

Cancer in: 1/2 or less of
one side

© 2018 Terese Winslow LLC
U.S. Govt. has certain rights.

©2018 Terese Winslow LLC, U.S. Govt. has certain rights.

BOO Process: Fluence Map Optimization

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation


iDG Results

Futures


References



Prostate CT slice



Prostate before
BOO



Fluence Map

BOO Workflow

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

Manually Selection/Protocols Adoption

Laborious process; could take up to 5 days for head and neck cancer treatment.

Pre-solve Large Sparse Dose Influence Matrix

Takes hours to solve for a single patient. Days/months for multiple patients.

Solve Fluence Map Optimization

Time-consuming: Often takes minutes.

Treatment Plan Flowchart

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs, RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Reprinted from "IMRT Optimization Algorithms. David Shepard. Swedish Cancer Institute. AAPM 2007."

Current Approaches and Limitations

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs, RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Innovation

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- A Tower Neural Network generates a policy that guides MCTS simulations for two players in a zero-sum Markov game
 - Produces a *utility (value) function* & a subjective *probability distribution*
- Each player in a two-player Markov game finds an alternating best response to the current player's average strategy
 - driving the neural network policy's weights toward an approximate **saddle equilibrium** [Heinrich et al. (2015)].
 - aids network in finding an *approximately optimal* beam angle candidate set that meets a dosimetric requirements.

State Encoding: Prostate Organ Masks

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



State Representation: Beam Angles

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



State Representation

Automating Treatment Planning in Radiation Therapy

Olakekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Two-player Fictitious network play with ResNet

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Tree Representation and Game Simulation

Automating Treatment Planning in Radiation Therapy


Olalekan
Ogunmolu

BOO
MCTS
BOO Workflow
Innovation

Head Stabilization

iDG

Futures



Tree Composition

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO
MCTS
BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs
Innovation

iDG

Robustness issues
Innovation
iDG Results

Futures

References

Every **node** of the tree, \mathbf{x} , has the following fields:

- a pointer to the parent that led to it, $\mathbf{x}.p$;
- the beamlets, \mathbf{x}_b , stored at that node; $b = \{1, \dots, m\}$;
- a set of move probabilities prior, $p(s, a)$;
- a pointer $\mathbf{x}.r$, to the reward r_t , for the state \mathbf{x}_t ;
- a pointer to the state-action value $Q(s, a)$ and its upper confidence bound $U(s, a)$;
- a visit count $N(s, a)$, that indicates the number of times that node was visited; and
- a pointer $\mathbf{x}.child$; to each of its children nodes.

Saddle Point Strategy Formulation

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- **Saddle point strategies** for optimal control sequence pair $\{a_t^{p_1^*}, a_t^{p_2^*}\}$ recursively obtained by optimizing, $V_t(s, a)$

$$V_t^*(s) = Q_t^*(s_t, \pi_t^{p_1}, \pi_t^{p_2}) = \min_{\pi^{p_1} \in \Pi^{p_1}} \max_{\pi^{p_2} \in \Pi^{p_2}} Q_t^*(s_t, \pi^{p_1}, \pi^{p_2})$$
$$\forall s_t \in \mathcal{S}, \pi^{p_1} \in \Pi^{p_1}, \pi^{p_2} \in \Pi^{p_2}.$$

such that

$$v_{p_1}^* \leq v^* \leq v_{p_2}^* \quad \forall \{\pi_t^{p_1}, \pi_t^{p_2}\}_{0 \leq t \leq T}.$$

- p_1, p_2 respectively generating a **mixed strategy** via **averaging the outcome** of individual plays.

Training and Validation Loss

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Average training (solid) and validation (dotted) loss function (MSE) values across six cross-validation folds for the network (blue) and full network.

BOO Results: Testing of self-play network

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Column Generation vs Neural Network

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Dose-Volume Histogram of CG vs DNN architectures [Sadeghnejad Barkousaraie, Azar and Ogunmolu, Olalekan and Jiang, Steve and Nguyen, Dan (2019)].

Conclusions

- Deep Neural Network optimizes network weights in a separate multiprocessing thread; Network outputs probabilities used to guide search;
- Sparse lookahead search builds tree with nodes labeled by state-action pairs in an alternating manner; sample rewards stored on edges connecting state-action with state nodes;
- Beam angles prediction takes between 2-3 minutes with MCTS vs. ~ 60 seconds with Column Generation Pre-training.

Head Stabilization in Radiation Therapy (RT)

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Head Stabilization in Cancer Radiation Therapy
 - Intensity-Modulated RT (IMRT)

Robotic Radiosurgery

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs
Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References



A Patient Head Motion-Correction Mechanism for MRI-LINAC RT

OLALEKAN OGUNMOLU

DEPARTMENT OF RADIATION ONCOLOGY, PENN SCHOOL OF MEDICINE

- Current Collaborators: Rodney Wiersma & Xinmin Liu (UChicago → UPenn)
- Past Collaborators: Steve Jiang, Xuejun Gu, (UT Southwestern); Nick Gans (UT Dallas, UT Arlington)

Correcting Head Motion: RT and MRI-LINACs

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu


BOO
MCTS
BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT
MRI-LINACs
Innovation

iDG
Robustness issues
Innovation
iDG Results

Futures
References



(a) The BRW SRS Frame [Chelvarajah et al. (2004)]



(b) Thermoplastic masks



(c) Frame With MRI Coils (PSOM)

4-D Motion Correction Stage

Automating
Treatment
Planning in
Radiation
Therapy

Olaelekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG




Robustness issues

Innovation

iDG Results

Futures

References



Liu et al. (2015)

4-DOF Motion Controller Block Diagram

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO
MCTS
BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT


MRI-LINACs
Innovation

iDG

Robustness issues
Innovation
iDG Results

Futures

References



Liu et al. (2015)

Phantom Feedback Motion Correction Results

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT


MRI-LINACs
Innovation

iDG

Robustness issues
Innovation
iDG Results

Futures

References



Time response of feedback control without (left) and with (right) decoupling control [Liu et al. (2015)].

Human Volunteer Feedback Motion Correction Results

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG



Robustness issues

Innovation

iDG Results

Futures

References



Head Motion Without and With Motion Correction. Left: Coupled Axes; Right: Decoupled Axes.

SRS: Wiersma Stewart-Gough Platform

Automating Treatment Planning in Radiation Therapy

Olakekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



6-DOF Motion Correction Results

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Drawbacks of current solutions

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Rigid patient's body assumption
- Non-compliant immobilization devices
- Invasiveness during radiosurgery/RT
- Attenuation of photon beams

Radiation Delivery Couch and Gantry

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References



Varian's TrueBeam Radiotherapy System.

Next-Gen RT Treatment with MRI-LINACs

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results


Futures

References

Elekta AB's (Sweden)




ViewRay's MRIdian



Elekta AB's (Sweden)



MagnetTx (Canada) Aurora RT



Soft Robots for Head Motion Compensation

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues


Innovation

iDG Results


Futures

References


IMRT Setup




IMRT/MRI



Head and Robot Panel



Head-Robot Closeup



Morphing in Cephalopods

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO
MCTS
BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT
MRI-LINACs
Innovation

iDG
Robustness issues
Innovation
iDG Results

Futures

References



Morph Stage 1



Morph Stage 2




Morph Stage 3



Morph Stage 4



Morph Stage 5



Morph Reversal

©Roger Hanlon, YouTube.

Cephalopods Neural-Controlled Physical Texture

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

Raises Periscope




Papillae



Papillae Up



Papillae × Papillae



Cuttlefish' Morphin ©Roger Hanlon, YouTube.

Cephalopods-inspired Actuator Design

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues



Innovation

iDG Results

Futures

References

Circumferentially Constrained And Radially Symmetric Elastomers (CCOARSE).



[Pikul et al. (2019)]

Nonlinear Elastic Deformation Analysis

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues



Innovation

iDG Results

Futures

References

IAB SHELLS AND AIR CAVITY/DEFORMATION ANALYSIS



$$r_i \leq r \leq r_o, \quad 0 \leq \theta \leq 2\pi, \quad 0 \leq \phi \leq \pi$$

Soft IK via Boundary Value Problem

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- With Cauchy's laws of motion, solve boundary value problem of traction, and find that

$$P(r) = \int_{r_i}^{r_0} \left[2C_1 \left(\frac{r}{R^2} - \frac{R^4}{r^5} \right) + 2C_2 \left(\frac{r^3}{R^4} - \frac{R^2}{r^3} \right) \right] dr \quad (1)$$

- i.e. Given a prescribed radius, find pressure to deform actuator between configurations

Volumetric Deformation Results (Simulation)

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation


iDG Results

Futures

References




(a) Left: Mesh model. Right: Stress distribution on outer skin.




(b) Displacement errors along x, y, z coordinates.

Inputs				Outputs			
C_1	C_2	R_i	r_i	R_o	r_o	P	ΔV
1.1e4	2.2e4	.027	.03	.03	.033	.76	≈ 0

Fig. 6: Volumetric Deformation (Expansion).



(a) Left: Mesh model. Right: Stress distribution on outer skin.



(b) Displacement errors along x, y, z coordinates.

Inputs				Outputs			
C_1	C_2	R_i	r_i	R_o	r_o	P	ΔV
1.1e4	2.2e4	.025	.03	.03	.028	-.34	≈ 0

Fig. 7: Volumetric Deformation (Compression).

Pneumatic Control and Deformation Scheme

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Volumetric Deformation Results (Actual)

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG







Robustness issues

Innovation

iDG Results

Futures

References



Actuator and Overall Mechanism with Phantom

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Head Motion Open Loop Control

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues



Innovation

iDG Results

Futures

References

Independent Actuation



Head Translation along x, y, z for a task of raising the head by a certain threshold above the table

Head rotation in Euler angles for a task of tilting the head about the x, y, z axes on the treatment table.

Ongoing Work: 6-DOF Closed-loop Control

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues



Innovation

iDG Results

Futures

References

Continuum Mechanical Model Validation/Differential Geometry/Newton-Euler Dynamics



3-DOF Simulation Testbed


Automating Treatment Planning in Radiation Therapy

Olalekan
Ogunmolu

B00

iDG

Futures



Model Reference Adaptive Control

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Indirect MRAC system. (Source mdpi.com)

3-DOF Model Reference Adaptive Control

Automating
Treatment
Planning in
Radiation
Therapy

Olaelekan
Ogunmolu

BOO
MCTS
BOO Workflow
Innovation

Head
Stabilization
Background –
LINACs RT
MRI-LINACs
Innovation

iDG
Robustness issues
Innovation
iDG Results
Futures

References

- Model head and bladder dynamics as
 - $\dot{x} = Ax + B\Lambda(u - f(x, u)) + w(k)$
- Approximate $f(x, u)$ by a neural network with continuous memory states
- Derive adaptive adjustment mechanism from Lyapunov analysis for Adaptive Control Parks (1966)
 - $u = \underbrace{\hat{K}_x^T x}_{\text{state feedback}} + \underbrace{\hat{K}_r^T r}_{\text{optimal regulator}} + \underbrace{\hat{f}(x, u)}_{\text{approximator}}$

Neural Network Architecture

Automating
Treatment
Planning in
Radiation
Therapy

Olakekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues


Innovation

iDG Results

Futures

References

Neural Net Architecture



Lyapunov Redesign: Theorem

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs
Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- For correct adaptive gains, $\hat{\mathbf{K}}_x$ and $\hat{\mathbf{K}}_r$, $\mathbf{e}(k)$ is ***uniformly ultimately bounded***, and the state \mathbf{x} converges to a neighborhood of \mathbf{r} .
- Choose a \mathbf{V} in terms of \mathbf{e} ; $\tilde{\mathbf{K}}_x^T$, $\tilde{\mathbf{K}}_r^T$; and parameter error $\varepsilon_f(\mathbf{x}(k))$ space

$$\mathbf{V}(\mathbf{e}, \tilde{\mathbf{K}}_x, \tilde{\mathbf{K}}_r^T) = \mathbf{e}^T \mathbf{P} \mathbf{e} + \text{tr}(\tilde{\mathbf{K}}_x^T \Gamma_x^{-1} \tilde{\mathbf{K}}_x^T |\Lambda|) + \text{tr}(\tilde{\mathbf{K}}_r^T \Gamma_r^{-1} \tilde{\mathbf{K}}_r^T |\Lambda|)$$

Results: Z and Pitch Motions

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Goal command: $(z, \theta, \phi) = (14\text{mm}, 1.6^\circ, 45^\circ)^T$.

Results: Roll Motion

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Conclusions

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs
Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Non-invasive soft robot for head motion compensation ✓
- Photons-transparent as opposed to rigid/electro-mechanical devices/robots ✓
- Adaptable under MRI coils for newer MRI-LINACs ✓

Part III: Robustness Margins and Robust Deep Policies

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Robustness Margins and Robust Deep Policies for Nonlinear Control

The robustness conundrum

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues


Innovation

iDG Results


Futures

References

- How to know *a priori* a policy's robustness limits?



- How to inculcate robustness into multistage decision policies?



Innovation

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –

LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- To quantify the brittleness, we optimize the stage cost

$$\max_{\mathbf{v}_t \sim \psi \in \Psi} \left[\sum_{t=0}^T \underbrace{c(\mathbf{x}_t, \mathbf{u}_t)}_{\text{nominal}} - \gamma \underbrace{g(\mathbf{v}_t)}_{\text{adversarial}} \right]$$

- To mitigate lack of robustness, we optimize the *cost-to-go*

$$\mathcal{J}_t(\mathbf{x}_t, \pi, \psi) = \min_{\mathbf{u}_t \sim \pi} \max_{\mathbf{v}_t \sim \psi} \left(\sum_{t=0}^{T-1} \ell_t(\mathbf{x}_t, \mathbf{u}_t, \mathbf{v}_t) + L_T(\mathbf{x}_T) \right),$$

- and seek a saddle point equilibrium policy that satisfies

$$\mathcal{J}_t(\mathbf{x}_t, \pi^*, \psi) \leq \mathcal{J}_t(\mathbf{x}_t, \pi^*, \psi^*) \leq \mathcal{J}_t(\mathbf{x}_t, \pi, \psi^*),$$

Results: Brittleness Quantification

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG



Robustness issues

Innovation

iDG Results

Futures

References



Results: Iterative Dynamic Game

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



Future Work: MRI/RT Immobilization

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs RT

MRI-LINACs
Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Explore multiple parallel robot mechanisms for head motion correction.
- Adopt iterative dynamic game approach [Ogunmolu et al. (2018)] for solving robust controller for head stabilization.
- Build on Freeman and Kokotovic's point-wise min-norm robust control lyapunov function to realize a meaningful value function in deep policies [Freeman and Kokotovic (1996)].

Constrained Robust Control Lyapunov Function (RCLF) Motion Planning

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO
MCTS

BOO Workflow
Innovation

Head Stabilization

Background – LINACs RT
MRI-LINACs
Innovation

iDG
Robustness issues
Innovation
iDG Results

Futures

References

Problem Definition

Find RCLF collision-free path $\sigma^* : [0, 1] \rightarrow \mathcal{M}_{free}$ given a path planning problem $(\mathcal{Q}_{free}, \xi_i, \mathcal{Q}_{goal})$, manipulation constraint, G , and cost function V such that $V(\sigma^*) = \min_{\sigma \in \Sigma_{\mathcal{M}_{free}}} V(\sigma)$ if one exists.

- Leverage [Freeman and Kokotovic (1996), Ogunmolu et al. (2018)].
- Greedy approach using L-BFGS optimization algorithm with box constraints, in contrast to the quadratic nonlinear constrained optimization e.g. Khansari-Zadeh and Billard (2014).

Reproducing a Nonlinear Motion with GMM

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



A CLF motion executor (red curves) that shows convergence to local attractors (green asterisks) and follows 3 different set trajectories (blue curves) for 2D nonlinear motion-trajectory problems on the WAM robot
[Reproduced from Ogunmolu et al. (2020)].

Reproducing a Nonlinear Motion with GMM

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG


Robustness issues

Innovation

iDG Results

Futures

References



A CLF motion executor (red curves) that shows convergence to local attractors (green asterisks) and follows 3 different set trajectories (blue curves) for 2D nonlinear motion-trajectory problems on the WAM robot
[Reproduced from Ogunmolu et al. (2020)].

Conclusions

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

BOO

MCTS

BOO Workflow
Innovation

Head
Stabilization

Background –
LINACs, RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Designed a non-invasive soft robot for head motion compensation in IMRT/emerging MRI-LINACs ✓
- Photons-transparent; Adaptable under MRI coils for newer MRI-LINACs ✓
- Fast inference of beam orientations in treatment planning:
Approx 60 secs beams prediction time✓
- Adapted H_{∞} control methods for quantifying the brittleness of deep policies✓
- Devised a min-max-trained deep saddle policy for mitigating model mismatch, transfer errors, and policy sensitivity e.t.c. ✓

Publications

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

BOO

MCTS

BOO Workflow

Innovation

Head Stabilization

Background – LINACs RT

MRI-LINACs

Innovation

iDG

Robustness issues

Innovation

iDG Results

Futures

References

- Johannes Heinrich, Marc Lanctot, and David Silver. Fictitious self-play in extensive-form games. In *International Conference on Machine Learning*, pages 805–813, 2015.
- Sadeghnejad Barkousaraie, Azar and Ogunmolu, Olalekan and Jiang, Steve and Nguyen, Dan. A Fast Deep Learning Approach for Beam Orientation Selection Using Supervised Learning with Column Generation on IMRT Prostate Cancer Patients. *Medical Physics, American Association of Physicists in Medicine*, 2019.
- Radhini Chelvarajah, Brigid Leighton, Linda Martin, Wayne Smith, and Rachael Beldham-Collins. Cranial immobilisation—is there a better way? *Radiographer*, 51(1):29–33, 2004.
- Xinmin Liu, Andrew H Belcher, Zachary Grellewicki, and Rodney D Wiersma. Robotic stage for head motion correction in stereotactic radiosurgery. In *2015 American Control Conference (ACC)*, pages 5776–5781. IEEE, 2015.
- Andrew Belcher. *Patient Motion Management with 6-DOF Robotics for Frameless and Maskless Stereotactic Radiosurgery*. PhD thesis, The University of Chicago, 2017.
- Olalekan Ogunmolu and Rodney Wiersma. A Real-Time Patient Head Motion Correction Mechanism for MRI-Linac Systems. In *Joint AAPM / COMP Meeting, Virtual*, 2020.
- James Pikul, Itai Cohen, and Robert Shepherd. Stretchable surfaces with programmable texture, May 2 2019. US Patent App. 16/161,029.
- PC Parks. Liapunov Redesign of Model Reference Adaptive Control Systems. *IEEE Transactions on Automatic Control*, 11(3):362–367, 1966.
- Olalekan Ogunmolu, Nicholas Gans, and Tyler Summers. Minimax Iterative Dynamic Game : Application to Nonlinear Robot Control. *IEEE International Conference on Intelligent Robots and Systems*, 2018.
- Randy A Freeman and Petar V Kokotovic. Inverse optimality in robust stabilization. *SIAM journal on control and optimization*, 34(4):1365–1391, 1996.
- S Mohammad Khansari-Zadeh and Aude Billard. Learning control Lyapunov function to ensure stability of dynamical system-based robot reaching motions. *Robotics and Autonomous Systems*, 62(6):752–765, 2014.
- Olalekan Ogunmolu, Rachel Skye Thompson, and Rodrigo Pérez Dattari. Learning Control Lyapunov Functions in Python. <https://github.com/lakehanne/lyapunovearner>, 2020. Accessed February 10, 2020.
- Olalekan Ogunmolu, Adwait Kulkarni, Yonas Tadesse, Xuejun Gu, Steve Jiang, and Nicholas Gans.

BOO: Search

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

- Formulated as a bandit search that imposes a regret term on the Q -value
- During the planning process, we estimate a *value*, $v(\mathbf{y}_t)$, that estimates the optimality of a beam block;
- In parallel, we refine the deep neural network policy by optimizing its weight in a separate thread.
 - Network parameters updated by a **mixed strategy** which combines its **pure strategy**,
 - It is a best response to the fictitious opponent's **average pure strategy**.

BOO: Fluence Map Optimization

Automating Treatment Planning in Radiation Therapy

Olalekan Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL Control

App C: 1DOF

Kalman Filtering

Sensor Fusion

System Identification

App D: CCOARSE

Actuator

- Q -value defined as

$$\bar{Q}(s, a) = Q_j(s, a) + c \sqrt{\frac{2 \ln n(s)}{N(s, a)}}, \quad (2)$$

$$a^* = \arg \max_a \bar{Q}(s, a) \quad (3)$$

Fluence Map Optimization

- $\mathcal{X} \implies$ total discretized voxels of interest (VOI 's) in a target volume
- $\mathcal{B}_1 \cup \mathcal{B}_2 \cup \dots \cup \mathcal{B}_n \subseteq \mathcal{B} \implies$ beam partition subset
- $\mathcal{D}_{ij}(\theta_k) \implies$ matrix that describes each dose influence, d_i
 - delivered to a discretized voxel, i , in a volume of interest, VOI_h ($h = 1, \dots, \mathcal{X}$), from a beam angle, θ_k , $k \in \{1, \dots, n\}$

BOO: Fluence Map Optimization

- Suppose further that $\mathcal{D}_{ij}(\theta_k)$ is the matrix that describes each dose influence, d_i
 - delivered to a discretized voxel, i , in a volume of interest, VOI_h ($h = 1, \dots, \mathcal{X}$), from a beam angle, θ_k ,
 $k \in \{1, \dots, n\}$
 - We compute the matrix $\mathcal{D}_{ij}(\theta_k)$ by calculating each d_i for every bixel, j , at every φ° , resolution, where $j \in \mathcal{B}_k$

BOO: FMO problem definition

- The fluence problem is to find the values of decision variables, x_j , for which d_i to the tumor is maximized, while simultaneously minimizing the d_i to critical structures
- For the voxels in a target volume,
 - let a weighted quadratic objective minimize the l_2 distance between a pre-calculated dose \mathbf{Ax} , and a doctor's prescribed dose, \mathbf{b}
 - let a weighted quadratic objective maximizes the l_2 distance between \mathbf{Ax} (where \mathbf{x} represents the vectorized bixels, x_j) and \mathbf{b}

BOO: FMO problem definition

- Cost

$$\frac{1}{v_s} \sum_{s \in \text{OARs}} \|(\underline{w}_s \mathcal{D}_{ij}^s \mathbf{x}_s)_+ - b_s\|_2^2 + \frac{1}{v_s} \sum_{s \in \text{PTVs}} \|(\bar{w}_s \mathcal{D}_{ij}^s \mathbf{x}_s - b_s)_+\|_2^2 \quad (4)$$

- Pre-calculated dose term:

$\mathbf{Ax} = \left\{ \sum_s \frac{w_s}{v_s} \mathcal{D}_{ij}^s \mathbf{x}_s \mid \mathcal{D}_{ij} \in \mathbb{R}^{n \times l}, n > l \right\}$, which is a combination of the dose components that belong to OARs and those that belong to PTVs.

- Let $w_s = \{\underline{w}_s, \bar{w}_s\}$ be the respective underdosing and overdosing weights for the OARs and PTVs

- v_s represents the total number of voxels in each structure.

BOO: FMO

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

- Rewriting the objective

$$\min \frac{1}{2} \|Ax - b\|_2^2 \quad \text{subject to } x \geq 0.$$

- With Lagrangian:

$$L(x, \lambda) = \min \frac{1}{2} \|Ax - b\|_2^2 - \lambda^T x.$$

- Introducing an auxiliary variable z , we have

$$\min_x \frac{1}{2} \|Ax - b\|_2^2, \quad \text{subject to } z = x, \quad z \geq 0,$$

BOO: FMO by way of ADMM

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF Kalman Filtering
Sensor Fusion
System Identification
App D: CCOARSE Actuator

- Solving either the \mathbf{x} and \mathbf{z} sub-problems, we have

$$\mathbf{x}^{k+1} = (\mathbf{A}^T \mathbf{A} + \rho \mathbf{I})^{-1} (\mathbf{A}^T \mathbf{b} + \rho \mathbf{z}^k - \boldsymbol{\lambda}^k). \quad (5)$$

- And using the soft-thresholding operator, $S_{\boldsymbol{\lambda}/\rho}$, we find that

$$\mathbf{z}^{k+1} = S_{\boldsymbol{\lambda}/\rho} (\mathbf{x}^{k+1} + \boldsymbol{\lambda}^k), \quad (6)$$

where $S_{\boldsymbol{\lambda}/\rho}(\tau) = (\mathbf{x} - \boldsymbol{\lambda}/\rho)_+ - (-\tau - \boldsymbol{\lambda}/\rho)_+$. $\boldsymbol{\lambda}$ is updated as

$$\boldsymbol{\lambda}^{k+1} = \boldsymbol{\lambda}^k - \gamma (\mathbf{z}^{k+1} - \mathbf{x}^{k+1}), \quad (7)$$

where γ is a parameter that controls the step length.

App A: 3-DOF Closed-loop Control

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: COARSE
Actuator

Closed-loop Phantom motion control along 3 DoFs with an adaptive neuro-controller.

3-DOF Simulation Testbed

Automating
Treatment
Planning in
Radiation
Therapy

Olaelekan
Ogunmolu

Appendices


App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering
Sensor Fusion

System
Identification

App D: CCOARSE
Actuator



Hardware Description

Control Design Goals

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: CCOARSE
Actuator

- Stabilize z , pitch, and roll states, *i.e.*

$$\mathbf{x} = \begin{pmatrix} z \\ \theta \\ \phi \end{pmatrix}$$

- By solving an adaptive state feedback controller, optimal regulation, and minimize parametric uncertainties

Control Design Goals


Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF Kalman Filtering
Sensor Fusion
System Identification
App D: CCOARSE Actuator

- Provide closed loop tracking given a desired trajectory, r
- Robustify system to (non-)parametric uncertainties



Indirect MRAC system. (Source mdpi.com)

Lyapunov Redesign: Theorem

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

- For correct adaptive gains, $\hat{\mathbf{K}}_x$ and $\hat{\mathbf{K}}_r$, $\mathbf{e}(k)$ is ***uniformly ultimately bounded***, and the state \mathbf{x} converges to a neighborhood of \mathbf{r} .
- Choose a \mathbf{V} in terms of \mathbf{e} ; $\tilde{\mathbf{K}}_x^T$, $\tilde{\mathbf{K}}_r^T$; and parameter error $\varepsilon_f(\mathbf{x}(k))$ space

$$\mathbf{V}(\mathbf{e}, \tilde{\mathbf{K}}_x, \tilde{\mathbf{K}}_r^T) = \mathbf{e}^T \mathbf{P} \mathbf{e} + \text{tr}(\tilde{\mathbf{K}}_x^T \Gamma_x^{-1} \tilde{\mathbf{K}}_x^T |\Lambda|) + \text{tr}(\tilde{\mathbf{K}}_r^T \Gamma_r^{-1} \tilde{\mathbf{K}}_r^T |\Lambda|)$$

Stability proof

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

$$\dot{V}(\mathbf{e}, \tilde{\mathbf{K}}_y, \tilde{\mathbf{K}}_r) = \dot{\mathbf{e}}^T \mathbf{P} \mathbf{e} + \mathbf{e}^T \mathbf{P} \dot{\mathbf{e}} + 2\text{tr}(\tilde{\mathbf{K}}_y^T \Gamma_y^{-1} \dot{\tilde{\mathbf{K}}}_y | \Lambda |) \\ + 2\text{tr}(\tilde{\mathbf{K}}_r^T \Gamma_r^{-1} \dot{\tilde{\mathbf{K}}}_r | \Lambda |)$$

$$= \left[\mathbf{A}_m \mathbf{e} + \mathbf{B} \Lambda [\Delta \hat{\mathbf{K}}_r^T \mathbf{r} + \Delta \hat{\mathbf{K}}_x^T \mathbf{x}] \right]^T \mathbf{P} \mathbf{e} + \dots$$

$$\mathbf{e}^T \mathbf{P} \left[\mathbf{A}_m \mathbf{e} + \mathbf{B} \Lambda [\Delta \hat{\mathbf{K}}_r^T \mathbf{r} + \Delta \hat{\mathbf{K}}_x^T \mathbf{x}] \right] + \dots$$

$$2\text{tr}(\Delta \mathbf{K}_x^T \Gamma_x^{-1} \dot{\hat{\mathbf{K}}}_x | \Lambda |) + 2\text{tr}(\Delta \mathbf{K}_r^T \Gamma_r^{-1} \dot{\hat{\mathbf{K}}}_r | \Lambda |)$$

Stability Analysis

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion
System
Identification

App D: COARSE
Actuator

$$= \mathbf{e}^T (\mathbf{P} \mathbf{A}_m + \mathbf{A}_m^T \mathbf{P}) \mathbf{e} + 2\mathbf{e}^T \mathbf{P} \mathbf{B} \boldsymbol{\Lambda} \left(\tilde{\mathbf{K}}_y^T \mathbf{y} + \tilde{\mathbf{K}}_r^T \mathbf{r} \right)$$

$$+ 2\mathbf{tr} \left(\tilde{\mathbf{K}}_y^T \boldsymbol{\Gamma}_y^{-1} \dot{\tilde{\mathbf{K}}}_y | \boldsymbol{\Lambda} | \right) + 2\mathbf{tr} \left(\tilde{\mathbf{K}}_r^T \boldsymbol{\Gamma}_r^{-1} \dot{\tilde{\mathbf{K}}}_r | \boldsymbol{\Lambda} | \right)$$

$$= -\mathbf{e}^T \mathbf{Q} \mathbf{e} - 2\mathbf{e}^T \mathbf{P} \mathbf{B} \boldsymbol{\Lambda} \boldsymbol{\varepsilon}_f(\mathbf{y}) + 2\mathbf{e}^T \mathbf{P} \mathbf{B} \boldsymbol{\Lambda} \tilde{\mathbf{K}}_y^T \mathbf{y}$$

$$+ 2\mathbf{tr} \left(\tilde{\mathbf{K}}_y^T \boldsymbol{\Gamma}_y^{-1} \dot{\tilde{\mathbf{K}}}_y \right) + 2\mathbf{e}^T \mathbf{P} \mathbf{B} \boldsymbol{\Lambda} \tilde{\mathbf{K}}_r^T \mathbf{r} + 2\mathbf{tr} \left(\Delta \mathbf{K}_r^T \boldsymbol{\Gamma}_r^{-1} \dot{\tilde{\mathbf{K}}}_r \right)$$

Notice $x^T y = \mathbf{tr} (y x^T)$ from trace identity

Stability Analysis Cont'd

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering

Sensor Fusion
System
Identification

App D: COARSE
Actuator

Therefore,

$$\begin{aligned}\dot{\mathbf{V}}(\cdot) = & -\mathbf{e}^T \mathbf{Q} \mathbf{e} - 2\mathbf{e}^T \mathbf{P} \mathbf{B} \boldsymbol{\Lambda} \varepsilon_f \\ & + 2 \mathbf{tr} \left(\tilde{\mathbf{K}}_y^T (\boldsymbol{\Gamma}_y^{-1} \dot{\tilde{\mathbf{K}}}_y + \mathbf{y} \mathbf{e}^T \mathbf{P} \mathbf{B} \text{sgn}(\boldsymbol{\Lambda})) \right) |\boldsymbol{\Lambda}| \\ & + 2 \mathbf{tr} \left(\tilde{\mathbf{K}}_r^T (\boldsymbol{\Gamma}_r^{-1} \dot{\tilde{\mathbf{K}}}_r + \mathbf{r} \mathbf{e}^T \mathbf{P} \mathbf{B} \text{sgn}(\boldsymbol{\Lambda})) \right) |\boldsymbol{\Lambda}|\end{aligned}$$

where for a real-valued x , we have $x = \text{sgn}(x)|x|$.

- first two terms will be negative definite for all $\mathbf{e} \neq 0$
 - since \mathbf{A}_m is Hurwitz
- other terms will be identically null if we choose the adaptation laws

$$\dot{\tilde{\mathbf{K}}}_y = -\boldsymbol{\Gamma}_y \mathbf{y} \mathbf{e}^T \mathbf{P} \mathbf{B} \text{sgn}(\boldsymbol{\Lambda}), \quad \dot{\tilde{\mathbf{K}}}_r = -\boldsymbol{\Gamma}_r \mathbf{r} \mathbf{e}^T \mathbf{P} \mathbf{B} \text{sgn}(\boldsymbol{\Lambda})$$

Stability Results: Ogunmolu et al. (2017)

Automating
Treatment
Planning in
Radiation
Therapy

Olakekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion

System
Identification

App D: COARSE
Actuator

$$\begin{aligned}\dot{\mathbf{V}}(\cdot) &= -\mathbf{e}^T \mathbf{Q} \mathbf{e} - 2\mathbf{e}^T \mathbf{P} \mathbf{B} \boldsymbol{\Lambda} \boldsymbol{\varepsilon}_f \\ &\leq -\lambda_{low} \|\mathbf{e}\|^2 + 2\|\mathbf{e}\| \|\mathbf{P} \mathbf{B}\| \lambda_{high}(\boldsymbol{\Lambda}) \boldsymbol{\varepsilon}_{max}\end{aligned}$$

- $\{\lambda_{low}, \lambda_{high}\} \equiv \min/\max \text{ eigenvalues of } Q \text{ and } \boldsymbol{\Lambda}$.
- $\dot{\mathbf{V}}(\cdot)$ is thus negative definite outside the compact set:
$$\chi = \left(\mathbf{e} : \|\mathbf{e}\| \leq \frac{2\|\mathbf{P} \mathbf{B}\| \lambda_{high}(\boldsymbol{\Lambda}) \boldsymbol{\varepsilon}_{max}(\mathbf{y})}{\lambda_{low}(\mathbf{Q})} \right)$$
 - i.e. \mathbf{e} is uniformly ultimately bounded, or $\mathbf{y}(t) \rightarrow 0$ as $t \rightarrow \infty$.

Appendix C: 1-DOF Closed-Loop Control

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion
System
Identification

App D: COARSE
Actuator

Left blank intentionally

Head Pose Estimation: Sensor Fusion

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control


App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: COARSE
Actuator



Kinect Xbox

Kinect v1

Sensors' Noise Floor

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices


App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: CCOARSE
Actuator



Case for Sensed Observation Filtering

Optimal head state estimation and sensor fusion

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

- Find the observation estimates $\hat{\mathbf{x}}(i)$ that minimize the mean-square error from true measurement, i.e.,

$$\hat{\mathbf{x}}(i|j) = \arg \min_{\hat{\mathbf{x}}(i|j) \in \mathbb{R}^n} \mathbb{E}\{(\mathbf{x}(i) - \hat{\mathbf{x}})^T (\mathbf{x}(i) - \hat{\mathbf{x}}) | z(1), \dots, z(j)\}$$

- Define the estimate error's covariance as

$$\mathbf{P}(i|j) \triangleq \mathbb{E}\{(\mathbf{x}(i) - \hat{\mathbf{x}}(i|j))^T (\mathbf{x}(i) - \hat{\mathbf{x}}(i|j)) | Z^j\}. \quad (8)$$

State estimation with Kalman filters

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: COARSE
Actuator

- Assume state model, $\mathbf{F}(k)$
- For a discretized time interval ΔT between measurements, we define the state

$$\mathbf{x}(k) = \mathbf{F}(k)\mathbf{x}(k-1) + \mathbf{B}(k)\mathbf{u}_k + \mathbf{G}_k\mathbf{w}_k \quad (9)$$

- with

$$\mathbf{F} = \begin{bmatrix} 1 & \Delta T \\ 0 & 1 \end{bmatrix} \quad (10)$$

Kalman filters

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

- (9) $\Rightarrow \mathbf{x}_k = \mathbf{F}_k \mathbf{x}_{k-1} + \mathbf{G}_k \mathbf{w}_k$
- \mathbf{G}_k := uncontrolled forces accelerating on head
- Head's acceleration, $a_k \sim \mathcal{N}(0, \sigma_a)$
- Setting $\mathbf{G}_k = \mathbf{I}_{2 \times 2}$ and $\mathbf{w}(k) \sim \mathcal{N}(0, \mathbf{Q}(k))$
 - set $\mathbf{Q}(k)$ to a random walk sequence, $\mathbf{W}_k = [\frac{\Delta T^2}{2}, \Delta T]^T$

Kalman Filters Design

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator

- Such that

$$\mathbf{Q} = \mathbf{W}\mathbf{W}^T \sigma_a^2 = \begin{bmatrix} \frac{\Delta T^4}{4} & \frac{\Delta T^3}{2} \\ \frac{\Delta T^3}{2} & \Delta T^2 \end{bmatrix} \sigma_a^2. \quad (11)$$

- Set the transfer matrix from the estimates, $\mathbf{x}(k)$, to observations, $z_1(k)$ and $z_2(k)$ according to

$$z_s = \mathbf{H}_s(k)\mathbf{x}(k) + v_s(k) \quad s = 1, 2 \quad (12)$$

- where $\mathbf{H}_s(k) = [1, \ 0]^T$ and $v_s(k) \sim \mathcal{N}(0, \sigma_{rs}^2)$

KF Priori and Posteriori State Estimates

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering

Sensor Fusion

System
Identification

App D: COARSE
Actuator

■ Prediction Phase:

$$\hat{\mathbf{x}}_{k|k-1} = \mathbf{F}\hat{\mathbf{x}}_{k-1|k-1} + \mathbf{B}_k \mathbf{u}_k$$

$$\mathbf{P}_{k|k-1} = \mathbf{F}_k \mathbf{P}_{k-1|k-1} \mathbf{F}_k^T + \mathbf{Q}_k \quad (13)$$

■ Update Phase:

$$\mathbf{K}(k) = \mathbf{P}(k|k-1) \mathbf{H}(k)^T [\mathbf{H}(k) \mathbf{P}(k|k-1) \mathbf{H}(k)^T + \mathbf{R}(k)]^{-1}$$

$$\hat{\mathbf{x}}(k|k) = \hat{\mathbf{x}}(k|k-1) + \mathbf{K}(k)(\mathbf{z}(k) - \mathbf{H}(k)\hat{\mathbf{x}}(k|k-1))$$

$$\mathbf{P}(k|k) = (\mathbf{I} - \mathbf{K}(k)\mathbf{H}(k))\mathbf{P}(k|k-1) \quad (14)$$

Xbox Filtering Results

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices


App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering
Sensor Fusion
System
Identification

App D: COARSE
Actuator



Kinect 2 Filtering Results

Automating
Treatment
Planning in
Radiation
Therapy


Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering
Sensor Fusion
System
Identification

App D: CCOARSE
Actuator



KF results of Kinect v1's observation


Filtering Results

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF
Kalman Filtering
Sensor Fusion
System Identification
App D: CCOARSE Actuator



Global fusion of local tracks

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL
Control
App C: 1DOF
Kalman Filtering

Sensor Fusion
System
Identification
App D: COARSE
Actuator

- Fuse both updates with a variance-weighted average of each local track as follows,

$$\hat{\mathbf{x}}(F)(k|k) = \mathbf{P}(F)(k|k) \sum_{s=1}^N \left[\mathbf{P}(s)^{-1}(k|k) \hat{\mathbf{x}}(s)(k|k) \right]$$

$$\text{where } \mathbf{P}(F)(k|k) = \left[\sum_{s=1}^N \mathbf{P}(s)^{-1}(k|k) \right]^{-1}.$$

Track-to-track Fusion

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control


App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: CCOARSE
Actuator



Track-to-track fusion of both sensors' local track estimates.

Vision-based Control

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: COARSE
Actuator

This page is left blank intentionally.

System Identification

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF
Kalman Filtering
Sensor Fusion
System
Identification
App D: CCOARSE
Actuator

- From I/O data, estimate a model
- Find set of optimal model parameters via the minimization,

$$G(t) = \arg \min_{\theta} V_N(\theta, Z^N)$$

- where $V_N(\theta, Z^N) = \sum_{k=1}^{\mathcal{K}} \sum_{i=1}^n \frac{1}{2} (\hat{y}_i(k) - y_i(k))^2$,
- and $Z^N = \{u(1) \cdots u(N), y(1) \cdots y(N)\}$
- After a least squares minimization, we derive a state-space realization,

$$\begin{aligned} \mathbf{x}(k + Ts) &= \mathbf{Ax}(k) + \mathbf{Bu}(k) + \mathbf{Ke}(k) \\ \mathbf{y}(k) &= \mathbf{Cx}(k) + \mathbf{Du}(k) + \mathbf{e}(k) \end{aligned} \quad (15)$$


PID-PI Cascaded Controller

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

- App A: FMO
- App B: 3DOF CL Control
- App C: 1DOF Kalman Filtering
- Sensor Fusion
- System Identification
- App D: CCOARSE Actuator




Open Loop Step Response

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF
Kalman Filtering
Sensor Fusion
System
Identification
App D: CCOARSE
Actuator




PI Closed-loop Step Response

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF
Kalman Filtering
Sensor Fusion
System
Identification
App D: CCOARSE
Actuator



Cascaded PID-PI Closed-loop Step Response

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices


App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering
Sensor Fusion

System
Identification

App D: CCOARSE
Actuator




PI-PD Cascaded Controller Experimental Results

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO
App B: 3DOF CL Control
App C: 1DOF
Kalman Filtering
Sensor Fusion
System
Identification
App D: CCOARSE
Actuator




Varying Setpoint Simulation

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

- App A: FMO
- App B: 3DOF CL Control
- App C: 1DOF Kalman Filtering
- Sensor Fusion
- System Identification
- App D: CCOARSE Actuator



Varying Setpoint Experiment

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control


App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: CCOARSE
Actuator



3-DOF Controller Design (IROS 2017)

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: COARSE
Actuator

■ Posing the cost

$$J = \sum_{k=0}^K x(k)^T Q x(k) + u(k)^T R u(k) + 2x(k)^T N u(k)$$

- we can obtain u as $\Delta u = \arg \min_{\Delta u} J$
- Δu is a future control sequence


Closed-loop control (Full state observer).

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

- App A: FMO
- App B: 3DOF CL Control
- App C: 1DOF Kalman Filtering
- Sensor Fusion
- System Identification
- App D: CCOARSE Actuator



$$\hat{x}(k+1) = A(k)\hat{x}(k) - K_{obs}[C(k)\hat{x}(k) - y(k)] + B(k)u(k).$$

1-DOF Control Results

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices



App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering
Sensor Fusion

System
Identification

App D: CCOARSE
Actuator



LQG Controller on mannequin head.

Appendix D: CCOARSE Actuator Schematic

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control


App C: 1DOF

Kalman Filtering

Sensor Fusion

System
Identification

App D: CCOARSE
Actuator



Soft IK via Boundary Value Problem

Automating
Treatment
Planning in
Radiation
Therapy

Olalekan
Ogunmolu

Appendices

App A: FMO

App B: 3DOF CL
Control

App C: 1DOF
Kalman Filtering
Sensor Fusion

System
Identification

App D: CCOARSE
Actuator

- With Cauchy's laws of motion, solve boundary volume problem of traction
- Using the following boundary conditions for the Cauchy Stress,
 - $\sigma_{rr}|_{R=R_0} = -P_{atm}, \sigma_{rr}|_{R=R_i} = -P_{atm} - P$
- And together with Cauchy's first law, we find that
 - $\sigma_{rr}(r) = - \int_{r_i}^{r_o} [2C_1\left(\frac{r}{R^2} - \frac{R^4}{r^5}\right) + 2C_2\left(\frac{r^3}{R^4} - \frac{R^2}{r^3}\right)] dr$
 - $\sigma_{rr}(r) = \int_{R_i}^{R_o} [2C_1\left(\frac{1}{r} - \frac{R^6}{r^7}\right) - 2C_2\left(\frac{R^4}{r^5} - \frac{r}{R^2}\right)] dr$
- With $\sigma_{rr}|_{R=R_i} = -P_{atm} - P$ and setting $P_{atm} = 0$, we find
 - $P(r) = \int_{r_i}^{r_o} [2C_1\left(\frac{r}{R^2} - \frac{R^4}{r^5}\right) + 2C_2\left(\frac{r^3}{R^4} - \frac{R^2}{r^3}\right)] dr$
 - $P(r) = \int_{R_i}^{R_o} [2C_1\left(\frac{1}{r} - \frac{R^6}{r^7}\right) - 2C_2\left(\frac{R^4}{r^5} - \frac{r}{R^2}\right)] dr$