Alliance in Healthcare

RGL BIO ALLIANCE CORPORATION
79-A Apolonio St. corner General Araneta St.
Project 4, Quezon City, 1109 Philippines
T: +63 2 738 6411, +63 2 617 8892
E: info@rglbioalliancecorp.com
W: www.rglbioalliancecorp.com

	Check Item	Acceptance Criteria	Result
Startup & Self-Test	Power-on self-test	No error codes displayed	■ Pass □ Fail
	Initialization sequence	Completes without interruption	■ Pass □ Fail
	System ready status	Analyzer reaches READY state	■ Pass □ Fail
	Startup noise	No abnormal sound or vibration	■ Pass □ Fail

	Check Item	Acceptance Criteria	Result
Sampling & Mechanics	Sample probe movement	Smooth, no obstruction	■ Pass □ Fail
	Probe alignment	Correctly aligned to sample position	■ Pass □ Fail
	Probe washing	No visible residue after wash	■ Pass □ Fail
	Clot detection behavior	No false clot alarms during test	■ Pass □ Fail
	Mechanical motion	No abnormal resistance or noise	■ Pass □ Fail

	Description	Reference	Result	Remarks
Analyzer Pressure	PC (50 kPa)	45 ~ 60 kPa	_____	■ Pass □ Fail
	SCI (40 kPa)	35 ~ 45 kPa	_____	■ Pass □ Fail
	CBC_W (-30 kPa)	-90 ~ 0 kPa	_____	■ Pass □ Fail
	WC2 (-40 kPa)	-45 ~ -35 kPa	_____	■ Pass □ Fail
	Fluid Pressure	50 ~ 120 kPa	_____	■ Pass □ Fail
	SPMT Temperature	31 ~ 38 °C	_____	■ Pass □ Fail
	Ambient Temperature	10 ~ 35 °C	_____	■ Pass □ Fail

	Description	Reference	Result	Remarks
Analyzer Temperature	SPMT Temperature	31 ~ 38 °C	_____	■ Pass □ Fail
	Ambient Temperature	10 ~ 35 °C	_____	■ Pass □ Fail
	DIFF / RET Reaction Bath	40 ~ 43 °C	_____	■ Pass □ Fail
	Preheating Bath	5 ~ 43 °C	_____	■ Pass □ Fail
	RBC Diluent	5 ~ 43 °C	_____	■ Pass □ Fail
	Instrument Temperature	10 ~ 41 °C	_____	■ Pass □ Fail
	CPU Temperature	0 ~ 100 °C	_____	■ Pass □ Fail
	ESR Assembly Temperature	34 ~ 43 °C	_____	■ Pass □ Fail

Background and Blank Counting Requirements (Blood)	Description	Reference	Value	Result
	WBC	**≤ 0.1** × 10^9 / L	_____	■ Pass □ Fail
	WBC-D	**≤ 0.2** × 10^9 / L	_____	■ Pass □ Fail
	RBC	**≤ 0.2** × 10^{12} / L	_____	■ Pass □ Fail
	RBC-O	**≤ 0.2** × 10^{12} / L	_____	■ Pass □ Fail
	HGB	**≤ 1** g / L	_____	■ Pass □ Fail
	PLT	**≤ 5** × 10^9 / L	_____	■ Pass □ Fail
	PLT-O	**≤ 5** × 10^9 / L	_____	■ Pass □ Fail

Repeatability Whole Blood	Description	Reference	Value	Result
	WBC	≤ 2.5 %	_____	■ Pass □ Fail
	RBC	≤ 1.5 %	_____	■ Pass □ Fail
	HGB	≤ 1.0 % (very poor ≤ 3)	_____	■ Pass □ Fail
	MCV	≤ 1.0 %	_____	■ Pass □ Fail
	HCT	≤ 1.5 %	_____	■ Pass □ Fail
	PLT	≤ 4.0 %	_____	■ Pass □ Fail

Remarks:

_____

_____

_____

``