


Respimatic 100



*An Affordable, Robust Emergency
Respiration Assist Device*

*Web Applications
Operating Manual*

Table of Contents

Introduction.....	3
Setting up Wi-Fi credentials.....	3
Launching the WEB Apps.....	6
WEB Dashboard.....	8
Dashboard Snapshots View	9
Dashboard Charts View	10
Dashboard Statistics View.....	11
Dashboard Alerts View.....	13
Dashboard Waveforms View	14
Dashboard Recording	15
WEB Analyzer.....	16
Analyzer Sidebar Menu	16
Analyzer Recording Selector.....	17
Analyzer Range Slider.....	17
Analyzer Recording Import Export	18
Analyzer Charts, Statistics, Breath Waveforms and Alerts Views.....	18
Updating System Firmware.....	19
FiO ₂ Calculator	21

Table of Figures

Figure 1: Enable/Disable Wi-Fi login.....	3
Figure 2: Wi-Fi Autoconnect	4
Figure 3: Connect to Respimatic W-Fi network.....	4
Figure 4: Wi-Fi Configuration Portal	5
Figure 5: Wi-Fi Login Successful.....	5
Figure 8: Web Apps Main Menu	6
Figure 9: Respimatic Systems Table.....	6
Figure 10: Respimatic Documentation.....	7
Figure 11: Breath Range Selector	8
Figure 12: Dashboard Menu	8
Figure 13: Dashboard Snapshots View	9
Figure 14: Dashboard Charts View	10
Figure 15: Charts Edit Menu	11
Figure 16: Dashboard Statistics View	12
Figure 17: Dashboard Alerts View	13
Figure 18: Breath Detailed Pressure and Flow Graphs	14
Figure 19: Breath Type Menu	14
Figure 20: Dashboard Recording View.....	15
Figure 21: Analyzer Sidebar Menu	16
Figure 22: Analyzer Selector Table.....	17
Figure 23: Analyzer Breath Range Selector.....	18
Figure 24: Analyzer Import View.....	18
Figure 25: Firmware Update Web Application	19
Figure 26: Step-by-step instructions on Updating Firmware	20

Introduction

In addition to the front panel, Respimatic 100 provides the ability to remotely monitor all respiration sessions via a WEB dashboard. Doctors and technicians can use the dashboard to connect to any Respimatic system using a unique system ID embedded in each system. This feature is useful to enable a remote specialist to observe the key system and patient parameters during a session and suggest a course of action for the local practitioners.

A brief overview of the process is as follows.

1. There must be a Wi-Fi network at the site where the Respimatic system is deployed. If required, use a 4G/5G dongle to establish a Wi-Fi network. One dongle can serve multiple systems at the same site at the same time.
2. There must be Wi-Fi or wired internet at the monitoring site.
3. Enable WEB dashboard monitoring on the Respimatic system.
4. Allow the Respimatic system to log on to the Wi-Fi network.
5. Visit the provided URL at the monitoring site.
6. Pair the WEB dashboard at the monitoring site with the Respimatic at the deployment site using the Respimatic System Unique ID.
7. WEB Dashboard allows monitoring only. It does not permit remote control of the system.




Figure 1: Enable/Disable Wi-Fi login

During the start-up sequence, the system gives an option to enable or disable remote monitoring for the system. Once enabled, the system guides the user to set up the remote WEB Apps.

Setting up Wi-Fi credentials

Respimatic system remembers the history of prior Wi-Fi networks that have been used by the system. The user is provided with an option to either auto-connect to a previously known Wi-Fi network or to configure a new one. If desired, the recorded Wi-Fi history can be erased at this time.




Figure 2: Wi-Fi Autoconnect

If Auto-connect is not enabled or if Auto-connect fails, the next option is to use a configuration portal to setup a new Wi-Fi network for the system to log in to. To enable login through a configuration portal, the system sets up a local, temporary Wi-Fi network named "Respimatic Wi-Fi". The user can choose to use either a wifi-enabled laptop or smartphone to log on to this network.





Figure 3: Connect to Respimatic W-Fi network

Upon login to this "Respimatic Wi-Fi" network, a portal screen is automatically presented on the laptop or the smartphone which guides the user step-by-step to enabling system to login to a desired Wi-Fi network. Choose from the list of available networks or manually enter the name of the network that the system should log in to. The portal time out in 2 minutes if unable to log on for whatever reason. If the system times out, the user can retry as many times as desired.

In case the portal does not automatically open, open a browser and navigate to 192.168.1.4 (URL) after connecting to the "Respimatic Wi-Fi" network.




The image shows a digital display for the 'Respimatic 100 Wi-Fi Login' portal. It has a dark blue background with white text. At the top, it says 'Respimatic 100' and 'Wi-Fi Login'. Below this is a horizontal line. Then, it asks for the 'Wi-Fi Name (SSID)' with a green input field containing '1077'. Below that, it asks for the 'Wi-Fi Password' with a green input field containing 'Act1077@home'. Another horizontal line follows. At the bottom, there is a green 'LOGIN' button.

Figure 4: Wi-Fi Configuration Portal

If the Wi-Fi login was unsuccessful, the user can still continue without enabling a WEB dashboard.

Upon a successful Wi-Fi login, the system is now ready to be connected to the WEB Apps using the unique system id (SYSUID). The SYSUID is a 20 character string starting with the prefix "RSP_" followed by 16 hexadecimal digits.



The image shows a yellow rectangular box with a black border containing green text. The text reads: 'Use Unique SysUID', 'UID_0123456789ABCDEF', 'for WEB access', and 'Continue (YES) ?'.

Figure 5: Wi-Fi Login Successful

Launching the WEB Apps

<https://www.respimatic.com>




Figure 6: Web Apps Main Menu





System Tag	System UID	FW Version	Actions
BANGALORE	UID_28615E07D6013C4A	2.2.1	 
DELHI	UID_28CFE43C4D200184	unknown	 
UDAIPUR	UID_28B1A879A2010387	unknown	 

Figure 7: Respimatic Systems Table

IMPORTANT

Use CTRL key and +/- keys to zoom in/out
or hold down the CTRL key and use the mouse wheel to zoom in/out
till the content fits well in the browser screen.

The screenshot above on the left is the main menu while the screenshot on the right above is the Systems Table.

Each unique system id (SYSUID) of the Respimatic 100 systems can be associated with an easy to remember name tag. The browser remembers the history of all the SYSUIDs that have been used and presents them in the dropdown list in the main menu box above.

The table of name tags and associated SYSUID (Systems Table) is accessed through the "Manage Known Systems" button. The + menu button on the Systems Table adds a system. To select a system to communicate with either double click on the appropriate row or use the checkmark button against the row. A system can be removed from the table using the trash menu button in the appropriate row. The trash button on the top right removes all system information. The systems table can also be exported as a JSON xml file and can be imported from a JSON xml file by clicking the appropriate icons on the table banner. Finally, the back arrow menu button on the top left can be used to navigate back to the main menu.

The following WebApps are available via the links on the main menu.

- **Launch Dashboard** - Monitor a session remotely. It requires a SysUID to be selected to determine the Respimatic system to connect to.
- **Launch Analyzer** - Analyse a previously recorded session. It requires a SysUID to be selected to determine the Respimatic system to connect to.
- **Update System Firmware** – Download and install a new release of firmware for the Respimatic system.
- **Calculate FiO2 Settings** – Calculate required Oxygen inflow rate. It does not require a SysUID.
- **View Documentation** – All the pdf documents are accessible through this link. It does not require a SysUID. If a "PDF Viewer" extension is loaded in your browser the documents can be read online or else, they can be downloaded. Figure below is a screenshot of the documents web page.



Figure 8: Respimatic Documentation

WEB Dashboard

The live WEB dashboard, when connected, offers a choice of six different views. The user can switch between these views at any time. Figure below shows the sidebar for the dashboard.



Figure 10: Dashboard Menu

The box on the top shows some relevant summary data for the current session.

The next box allows for selecting the type of view to display. Available views are as below.


1. Snapshots view
2. Charts view
3. Statistics view
4. Alerts view
5. Breath Shapes View
6. Recording View

The next box allows for turning on/off session recording and dashboard updates. Pausing dashboard updates only stops the display from updating, fresh data continues to get collected and will get displayed once the 'Pause Dashboard' is turned OFF.

The box at the bottom displays all FiO₂ related settings. These settings are only for monitoring, the user must use the physical system's control panel to change these settings like all other settings.

The Figure below shows a Slider that can be used to select the range of breath numbers to display the required data for. The breath number range can be selected by grabbing the slider handles and sliding then to the required breath numbers. The range can then be committed using the checkmark button. Once committed the display is frozen to the selected range. The rolling button on the extreme left discards the selected range and enables a fresh display update on each new breath. The double arrow button selects the entire range and freezes the display to that.

Figure 9: Breath Range Selector



Dashboard Snapshots View

Below is a screenshot of the WEB Dashboard "Snapshots View". All the Respiratic parameters, both input and output, are presented on the dashboard for easy viewing as a cohesive whole.




Figure 11: Dashboard Snapshots View

Note that for safety reasons, the Dashboard only allows monitoring of a remote system. The dashboard cannot be used to control the remote system. That must be done using the front panel of the physical Respimatic system.

Dashboard Charts View

The Dashboard also provides an option for a "Charts View". A screenshot of the charts view is shown below. This screenshot shows three chart boxes. A chart box can be added at any time using the + menu button on the top left of the chart box. Use the trash menu button on the top right to delete a chart box.

The parameters to chart can be selected using the checkboxes on the edit menu. The edit menu button is also on the top right of each chart box. The charts are updated after every breath. The X-axis can be selected as breath number or as elapsed time between breaths.

By default, the system charts the selected parameters for the past 60 breaths on a rolling basis, the charts are updated after each breath.




Figure 12: Dashboard Charts View

There is a Range Slider above the chart boxes to pan or zoom to any desired range of breath numbers. Use the handles on the Slider to select a particular range. If a particular range is selected, further chart updates are disabled till the range is reset using the RESET button on the range slider, the new parameter values are still stored for later display. Upon RESET, the chart boxes return to their default display mode.

The chart box edit menu is shown below. Any collection of the shown parameters can be displayed in any chart box or in multiple chart boxes.

Title

Y-axis Parameters

- ☐ Pressures
 - ☐ Peak Pressure
 - ☐ Plateau Pressure
 - ☐ PEEP Pressure
- ☐ Volumes
 - ☐ Tidal Volume
 - ☐ Minute Volume
- ☐ Breaths per minute
 - ☐ Mandatory BPM
 - ☐ Spontaneous BPM
- ☐ Breath Type
 - ☐ Mandatory Breath
 - ☐ Spontaneous Breath
 - ☐ Maintenance Breath
- ☐ Instant Lung Compliances
 - ☐ Instant Dynamic Lung Compliance
 - ☐ Instant Static Lung Compliance
- ☐ Oxygen Related
 - ☐ FI02
 - ☐ Incoming Oxygen Flow Required
 - ☐ Incoming Oxygen Purity
- ☐ Miscellaneous
 - ☐ System Errors
 - ☐ System Warnings
 - ☐ System Notifications
 - ☐ System Temperature

X-axis Units

☒ Breath Number ☐ Breath Times

Figure 13: Charts Edit Menu

The Dashboard also provides an option for a "Statistics View". A screenshot of the statistics view is shown below.

The statistics are collected for the entire range updated after every breath by default. The range slider can be used to gather statistics for any range of breath numbers. The RESET menu button on the range slider causes the updates and display to go back to their default mode.




Figure 14: Dashboard Statistics View

Dashboard Alerts View

The Dashboard also provides an option for a "Alerts View". A screenshot of the alerts view is shown below.

By default, it displays the complete history of errors and warnings encountered in the current session. The range slider on top can be used to see the alerts for any range of breath numbers. The RESET menu button causes the updates to go back to their default mode.



Figure 15: Dashboard Alerts View

Dashboard Waveforms View

The Dashboard also provides an option to view the detailed pressure and flow waveforms for selected breaths. Figure below shows a sample of such a view.




Figure 16: Breath Detailed Pressure and Flow Graphs

Title

Y-axis Parameters

☐ All

☐ Mandatory Breaths
☐ Volume Controlled

☐ Spontaneous Breaths
☐ Volume Controlled
☐ Pressure Supported

☐ Include Maintenance Breaths
☐ Include Error Breaths
☐ Include Abnormal Breaths

Figure 17: Breath Type Menu

The EDIT icon on each box allows the selection of the kinds of breaths to display. The graphs are color coded as per the legend at the bottom of the page. The breath selection menu is shown in the Figure on the left.

Dashboard Recording

The Dashboard also provides an option to record any part of the current session using the "Start Recording" menu button on the sidebar menu. The recording can be paused at any time causing that paused window to not be recorded. The Recording at the bottom left of the Snapshots view indicates whether the recording is currently active.

The recording is stored in a JSON database on the disk of the laptop or the desktop that the browser is running on. This recording can be analysed at any future time using the Analyzer WebApp. A screenshot of the recording view is shown below.





Figure 18: Dashboard Recording View

WEB Analyzer

The Analyzer enables the analysis of a previously recorded session with a patient. The process starts with selecting a session recording to analyse.



Analyzer Sidebar Menu

The main menu for the Analyzer is presented as a sidebar menu as shown in the Figure on the left.

The box on the top provides a summary of the currently selected recording for analysis.

Figure 19: Analyzer Sidebar Menu

Analyzer Recording Selector

Each previously recorded session is presented in a Selector table. Select a recording for analysis either by double-clicking on a row or using the appropriate checkmark menu button.

Mickey Mouse [27-06-2023 12:59:55]		
RESPIMATIC-100 Recordings		
Recording Name	Created	Actions
New Recording	06-07-2023 09:06:04	✓ ↗ 🗑
Mickey Mouse	27-06-2023 12:59:55	✓ ↗ 🗑
Demo Recording	24-06-2023 10:37:09	✓ ↗ 🗑

SYSTEM UID
 UID_28615E07D6013C4A
 (BANGALORE)

RESPIMATIC 100 WEB ANALYZER




Figure 20: Analyzer Selector Table

After a database is selected, a summary of the recording data is displayed in the top box of the sidebar and the selected table row is highlighted.

Analyzer Range Slider

By default, the analysis window is the entire recording. To select a particular breath number range to analyse, use the handles of the Range Slider shown below to zero in on




Figure 21: Analyzer Breath Range Selector

the interval of interest. All Analysis actions use the currently selected analysis range. The range selector works in the same fashion as described earlier in the Dashboard section.

Analyzer Recording Import Export

The EXPORT menu button on each row enables the user to export the database to a text file that can be sent to others for analysis. The IMPORT button on the top left allows the user to import a previously exported text file as a new session available for analysis. Below is a screenshot of the Import screen.




Figure 22: Analyzer Import View

Analyzer Charts, Statistics, Breath Waveforms and Alerts Views

Finally, the Charts, Statistics, Breath Shapes and Alerts views work in the same fashion as described in the Dashboard section above.

Updating System Firmware

Occasionally the system may need to be upgraded in the field for new features or bugs. The new firmware releases are made available on the web for downloading and installing. The following menu interface is a part of the Web Applications to enable the end user to accomplish these upgrades.

For further details click on the Step-by-step Instructions on this application.



The screenshot displays a web application interface for updating the RESPIMATIC 100 firmware. The interface is dark-themed with white and light blue text. At the top, a blue box contains the title "RESPIMATIC 100 Update Firmware". Below this, the text "Step-by-step Instructions" is displayed. A horizontal line separates this from the next section, "One-time Download Arduino Builder". Another horizontal line follows. Below that, a light blue box contains the text "Select and Download Release". Underneath this is a table with three columns: "Version", "Release Date", and "Get". The table has one data row showing "1.0.1", "16-May-2023", and a download icon. At the bottom of the interface, there are three logos: "TECHNO INDIA NIR INSTITUTE OF TECHNOLOGY", "CLEARPACK", and "PILAM ATMANANDRA RESEARCH CENTER".

Version	Release Date	Get
1.0.1	16-May-2023	

Figure 23: Firmware Update Web Application

Below is a screenshot of step-by-step instructions.

Slide 1: How to Update Firmware Re...
Slide 2: Equipment needed
Slide 3: Firmware Update Overview
Slide 4: Download Arduino Builder
Slide 5: Download Arduino Builder S...
Slide 6
Slide 7
Slide 8: Install Arduino Builder
Slide 9: Install Arduino Builder Step 1
Slide 10: Install Arduino Builder Sta...
Slide 11: Install Arduino Builder Step 3
Slide 12: Install Arduino Builder Step 4
Slide 13: Install Arduino Builder Step 5
Slide 14: Download a Firmware Rele...
Slide 15: Download Firmware Releas...
Slide 16
Slide 17
Slide 18
Slide 19
Slide 20: Install Firmware Release
Slide 21: Install Firmware Release St...
Slide 22
Slide 23
Slide 24
Slide 25
Slide 26
Slide 27
Slide 28
Slide 29
Slide 30
Slide 31
Slide 32
Slide 33: DONE

HOW TO UPDATE FIRMWARE RESPIMATIC 100

STEP-BY-STEP PROCEDURE

EQUIPMENT NEEDED

USB Cable	Windows Laptop	Respimatic 100
		
One end with micro-USB connector		

Figure 24: Step-by-step instructions on Updating Firmware

FiO₂ Calculator

FiO₂ is controlled externally by setting an appropriate Oxygen input flow rate from the external Oxygen source.

This stand-alone calculator enables calculating the required Oxygen flow rate given the following parameters.

- Deployment Altitude
- Desired FiO₂
- Purity of the Oxygen Source
- Tidal Volume
- Respiration Rate

Simply position the mouse over any gauge and turn the scroll wheel to change its value. Else, the required value can be typed in the center of each gauge.



Figure 25: FiO₂ Calculator

Index Table

4G/5G dongle.....	3	Range Slider	10, 17
Alerts View.....	13	recording	15
auto-connect.....	3	Respimatic Wi-Fi	4
Charts View	10	Selector table.....	17
configuration portal	4	Snapshots View.....	9
export	18	Statistics View	11, 12
FiO2	7, 21	Systems Table.....	7
import	18	SYSUID.....	5, 6
JSON.....	7, 15	X-axis.....	10
main menu	6	xml.....	7
portal time out.....	4		