



**INSPIRE-100** 

# Setting the Context

The Motivation

The Problem Statement

INSPIRE-100 (Patent Pending)

Is it right for you?

Need adult respiratory support?

Support from Initiation to Weaning?

No compressed air or piped Oxygen?

Connect to O, Cylinder or Concentrator?

Full range of Respiration parameters?

Breath Synchronization for Patient Comfort?

Remote monitoring capability?

Handle harsh-uncontrolled Environment?

Easy-to-use System?

**Budget Friendly?** 

### Respiration Assist Devices

Categories - Usage and Pricing











BiPAP Big Hole

**ICU Ventilator** 

### CPAP v/s BiPAP v/s INSPIRE-100 v/s ICU-VENTILATOR

CPAP	BiPAP	Respimatic 100	ICU Ventilator		
Continuous Positive Airway Pressure	Continuous Bi-Level Airway Positive	Mechanical Ventilation with 4 most-	Mechanical Ventilation with very		
Continuous i ositive All way i ressure	Pressure	used ventilation modes and controls	sophisticated modes and controls		
Non-invasive	Non-invasive	Non-invasive + Invasive	Non-invasive + Invasive		
High Flow + PEEP	Inspiratory Pressure + PEEP	Tidal Volume + Support Pressure + PEEP	Tidal Volume + Support Pressure + PEEP		
Useful for Type 1 respiratory Failure (Hypoxemic)	Useful for Type 2 respiratory Failure (Hypercopnic)	Useful for Hypoxemic and Hypercopnic respiratory failure	Useful for Hypoxemic and Hypercopnic respiratory failure		
Continuous flow of air at a constant pressure. Increases mean airway pressure to recruit collapsed alveoli	Continuous flow of air at different constant pressures during inspiration and expiration breathing phase	Independent control over the volume, the respiration rate and pressure	Independent control over the volume, the respiration rate and pressure		
Useful only when patient can breathe on his own	Useful only when patient can breathe on his own	Useful when patient can or CANNOT breathe on his own	Useful when patient can or CANNOT breathe on his own		
Only Spontaneous breaths that are patient triggered.	Only Spontaneous breaths that are patient triggered.	Spontaneous breaths + Mandatory breaths controlled by RR and I:E	Spontaneous breaths + Mandatory breaths controlled by RR and I:E		
External FiO2 control	External FiO2 control	System assisted FiO2 control	Direct FiO2 control		
Breath Synchronization N/A	Breath Synchronization N/A	Full Breath Synchronization	Full Breath Synchronization		
No Tidal Volume control	Indirect Tidal Volume control (IPAP-EPAP)	Direct Tidal Volume control	Direct Tidal Volume control		
No Respiration Rate control	No Respiration Rate control	Direct Respiration Rate control	Direct Respiration Rate control		
No Inspiration: Expiration ratio control	No Inspiration: Expiration ratio control	Direct Inspiration: Expiration control	Direct Inspiration:Expiration control		
External Humidity control	External Humidity control	External Humidity control	Direct Humidity control		
No display of Peak, Plateau or PEEP	No display of Peak, Plateau or PEEP	Full display of Peak, Plateau and PEEP	Full display of Peak, Plateau and PEEP		
Minimal alarm signals	Minimal alarm signals	Full set of Alarm signals	Full set of Alarm signals		
No remote monitoring	No remote monitoring	Sophisticated Remote WEB Dashboard	Minimal Remote monitoring (if any)		

## Observations on Ventilator Evolution



Iron Lung Age



Pneumatic Age



μController Age



Smart "E"-Age

# Most-used Ventilation Modes have not changed

- Volume and Pressure Control
- Control BPM, I/E, VT and PS
- Monitor pressures and flow
- Safety Alarm systems

# Diminishing Returns from what has evolved ...

- Exotic Ventilation modes
- Multitude of Sensors
- Fancy Touch-screen LCD Displays

# INSPIRE-100 Details

**System Components** 

Technical Details

# US and INDIA IP Protection



(19) United States

#### (12) Patent Application Publication (10) Pub. No.: US 2023/0001126 A1 Nanda et al.

(43) Pub. Date: Jan. 5, 2023

(54) VENTILATOR

- (71) Applicants: Sunil Nanda, Bangalore (IN); Pankaj Kumar Porwal, Udaipur (IN)
- (72) Inventors: Sunil Nanda, Bangalore (IN); Pankaj Kumar Porwal, Udaipur (IN)
- Dec. 21, 2021
- Foreign Application Priority Data

#### Publication Classification

(51) Int. Cl. A61M 16/20 A61M 16/00


(2006.01) (2006.01)

(52) U.S. Cl.

A6IM 16/204 (2014.02); A6IM 16/0078 (2013.01); A6IM 16/0003 (2014.02); A6IM 2205/52 (2013.01); A6IM 2016/0027 (2013.01); A6IM 2205/70 (2013.01)

#### ABSTRACT

Provided is a ventilator that includes a breathing system, a mechanical system coupled to breathing system, and a control system coupled to breathing system and mechanical system. The control system includes pressure sensors, processing circuitry, and memory configured to store a look-up table. The processing circuitry receives a set of values for plurality of parameters, identifies a compression value from a plurality of compression values in the look-up table based on the received set of values. The processing circuitry causes the mechanical system to compress a bag valvabreathing system in accordance with the identify sion value. The compression of the bag gaseous inhalant to flow through the breathi a time-interval. The processing circuitry actual volume of the gaseous inhalant and iter fies the compression value of the bag valve desired volume of the gaseous inhalant.



An important and necess and F 62 encountered in our system are of the order of tens of cmH<sub>2</sub>O wh is of the of a thousand cmH2O of pressure. At sea level, Patmosphere is O. Even at an altitude of 15,000 feet, . On the other hand, the  $P_{G1}$  and  $P_{G2}$  in the system

Patmosphere is appro range from 1

term is negligible compared to (2\*Patmosphere), even more so since it ued a square root. The flow equation can be simplified to the one below.

$$Q \propto \sqrt{\frac{(P_{G1} - P_{G2})}{Patmosphere}}$$


Recalling Equation 2 from the theory section above, this equation can be recast as below given that the orifice characteristics and pressure tap locations are the same for every system.

$$Q = C * \frac{\sqrt{(P_{G1} - P_{G2})}}{\sqrt{(Patmosphere)}}$$
 where  $\underline{C} = f(Re)$  Reynold's number

At a given geographical location, *Patmosphere* is also a constant. So, the above equation further reduces to the one below.

$$Q = \left(\frac{C}{\sqrt{Patmoshpere}}\right) * \sqrt{(P_{G1} - P_{G2})}$$

The equation needs further simplification to ease the computation burden of the square root computation for and inexpensive micro-controller. The constraints are as below.



# Our Solution INSPIRE-100

4 Commonly Used Ventilation Modes CMV, ACV, SIMV, PSV

Respiration Rate, Tidal Volume, PEEP, Pressure Support & FiO2 Controls

Volume Controlled and Pressure Supported
Breaths


Mandatory &
Spontaneous Breaths
with Full Breath
Synchronization

Complete set of WEB Apps
Remote Dashboard
Remote Recorder
Remote Analyzer

Low-speed Wi-Fi sufficient Phone Hot-spot sufficient

Uses secure HTTPS protocol

Field upgradable with new Firmware releases



# Breathing Circuit

Proprietary, <u>patent-pending</u> Pressure line connector with Orifice plate

COTS single-limb Breathing Circuit with NRBM

BVM or Ambu Bag with Reservoir

Pressure sensing lines

PEEP valve

**HME** Filter

Oxygen Source






Off-the-Shelf Single limbed Circuit with NRBM



Proprietary Dual
Pressure line Connector



# Front Panel The Human-Machine Interface

No delicate touch screen etc.

Easy to read 7-seg
LED Parameters
Display

Parameter selection using navigation arrow buttons

4-line LCD Display for displaying Messages and Menus

Peak, Plateau, PEEP pressures displayed after each breath

Shows Delivered
Volumes, Lung
Compliances, Breath
types etc.



# Dashboard Snapshot View


Anyone can monitor any patient via the WEB One-to-many and many-to-one monitoring

#### 5 Dashboard views

- Snapshots
- Breath Waveforms
- Charts
- Statistics
- Alerts (Audible or not)

Breath Range Selector on every view to display data for different breath ranges

 For instance, use to compare the statistics for the first hour of ventilation against the second hour.




## Dashboard Waveforms View

Pressure and Flow Graphs


For Selected Breaths


Periodic Display

Display on demand




# Dashboard Charts View





# Dashboard Statistics View





#### Parameters Measured

Parameter	Units	Min	Max	Avg
Peak Pressure	cmH20	27.0	30.0	28.6
Plateau Pressure	cmH20	17.0	29.0	27.1
PEEP Pressure	cmH20	5.0	5.0	5.0
Tidal Volume Delivered	ml	384.0	412.0	399.8
Total Minute Volume	litres/min	8.0	8.1	8.0
Mandatory Minute Volume	litres/min	8.0	8.1	8.0
Spontaneous Minute Volume	litres/min			
Mandatory BPM	bpm	20.0	20.0	20.0
Spontaneous BPM	bpm			
FIO2	%	21.0	21.0	21.0
Static DeltaV/DeltaP	ml/cmH20	17.0	32.0	18.3
Dynamic DeltaV/DeltaP	ml/cmH20	16.0	18.0	17.0
System Temperature	degC	27.0	27.0	27.0

#### **Miscellaneous Information**

Information	Value
Number of Breaths	73
Number of Mandatory Breaths	73
Number of Spontaneous Breaths	C
Number of Maintenance Breaths	C
Number of CMV-mode Spontaneous Breaths	C
Number of Missing Intervals (Packet loss)	C
Number of WiFi or Server Disconnects	C
Number of Notifications	C
Number of Warnings	C
Number of Errors	C

#### **Static Information**

Patient Name: Rajnikanth Bond Gender: Male Age: 69yr Weight: 74kg Height: 181cm

System Location: Namma Bengaluru Location Altitude: 3000 ft (915 mtrs) Location Atmospheric Oxygen: 19%

#### **Parameter Settings Used**

Parameter	Units	Values	
Ventilation Mode	mode	ACV	
Tidal Volume	ml	400	
Minute Volume	l/min	10	
Respiration Rate	bpm	20	
I:E Ratio	ratio	1:2	
PEEP Pressure	cmH20	5	
Maximum Pressure	cmH20	50	
Support Pressure	cmH20	20	
Support Pressure Termination	%flow,secs	20%	
FIO2	%	21	

#### **Sequence of Parameter Combinations**

) )	MODE	VT/MV	RR	I:E	PEEP	PMAX	PS	TPS	FIO2	# of BREATHS	Before BREATH#
)	?	?	?	?	?	?	?	?	?	1	0
)	ACV	400	20	1:2	5	50	20	20%	?	2	2
)	ACV	400	20	1:2	5	50	20	20%	21	70	4


# Dashboard Alerts View




# Analyzer

Any patient Session can be recorded locally or remotely.

The recorded Session can then be analyzed off-line using the Analyzer.





### Nurses' Station


# Synopsis of all systems

Online Status
Current State
Alarms

One System Active. Two Systems Inactive.



Alternating Views for Active Systems




Warning Alternating Views for Systems



# Updating Firmware

Firmware releases available on the WEB.

Step-by-step menu driven update procedure



### Ventilation Modes

### The 4 most frequently used

#### Continuous Mandatory Ventilation (CMV)

Volume Controlled Mandatory Breaths

Ignore spontaneous breaths

#### Synchronized Assist Control Ventilation (Sync ACV)

Volume Controlled Mandatory Breaths

Volume controlled breaths in response to spontaneous breaths

Breath Synchronization

#### Synchronized Intermittent Mandatory Ventilation (SIMV)

Volume Controlled Mandatory Breaths

Pressure supported breaths in response to spontaneous breaths

Breath Synchronization

## Pressure Support Ventilation (PSV)

Pressure supported breaths in response to spontaneous breaths

Monitoring of Minute Volume

Fallback to SIMV if insufficient Minute volume

# Volume Controlled Breaths (All modes)

#### Tidal Volume (ml)

200 to 600 ml increments of 50 ml

#### Respiratory Rate (bpm)

10 to 30 bpm increments of 1 bpm

#### Inspiration/Expiration Ratio (I:E)

1:1 1:2 1:3

#### PEEP (cmH<sub>2</sub>O)

4 to 15 cmH<sub>2</sub>O increments of 1 cmH<sub>2</sub>O

#### Max Pressure (cmH<sub>2</sub>O)

15 to 50 cmH<sub>2</sub>O increments of 5 cmH<sub>2</sub>O

#### FiO<sub>2</sub> Support

System Managed Externally Controlled 21% to 100%

# Pressure Supported Breaths

(SIMV & PSV modes)

#### Support Pressure (PS)

5 cmH<sub>2</sub>O to 35 cmH<sub>2</sub>O in increments of 5 cmH<sub>2</sub>O

#### Support Pressure Termination (TPS)

#### Flow-dependent

Terminate when flow falls to 10%, 20%, 30%, 40%, 50% or 60% of peak flow

#### Time dependent

Terminate after 1.0 to 2.5 secs in increments of 0.5 secs


# Breath Synchronization

#### **Both ACV and SIMV modes**

- A must for patient comfort
- Synchronize Mandatory breaths with Spontaneous breaths
- Prevent breath stacking

#### **Breath Syncing in SIMV mode**


There is a sync-window – the next mandatory breath is rescheduled only if spontaneous breath within the sync-window



#### Breath Syncing in Synchronized AC Mode

There is no sync-window - the next mandatory breath is always rescheduled after a spontaneous breath

Example below: Tidal Volume = 500ml Respiration Rate = 15 bpm



# FiO<sub>2</sub> Settings


Without the Reservoir bag, FiO<sub>2</sub> delivered is the Atmospheric O<sub>2</sub> content at site


FiO<sub>2</sub> delivery with the Reservoir bag is mathematically modelled, calibrated and verified in the Lab to provide +/- 5% accuracy

Front-panel guides the user in setting the appropriate input  $O_2$  flow rate from the  $O_2$  source for a given  $FiO_2$ 

The mathematical model provides for a possible  $O_2$  concentrator as an  $O_2$  source (purity < 100%)


Online Web-accessible FiO<sub>2</sub> calculator is also provided for exploration purposes





### Alarms and Safety Features


- Enforcement of Pre-use checks
- Maintenance Breaths till Alarm situation rectified
- Alarms, Warnings and Notifications
  - Max Pressure Alarm
  - Pressure Leak Alarm
  - Pressure Loss Alarm
  - Airway Blockage Alarm
  - System Temperature Alarm
  - Sensor Failure Alarm
  - Breathing Circuit Failure Alarm
  - Detect coughing/hiccupping fits
  - Inconsistent input parameters
  - Extreme parameter combination warnings
  - And many more ...




# Automated Testing

#### **AutoTest Controls**

- Select PEEP
- Run all combinations of Selected ranges of
  - Ventilation Modes
  - Respiration Rates
  - I:E ratios
- 3. # Breaths per Combo
- 4. % Spontaneous Breaths




### Some Results




Sample FiO2 Data									
	PEEP=6,EI=2	2	Measured	Calculated	ERROR				
VT	RR	O2 (I/min)	FiO2(%)	FiO2(%)	%				
400	20	2	40.5	39	4%				
400	20	2.5	45.5	43.5	4%				
400	20	3 3.5	49	48	2%				
400	20		54	53.5	1%				
400	20	4	59.8	57	5%				
400	20	4.5	63.5	61.5	3%				
400	400     20       400     30       400     30       400     30		67.5	66	2%				
400			33	33	0%				
400			36.5	36	1%				
400			38.5	39	1%				
400	30	3.5	41.5	42	1%				
400	30	4	44.5	45	1%				
400	30	4.5	47.1	48	2%				
400	30	5	49.8	51	2%				

#### VT Titration for VT = 200, 300 and 400ml



#### A Mix of VC and PS Breaths



# Thank you