


INSPIRE-100

Emergency Respiration Assist Device



Web Applications Manual



Table of Contents

Introduction	3
Setting up Wi-Fi credentials	3
Launching the WEB Apps	7
WEB Dashboard	11
Dashboard Audible Alarms	12
Dashboard Snapshots View	13
Dashboard Statistics View	16
Dashboard Alerts View	17
Dashboard Waveforms View	18
Dashboard Recording	19
WEB Analyzer	20
Analyzer Sidebar Menu	20
Analyzer Recording Selector	21
Analyzer Range Slider	22
Analyzer Recording Import Export	22
Analyzer Charts, Statistics, Breath Waveforms and Alerts Views	23
Updating System Firmware	24
FiO ₂ Calculator	26

Table of Figures

Figure 1: Enable/Disable Wi-Fi login.....	3
Figure 2: Wi-Fi Autoconnect	4
Figure 3: Connect to INSPIRE-100 Wi-Fi network.....	4
Figure 4: Login OTP Selection	5
Figure 5: Wi-Fi Configuration Portal	5
Figure 6: Wi-Fi Login Successful.....	6
Figure 8: INSPIRE-100 Systems Table.....	7
Figure 7: Web Apps Main Menu	7
Figure 9: Login Message Popup.....	8
Figure 10: Add a new system to Systems Table	9
Figure 11: Add a new System using OTP.....	9
Figure 12: INSPIRE-100 Documentation	10
Figure 14: Dashboard Menu	11
Figure 13: Breath Range Selector	11
Figure 15: Setting Audio Alarms	12
Figure 16: Dashboard Snapshots View	13
Figure 17: Dashboard Charts View	14
Figure 18: Charts Edit Menu	15
Figure 19: Dashboard Statistics View	16
Figure 20: Dashboard Alerts View	17
Figure 21: Breath Detailed Pressure and Flow Graphs	18
Figure 22: Breath Type Menu	18
Figure 23: Dashboard Recording View.....	19
Figure 24: Analyzer Sidebar Menu.....	20
Figure 25: Analyzer Selector Table.....	21
Figure 26: Analyzer Breath Range Selector.....	22
Figure 27: Analyzer Import View.....	22
Figure 28: Firmware Update Web Application	24
Figure 29: Step-by-step instructions on Updating Firmware	25

Introduction

In addition to the front panel, INSPIRE-100 provides the ability to remotely monitor all respiration sessions via a WEB dashboard. Doctors and technicians can use the dashboard to connect to any INSPIRE-100 system using a unique system ID embedded in each system. This feature is useful to enable a remote specialist to observe the key system and patient parameters during a session and suggest a course of action for the local practitioners.

A brief overview of the process is as follows.

1. There must be a Wi-Fi network at the site where the INSPIRE-100 system is deployed. If required, use a 4G/5G dongle to establish a Wi-Fi network. One dongle can serve multiple systems at the same site at the same time.
2. There must be Wi-Fi or wired internet at the monitoring site.
3. Enable WEB dashboard monitoring on the INSPIRE-100 system.
4. Allow the INSPIRE-100 system to log on to the Wi-Fi network.
5. Visit the provided URL at the monitoring site.
6. Pair the WEB dashboard at the monitoring site with the INSPIRE-100 at the deployment site using the INSPIRE-100 System Unique ID.
7. WEB Dashboard allows monitoring only. It does not permit remote control of the system.

During the start-up sequence, the system gives an option to enable or disable remote monitoring for




Figure 1: Enable/Disable Wi-Fi login

the system. Once enabled, the system guides the user to set up the remote WEB Apps.

Setting up Wi-Fi credentials

INSPIRE-100 system remembers the history of prior Wi-Fi networks that have been used by the system. The user is provided with an option to either auto-connect to a previously known Wi-Fi network or to configure a new one or to use an OTP. If desired, the recorded Wi-Fi history can be erased at this time.




Figure 2: Wi-Fi Autoconnect

If Auto-connect is not enabled or if Auto-connect fails, the next option is to use a configuration portal to setup a new Wi-Fi network for the system to log in to. To enable login through a configuration portal, the system sets up a local, temporary Wi-Fi network named “INSPIRE-100 Wi-Fi”. The user can use either a wifi-enabled laptop or smartphone to log on to this network.




Figure 3: Connect to INSPIRE-100 Wi-Fi network

Upon login to this “INSPIRE-100 Wi-Fi” network, a portal screen is automatically presented on the laptop or the smartphone which guides the user step-by-step to enabling system to login to a desired Wi-Fi network. The portal time out in 2 minutes if unable to log on for whatever reason. If the system times out, the user can retry as many times as desired.

In case the portal does not automatically open, open a browser and navigate to 192.168.1.4 (URL) after connecting to the “INSPIRE-100 Wi-Fi” network.

WiFi login has two options as below.




Figure 4: Login OTP Selection

1. Anonymous

No message is sent to the Web Apps portals to announce the login. Only those browsers with prior knowledge of the particular system can connect to it.

(See section on Known Systems Table)

2. Broadcast

A message is sent to all Web Apps portals to announce the login. An OTP will be generated and displayed on the system's Front-panel. Only those browsers with knowledge of the OTP (One Time Password) system can connect to it.

(See section on Known Systems Table)

INSPIRE-100
Wi-Fi Login

Wi-Fi Name (SSID)


Wi-Fi Password

LOGIN

Figure 5: Wi-Fi Configuration Portal

If the Wi-Fi login was unsuccessful, the user can still continue without enabling a WEB dashboard.

Upon a successful Wi-Fi login, the system is now ready to be connected to the WEB Apps using the unique system id (SYSUID). The SYSUID is a 20 character string starting with the prefix "UID_" followed by 16 hexadecimal digits. Another option is to connect using the OTP displayed as below. Further details can be found in the next section.



Use Unique SysUID
UID_0123456789ABCDEF
OTP is 4728
Noted (YES) ?

Figure 6: Wi-Fi Login Successful

Launching the WEB Apps

<https://www.inspire-100.com>




Figure 8: Web Apps Main Menu







System Tag	System UID	FW Version	Actions
BANGALORE	UID_28615E07D6013C4A	3.1.1	 
PROTOTYPE 1	UID_28705904000000EE	3.1.1	 
UDAIPUR	UID_28B1A879A2010387	unknown	 

Figure 7: INSPIRE-100 Systems Table

IMPORTANT

Use CTRL key and +/- keys to zoom in/out
or hold down the CTRL key and use the mouse wheel to zoom in/out
till the content fits well in the browser screen.

The screenshot above on the left is the main portal menu while the screenshot on the right above is the Systems Table.

All the Web Apps communicate with a particular INSPIRE-100 system via a Unique System ID (UID). Each INSPIRE-100 system has a built-in UID which is 20 characters long (e.g. UID_AAAABBBBCCCCDDDD). This UID is displayed on the system's front panel upon a Wi-Fi login. It can also be accessed via the STANDBY menu on the system.

Each time a system logs in, the main portal page displays a popup message as below.




Figure 9: Login Message Popup


The main Inspire-100.com portal holds a table of the INSPIRE-100 systems it recognizes. Initially this table is empty. Once populated, it is accessible for all times after that. The table is accessed via the “Manage Known Systems” button.

Each unique system id (SYSUID) of the INSPIRE-100 systems can be associated with an easy to remember name tag. The browser remembers the history of all the SYSUIDs that have been used and presents them in the dropdown list in the main menu box above.

The table of name tags and associated SYSUID (Systems Table) is accessed through the “Manage Known Systems” button. The + menu button on the Systems Table adds a system. To select a system to communicate with either double click on the appropriate row or use the checkmark button against the row. A system can be removed from the table using the trash menu button in the appropriate row. The trash button on the top right removes all system information. The systems table can also be exported as a JSON xml file and can be imported from a JSON xml file by clicking the appropriate icons on the table banner. Finally, the back arrow menu button on the top left can be used to navigate back to the main menu.

There are two ways to populate the systems table.

1. If you know the UID of your system(s), simply add the information using the “Add New System” button. Each system can be assigned a tag name so that it is easy to remember and access. The popup for “Add New System” is shown below.



ADD new System


System UID

System TAG

ADD System CANCEL

Figure 10: Add a new system to Systems Table

2. Every time a recognized or unrecognized system logs in, there is a message displayed on the portal web page. In case of an unrecognized system, the portal popup message provides an option to add the unrecognized system to the systems table using an OTP displayed on the system's front panel.



ADD Detected System using OTP

OTP from System Front Panel

ENTER OTP CANCEL

Figure 11: Add a new System using OTP

The following WebApps are available via the links on the main menu.

- **Launch Dashboard** - Monitor a session remotely. It requires a SysUID to be selected to determine the INSPIRE-100 system to connect to.
- **Launch Analyzer** - Analyse a previously recorded session. It requires a SysUID to be selected to determine the INSPIRE-100 system to connect to.
- **Update System Firmware** – Download and install a new release of firmware for the INSPIRE-100 system.

- **Calculate FiO2 Settings** – Calculate required Oxygen inflow rate. It does not require a SysUID.
- **View Documentation** – All the pdf documents are accessible through this link. It does not require a SysUID. If a “PDF Viewer” extension is loaded in your browser the documents can be read online or else, they can be downloaded. Figure below is a screenshot of the document’s web page.



Figure 12: INSPIRE-100 Documentation

WEB Dashboard

The live WEB dashboard, when connected, offers a choice of six different views. The user can switch between these views at any time. Figure below shows the sidebar for the dashboard.




Figure 13: Dashboard Menu

The box on the top shows some relevant summary data for the current session.

The next box allows for selecting the type of view to display. Available views are as below.


1. Snapshots view
2. Charts view
3. Statistics view
4. Alerts view
5. Breath Shapes View
6. Recording View

The next box allows for turning on/off session recording and dashboard updates. Pausing dashboard updates only stops the display from updating, fresh data continues to get collected and will get displayed once the 'Pause Dashboard' is turned OFF.

The box at the bottom displays all FiO₂ related settings. These settings are only for monitoring, the user must use the physical system's control panel to change these settings like all other settings.

The Figure below shows a Slider that can be used to select the range of breath numbers to display the required data for. The breath number range can be selected by grabbing the slider handles and sliding then to the required breath numbers. The range can then be committed using the checkmark button. Once committed the display is frozen to the selected range. The rolling button on the extreme left discards the selected range and enables a fresh display update on each new breath. The double arrow button selects the entire range and freezes the display to that.

Figure 14: Breath Range Selector



Dashboard Audible Alarms

By default, all audible alarms are turned OFF. The audio alarms can be selectively turned ON/OFF using the “Control Audio Alarms” button.

Below is a screenshot of the Audio alarm settings.




Figure 15: Setting Audio Alarms

Dashboard Snapshots View

Below is a screenshot of the WEB Dashboard “Snapshots View”. All the INSPIRE-100 parameters, both input and output, are presented on the dashboard for easy viewing as a cohesive whole.

Note that for safety reasons, the Dashboard only allows monitoring of a remote system. The dashboard cannot be used to control the remote system. That must be done using the front panel of the physical INSPIRE-100 system.




Figure 16: Dashboard Snapshots View

Dashboard Charts View

The Dashboard also provides an option for a “Charts View”. A screenshot of the charts view is shown below. This screenshot shows three chart boxes. A chart box can be added at any time using the + menu button on the top left of the chart box. Use the trash menu button on the top right to delete a chart box.

The parameters to chart can be selected using the checkboxes on the edit menu. The edit menu button is also on the top right of each chart box. The charts are updated after every breath. The X-axis can be selected as breath number or as elapsed time between breaths.

By default, the system charts the selected parameters for the past 60 breaths on a rolling basis, the charts are updated after each breath.




Figure 17: Dashboard Charts View

There is a Range Slider above the chart boxes to pan or zoom to any desired range of breath numbers. Use the handles on the Slider to select a particular range. If a particular range is selected, further chart updates are disabled till the range is reset using the RESET button on the range slider, the new parameter values are still stored for later display. Upon RESET, the chart boxes return to their default display mode.

The chart box edit menu is shown below. Any collection of the shown parameters can be displayed in any chart box or in multiple chart boxes.

Title

Y-axis Parameters

- ☐ Pressures
 - ☐ Peak Pressure
 - ☐ Plateau Pressure
 - ☐ PEEP Pressure
- ☐ Volumes
 - ☐ Tidal Volume
 - ☐ Total Minute Volume
 - ☐ Mandatory Minute Volume
 - ☐ Spontaneous Minute Volume
- ☐ Breaths Information
 - ☐ Mandatory BPM
 - ☐ Spontaneous BPM
- ☐ Breath Type
 - ☐ Mandatory Breath
 - ☐ Spontaneous Breath
 - ☐ Maintenance Breath
- ☐ Instant Lung DeltaV/DeltaP
 - ☐ Dynamic Lung DeltaV/DeltaP
 - ☐ Static Lung DeltaV/DeltaP
- ☐ Oxygen Related
 - ☐ FI02
 - ☐ Incoming O2 Flow Required
 - ☐ Incoming O2 Purity
- ☐ Miscellaneous
 - ☐ System Errors
 - ☐ System Warnings
 - ☐ System Notifications
 - ☐ System Temperature

X-axis Units

☒ Breath Number ☐ Breath Times

Figure 18: Charts Edit Menu

Dashboard Statistics View

The Dashboard also provides an option for a “Statistics View”. A screenshot of the statistics view is shown below.

The statistics are collected for the entire range updated after every breath by default. The range slider can be used to gather statistics for any range of breath numbers. The RESET menu button on the range slider causes the updates and display to go back to their default mode.

Breath Range Selector

Parameters Measured

Parameter	Units	Min	Max	Avg
Peak Pressure	cmH2O	27.0	30.0	28.6
Plateau Pressure	cmH2O	17.0	29.0	27.1
PEEP Pressure	cmH2O	5.0	5.0	5.0
Tidal Volume Delivered	ml	384.0	412.0	399.8
Total Minute Volume	litres/min	8.0	8.1	8.0
Mandatory Minute Volume	litres/min	8.0	8.1	8.0
Spontaneous Minute Volume	litres/min	----	----	----
Mandatory BPM	bpm	20.0	20.0	20.0
Spontaneous BPM	bpm	----	----	----
FIO2	%	21.0	21.0	21.0
Static DeltaV/DeltaP	ml/cmH2O	17.0	32.0	18.3
Dynamic DeltaV/DeltaP	ml/cmH2O	16.0	18.0	17.0
System Temperature	degC	27.0	27.0	27.0

Miscellaneous Information

Information	Value
Number of Breaths	73
Number of Mandatory Breaths	73
Number of Spontaneous Breaths	0
Number of Maintenance Breaths	0
Number of CMV-mode Spontaneous Breaths	0
Number of Missing Intervals (Packet loss)	0
Number of WiFi or Server Disconnects	0
Number of Notifications	0
Number of Warnings	0
Number of Errors	0

Static Information

Patient Name: Rajnikanth Bond

Gender: Male Age: 69yr

Weight: 74kg Height: 181cm

System Location: Namma Bengaluru

Location Altitude: 3000 ft (915 mtrs)

Location Atmospheric Oxygen: 19%

Parameter Settings Used

Parameter	Units	Values
Ventilation Mode	mode	ACV
Tidal Volume	ml	400
Minute Volume	l/min	10
Respiration Rate	bpm	20
I:E Ratio	ratio	1:2
PEEP Pressure	cmH2O	5
Maximum Pressure	cmH2O	50
Support Pressure	cmH2O	20
Support Pressure Termination	%flow,secs	20%
FIO2	%	21

Sequence of Parameter Combinations

MODE	VT/MV	RR	I:E	PEEP	PMAX	PS	TPS	FIO2	# of BREATHS	Before BREATH#
?	?	?	?	?	?	?	?	?	1	0
ACV	400	20	1:2	5	50	20	20%	?	2	2
ACV	400	20	1:2	5	50	20	20%	21	70	4

Figure 19: Dashboard Statistics View

Dashboard Alerts View

The Dashboard also provides an option for a “Alerts View”. A screenshot of the alerts view is shown below.

By default, it displays the complete history of errors and warnings encountered in the current session. The range slider on top can be used to see the alerts for any range of breath numbers. The RESET menu button causes the updates to go back to their default mode.



Figure 20: Dashboard Alerts View

Dashboard Waveforms View

The Dashboard also provides an option to view the detailed pressure and flow waveforms for selected breaths. Figure below shows a sample of such a view.




Figure 21: Breath Detailed Pressure and Flow Graphs

The EDIT icon on each box allows the selection of the kinds of breaths to display. The graphs are color coded as per the legend at the bottom of the page. The breath selection menu is shown in the Figure on the left.

Figure 22: Breath Type Menu

Dashboard Recording

The Dashboard also provides an option to record any part of the current session using the “Start Recording” menu button on the sidebar menu. The recording can be paused at any time causing that paused window to not be recorded. The Recording at the bottom left of the Snapshots view indicates whether the recording is currently active.

The recording is stored in a JSON database on the disk of the laptop or the desktop that the browser is running on. This recording can be analysed at any future time using the Analyzer WebApp. A screenshot of the recording view is shown below.





Figure 23: Dashboard Recording View

WEB Analyzer

The Analyzer enables the analysis of a previously recorded session with a patient. The process starts with selecting a session recording to analyse.



Analyzer Sidebar Menu

The main menu for the Analyzer is presented as a sidebar menu as shown in the Figure on the left.


The box on the top provides a summary of the currently selected recording for analysis.




Figure 24: Analyzer Sidebar Menu

Analyzer Recording Selector

Each previously recorded session is presented in a Selector table. Select a recording for analysis either by double-clicking on a row or using the appropriate checkmark menu button.


After a database is selected, a summary of the recording data is displayed in the top box of the sidebar and the selected table row is highlighted.



INSPIRE-100 Recordings		
Recording Name	Created	Actions
Bugs Bunny	08-01-2024 11:09:03	  

SYSTEM UID
UID_28615E07D6013C4A
(BANGALORE)

INSPIRE-100
Analyzer



tekMedika

Figure 25: Analyzer Selector Table

Analyzer Range Slider




Figure 26: Analyzer Breath Range Selector

By default, the analysis window is the entire recording. To select a particular breath number range to analyse, use the handles of the Range Slider shown below to zero in on the interval of interest. All Analysis actions use the currently selected analysis range. The range selector works in the same fashion as described earlier in the Dashboard section.

Analyzer Recording Import Export

The EXPORT menu button on each row enables the user to export the database to a text file that can be sent to others for analysis. The IMPORT button on the top left allows the user to import a previously exported text file as a new session available for analysis. Below is a screenshot of the Import screen.




Figure 27: Analyzer Import View

Analyzer Charts, Statistics, Breath Waveforms and Alerts Views

Finally, the Charts, Statistics, Breath Shapes and Alerts views work in the same fashion as described in the Dashboard section above.

CONFIDENTIAL

Updating System Firmware

Occasionally the system may need to be upgraded in the field for new features or bugs. The new firmware releases are made available on the web for downloading and installing. The following menu interface is a part of the Web Applications to enable the end user to accomplish these upgrades.

For further details click on the Step-by-step Instructions on this application.



Version	Release Date	Get
3.1.1	13-Dec-2023	
1.1.1	16-Oct-2023	
1.0.1	16-May-2023	

Figure 28: Firmware Update Web Application

Below is a screenshot of step-by-step instructions.

The screenshot displays a presentation interface. On the left is a vertical table of contents with 32 slides. The main content area shows the title slide and the 'EQUIPMENT NEEDED' slide.

Table of Contents (Left Sidebar):

- Slide 1: INSPIRE-100 How to Updat...
- Slide 2: Equipment needed
- Slide 3: Firmware Update Overview
- Slide 4: Download Arduino Builder
- Slide 5: Download Arduino Builder S...
- Slide 6
- Slide 7
- Slide 8: Install Arduino Builder
- Slide 9: Install Arduino Builder Step 1
- Slide 10: Install Arduino Builder Ste...
- Slide 11: Install Arduino Builder Step 3
- Slide 12: Install Arduino Builder Step 4
- Slide 13: Install Arduino Builder Step 5
- Slide 14: Download a Firmware Rele...
- Slide 15: Download Firmware Releas...
- Slide 16
- Slide 17
- Slide 18
- Slide 19: Install Firmware Release
- Slide 20: Install Firmware Release St...
- Slide 21
- Slide 22
- Slide 23
- Slide 24
- Slide 25
- Slide 26
- Slide 27
- Slide 28
- Slide 29
- Slide 30
- Slide 31
- Slide 32: DONE

Title Slide:

INSPIRE-100

How to Update Firmware

STEP-BY-STEP PROCEDURE

tekMedika

EQUIPMENT NEEDED


USB Cable	Windows Laptop	INSPIRE-100
 One end with micro-USB connector		

Figure 29: Step-by-step instructions on Updating Firmware

FiO₂ Calculator

FiO₂ is controlled externally by setting an appropriate Oxygen input flow rate from the external Oxygen source.

This stand-alone calculator enables calculating the required Oxygen flow rate given the following parameters.

- Deployment Altitude
- Desired FiO₂
- Purity of the Oxygen Source
- Tidal Volume
- Respiration Rate

Simply position the mouse over any gauge and turn the scroll wheel to change its value. Else, the required value can be typed in the center of each gauge.

INSPIRE-100 FiO₂ Calculator

Required Incoming O₂ Flow

0.0 (litres/min)

Altitude: feet

Desired VT (ml):

Desired RR (bpm):

O₂ Purity (%):

Desired FiO₂ (%):


 **tekMedika**

Figure 25: FiO₂ Calculator

Index Table

4G/5G dongle.....	3	portal time out.....	4
Alerts View	17	Range Slider	14, 22
auto-connect.....	3	recording.....	19
Charts View	14	Respimatic Wi-Fi	4
configuration portal	4	Selector table.....	21
export.....	22	Snapshots View.....	13
FiO2	10, 26	Statistics View.....	16
import	22	Systems Table	8
JSON	8, 19	SYSUID.....	6, 8
main menu	7	X-axis.....	14
OTP.....	6, 9	xml	8