27/1/25.

Laser



Grating Spectrometer.



? पर साइड (-ve)

SAME side पर (+ve)

grating property

related to λ

Modern spectrometer.

Reflection grating is happening.

# of groves: 100/mm.



Path $\Delta = d(\sin\alpha \pm \sin\beta) = m\lambda$

Maximum reflection

Specular reflection.

$i = r.$

$\alpha - \theta = \beta + \theta.$

$2\theta = \alpha - \beta.$

$\theta = \left(\frac{\alpha - \beta}{2}\right)$

Blaze angle

Intensity Distribution.

$$I = RI_0 \cdot \frac{\sin^2 \frac{N\phi}{2}}{\sin^2 \phi/2} \quad \phi = \frac{2\pi}{\lambda} d(\sin\alpha \pm \sin\beta).$$

Maximum condition: $\phi = 2m\pi.$ $\left(\lim \frac{0}{0}\right).$

$$\frac{\left(\frac{N\phi}{2}\right)^2}{\left(\phi/2\right)^2}$$

Minimum condition: $\frac{N\phi}{2} = l\pi$

$l = \{1, 2, 3 \ldots N-1\}$ $\approx I = RI_0 N^2$

For $N = 5.$

# of minima = 4.

1, 2, 3, 4



$\beta \Rightarrow$ Diffraction $\angle$.



$$\sin(\beta_m + \epsilon) = \sin\beta_m \cos\epsilon + \cos\beta_m \sin\epsilon.$$

$$\cancel{= \beta_m \cos\epsilon + \sin\epsilon \text{ for small}}$$

$$= \epsilon \cos\beta_m + \sin\beta_m$$

$$\phi = \frac{2\pi}{\lambda} d\left[\sin\alpha + \sin\beta_m + \epsilon\cos\beta_m\right].$$

$$I = I_0 R_0 \frac{\sin^2 \frac{N\phi}{2}}{\sin^2 \phi/2}$$

for max $\Rightarrow \phi = 2m\pi + \frac{2\pi}{\lambda} d\,\epsilon \cos\beta_m$

$= 2m\pi + \delta$

$$\boxed{\delta = \frac{2\pi d}{\lambda}\epsilon \cos\beta_m.}$$

For max: $(N\phi/2)$

$$I = I_0 R_0 N^2 \frac{\sin^2 \frac{N\delta}{2}}{\left(\frac{N\delta}{2}\right)^2}$$

$\frac{N\delta}{2} = \pi$.

Total reflected power from grating.

Depends on reflectivity only.

$$\epsilon = \frac{\lambda}{N d\cos\beta_m}$$

$\frac{N\pi}{\lambda} d\epsilon\cos\beta_m = \pi$.

$\epsilon = \frac{\lambda}{N d\cos\beta_m}$

For large N, per group power dist more & less each one gets.
But total power is invariant of N

Power of resolution:

Rayleigh's criterion:



Resolution / Resolving power.

$d(\sin\alpha + \sin\beta) = m\lambda.$

Angular dispersion: $\frac{d}{d\lambda}\beta \Rightarrow \left(d\cos\beta \frac{d\beta}{d\lambda}\right) = m.$

$$\frac{d\beta}{d\lambda} = \frac{m}{d\cos\beta}.$$



$$\frac{d\beta}{d\lambda}\Delta\lambda = \frac{\lambda}{N d\cos\beta}$$

→ # grooves that are illuminated

$$\frac{\lambda}{\Delta\lambda} = \frac{m N d\cos\beta}{d\cos\beta}$$

$$\boxed{\frac{\lambda}{\Delta\lambda} = mN}$$

(*) Intensity – Resolution tradeoff

Grating Spectrometer. 0.8 | 1 | 0.5

Types of grating:



Δ path

$$d(\sin\alpha \pm \sin\beta) = m\lambda.$$

1. Littrow grating [design]. → Not isolated occurrence

$\alpha = \beta.$

$\therefore 2d\sin\alpha = m\lambda.$

[wavelength selective grating].

2. Echelle grating: (for higher order 'm').



$$\boxed{2d\sin\alpha = m\lambda.}$$

$$\alpha = 90 - \theta.$$

$$\boxed{2d\cos\theta = m\lambda}$$

(Ex) Grating size : 10×10 cm^2.

N = 10^3 grooves/mm.

m(order) = 2.

$\frac{\lambda}{\Delta\lambda} = 2 \times 10^{6}$

$\lambda = 500$ nm.

$\Delta\lambda = \cancel{2.5 \times 10^{-?}}$ ~~nm~~,
2.5×10^{-8} nm.

$\cancel{0.5 \times 10^{-8} \times 10}$

Littrow grating

$$\Delta\lambda = \left(\frac{\lambda}{a} + \frac{b}{f}\right)\left(\frac{d\theta}{d\lambda}\right)^{-1}.$$

$\lambda = 500$ nm.
$\beta = \alpha = 30^\circ$
$f = 1$ m.
$b = 50$ μm
$m = 2$, $\cancel{d} = 10^{3}$ ~~mm~~

$\frac{d\beta}{d\lambda}$ → corresponds to δ (deviation) as diffraction deviation.

$$\frac{d\beta}{d\lambda} = \frac{m}{d\cos\beta} = \frac{2\,d\sin\beta}{d\cos\beta\,\lambda}$$

$$= \frac{2\tan\beta}{\lambda}$$

$$= \frac{2}{\sqrt{3}}$$

$$\frac{500 \times 10^{-9}}{10^{-3}}\ \text{m}^2.$$

$\cancel{+50 \times 10^{-6}}$

$$\left[\frac{500 \times 10^{-9}}{10^{-3}} + 50 \times 10^{-6}\right]\left(\frac{2}{\sqrt{3} \times 500 \times 10^{-9}}\right)^{-1}$$

$$\cancel{\frac{550 \times 500\sqrt{3} \times 10}{2}} = 238156.98 \times 10^{-56} \times 10$$

$= 0.23$.

$$\left(550 \times 10^{-6}\right)\left[\frac{2}{\sqrt{3} \times 500 \times 10^{-9}}\right]^{-1}.$$

$$= \frac{550 \times 500 \times \sqrt{3} \times 10^{-?}}{2}\ \text{nm}$$

$$= 238156.98 \times 10^{-7}$$

$\Delta\lambda$ = $\boxed{\cancel{0.0228\ \text{nm}}}$ → $\boxed{0.0238\ \text{nm}}$

Angular dispersion

focal length = 1000 mm.

$$\frac{dx}{d\lambda} = f\,\frac{d\beta}{d\lambda}.$$

$$\frac{dx}{d\lambda} = f\,\frac{2\tan\beta}{\lambda}$$

$$= \frac{2f\tan\beta}{\lambda}.$$

More focal length (f), less resolution ($\Delta\lambda$).

$$\frac{dx}{d\lambda} = \frac{2 \times 10^{3}}{\sqrt{3} \times 500 \times 10^{-6}}$$

$= .0023 \times 10^{9}$
$= 2.31 \times 10^{?}$

$\frac{d\lambda}{dx}$ ⇒ কতটা λ change হবে যদি x 1 unit change হয়।

Detector.

(1) Thermal Detector

* Resistance changes with temperature
* λ insensitive

(2) Photodetector:

* change in band gap.
* λ sensitive.
* current starts when e^- goes from valence to conduction band.

29/9/25

Characteristics of Detectors

(1.) Spectral response $(R(\lambda))$: (1.) spectral ~~range~~ range.
(2.) Two relative intensity

$$= \frac{I(\lambda_1)}{I(\lambda_2)} = \frac{R(\lambda_1)}{R(\lambda_2)}$$

(2.) Sensitivity :

$\left(\frac{\text{Volt}}{\text{watt}}, \frac{\text{Amp}}{\text{watt}}, \frac{\text{Volt}}{\text{Area} \times \text{irradiance}}\right)$

(3.) Signal to noise ratio (S/N):

i. Electronic noise
ii. Thermal noise.

* NOISE equivalent of input power → power level for which $S/N = 1$. [NEP].

(4.) Time Response : Time dependent measurements

$R_2 C \ll \tau$

$R_2 C \ll \frac{1}{f}$

$f \gg [R_2 C]^{-1}$

τ : pulse time.

(5.) Cost of the Detector :

THERMAL DETECTOR

* Power measurement of CW laser.
* output energy of pulse laser.



H: Heat capacity of detector.
G: Thermal conductance to sink

$$\beta P = H\frac{dT}{dt} + G(T - T_s).$$

$$\frac{\beta P + G T_s}{H} = \frac{H}{H}\frac{dT}{dt} + \frac{G}{H}T.$$

$$\frac{dT}{dt} + \left(\frac{G}{H}\right)T = \left(\frac{\beta P + G T_s}{H}\right)$$

$$\ln(IF) = \int \frac{G}{H}\, dt \Rightarrow IF = e^{Gt/H}$$

$$T e^{Gt/H} = \int \left(\frac{\beta P + G T_s}{H}\right) e^{Gt/H}\, dt.$$

$$\left[T e^{Gt/H}\right]_{u=T_s}^{u=T} = \left[\frac{\beta P + G T_s}{\cancel{H}}\cdot\frac{\cancel{H}}{G}\, e^{Gt/H}\right]_{t=T_s}^{u=T}$$

$$\cancel{T(t) = b/a + \ldots e^{-Gt/H}}$$

$$T(t) = b/a + (T_s - b/a)e^{-at}$$

$$T(t) = \frac{\beta P + G T_s}{\cancel{H}}\cdot\frac{\cancel{H}}{G} + \left[T_s - \left(\frac{\beta P + G T_s}{H}\cdot\frac{H}{G}\right)\right]e^{-Gt/H}$$

$$T(t) = \left(\frac{\beta P + G T_s}{G}\right) + \left[T_s - \frac{\beta P + G T_s}{G}\right]e^{-Gt/H}.$$

$u = T$. $T = T_s + \frac{\beta P + G T_s}{G}\left(1 - e^{-Gt/H}\right).$

$$T = T_s + \frac{\beta P + G T_s}{G}\left(1 - e^{-Gt/H}\right)$$

Response time = G/H.

Rise in temp gives sensitivity.

$$\cancel{H\frac{dT}{dt} + T = \left(\frac{\beta P + G T_s}{G}\right)}.$$

$$P = P_0 (1 + a\cos(\Omega t)).$$

$$T = T_S + \Delta T [1 + a\cos(\Omega t + \phi)].$$

$$\Delta T = \frac{a\beta P}{\sqrt{1 + \frac{\Omega^2 H^2}{G^2}}}$$

$$\tan\phi = \frac{\Omega H}{G}.$$

$$\phi = \tan^{-1}\left(\frac{\Omega H}{G}\right).$$

Sd dt 8

For continuous laser

$$\int_0^t \beta P \, dt = H \int_0^t dT$$

$$\beta P t = H \Delta T$$

$$\Delta T = \left(\frac{\beta P t}{H}\right).$$

Some examples of Thermal detector

(1.) Thermistor.
(2.) Bolometer.
(3.) Golay cell
(4.) Thermocouple detector