

产品图片 Product show



产品包装Packing

Carton size (纸箱尺寸) : 58cmx38cmx43cm
50pcs/cartons 每箱50件
Gross Weight/Carton(毛重/每箱) : 12.5kgs/箱
空白纸箱 , 根据客户要求贴箱贴



Plastic bag size(胶袋尺寸) : 30CMX45CM



Enterprise qualification

International Certification Registrar International Certification Registrar



Certificate

No. ICR Polska/P4440131



Name and address of
certificate owner:

Anhui Me In Garments Co., Ltd.
Er Huang Road, Xian Sheng Dian Township, Jin'an District, Lu'an City, An hui Province.

Name and address of
manufacturer:

Anhui Me In Garments Co., Ltd.
Er Huang Road, Xian Sheng Dian Township, Jin'an District, Lu'an City, An hui Province.

Product name: Protective Clothing

Product types: S.M.J.

Product trademark: n/a

This certificate confirms that the product meets the requirements of the following standards and within limits of its standards gives presumption of conformity with essential requirements of Regulation 2016/425

EN14126:2003 + AC:2004

The certification process has been carried out in accordance with the program PC-P-07-07.

Evaluation has been carried out in accordance with test reports made by UAC Quality Technology Service (UK) Ltd

No. of test reports: TCF_UAC_20200322011PPPE

Certificate issue date: 25.03.2020

Expiration date: 24.03.2025

The mutual obligations and rights of the certification are regulated by the contract
No. ICR Polska/2020-3109.

This certificate applies to products having the same attributes (parameters), intended use, that have been evaluated and meet the requirements of the aforementioned standard.

Director: Rafal Kajrowski

Warsaw, 25. 03. 2020



ICR Polska Co. Ltd.
ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com e-mail: icrpolska@crqa.com



Enterprise qualification企业资质



一次性隔离衣

Disposable Isolation Coveralls

安徽海能半影服饰有限公司 ANHUI ME IN GARMENTS CO., LTD

产品信息 Product Information

产品名称 Product Name:	一次性隔离衣
	Disposable Isolation Coveralls
产品规格 Product Specification	连身式带帽拉链形
产品标准 Product Standard	GB19082-009
产品描述 Product Description	采用非织造布为主要原料，经裁剪，缝纫制成。非无菌提供，一次性使用。 Made of non-woven fabric after cutting and stitching, non-sterile and disposable.
产品组成 Construction	PP无纺布+PE 复合膜 PP non-woven + PE with membrane
产品用途 Product Function	一般防护 General Protection

Enterprise qualification企业资质

海关进出口货物收发货人备案回执

第一类医疗器械备案凭证

安徽海能半影服饰有限公司：

根据相关法规要求，对你单位第一类医疗器械：隔离衣予以备案，备案号：皖六械备 20200010 号。

企业名称	安徽海能半影服饰有限公司
统一社会信用代码	91341502MA2TFJNR8K
海关编码	34139609BY
检验检疫备案号	3464300007
有效期	长期



Enterprise qualification企业资质

第一类医疗器械备案信息表

备案号: 皖六械备 20200010 号

备案人名称	安徽海能半影服饰有限公司
备案人组织机构代码	91341502MA2TFJNR8K
备案人注册地址	安徽省六安市金安区先生店乡二黄路
生产地址	安徽省六安市金安区先生店乡二黄路
代理人	/
代理人注册地址	/
产品名称	隔离衣
型号/规格	连身式, S (身高 160-165)、M (身高 170-175)、L (身高 180-185)
产品描述	采用非织造布为主要原料, 经裁剪、缝纫制成。非无菌提供, 一次性使用。
预期用途	用于医疗机构门诊、病房、检验室等作普通隔离。
备注	
备案单位和日期	六安市市场监督管理局 备案日期: 2020年3月6日 
变更情况	



Enterprise qualification企业资质



TEST REPORT



No: 200047244

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: IGOH-2562-54

ISSUE DATE: 2020-03-21



APPLICANT: Wujiang Tutaike Textiles & Finishing Co., LTD
ADDRESS: NO. 1599 South 3rd Ring Road Shengze Wujiang Jiangsu

APPLICANT PROVIDED SAMPLE DESCRIPTION:

ONE(1) PIECE OF FABRIC (DISPOSABLE PROTECTIVE COVER)

COLOUR: WHITE

MANUFACTURE'S NAME: Wujiang Tutaike Textiles & Finishing Co., LTD

DATE RECEIVED/DATE TEST STARTED: 2020-03-20

CONCLUSION:

RESISTANCE TO WATER PENETRATION M

SURFACE WATER REPELLENCY M

BREAKING STRENGTH M

ELONGATION AT BREAK M

ELECTROSTATIC DECAY M

RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD M

FILTRATION EFFICIENCY M

WATER-VAPOUR TRANSMISSION RATE M

NOTE: "M" - MEET THE STANDARD'S REQUIREMENT "F" - FAIL TO MEET THE STANDARD'S REQUIREMENT

REMARK:

THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200019288.

ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).

COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.

THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHEJIANG ROAD, PANNU DISTRICT, GUANGZHOU, GUANGDONG, P.R. CHINA.



APPROVED BY:
ZiShan Guo SENIOR ENGINEER

郭子山

Enterprise qualification企业资质



TEST REPORT

No:200047244



Enterprise qualification企业资质



TEST REPORT

No:200047244

RESISTANCE TO WATER PENETRATION(kPa)

(GB/T 4744-1997, STEADILY INCREASING WATER PRESSURE:6.0kPa/min, TEMPERATURE OF THE WATER:20.0°C , FACE DOWN AND FACE SIDE TESTED)

		REQUIREMENT
1#	26.0	≥1.67
2#	25.6	(GB 19082-2009)
3#	25.8	
4#	27.0	
5#	26.2	
AVERAGE	26.2	

SURFACE WATER REPELLENCY(GRADE)

(GB/T 4745-1997)
(GB/T 4745-1997 EQV ISO 4920-1981)

	REQUIREMENT
1# 4	≥3
2# 4	(GB 19082-2009)
3# 4	

BREAKING STRENGTH

(GB/T 3923.1-1997, THE DISTANCE BETWEEN THE CLAMPS:100mm, RATE:100mm/min)

	REQUIREMENT
LENGTH	123N
WIDTH	59N ≥45N ≥45N (GB 19082-2009)

ELONGATION AT BREAK

(GB/T 3923.1-1997, THE DISTANCE BETWEEN THE CLAMPS:100mm, RATE:100mm/min)

	REQUIREMENT
LENGTH	93%
WIDTH	99% ≥15% ≥15% (GB 19082-2009)

ELECTROSTATIC DECAY(s)

(GB 19082-2009 5.10, TEMP:23.0°C, RH:50.1%, VOLTAGE APPLIED: ±5000V)

MAXIMUM	0.01	≤0.5
		(GB 19082-2009)

RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD(GRADE)

(GB 19082-2009 ANNEX A, SURFACE TENSION OF SYNTHETIC BLOOD: 0.042N/m)

6	≥2	REQUIREMENT
		(GB 19082-2009)

FILTRATION EFFICIENCY(%)

(GB 19082-2009 5.7, AIR FLOW:15L/min, AEROSOL:NaCl, AEROSOL CONCENTRATION:15mg/m³, TEMP:23.0°C, RH:35.6%)

	REQUIREMENT
MINIMUM	100 ≥70 (GB 19082-2009)



Enterprise qualification企业资质



TEST REPORT

No:200047244

WATER-VAPOUR TRANSMISSION RATE(g/(m² · d))
(GB/T 12704-1991, TEST CONDITION:38.0°C, 90.0%, REACTION TIME:1h, METHOD:DESICCANT
METHOD)

4790

REQUIREMENT
≥2500
(GB 19082-2009)

