

htl[®] | 鑫思特[®]
专业仪器仪表制造厂商

非接触红外体温计
型号：HT-820D



CC SHOT ON MI CC9 PRO
AI PENTA CAMERA



Specification	
Storage temperature	-10~60°C(14~140°F)
Usage temperature	10~40°C(50~104°F)
Usage battery	3V
Weight	122g
Size	145 x 80 x 40mm
Technical Parameters	
Temperature measuring range	32~43°C(90~109°F)
temperature measuring accuracy	$\pm 0.3^{\circ}\text{C}$ ($\pm 0.6^{\circ}\text{F}$)
Surface Temperature measuring range	0~100°C(32~212°F)
Surface temperature measuring accuracy	$\pm 1^{\circ}\text{C}$ ($\pm 1.8^{\circ}\text{F}$)
Minimum readings	$\pm 0.1^{\circ}\text{C}$ ($\pm 0.1^{\circ}\text{F}$)
Response time	0.5s
Measuring distance	5 ~ 8 cm
Auto power off	7s

MADE IN CHINA

Shenzhen BCTC Technology Co., Ltd.
No. 101, Yousong Road, Longhua New District,
Shenzhen, Guangdong, P.R. China



Certificate of Compliance

Certificate Number: BCTC-160606825C

Applicant

Manufacturer

Product

M/N : HT-860D
HT-820D

Test Standard : EN 55022: 2010+AC:2011
EN 61000-3-2: 2014, EN 61000-3-3: 2013
EN 55024: 2010
EN 61000-4-2: 2009, EN 61000-4-3: 2006+A1:2008+A2:2010
EN 61000-4-4: 2012, EN 61000-4-5: 2014
EN 61000-4-6: 2014, EN 61000-4-8: 2010, EN 61000-4-11: 2004

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC Directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. It is only valid in connection with the test report number: BCTC-160606825E.



Jun. 13, 2016

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole product and relevant. Directives have to be observed.

Tel: 400-786-9558 0755-33019988
[Http://www.bcto-lab.com](http://www.bcto-lab.com) [Http://www.bcto-lab.com.cn](http://www.bcto-lab.com.cn)



医疗器械生产许可证

许可证编号: 粤食药监械生产许20162810号

企业名称: 东莞市鑫泰仪器仪表有限公司

生产地址: 东莞市塘厦镇龙背岭村春风路149-102号

法定代表人: 罗国栋

生产范围: II类6820普通诊察器械

企业负责人: 罗国栋

住 所: 东莞市塘厦镇龙背岭村春风路149-102号

发证部门: 广东省食品药品监督管理局

有效期限: 至 2021 年 06 月 06 日

发证日期: 2017 年 07 月 24 日

国家食品药品监督管理总局制

中华人民共和国医疗器械注册证

注册证编号: 粤械注准20162200045

注册人名称	东莞市鑫泰仪器仪表有限公司
注册人住所	东莞市塘厦镇龙背岭村春风路149-102号
生产地址	东莞市塘厦镇龙背岭村春风路149-102号(生产地址102为1栋2层)
产品名称	非接触红外体温计
型号、规格	HT-820, HT-820D, HT-860D, HT-880
结构及组成	由红外探头主件、主线路板组件、LCD显示组件、外壳组件组成。
适用范围	通过测量额头的热辐射来显示被测对象的体温。
附件	产品技术要求。
其他内容	无
备 注	

审批部门: 广东省食品药品监督管理局

批准日期: 2016年8月8日
有效期至: 2021年8月8日

编号: N° 0466401



营业执照

(副本) (副本号:1-1)

统一社会信用代码 91441900555608407H

名称 东莞市鑫泰仪器仪表有限公司
类型 有限责任公司(自然人投资或控股)
住所 东莞市塘厦镇塘厦社区宏业工业区13-16号F栋
法定代表人 罗国栋
注册资本 人民币伍佰万元
成立日期 2010年05月06日
营业期限 长期
经营范围 产销:仪器仪表、电子产品、医疗器械(第二类医疗器械);货物进出口、技术进出口。(依法须经批准的项目,经相关部门批准后方可开展经营活动。)



登记机关



2018年8月2日

请于每年6月30日前报送年度报告,逾期将受到信用惩戒和处罚。
途径:登录企业信用信息公示系统,或“东莞工商”微信公众号。

企业信用信息公示系统网址:
<http://gsxt.gdgs.gov.cn/>

中华人民共和国国家工商行政管理总局监制