







BST

A RELIABLE TESTING FOR TRUST

GLOBAL TESTING AND CERTIFICATION PROVIDER SERVICE GLOBAL FACTORY

Certificate of Compliance

Certificate No. : BSTXD200313687901EC

Applicant : HANGZHOU QINGYUAN MEDICAL EQUIPMENT TECHNOLOGY CO., LTD
NO 688, Dong'an First Rd., Xintang Street, Xiaoshan, Hangzhou, Zhejiang, China 311215

Manufacturer : HANGZHOU QINGYUAN MEDICAL EQUIPMENT TECHNOLOGY CO., LTD
NO 688, Dong'an First Rd., Xintang Street, Xiaoshan, Hangzhou, Zhejiang, China 311215

Product Name : FOREHEAD THERMOMETER
Main Test Model : QY-EWQ-01

Test Standard : EN 55014-1:2017
EN 61000-3-2:2014
EN 61000-3-3:2013
EN 55014-2:2015

As shown in the Test Report No. : BSTXD200313687901ER

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.
The certificate applies to the tested sample above mentioned only and shall not imply an assessment of the whole production.



CE



Christina Deng
Manager
Mar. 05, 2020

BST Testing (Shenzhen) Co.,Ltd.

Add: No.7, New Era Industrial Zone, Guantian, Bao'an District, Shenzhen, Guangdong, China

Certificate Search: <http://www bst-lab.com>, Tel: 400-882-8828, 8009990305, E-mail: christina@bst-lab.com

中华人民共和国医疗器械注册证

注册证编号: 湖械注准20202071070

注册人名称	杭州氟源医疗器械科技有限公司
注册人住所	浙江省杭州市萧山区新塘街道东安一路688号
生产地址	浙江省杭州市萧山区新塘街道东安一路688号
代理人名称	不适用
代理人住所	不适用
产品名称	红外额温计
型号、规格	QY-EWQ-01
结构及组成	产品由测温头、按钮、显示屏、电源部分、机身组成。
适用范围	产品通过测量额头的热辐射来显示被测对象的体温。
附件	产品技术要求(除YY 0505-2012和GB/T 14710-2009外)。
其他内容	/
备注	本产品为防控新型冠状病毒感染的肺炎疫情应急审批产品, 注册证有效期为6个月, 产品标签和说明书上应醒目标注“仅供防控疫情应急使用”

审批部门: 浙江省药品监督管理局

批准日期: 2020年03月13日
有效期至: 2020年09月11日





国家食品药品监督管理总局制

