Designing Patient-Specific Optimal Neurostimulation Patterns for Seizure Suppression

Roman A. Sandler *[1,2], Kunling Geng [3], Dong Song [3], Robert E. Hampson [4], Mark R. Witcher [5], Sam A. Deadwyler [4], Theodore W. Berger [3] & Vasilis Z. Marmarelis[3]

[1]Department of Physics & Astronomy, University of California, Los Angeles, Los Angeles, CA, USA
[2]W. M. Keck Center for Neurophysics, University of California, Los Angeles, Los Angeles, CA, USA
[3]Department of Biomedical Engineering, University of Southern California, Los Angeles, CA, USA
[4]Department of Physiology & Pharmacology, Wake Forest University, Winston-Salem, NC, USA
[5]Department of Neurosurgery, Wake Forest University, Winston-Salem, NC, USA

October 31, 2017

Abstract

Neurostimulation is a promising therapy for abating epileptic seizures. However, it is extremely difficult to identify optimal stimulation patterns experimentally. In this study human recordings are used to develop a functional 24 neuron network statistical model of hippocampal connectivity and dynamics. Spontaneous seizure-like activity is induced in-silico in this reconstructed neuronal network. The network is then used as a testbed to design and validate a wide range of neurostimulation patterns. Commonly used periodic trains were not able to permanently abate seizures at any frequency. A simulated annealing global optimization algorithm was then used to identify an optimal stimulation pattern which successfully abated 92% of seizures. Finally, in a fully responsive, or "closed-loop" neurostimulation paradigm, the optimal stimulation successfully prevented the network from entering the seizure state. We propose that the framework presented here for algorithmically identifying patient-specific neurostimulation patterns can greatly increase the efficacy of neurostimulation devices for seizures.

Publication Statement

None of this material has been published elsewhere

*Corresponding Author: rsandler00@gmail.com