


# Surface Tension Calibration Ethanol X-Bar Chart, Aug-Dec, 2011



Number of groups = 20

Center = 22.38

StdDev = 0.1268272

LCL = 22.20984

UCL = 22.55016

Number beyond limits = 2

Number violating runs = 0