


Surface Tension Calibration Ethanol R Chart, Aug–Dec, 2011



Number of groups = 20

Center = 0.295

StdDev = 0.1268272

LCL = 0

UCL = 0.6237686

Number beyond limits = 0

Number violating runs = 0