



IIT ROORKEE



NPTEL ONLINE
CERTIFICATION COURSE

FELIX ORLANDO MARIA JOSEPH
DEPARTMENT OF ELECTRICAL ENGINEERING



Experiments on Robot Assisted Percutaneous Interventions




Outline


1. Introduction
2. Two Loop Coordinated Control
3. Effect of Tissue Interference
4. Conclusion




Introduction



Robot Aided Needling System



Coordinated Control



$$u_{PID}(k) = u_{PID}(k-1) + K_P[e(k) - e(k-1)] + K_I T_S e(k) + K_D[e_f(k) - 2e_f(k-1) + e_f(k-2)]$$

$$u_{PID-P^3}(k) = u_{PID-P^3}(k-1) + K_P[e(k) - e(k-1)] + K_I T_S e(k) + K_D[e_f(k) - 2e_f(k-1) + e_f(k-2)] + K_T[e(k) - e(k-1)]^3$$


Ref: Felix Orlando et al, "Development of Closed Loop Coordinated Control of a Robot Guided SMA Actuated Flexible Active Needle with Multimodal Sensory Feedbacks," IEEE-IECON 2017, pp.(2846-2851)

SMA Actuated Smart Needle




Ref: Felix Orlando et al, "Development of Closed Loop Coordinated Control of a Robot Guided SMA Actuated Flexible Active Needle with Multimodal Sensory Feedbacks," IEEE-IECON 2017, pp.(2846-2851)

Experiment with EM sensory modality




Ref: Felix Orlando et al, "Development of Closed Loop Coordinated Control of a Robot Guided SMA Actuated Flexible Active Needle with Multimodal Sensory Feedbacks," IEEE-IECON 2017, pp.(2846-2851)

Experiment with Vision sensory modality



Experiment with U/S sensory modality




Result Analysis

Feedback Modalities	RMSE values (Mean \pm SD in mm)	
	Curvilinear	Rectilinear
EM	0.08 \pm 0.01	0.07 \pm 0.01
Vision	0.13 \pm 0.02	0.12 \pm 0.01
US	0.17 \pm 0.02	0.15 \pm 0.02




Ref: Felix Orlando et al, "Development of Closed Loop Coordinated Control of a Robot Guided SMA Actuated Flexible Active Needle with Multimodal Sensory Feedbacks," IEEE-IECON 2017, pp.(2846-2851)



With EM Sensory Feedback



RMSE 0.1 mm



Ref: Felix Orlando et al, "Development of Closed Loop Coordinated Control of a Robot Guided SMA Actuated Flexible Active Needle with Multimodal Sensory Feedbacks," IEEE-IECON 2017, pp.(2846-2851)

Limitations

- No 3D movement of the needle – only 2D is considered.
- As the needle goes through the body more, the exerted force may increase. But, here in this study, no relation between amount of needle entrance, tissue-needle interaction dynamics and tracking performance are done.
- Adaptive control algorithm to overcome the disturbance while insertion through the tissue is not performed.

Ref: Felix Orlando et al, “Development of Closed Loop Coordinated Control of a Robot Guided SMA Actuated Flexible Active Needle with Multimodal Sensory Feedbacks,” IEEE-IECON 2017, pp.(2846-2851)



Effects of Tissue Interference on Sensing Modalities for Robot Guided Needle Intervention (cont'd)

- ❖ **Compromising factors** – Needle placement accuracy:
 - Needle design (active or passive, rigid or flexible).
 - Tissue heterogeneities, obstacles, and critical structures.


- ❖ **Robotic assistance** – a potential solution.

- ❖ **Challenging factors:**
 - Sensory feedback signal attenuation and noise.
 - Needle-tissue interaction.

Ref: Felix Orlando et al., 'Effects of Tissue Interference on Sensing Modalities for Robot Guided Needle Intervention', AAPM Meeting, Anaheim, LA, July 2015.



Effects of Tissue Interference on Sensing Modalities for Robot Guided Needle Intervention (cont'd)




18-gauge, 200 mm, Aurora tracking needle

Ref: Felix Orlando et al., 'Effects of Tissue Interference on Sensing Modalities for Robot Guided Needle Intervention', AAPM Meeting, Anaheim, LA, July 2015.

Effects of Tissue Interference on Sensing Modalities for Robot Guided Needle Intervention (cont'd)

US Experimental Setup



EM Experimental Setup




Table: RMSE Comparison

	EM Feedback Mean \pm SD (in mm)	US Feedback Mean \pm SD (in mm)
No Tissue Interference (Distance: EM = 240mm; US = 95 mm)	0.13 \pm 0.03	0.38 \pm 0.07
Thin Tissue Interference (Tissue: EM = 25 mm; US = 15 mm)	0.16 \pm 0.03	1.00 \pm 0.11
Thick Tissue Interference (Tissue: EM = 90 mm; US = 25 mm)	0.19 \pm 0.04	1.82 \pm 0.12

Ref: Felix Orlando et al., 'Effects of Tissue Interference on Sensing Modalities for Robot Guided Needle Intervention', AAPM Meeting, Anaheim, LA, July 2015.

Needle Tip Prediction using Kalman Filter

- We use an ultrasound imaging feedback to estimate position of operating needle from an EM sensor in the needle.
- Error between present and previous positions obtained is given as feedback input.
- Random artifacts may lead to huge errors resulting in corresponding feedback input.
- Hence, the use of Kalman filter to estimate the position of the needle tip and use of weighted average between the estimated position and obtained position.

Ref: Nisha Agarwal, Ankur Yadav, Arjun Gupta and **M. Felix Orlando**, "Real-time Needle Tip Localization in 2D Ultrasound Images using Kalman Filter," IEEE/ASME Advanced Intelligent Mechatronics, (**IEEE/ASME-AIM**) 2019, HongKong, 8-12, July 2019, pp. 1008-1012.



Needle Tip Prediction using Kalman Filter


This research study proposes a method to track the brachytherapy needle tip position in a real-time ultrasound imaging using image processing technique and Kalman filter. It involves the following steps:

- Data collection
- Preprocessing
- Needle detection
- Kalman filter

Ref: Nisha Agarwal, Ankur Yadav, Arjun Gupta and **M. Felix Orlando**, "Real-time Needle Tip Localization in 2D Ultrasound Images using Kalman Filter," IEEE/ASME Advanced Intelligent Mechatronics, (**IEEE/ASME-AIM**) 2019, HongKong, 8-12, July 2019, pp. 1008-1012.




Needle Tip Prediction using Kalman Filter




Ref: Nisha Agarwal, Ankur Yadav, Arjun Gupta and **M. Felix Orlando**, "Real-time Needle Tip Localization in 2D Ultrasound Images using Kalman Filter," IEEE/ASME Advanced Intelligent Mechatronics, (**IEEE/ASME-AIM**) 2019, HongKong, 8-12, July 2019, pp. 1008-1012.

Needle Tip Prediction using Kalman Filter




Ref: Nisha Agarwal, Ankur Yadav, Arjun Gupta and **M. Felix Orlando**, "Real-time Needle Tip Localization in 2D Ultrasound Images using Kalman Filter," IEEE/ASME Advanced Intelligent Mechatronics, (**IEEE/ASME-AIM**) 2019, HongKong, 8-12, July 2019, pp. 1008-1012.

Needle Tip Prediction using Kalman Filter



Ref: Nisha Agarwal, Ankur Yadav, Arjun Gupta and **M. Felix Orlando**, "Real-time Needle Tip Localization in 2D Ultrasound Images using Kalman Filter," IEEE/ASME Advanced Intelligent Mechatronics, (**IEEE/ASME-AIM**) 2019, HongKong, 8-12, July 2019, pp. 1008-1012.

Needle Tip Prediction using Kalman Filter



Ref: Nisha Agarwal, Ankur Yadav, Arjun Gupta and **M. Felix Orlando**, "Real-time Needle Tip Localization in 2D Ultrasound Images using Kalman Filter," IEEE/ASME Advanced Intelligent Mechatronics, (**IEEE/ASME-AIM**) 2019, HongKong, 8-12, July 2019, pp. 1008-1012.

Conclusions and Future Works

- Inverse kinematic relationship or a path planning algorithm must be included in the study.
- For real applications, how can someone use the sensory systems utilized in this study must be worked on more.
- Modeling of the systems, performance evaluation of the controller needs to be done.

Thank You!

