Procalb Instrument Calibrations
Division of RTNTS GP Ltd

Procalb Instrument

Customer: Rare Snacks Ltd
Address: Unit 5, First Avenue, Globe Business Park SL7 1YA Marlow UK
Job No.: AK070524-3

Test Certificate Report			
Date of Issue:	03-Jan-2026	**Equipment Id:**	EQP7105260103129
Equipment Schedule:	Scale	**Location:**	2
Serial Number:	PW01		
Manufactured by:	Ohaus		

Equipment Testing Details

		Applied Load	As Found	Deviation	Definitive Reading	Deviation	Notes
Load and Linearity Test		0.00	0.00	0.00			
		10.00	9.99	-0.01			
		15.00	14.98	-0.02			
		20.00	19.98	-0.02			
		30.00	29.98	-0.02			
Repeatability (all centre)		15.00	14.98	-0.02			
		15.00	14.98	-0.02			
		15.00	14.98	-0.02			
		15.00	14.98	-0.02			
Eccentricity	Front Left	10.00	9.99	-0.01			
	Front Right	10.00	9.99	-0.01			
	Centre	10.00	9.98	-0.02			
	Rear Left	10.00	9.98	-0.02			
	Rear Right	10.00	9.98	-0.02			

Test Equipment Used:

Test Equipment Used:	Serial No.	UKAS Certificate No.
name1	serial1	ukas1
name2	serial2	ukas2

Conclusion

We certify that the above instrument has been checked and tested for accuracy and passed within the limits as stated to alarm/fail.

Calibrated by: Usama Mustafa
Signature:
Date of Calibration: 03-Jan-2026
Next Calibration Due Date: 02-Jan-2027

Scan QR For More