INTERNAL CALIBRATION LOG FOR SCALES

MONTHLY CALIBRATION										
Follow the diagram to take a weight reading from each of the five points on the weighing surface. Record the average weight of the five checks (g) on the internal calibration log for weighing scales document.										
Month	**Scale No.**	**Ohaus T32XW B315215998**							**Corrective Actions**	**Sign**
WC	**Location**	**HCA (Big one)**					**Avg.**	**Diff.**		
14-05-2026	**100g**	1	1	1	1	1	1	-99		
	1kg	2	2	2	2	2	2	-998		
	5kg	3	3	3	3	3	3	-4997		

Verified by:

Date: 2026-05

INTERNAL CALIBRATION LOG FOR SCALES

Tolerance =	**100g**	± 0.0 g
	1kg	± 0.10 Kg
	5kg	± 0.20 Kg

Verified by:

Date: 2026-05