

Certificate Customer Measurement Report

Generated on: 12/02/2025 01:37 by Loghid SPAIN

This document provides a Certificate Customer Measurement Report - Loghid H2 Fingerprint, detailing the quality of hydrogen gas based on the ISO 14687 standard.

ISO Threshold: The maximum allowable concentration for a given substance, as defined by ISO 14687.

Measurement Method: The technique used to measure the substance (e.g., GC, OFCEAS, NDIR).

Measurement Range: Testing range to ensure the instrument is correctly calibrated and capable of detecting contamination.

Parameter	Value
Measurement ID	1
H2 Fingerprint	A2DG-UXM2-6KMQ-SOSN
Date	11/02/2025
Customer ID	8
Customer	Zoilo Ríos
Supplier ID	13
Supplier	Ruben-Vasile M. Ungureanu
Location	Gasolinera Aragón
Notes	N/A
Water	H2O
Water ISO Threshold	5 $\mu\text{mol/mol}$
Water Measurement Method	OFCEAS
Water Measurement Range	N/A
Methane	CH4
Methane ISO Threshold	2 $\mu\text{mol/mol}$
Methane Measurement Method	GC-PDHID
Methane Measurement Range	N/A
Non-CH4 Hydrocarbons	NoCH4
Non-CH4 Hydrocarbons ISO Threshold	2 $\mu\text{mol/mol}$

Non-CH4 Hydrocarbons Measurement Method	2 µmol/mol
Non-CH4 Hydrocarbons Measurement Range	N/A
Oxygen	O2
Oxygen ISO Threshold	5 µmol/mol
Oxygen Measurement Method	OFCEAS
Oxygen Measurement Range	N/A
Helium	He
Helium ISO Threshold	2 µmol/mol
Helium Measurement Method	GC-TCD
Helium Measurement Range	N/A
Nitrogen	N2
Nitrogen ISO Threshold	2 µmol/mol
Nitrogen Measurement Method	OFCEAS
Nitrogen Measurement Range	N/A
Argon	Ar
Argon ISO Threshold	2 µmol/mol
Argon Measurement Method	OFCEAS
Argon Measurement Range	N/A
Carbon Dioxide	CO2
Carbon Dioxide ISO Threshold	2 µmol/mol
Carbon Dioxide Measurement Method	NDIR
Carbon Dioxide Measurement Range	N/A
Carbon Monoxide	CO
Carbon Monoxide ISO Threshold	2 µmol/mol
Carbon Monoxide Measurement Method	OFCEAS
Carbon Monoxide Measurement Range	N/A
Sulphur Compounds	S
Sulphur Compounds ISO Threshold	0.004 µmol/mol
Sulphur Compounds Measurement Method	GC-PFPD
Sulphur Compounds Measurement Range	N/A
Formaldehyde	CH2O
Formaldehyde ISO Threshold	0.01 µmol/mol
Formaldehyde Measurement Method	OFCEAS
Formaldehyde Measurement Range	N/A

Formic Acid	HCOOH
Formic Acid ISO Threshold	0.01 µmol/mol
Formic Acid Measurement Method	OFCEAS
Formic Acid Measurement Range	N/A
Ammonia	NH ₃
Ammonia ISO Threshold	0.1 µmol/mol
Ammonia Measurement Method	OFCEAS
Ammonia Measurement Range	N/A
Halogenated Compounds	Halogenados
Halogenated Compounds ISO Threshold	0.3 µmol/mol
Halogenated Compounds Measurement Method	GC-ECD
Halogenated Compounds Measurement Range	N/A
Hydrocarbons	HC
Hydrocarbons ISO Threshold	2 µmol/mol
Hydrocarbons Measurement Method	GC-FID
Hydrocarbons Measurement Range	N/A

SHA-256: 907c0e65a06a319c5e8ab01ebac13b145e14403c600e85b46c1587c11e3bf81c

More info: info@loghid.com