

# FEEDBACK

We tested the prototype in Rehabilitation centre for Blind Women in Trichy.



**NAME :** Regina Mary

**AGE :** 50+

**Test 1:** Since the product has only a single braille cell, the test subject found it difficult to identify the relative position of the embossed dots. She was also was not comfortable with the speed.

**Test 2:** Adjustable speed control was made and all the dots were embossed a little in normal state while the required dots were actuated above it. Hence the person felt it a lot more easier to recognise the characters.

## ***Her feedback:***



*She was not able to read as fast as she reads in braille sheet as this was the first time, but reading felt a lot more comfortable and easier. She also said that with practice the reading speed can be improved a lot. She thanked us for creating such a product to help the blind community.*