

Finger Lakes Instrumentation Quality Assurance Test Summary  
FLI Library Version: Software Development Library for Windows 1.98  
FLITest\_SCX Version: 3.31 SCX, Tester SN: 6

Serial Number: PL2300421  
Model: PROLINE PL230  
Sensor designation: CCD230-42-1-143 16423-09-06  
Sensor Type: Single Channel  
Lab ambient temp: 24.0  
CCD Test Temp: -25

Basic Functions

CCD temp sensor: OK.  
Base temp sensor: OK.

CCD Chamber

Desiccant installed: OK.  
Purged: OK.  
Noble Gas Back-fill: OK.  
Window clean on both surfaces: OK.  
CCD free of dust: Yes

Performance

Cooler can achieve dT greater than: -45.0C (measured -53.2)  
Cooler is rate limited: Yes (-4.95 C/min)  
Cooler test time: 872.57  
Cooler Tests: Pass  
Noise distribution is random: Yes  
Bias frame histogram is Gaussian: Yes  
Standard test target appearance OK: Yes

Specific Tests

Amplifier Glow Test: Pass  
Amplifier Glow Auxiliary Test: Pass  
Power Supply voltage: 11.7  
AUX Connector Pin 4 voltage: -0.0  
Camera Current: 819.5 mA (Pass)  
Cooler Current: 3899.8 mA (Pass)  
External Trigger: Working  
AUX Pin 1: Working OK.  
AUX Pin 2: Working OK.  
AUX Pin 3: Working OK.

Gain and Noise Measurement

3 MHz Main

Conversion factor (gain): 1.81  
Bias: 974.3  
Readout noise: 20.1 (e-)  
Saturation: 116716.6 (e-)  
Horizontal Binning Test: Pass  
Bias Drift Test: Pass  
100s Dark Current Growth: 18.5 ADU (Pass) @ -25.0 C  
Maximum Linearity Error: 0.287 (%) (Pass)

500KHz Main

Conversion factor (gain): 1.80  
Bias: 995.7  
Readout noise: 14.0 (e-)  
Saturation: 115873.7 (e-)  
Horizontal Binning Test: Pass  
Bias Drift Test: Pass  
100s Dark Current Growth: 18.2 ADU (Pass) @ -25.0 C  
Maximum Linearity Error: 0.014 (%) (Pass)

Camera Passed.

 APPROVED