


Raster scan overlay on real CT — coloured circles = per-shot cavitation-cloud radius (not FWHM)


(cyan=GTV yellow=CTV +5 mm [HOPE4LIVER] white=PTV +8 mm total)

OAR contours: red=vessels, orange=gallbladder, magenta=capsule (dashed/flag), amber=extrahepatic, darkred=bone

us intrinsic-threshold (1 MHz, 30 MPa)
raster: 28 pts $\times$ 4.8 mm cloud r, 0.9 min
(cyan=GTV yellow=CTV white=PTV)
OAR footprint overlap: 25 pts (dose monitored)



ms shock-vapor (1 MHz, 15 MPa, 10 ms)
raster: 47 pts $\times$ 7.2 mm cloud r, 1.9 min
(cyan=GTV yellow=CTV white=PTV)
OAR footprint overlap: 45 pts (dose monitored)



ms sub-threshold cav (500 kHz, 18 MPa, 5 ms)
raster: 47 pts $\times$ 7.2 mm cloud r, 1.5 min
(cyan=GTV yellow=CTV white=PTV)
OAR footprint overlap: 47 pts (dose monitored)

