

# Contents

Contents	I
List of Figures	II
List of Tables	III
Abbreviations	IV
Glossary	V
Nomenclature	VI
1 Introduction	1
1.1 Project scope . . . . .	2
2 Theory	5
2.1 Ultrasound . . . . .	5
2.1.1 Scattering . . . . .	6
2.1.2 Attenuation . . . . .	7
2.1.3 Transducer . . . . .	8
2.1.4 Doppler effect . . . . .	8
2.2 Flow physics . . . . .	10
2.2.1 Blood flow . . . . .	12
2.3 Devices . . . . .	12
2.3.1 Continuous-wave flowmeter . . . . .	13
2.3.2 Pulsed-wave flowmeter . . . . .	15
Bibliography	19
A Bill of Materials	22
B Instruments	25

# List of Figures

2.1	Particle displacement for a propagating ultrasound wave . . . . .	5
2.2	Single element ultrasound transducer construction . . . . .	8
2.3	Transducer types for acquiring B-mode images . . . . .	9
2.4	Doppler effect diagram . . . . .	10
2.5	Circulatory system of the human body . . . . .	11
2.6	Diagram of ultrasound wave transmitted and reaching blood vessel with incident angle $\theta$ . . . . .	12
2.7	Block diagram of continuous-wave flowmeter . . . . .	13
2.8	Doppler signals in time and frequency domain showing demodulation effects	14
2.9	Block diagram of pulsed-wave flowmeter . . . . .	15
2.10	Sampling for a gate pulsed wave system with a single range . . . . .	16

# List of Tables

2	.....	VI
1.1	Comparison of medical imaging modalities . . . . .	1
1.2	Project specification table . . . . .	2
2.1	Approximate density, sound speed, and acoustic impedance of human tissue types . . . . .	6
2.2	Approximate attenuation values for human tissue . . . . .	7
2.3	Measured frequency shifts with a Doppler <b>3 MHz</b> transducer at various velocities at a <b>45°</b> incident angle . . . . .	15
A.1	BOM . . . . .	22
B.1	List of instruments used for solder work . . . . .	25
B.2	List of instruments used for testing . . . . .	25

# Abbreviations

Notation	Description
ADC	analogue-to-digital converter
BP	band-pass
CT	computed tomography
CW	continuous-wave
DTU	Danmarks Tekniske Universitet (Technical University of Denmark)
KAIST	Korea Advanced Institute of Science and Technology
low-res	low resolution
MRI	magnetic resonance imaging
PSD	power spectral density
PW	pulsed-wave
US	ultrasound

# Glossary

Notation	Description
adiabatic	Any process that happens without heat gain or loss is considered adiabatic
Doppler effect	A change in frequency of a wave in relation to an observer who is moving relative to the wave source
transcutaneous	Applied across the depth of the skin without invasive penetration

# Nomenclature

Name	Unit	Description
Input filter		
$V_{\text{ref}}$	V	Reference voltage
$C_{\text{hp}}$	F	High pass filter capacitor
$R_{\text{hp}}$	$\Omega$	High pass filter resistor
$f_{\text{hp}}$	Hz	High pass cut-off frequency
$C_{\text{lp}}$	F	Low pass filter capacitor
$R_{\text{lp}}$	$\Omega$	Low pass filter resistor
$f_{\text{lp}}$	Hz	Low pass cut-off frequency
$A_v$	1	Amplification factor
Modulator		
$R_1$	$\Omega$	AIM voltage divider resistor
$R_2$	$\Omega$	AIM voltage divider resistor
$R_{\text{in}}$	$\Omega$	AIM input resistor
$R_{\text{fb}}$	$\Omega$	AIM feedback resistor
$C_1$	F	AIM capacitor
$V_{\text{in}}$	V	Input signal voltage
$V_{\text{span}}$	V	Voltage range of input signal
$V_{\text{pwm}}$	V	PWM signal
$V_H$	V	$V_{\text{pwm}}$ high level voltage
$V_L$	V	$V_{\text{pwm}}$ low level voltage
$V_{\text{out}}$	V	$V_{\text{pwm}}$ voltage range ( $V_H - V_L$ )
$V_{\text{hys}}$	V	Hysteresis voltage
$V_{\text{hw}}$	V	Hysteresis width
$V_{\text{th}_H}$	V	Hysteresis threshold upper voltage
$V_{\text{th}_L}$	V	Hysteresis threshold lower voltage
$V_c$	V	PWM carrier voltage
$V_{c_H}$	V	PWM carrier upper voltage
$V_{c_L}$	V	PWM carrier lower voltage
$f_{sw}$	Hz	PWM signal frequency
$D$	1	PWM signal duty cycle
$t_H$	s	PWM carrier charge time

Continued on next page

Table 2: (Continued)

Name	Unit	Description
$t_L$	s	PWM carrier discharge time
$\tau$	1	PWM carrier charge constant
$R_{th}$	$\Omega$	PWM carrier thevenin resistance
$f_{idle}$	Hz	PWM signal idle switching frequency
$k_2$	1	$R_{fb}$ , $R_{in}$ voltage divider
Gate driver		
$D_{dt}$	1	Dead-time circuit diode
$R_{dt}$	$\Omega$	Dead-time circuit resistor
$C_{dt}$	F	Dead-time circuit capacitor
$V_C$	V	Dead-time circuit supply voltage
$V_s$	V	IC supply voltage
$t_c$	s	Charging circuit time
Power stage		
$V_{DD}$	V	Power supply voltage
$Q_1, Q_2, Q_3, Q_4$	1	Power stage switches
$V_g$	V	Gate driver signal
$V_o$	V	Output voltage
$I_o$	A	Output current
$R_{BTL}$	$\Omega$	Speaker equivalent load resistance
Output filter		
$R_f$	$\Omega$	Output filter single-ended load
$C_{BTL}$	F	Output filter differential capacitance
$C_f$	F	Output filter single-ended capacitance
$L_f$	H	Output filter inductance
$\Delta I_L$	A	Output filter ripple current
$Q$	1	Output filter quality factor
$f_c$	Hz	Output filter cut-off frequency
$\omega_n$	$\text{rads}^{-1}$	Output filter natural frequency
$\zeta$	1	Output filter damping ratio

Continued on next page

---

Table 2: (Continued)

Name	Unit	Description
Shunt regulator		
$R_{sh}$	$\Omega$	Shunt current limiting resistor
$I_K$	A	Shunt cathode current
$I_{K_{\max}}$	A	Shunt maximum cathode current
$I_{K_{\min}}$	A	Shunt minimum cathode current
$I_{su}$	A	Shunt supply current
$R_{A1}$	$\Omega$	Shunt adjust resistor 1
$R_{A2}$	$\Omega$	Shunt adjust resistor 2
$C_L$	F	Shunt load capacitance
$V_{\text{ref}_{\text{IC}}}$	V	Shunt internal reference voltage



# 1 Introduction

The progress of diagnostic imaging has advanced significantly during the 20th century. As the cost of high speed computational systems has grown increasingly accessible, so has the use of medical imaging become prominent. Potentially millions of people have been spared painful exploratory surgery through non-invasive diagnostic imaging. And thus, lives can be saved by early diagnosis and intervention through medical imaging. Advancement in scientific visualization have in turn generated more complex datasets of increased size and quality. Four major technologies used are ultrasound (US), X-ray, computed tomography (CT), and magnetic resonance imaging (MRI). Each of the technologies have distinct advantages and disadvantages in biomedical imaging, thus each are still relevant for modern medicine. Table 1.1 contains a comparison and summary of the various fundamental diagnostic imaging modalities.

Table 1.1: Comparison of medical imaging modalities [18]

Modality	Ultrasound	X-ray	CT	MRI
Topic	Longitudinal, shear, mechanical properties	Mean X-ray tissue absorbtion	Local tissue X-ray absorbtion	Biochemistry (T1 and T2)
Access	Small windows adequate	2 sides needed	Circumferential around body	Circumferential around body
Spatial resolution	0.2 mm to 3 mm <sup>a</sup>	~ 1 mm	~ 1 mm	~ 1 mm
Penetration	3 cm to 25 cm <sup>b</sup>	Excellent	Excellent	Excellent
Safety	Excellent	Ionizing radiation	Ionizing radiation	Very good
Speed	Real-time	Minutes	20 minutes	Varies <sup>†</sup>
Cost	\$	\$	\$\$	\$\$\$
Portability	Excellent	Good	Poor	Poor
Volume coverage	Real-time 3D volumes, improving	2D	Large 3D volume	Large 3D volume
Contrast	Increasing (shear)	Limited	Limited	Slightly flexible
Intervention	Real-time 3D increasing	No <sup>c</sup>	No	Yes, limited
Functional	Functional ultrasound	No	No	fMRI

<sup>a</sup> Frequency and axially dependent.

<sup>b</sup> Frequency dependent.

<sup>c</sup> Fluoroscopy limited.

<sup>†</sup> Typical: 45 minutes, fastest: Real-time (low-res).

Since medical imaging has been reportedly performed over 5 billion times as of 2004 [8],

and later numbers from 2011 show a general doubling, and particularly a ten-fold increase in ultrasound examinations between year 2000 and 2011 [18]. Recent data reveals that this trend of doubling has continued through the years 2010 to 2020 [26], and reveals that even though patient processes were disrupted during the global SARS-CoV-2 pandemic, the number of medical imaging examinations per 1000 patients, still increased. The reasons for this, and particularly why ultrasound has been notably increased in use, can be attributed to its high, but inconsistent, resolution, cost effectiveness, portability, and real-time interventional imaging. The downside of ultrasound is its limited penetration and restrictions for use in certain body parts. When just comparing soft-tissue examinations, which ultrasound is limited to, both CT and MRI can image the entire body with consistent resolution and contrast, but are more expensive and has poor portability due to immense size of its hardware.

The cardiovascular system, which transports oxygen and nutrients to tissue, produces a complex flow pattern that cause velocity fluctuations. Several cardiovascular diseases are also known to cause abnormal blood flow. As mentioned earlier, ultrasound is a powerful tool for conducting non-invasive imaging of the cardiovascular system [14], [17], and has no adverse risk to patients. Determining power spectral density (PSD) of a received signal is a common way to estimate blood velocity. A processed image of the PSD over time is commonly known as a sonogram, where changes in blood velocity over time can be seen.

## 1.1 Project scope

The aim of this project is to study the application of ultrasound in the context of blood flow measurements. Various scientific articles have been studied to gain knowledge of prior research [4]–[6], [9]–[13], [15], [16], [20]–[25]. In addition, the textbooks [17]–[19] have also been instrumental in forming a solid knowledge base for the thesis. The desire is to build upon the vast knowledge already gathered by prominent researchers in the field of ultrasound systems for blood velocity estimation. Finally, using the knowledge gained, to design and implement an electronic device capable of performing these measurements using a novel approach. The project specifications are written in table 1.2.

Table 1.2: Project specification table

Project specification
Study and research ultrasound and its principles and applications
Design and implement a device for ultrasound blood velocity estimation
Investigate and test the device in an experimental setting
Validate results with commercial equipment
Make quantifiable performance measurements on system
Write a technical report documenting the project work

The project is conducted under the guidance of advisors from the affiliated institutions Danmarks Tekniske Universitet (Technical University of Denmark) (DTU), Department of Electrical Engineering, Department of Applied Mathematics and Computer Science, and Korea Advanced Institute of Science and Technology (KAIST) at the Brain/Bio Medical Microsystems Laboratory.

The report is divided into five chapters, where the first part is an introduction to the project. The second chapter will focus on explaining the theory of the topic of the project.

The third chapter focuses on the synthesis of a system for experimental testing. The fourth chapter explains the production of the hardware. The fifth chapter will explain the testing methodology performed on the hardware. Finally, additional documentation of testing, code, circuit diagrams, and of laboratory setups can be found in the appendix.



## 2 Theory

### 2.1 Ultrasound




Figure 2.1: Particle displacement for a propagating ultrasound wave [17]

US is a technology that transmit sound wave with frequencies above the audible range (20 to 20 000 Hz) to mechanically vibrate matter. The particles in the medium would be at rest and distributed uniformly before any disturbance. The wave propagates as a disturbance and the particles oscillate around their mean position due to the presence of the ultrasonic wave. Typically the US frequency band used in clinical settings are from 1 to 15 MHz [18]. Figure 2.1 visualizes the propagation of a plane wave in matter. The oscillation occurs parallel to the wave's direction, making it longitudinal, and the disturbance will propagate with the variable  $c$ , which is determined by the medium and is given by eq. (2.1).

$$c = \sqrt{\frac{1}{\rho_0 \kappa_S}} \quad (2.1)$$

Where  $\rho_0$  is the mean density ( $\text{kgm}^{-3}$ ) and  $\kappa_S$  is the adiabatic compressibility ( $\text{m}^2\text{N}^{-1}$ ). Since in the majority of cases, the propagation of ultrasound is linear, it is assumed in this work. The acoustic pressure of the harmonic plane wave is expressed by eq. (2.2)

$$p(t, z) = p_0 e^{j(\omega t - kz)} \quad (2.2)$$

And propagates along the  $z$ -axis.  $\omega$  is the angular frequency,  $k$  is the wave number and is expressed by  $k = \omega/c = 2\pi/\lambda$ , and  $p_0$  is the acoustic pressure amplitude. A spherical wave is expressed by eq. (2.3)

$$p(t, r) = p_0 e^{j(\omega t - kr)} \quad (2.3)$$

Where  $r$  is radial distance, and is defined in a polar coordinate system. For each time instance, the acoustic pressure  $p(t, r)$  is constant over a fixed radial position. In this scenario, the pressure amplitude is given by  $p_0(r) = k_p/r$ , where  $k_p$  is a constant since

the energy of the outgoing wave must be constant. Particle speed  $u$  is dependent on the pressure caused by a wave expressed by eq. (2.4)

$$u = \frac{p}{Z} \quad (2.4)$$

Where  $Z$  is the characteristic acoustic impedance, defined as the ratio of acoustic pressure to particle speed at a given position in the medium and is expressed by eq. (2.5).

$$Z = \rho_0 c \quad (2.5)$$

Characteristic acoustic impedance  $Z$  is one of the most significant variables in the characterization of propagating plane waves. Reference values for density, speed of sound, and characteristic acoustic impedance can be seen in table 2.1.

Table 2.1: Approximate density, sound speed, and acoustic impedance of human tissue types [17]

Medium	Density ( $\rho_0$ ) kg/m <sup>3</sup>	Speed of sound ( $c$ ) m/s	Acoustic impedance ( $Z$ ) kg/(m <sup>2</sup> s)
Air	1.2	333	$0.4 \times 10^3$
Blood	$1.06 \times 10^3$	1566	$1.66 \times 10^6$
Bone	1.38– $1.81 \times 10^3$	2070–5350	$3.75\text{--}7.38 \times 10^6$
Brain	$1.03 \times 10^3$	1505–1612	$1.55\text{--}1.66 \times 10^6$
Fat	$0.92 \times 10^3$	1446	$1.33 \times 10^6$
Kidney	$1.04 \times 10^3$	1567	$1.62 \times 10^6$
Lung	$0.4 \times 10^3$	650	$0.26 \times 10^6$
Liver	$1.06 \times 10^3$	1566	$1.66 \times 10^6$
Muscle	$1.07 \times 10^3$	1542–1626	$1.65\text{--}1.74 \times 10^6$
Spleen	$1.06 \times 10^3$	1566	$1.66 \times 10^6$
DI	$1 \times 10^3$	1480	$1.48 \times 10^6$

In the following sections, various acoustic wave phenomena will be briefly described.

### 2.1.1 Scattering

A wave propagating across a medium continues in the same direction until it encounters a new medium. When this occurs, a portion of the wave is transmitted into the new medium, with a change in direction. Because the scattered wave is the result of several contributors, it is necessary to define it statistically. The amplitude distribution is Gaussian [17] and can thus be fully described by its mean and variance. The mean value is zero because the dispersed signal is caused by variances of the acoustic characteristics in tissue. The correlation between multiple data is what allows ultrasound to determine blood velocities. Because minor movements have a significant correlation, it is feasible to discover alterations in location by comparing sequential measurements of moving structure, such as blood cells. In medical ultrasound, just one transducer is utilised for transmitting and receiving, and only the backscattered signal is analysed. The power of scattered signal is defined by the scattering cross-section, which in small cases mean a uniform intensity  $I_i$ , and is expressed by eq. (2.6).

$$P_s = I_i \sigma_{sc} \quad (2.6)$$

Where  $\sigma_{sc}$  is the scattering cross-section in square meters. The backscattering cross section is material dependant and determines the intensity of the scattering. If the dispersed energy is evenly emitted in all directions, the scattered intensity is given by eq. (2.7).

$$I_s = \frac{P_s}{4\pi R^2} = \frac{\sigma_{sc}}{4\pi R^2} \cdot I_i \quad (2.7)$$

Where  $R$  is distance to the scattering region [17]. This results in a spherical wave. A transducer with radius  $r$  gives the power  $P_r$ , presuming the attenuation and focus is neglected, and is expressed by eq. (2.8).

$$P_r = I_s \pi r^2 = \sigma_{sc} \frac{r^2}{4R^2} \cdot I_i \quad (2.8)$$

The backscattering coefficient, which characterizes scattering from a volume of scatterers, is another measure of scattering strength. It is defined as average received power per steradian volume of scatterers when flooded with plane waves of unit amplitude and the unit is 1/**cmsr**. Back scattering coefficients in blood are significantly lower than the back scattering coefficients from various tissue types. This poses a challenge when estimating blood flow close to tissue vessel walls [3], [17].

### 2.1.2 Attenuation

The ultrasonic wave will be reduced as it propagates through tissue due to absorption and scattering. The attenuation in tissue is frequency dependent, with greater attenuation with increasing frequency. Because of absorption and dispersion, the ultrasonic wave will be attenuated as it travels through tissue. The relationship between attenuation, distance traveled, and frequency is frequently linear. Attenuation in tissue occurs as a result of both dispersion, which spreads energy in all directions, and absorption, which turns it to thermal energy.

Table 2.2: Approximate attenuation values for human tissue [17]

Tissue	Attenuation dB/(MHz · cm)
Liver	0.6–0.9
Kidney	0.8–1
Spleen	0.5–1
Fat	1–2
Blood	0.17–0.24
Plasma	0.01
Bone	16–23

The pressure of a wave propagating in  $z$ -direction decreases exponentially expressed by eq. (2.9)

$$p(z) = p(z = 0)e^{-\alpha z} \quad (2.9)$$

Where  $p(z = 0)$  is the pressure in the point of origin and  $\alpha$  is the attenuation coefficient. The attenuation coefficient unit is **Npcm**<sup>-1</sup> and alternatively, **dBcm**<sup>-1</sup> with the relationship described in eq. (2.10).

$$\alpha = \frac{1}{z} \ln \frac{p(z=0)}{p(z)} \quad (2.10a)$$

$$\alpha(\text{dBcm}^{-1}) = 20(\log_{10} e) \alpha(\text{Npcm}^{-1}) = 8.68 \alpha(\text{Npcm}^{-1}) \quad (2.10b)$$

The significance of absorption and scattering in ultrasonic attenuation in biological tissues is a point of contention. Scattering adds just a few percent to attenuation in most soft tissues. As a result, it is fair to conclude that absorption is the primary mechanism for ultrasonic attenuation in biological tissues [19].

### 2.1.3 Transducer




Figure 2.2: Single element ultrasound transducer construction [17]

A layperson knows transducers as speakers and microphones in the context of PA systems. In the case of medical US it is the device that generates the acoustic pressure field, which is emitted into tissue. The transducer has a piezoelectric crystal inside the housing. When excited, this crystal emits ultrasound waves toward flowing blood. The red blood cells will reflect a fraction of the emitted waves. These reflected waves are of a different frequency than the transmitted wave. If the red blood cells are moving away from the transducer, the frequency will be lower. If the red blood cells are moving towards the transducer, the frequency will be higher. This is caused by the Doppler effect. The reflected ultrasonic waves return to the crystal and are converted back into electrical signals. The single element transducer shown in fig. 2.2 has a minimal imaging window and has to be mechanically manipulated to get a wide window, which is unfeasible for responsive high-frequency imaging. Thus, usually an array transducer is used. Various US transducer types exist with different strengths and weaknesses, shown in fig. 2.3.

### 2.1.4 Doppler effect

The Doppler effect is a phenomena in which an observer perceives a shift in the frequency of sound emitted from a source when either the source or the observer is moving, or both are moving. The reason for the perceived change in frequency is visualised in fig. 2.4. In diagram (a), the source  $S_p$  is stationary and producing a spherical distribution pattern of the wave with the perceived frequency of the observer is given by  $f = c/\lambda$ , where  $c$  is the velocity of the wave in the medium and  $\lambda$  is the wavelength. In diagram (b), the sound source is moving towards the right with a velocity  $v$ . The locomotion of the source changes the distribution pattern and causes a longer wavelength on the left, indicating a lower perceived frequency, and a shorter wavelength on the right, indicating a higher perceived frequency, both denoted as  $\lambda'$  in the diagram. In the case of the observer on the right side, the perceived frequency becomes eq. (2.11).






Figure 2.3: Transducer types for acquiring B-mode images [17]

$$f' = \frac{c}{\lambda} = \frac{c}{\lambda - vT} = \frac{c}{(c - v)T} = \frac{c}{c - v} \cdot f_0 \quad (2.11)$$

And viceversa, on the left side, the perceived frequency becomes eq. (2.12).

$$f' = \frac{c}{c + v} \cdot f_0 \quad (2.12)$$

Where

This perceived difference between the frequency that is transmitted from the source  $f_0$ , and the perceived frequency  $f'$  is also called the Doppler frequency,  $f_d$ . When these connections are combined, the Doppler frequency for a source moving with velocity  $v$  and an observer traveling with velocity  $v'$  is given by eq. (2.13).

$$f_d = f' - f = \left( \frac{c + v'}{c - v} - 1 \right) \quad (2.13)$$

If both source and observer are moving with the same velocity,  $v$ , assuming  $c \gg v$ , the  $v$  cancels out and the expression is reduced to eq. (2.14).

$$f_d = \frac{2vf}{c} \quad (2.14)$$

If the velocity of the moving source is traveling with an incident angle  $\theta$ , the  $v$  in eq. (2.14) is replaced with  $v(\cos \theta)$ . This results in the expression found in eq. (2.15) and forms the




Figure 2.4: Doppler effect diagram. A stationary observer perceives a change in frequency of a wave generated by a moving source toward the observer as a result of a wavelength shift from  $\lambda$  to  $\lambda'$ . In (a), the source is still. In (b), the source is moving at a velocity  $v$ . [19]

basis for applied Doppler effect measurements.

$$f_d = \frac{2v(\cos \theta)f}{c} \quad (2.15)$$

The Doppler effect is utilised in ultrasonic Doppler devices used to image blood flow transcutaneously. An ultrasonic transducer in these devices sends ultrasonic waves into a blood artery, and the scattered radiation from moving red cells is measured by either the same transducer or a second transducer. The Doppler frequency, which is determined by the red blood cell velocity, is extracted using modern electronic demodulation techniques.

## 2.2 Flow physics

The flow physics of the human circulatory system are sophisticated and numerous non-stationary flow patterns emerge. The human circulatory system takes care of transporting oxygen and nutrients to the organs as well as disposing of waste products produced by metabolism. It is possible because the blood within the circulatory system contains several smaller subcomponents such as plasma and formed cellular elements that perform these vital functions. Initially, blood is discharged from the left ventricle of the heart via the aorta and travels to all areas of the body via multiple branches of the arterial tree. When the blood flows through the arteries, they into smaller channels known as arterioles. These arterioles lead into a network of tiny capillaries via which nutrients and waste materials are exchanged between the blood and the organs. The capillaries connect to form a network of venules, which supply the veins, which deliver blood back to the heart. This system, in its totality, is called the systemic circulation. A diagram of the circulatory system as described above, can be seen in fig. 2.5. In summary, when examining the elements that comprise the circulatory system, it consists of several components:

- Heart, primary organ of the circulatory system that maintains blood pressure and controls blood velocity.
- Blood, and its sub-components
  - Plasma, which forms the primary volume and contains nutrients and formed cellular elements.




Figure 2.5: Circulatory system of the human body [17]

- Red and white blood cells, which carry oxygen and fight off infections, respectively.
- Platelets, which are also known as thrombocytes, with the function to clot during a blood vessel injury.

- Blood vessels
  - Arteries (and arterioles), that transport oxygenised blood to organs and tissues at high pressure and velocity.
  - Capillaries, thin but wide ranging blood vessels that performs the exchange of matter between the circulatory system and tissue.
  - Veins (and venules), carries blood back to the heart at low pressure and velocity.

### 2.2.1 Blood flow

Blood flow, which is the amount of blood that goes through a blood vessel in a particular period of time, and has a complicated flow pattern due to its pulsing flow. An advanced analysis of haemodynamics is not within the scope of this report, so the explanation will be brief. The primary forces that determines the blood flow  $F$  are the pressure difference across a blood vessel and vascular resistance. It is determined by Ohm's law as in eq. (2.16).

$$F = \frac{\Delta P}{R} \quad (2.16)$$

Where  $\Delta P$  is the pressure difference across the blood vessel and  $R$  is the vascular resistance. The pressure difference  $\Delta P$  is calculated with eq. (2.17).

$$\Delta P = P_1 - P_2 \quad (2.17)$$

Where  $P_1$  and  $P_2$  are the blood pressures measured at each end of the blood vessel. Pressure has a significant importance on blood flow because an increase in arterial pressure not only increases the force that pushes blood through the capillaries, but it also expands the vessels, lowering vascular resistance.

## 2.3 Devices




Figure 2.6: Diagram of US wave transmitted and reaching blood vessel with incident angle  $\theta$  [19]

A device that measures the flowing of blood is called a flowmeter. Flowmeters may be used both inside and outside of vessels. One of the flowmeters that may be used outside the vessel to monitor flow is US. Figure 2.6 depicts an ultrasonic wave of frequency  $f$  insonifying a blood artery, resulting in an angle of  $\theta$  relative to velocity  $v$ . For simplicity,

it is assumed that blood flows in a vessel at a constant velocity  $v$ . The echoes returned are shifted in frequency as described in eq. (2.15) in earlier in the chapter. The echoes scattered by blood after being insonified by an ultrasonic wave convey information about the velocity of blood flow. Blood flow measurements are often used in clinical settings to determine the status of blood vessels and organ functioning. The two commonly used fundamental techniques for ultrasound Doppler flow measurements are continuous-wave (CW) and pulsed-wave (PW). Both will be explained.

### 2.3.1 Continuous-wave flowmeter




Figure 2.7: Block diagram of CW flowmeter [17]

The earliest non-invasive cardiovascular diagnostic technologies relied heavily on CW Doppler flowmeters. One of the earliest concepts for a device to estimate and study blood flow was proposed by Satomura, Yoshida, Mori, et al.[1] during the 1950's in Japan. To continuously transmit waves and receive signals from moving reflectors, the CW flowmeter uses two transducers. CW flowmeters use less sophisticated electronics than PW flowmeters. A drawback to the CW flowmeter is the lacking depth discrimination due to the continuous characteristic of this device type. A block diagram of a typical CW flowmeter can be seen in fig. 2.7. The basic principles of the device is previously explained in section 2.1.4, and the measurement of the device is described in eq. (2.11). The device continuously emits an ultrasonic wave in the first transducer expressed as a function of time by eq. (2.18) [17].

$$e(t) = \cos(2\pi f_0 t) \quad (2.18)$$

While receiving the backscattered signal on the second transducer expressed by eq. (2.19) [17].

$$r_s(t) = a \cos(2\pi f_0 \alpha(t - t_0)) \quad (2.19)$$

$$\alpha \approx 1 - \frac{2v_z}{c} \quad (2.20)$$

$$\alpha t_0 \approx \frac{2d_0}{c} \quad (2.21)$$

Where  $v_z$  indicates the velocity in the  $z$ -direction. Applying the Fourier transform, the expression yields eq. (2.22).

$$r_s(t) \cdot e^{j2\pi f_0 t} \Longleftrightarrow R_s(f - f_0) \quad (2.22)$$

Where  $R_s(f - f_0)$  is the Fourier transform of  $r_s(t)$ . The received signal is then multiplied

with a quadrature signal of frequency  $f_0$  to find the Doppler frequency in eq. (2.23).

$$m(t) = a [\cos(2\pi f_0 t) + j \sin(2\pi f_0 t)] \cos(2\pi f_0 \alpha(t - t_0)) \quad (2.23)$$

$$= \frac{a}{2} \left\{ \cos(2\pi f_0 [(1 - \alpha)t - \alpha t_0]) + \cos(2\pi f_0 [(1 + \alpha)t - \alpha t_0]) \right. \\ \left. + j \sin(2\pi f_0 [(1 - \alpha)t - \alpha t_0]) + j \sin(2\pi f_0 [(1 + \alpha)t - \alpha t_0]) \right\} \quad (2.24)$$

As is general for quadrature demodulation, the resulting signal contains the frequency components of the sum and difference of the emitted and received signals' frequencies shown in fig. 2.8, where the signals are shown in time and frequency domains.




Figure 2.8: Doppler signals in time and frequency domain showing demodulation effects [19]

Generally a band-pass (BP) filter is used on the demodulated signal to remove the high frequency summed signal at twice the frequency of  $f_0$ . The filtered signal after the BP filter is expressed by eq. (2.25) and contains the Doppler shift of the emitted signal.

$$m_f(t) \approx \frac{a}{2} e^{(j2\pi f_0 \frac{2v_z}{c} t)} e^{(-j2\pi f_0 \alpha t_0)} \quad (2.25)$$

Where the second exponential term is the delay proportional to the time between transmission and receiving of the signal. The selected cutoff frequency is chosen to be much lower than the carrier frequency to remove the carrier wave. One issue with ultrasonic Doppler blood flow monitoring is that the blood vessels that generate large reflected echoes are

also moving with a low velocity. These big, slow-moving echoes are referred to as clutter signals in Doppler nomenclature. The band pass filter's low-end cutoff frequency must be designed to minimize interference from these clutter signals. The design of this band pass filter in the low-frequency region, which serves the function of high pass, also known as a clutter rejection filter, has proven troublesome since the magnitude of clutter signals is many orders greater than that of blood and may obfuscate those from slow-moving blood.

Table 2.3: Measured frequency shifts with a Doppler **3 MHz** transducer at various velocities at a  $45^\circ$  incident angle [17]

Velocity ( $v$ ) m/s	Doppler frequency ( $f_d$ ) Hz
0.01	28
0.1	276
0.5	1377
1	2755
2	5510
5	13 770

Seen in table 2.3 is an example of measured Doppler frequencies using a **3 MHz** transducer using the method shown in fig. 2.6. Note that the frequencies measured are all within the audible range.

### 2.3.2 Pulsed-wave flowmeter




Figure 2.9: Block diagram of PW flowmeter [17]

The concept of a pulsed-wave flowmeter was proposed in [2] and other related articles. In this type of flowmeter is periodically changing from a transmitter to a receiver. In the transmitting mode, the transducer emits a series of pulses. When in the receiving mode, the transducer is listening for the back-scattered signal. A simplified block diagram can be seen in fig. 2.9. The movement of particles within the blood cause a displacement in the back-scattered signal. These systems are commonly referred to as “Doppler systems” even though it is somewhat misleading. The effects of attenuation is also causing a shift in frequency of a higher magnitude than the velocity of particles in blood. This is because the conventional Doppler effect is not the straight forward methodology that is applied to the analysis of the back-scattered signal. It is, in fact, an artifact. It is the shift in the location of the scatters that is observed, not the shift in the transmitted frequency. Figure 2.10 shows the received signal after demodulation and filtering; the depth in tissue is fixed here, and the signals displayed on the left side of the figure are the result of a pulse sequence. Each line represents a single pulse, and each pulse is emitted at a pulse repetition frequency,  $f_{prf}$ . Instead, on the right side, the dotted line shows the sampled

signal formed by taking into account the amplitude of each pulse after a specified time period.




Figure 2.10: Sampling for a gate pulsed wave system with a single range. To depict the signals on the graph, a single pulse is emitted for each line, and the signals are displaced in amplitude. The sampled signal is displayed on the right. [17]

After the back-scattered signal is received it is multiplied by the center frequency of the emitted pulse and filtered to remove the sum frequency [17]. A analogue-to-digital converter (ADC) quantifies the signal for further signal processing. Referring to displacement fig. 2.10 again, the dashed vertical line represents the sample of each pulse that is taken. If sampling is done  $T_s$  after pulse emission, the measurement depth is expressed by eq. (2.26).

$$d_0 = \frac{T_s c}{2} \quad (2.26)$$

Hypothetically, if the velocity of stationary scatterers in blood was measured, a constant amplitude would be measured. A change in sample value is observed when movement is present. Between two pulses the scatterer movement is proportional to the velocity  $v_z$  in the direction of the ultrasound beam. Time shift of the  $t_s$  is expressed as eq. (2.27).

$$t_s = \frac{2v_z}{c} \cdot T_{\text{prf}} \quad (2.27)$$

Where  $c$  is the speed of sound, and  $T_{\text{prf}}$  is the timespan between each pulse emission. Taking one sample from each line at a certain depth yields a sampled signal with a frequency proportional to the scatter velocity. Thus, if a sample is taken at the same depth for each line, resulting in a sinusoidal signal proportional in frequency to the scatter velocity [7] and that signal is expressed by eqs. (2.28a) and (2.28b).

$$r(i) = a(i) \sin(2\pi f_p T_{\text{prf}} \cdot i) \quad (2.28a)$$

$$f_p = \frac{2v_z}{c} f_0 \quad (2.28b)$$



Where  $a(i)$  is the amplitude,  $f_0$  is the emitted frequency, and  $\theta$  is the phase factor in the depth of interest. This technique improved the accuracy of the investigations of blood vessels and facilitated the display of velocity profiles. Furthermore, employing two transducers or a multi-element transducer, duplex mode imaging (displaying both a B-mode picture and a blood velocity estimate) became feasible. Two-transducer systems are no longer utilised, since it is easier to create a duplex picture with a multi-element transducer.

Hvor er  $\theta$  i  
udtrykket?  
Omskriv fra  
Jensen2012



# Bibliography

- [1] S. Satomura, T. Yoshida, M. Mori, Y. Nimura, G.-i. Hikita, S. Takagishi, and K. Nakanishi, "Analysis of heart motion with ultrasonic Doppler method and its clinical application," *American Heart Journal*, vol. 61, pp. 61–75, 1 January 1961, ISSN: 00028703. DOI: 10.1016/0002-8703(61)90517-8. [Online]. Available: <https://linkinghub.elsevier.com/retrieve/pii/0002870361905178>.
- [2] D. Baker, "Pulsed Ultrasonic Doppler Blood-Flow Sensing," *IEEE Transactions on Sonics and Ultrasonics*, vol. 17, pp. 170–184, 3 Jul. 1970, ISSN: 0018-9537. DOI: 10.1109/t-su.1970.29558. [Online]. Available: <http://ieeexplore.ieee.org/document/1538563/>.
- [3] K. K. Shung and G. A. Thieme, *Ultrasonic scattering in biological tissues*. CRC Press, 1993, p. 499, ISBN: 9780849365683. [Online]. Available: <https://www.routledge.com/Ultrasonic-Scattering-in-Biological-Tissues/Shung-Thieme/p/book/9780849365683>.
- [4] J. A. Jensen, "An Analysis of Pulsed Wave Ultrasound Systems for Blood Velocity Estimation," *Acoustical Imaging*, L. T. Piero and Masotti, Eds., pp. 377–384, 1996. DOI: 10.1007/978-1-4419-8772-3\_\_61. [Online]. Available: <http://link.springer.com/10.1007/978-1-4419-8772-3%5C%5F61>.
- [5] P. J. Fish, "Ultrasonic investigation of blood flow," *Proceedings of the Institution of Mechanical Engineers, Part H: Journal of Engineering in Medicine*, vol. 213, pp. 169–180, 3 March 1999, Pmid: 10420772. DOI: 10.1243/0954411991534898. [Online]. Available: <https://doi.org/10.1243/0954411991534898>.
- [6] T. Jansson, H. W. Peresson, and K. Lindström, "Estimation of blood perfusion using ultrasound," *Proceedings of the Institution of Mechanical Engineers, Part H: Journal of Engineering in Medicine*, vol. 213, pp. 193–193, 3 March 1999, ISSN: 0954-4119. DOI: 10.1243/0954411991534915. [Online]. Available: <http://journals.sagepub.com/doi/10.1243/0954411991534915>.
- [7] P. Munk, "Estimation of blood velocity vectors using ultrasound," *Technical University of Denmark*, 2000. [Online]. Available: <https://findit.dtu.dk/en/catalog/537f0d537401dbcc1200be04%20http://server.oersted.dtu.dk/publications/views/publication%5C%5Fdetails.php?id=827>.
- [8] E. Picano, "Sustainability of medical imaging," *Bmj*, vol. 328, pp. 578–580, 7439 March 2004, ISSN: 0959-8138. DOI: 10.1136/bmj.328.7439.578. [Online]. Available: <https://www.bmj.com/lookup/doi/10.1136/bmj.328.7439.578>.
- [9] E. Chérin, R. Williams, A. Needles, G. Liu, C. White, A. S. Brown, Y.-Q. Zhou, and F. S. Foster, "Ultrahigh frame rate retrospective ultrasound microimaging and blood flow visualization in mice in vivo," *Ultrasound in Medicine & Biology*, vol. 32, pp. 683–691, 5 May 2006, ISSN: 03015629. DOI: 10.1016/j.ultrasmedbio.2005.12.015. [Online]. Available: <https://linkinghub.elsevier.com/retrieve/pii/S0301562906000172>.
- [10] X. Xu, J. Yen, and K. K. Shung, "A low-cost bipolar pulse generator for high-frequency ultrasound applications," *IEEE Transactions on Ultrasonics, Ferroelectrics and Frequency Control*, vol. 54, pp. 443–447, 2 February 2007, ISSN: 0885-3010. DOI: 10.1109/tuffc.2007.259. [Online]. Available: <http://ieeexplore.ieee.org/document/4107704/>.
- [11] I. K. H. Tsang, B. Y. S. Yiu, D. K. H. Cheung, H. C. T. Chiu, C. C. P. Cheung, and A. C. H. Yu, "Design of a multi-channel pre-beamform data acquisition system for an ultrasound research scanner," *Ieee*, Sep. 2009, pp. 1840–1843, ISBN: 978-1-4244-

- 4389-5. DOI: 10.1109/ultsym.2009.5441869. [Online]. Available: <http://ieeexplore.ieee.org/document/5441869/>.
- [12] E. A. Aydın and İ. Güler, "Design Of Pic-controlled Pulsed Ultrasonic Transmitter For Measuring Gingiva Thickness," *Instrumentation Science & Technology*, vol. 38, pp. 411–420, 6 October 2010, ISSN: 1073-9149. DOI: 10.1080/10739149.2010.509149. [Online]. Available: <http://www.tandfonline.com/doi/abs/10.1080/10739149.2010.509149>.
- [13] N. Matsuoka, D.-G. Paeng, R. Chen, H. Ameri, W. Abdallah, Q. Zhou, A. Fawzi, K. K. Shung, and M. Humayun, "Ultrasonic Doppler measurements of blood flow velocity of rabbit retinal vessels using a 45-MHz needle transducer," *Graefe's Archive for Clinical and Experimental Ophthalmology*, vol. 248, pp. 675–680, 5 2010, ISSN: 1435-702x. DOI: 10.1007/s00417-009-1298-9. [Online]. Available: <https://doi.org/10.1007/s00417-009-1298-9>.
- [14] K. L. Hansen, M. B. Nielsen, and J. A. Jensen, "In-vivo studies of new vector velocity and adaptive spectral estimators in medical ultrasound," 2011. [Online]. Available: <https://orbit.dtu.dk/en/publications/in-vivo-studies-of-new-vector-velocity-and-adaptive-spectral-esti>.
- [15] C.-C. Huang, P.-Y. Lee, P.-Y. Chen, and T.-Y. Liu, "Design and implementation of a smartphone-based portable ultrasound pulsed-wave doppler device for blood flow measurement," *IEEE Transactions on Ultrasonics, Ferroelectrics and Frequency Control*, vol. 59, pp. 182–188, 1 January 2012, ISSN: 0885-3010. DOI: 10.1109/tuffc.2012.2171. [Online]. Available: <http://ieeexplore.ieee.org/document/6138742/>.
- [16] C. A. Winckler, P. R. Smith, D. M. J. Cowell, O. Olagunju, and S. Freear, "The design of a high speed receiver system for an ultrasound array research platform," *Ieee*, October 2012, pp. 1481–1484, ISBN: 978-1-4673-4562-0. DOI: 10.1109/ultsym.2012.0370. [Online]. Available: <http://ieeexplore.ieee.org/document/6562380/>.
- [17] J. A. Jensen, *Estimation of Blood Velocities Using Ultrasound: A Signal Processing Approach*, Third Edition. Department of Electrical Engineering, Technical University of Denmark, August 2013, ISBN: 9780521464840.
- [18] T. L. Szabo, *Diagnostic Ultrasound Imaging: Inside Out*, Second Edition. Elsevier, 2014, ISBN: 9780123964878. DOI: 10.1016/c2011-0-07261-7. [Online]. Available: <https://linkinghub.elsevier.com/retrieve/pii/C20110072617>.
- [19] K. K. Shung, *Diagnostic Ultrasound: Imaging and Blood Flow Measurements*, Second Edition. CRC Press, December 2015, ISBN: 978-1-4665-8264-4.
- [20] P. Govindan, B. Wang, P. Ravi, and J. Saniie, "Hardware and software architectures for computationally efficient three-dimensional ultrasonic data compression," *IET Circuits, Devices & Systems*, vol. 10, pp. 54–61, 1 January 2016, ISSN: 1751-858x. DOI: 10.1049/iet-cds.2015.0083. [Online]. Available: <https://onlinelibrary.wiley.com/doi/10.1049/iet-cds.2015.0083>.
- [21] B. Wang and J. Saniie, "Ultrasonic Signal Acquisition and Processing platform based on Zynq SoC," vol. 2016-August, *Ieee*, May 2016, pp. 0448–0451, ISBN: 978-1-4673-9985-2. DOI: 10.1109/eit.2016.7535282. [Online]. Available: <http://ieeexplore.ieee.org/document/7535282/>.
- [22] B. Wang and J. Saniie, "A High Performance Ultrasonic System for Flaw Detection," *Ieee*, October 2019, pp. 840–843, ISBN: 978-1-7281-4596-9. DOI: 10.1109/ultsym.2019.8926280. [Online]. Available: <https://ieeexplore.ieee.org/document/8926280/>.
- [23] H. Ding, D. Yang, J. Xu, X. Chen, X. Le, Y. Wang, L. Lin, and J. Xie, "Pulsed Wave Doppler Ultrasound Using 3.7 MHz Pmutts Toward Wearable Blood Flow Measurements," *Ieee*, January 2020, pp. 400–403, ISBN: 978-1-7281-3581-6. DOI:

- 10.1109/mems46641.2020.9056430. [Online]. Available: <https://ieeexplore.ieee.org/document/9056430/>.
- [24] B. Jana, R. Biswas, P. K. Nath, G. Saha, and S. Banerjee, "Smartphone-Based Point-of-Care System Using Continuous-Wave Portable Doppler," *IEEE Transactions on Instrumentation and Measurement*, vol. 69, pp. 8352–8361, 10 October 2020, ISSN: 0018-9456. DOI: 10.1109/tim.2020.2987654. [Online]. Available: <https://ieeexplore.ieee.org/document/9064842/>.
- [25] H. Ding, D. Yang, M. Qu, et al., "A Pulsed Wave Doppler Ultrasound Blood Flowmeter by PMUTs," *Journal of Microelectromechanical Systems*, vol. 30, pp. 680–682, 5 October 2021, ISSN: 1941-0158. DOI: 10.1109/jmems.2021.3103756. [Online]. Available: <https://ieeexplore.ieee.org/abstract/document/9515180>.
- [26] M. Winder, A. J. Owczarek, J. Chudek, J. Pilch-Kowalczyk, and J. Baron, "Are We Overdoing It? Changes in Diagnostic Imaging Workload during the Years 2010–2020 including the Impact of the SARS-CoV-2 Pandemic," *Healthcare*, vol. 9, p. 1557, 11 November 2021, ISSN: 2227-9032. DOI: 10.3390/healthcare9111557. [Online]. Available: <https://www.mdpi.com/2227-9032/9/11/1557>.

# A Bill of Materials

Table A.1: Bill of Materials for the entire system

Component	Value	Footprint	Classification	Description
IO				
C1	1n	0603	X7R50V	Preamp
C13	1u	RAD-0.3in	Film 40V	Preamp
C14	1.5n	0603	X7R50V	PI controller
C15	NC	0603	X7R50V	PI controller
C2, C3, C4, C9, C10	100n	0603	X7R16V	Decoupling
C5, C11	1u	0603	X5R6.3V	Decoupling
C6, C7	100n	0805	X7R50V	Decoupling
C8	1500u	RAD-0.3in	63Vdc	Decoupling
P10	3-pin	Male	250V	Controller by-pass
P11, P12	10-pin	Female	250V	InterPCB con.
P2, P3	1-pin header	Male	250V	Measurements
P4, P5	2-way screw	NA	300V 15A	30V Power, Output
P6	2-pin header	Molex KK254	500V 4A	Audio in
P7	4-pin header	Female	250V	5V Power
P8, P9	BNC	BNC PCB	500V	Audio in, Output
R2	16.2k	0603	1%	Preamp
R1, R3	2k	0603	1%	Preamp
R4, R7	4.75k	0603	1%	Voltage ref
R5	33.2k	0603	NA	PI controller
R6	0 (short)	0603	NA	PI controller
R9	0 (short)	0603	NA	PI controller
R10	NC	0603	NA	PI controller
R8	2.49k	0603	1%	Voltage ref
U1	OPA2365	SOIC-8	NA	Preamp, PI
U2	TLV431A	SOT-23	1%	Voltage ref

Table A.1: Bill of Materials for the entire system (Continued)

Component	Value	Footprint	Classification	Description
Power stage				
C1, C11, C15, C19	100n	0603	X7R16V	Gate driver
P1, P2	10-pin header	Male	250V	InterPCB con.
C2, C16	150n	0603	X5R10V	Gate driver
C3, C10	100p	0603	NPO50V	Decoupling
L1, L2	1.768u	Radial	NA	Output filter
U1, U3	LM5113	WSO-10	NA	Gate driver
R20	15m sense	1210	1% 1W	Output filter
R1, R6	500	SMDtrim 3213	NA	Gate driver
Q1, Q2, Q3, Q4	BSZ097-N10NS5	TSDSON-8	NA	Power stage
R2, R5, R8, R10	5	0603	1%	Gate driver
R4, R9	0	0603	1%	Gate driver
D1, D2	Diode	NA	85V 0.25A	Gate driver
C5, C6, C9, C17	10u	1210	X7R50V	Decoupling
C12, C13, C14	680n	1210	X7R100V	Output filter
AIM regulator				
C18,C20,C27	100n	0603	X7R16V	Decoupling
C28, C29, C30, C31	100n	0603	X7R16V	Decoupling
U6	LT1999	MSOP-8	NA	Current acq.
U4	AD8274	MSOP-8	NA	Volt acq.
U2	LT1711	MSOP-8	NA	AIM
C4	NC	0603	X7R16V	AIM
C7	NC	0603	X7R16V	AIM
C8	NC	0603	X7R16V	Cur. acq.
CA1	1.5n	0603	X7R50V	AIM
R14	8.66k	0603	1%	AIM
R18	16.2k	0603	1%	AIM
R23	1.5k	0603	1%	AIM
R3, R11	120k	0603	1%	Voltage acq.
RA1	2.2k	0603	1%	AIM
RA2	20k	0603	1%	AIM

Continued on next page

Table A.1: Bill of Materials for the entire system (Continued)

Component	Value	Footprint	Classification	Description
RA3	2.74k	0603	1%	AIM
RA4	0	0603	NA	AIM



## B Instruments

Table B.1: List of instruments used for solder work

Function	Manufacturer	Model
Visual inspection microscope	Leica	A60
Manual soldering	Weller	WX2
Heat gun	Thermaltronics	TMT-HA600-2
SMT solder paste dispensing	Fisnar	SL101N
Reflow oven	Mistral	260
DMM	Meterman	38XR

Table B.2: List of instruments used for testing

Function	Manufacturer	Model
Power supply	RIGOL	DP832A
DMM	Agilent	34410A
Frequency response analysis	N4L	PSM1735
Audio analysis	Audio Precision	APx500
Audio analysis filter	Audio Precision	AUX-0025
Control loop analyzer	OMICRON Lab	BODE 100
Control loop injection transformer	OMICRON Lab	B-WIT 100
Passive load	Unknown	Power resistor $4\Omega$
Passive load	Unknown	Loudspeaker $4\Omega$
Active cooling	ebm	Test of bold

Lorem ipsum dolor sit amet, consectetur adipiscing elit. Etiam lobortis facilisis sem. Nullam nec mi et neque pharetra sollicitudin. Praesent imperdiet mi nec ante. Donec ullamcorper, felis non sodales commodo, lectus velit ultrices augue, a dignissim nibh lectus placerat pede. Vivamus nunc nunc, molestie ut, ultricies vel, semper in, velit. Ut porttitor. Praesent in sapien. Lorem ipsum dolor sit amet, consectetur adipiscing elit. Duis fringilla tristique neque. Sed interdum libero ut metus. Pellentesque placerat. Nam rutrum augue a leo. Morbi sed elit sit amet ante lobortis sollicitudin. Praesent blandit blandit mauris. Praesent lectus tellus, aliquet aliquam, luctus a, egestas a, turpis. Mauris lacinia lorem sit amet ipsum. Nunc quis urna dictum turpis accumsan semper.

DTU Electrical Engineering  
Technical University  
of Denmark

Ørstedes Plads, Building 348  
2800 Kgs. Lyngby  
Tel. (+45) 4525 3800

[www.elektro.dtu.dk](http://www.elektro.dtu.dk)

KAIST EE  
Korea Advanced Institute  
of Science & Technology

E3-2, 291  
34141  
Tel. (+82) 042-350-2114

[www.ee.kaist.ac.kr](http://www.ee.kaist.ac.kr)