

A Piezoelectric-Driven High-Throughput Microfluidic Platform for Optimizing Biomanufacturing Process for Cell-based Immunotherapy Cancer Treatment

Mohammad Kader^{1,2}, Matthew Kuan^{1,2}, Aldey Brutus^{1,3}, Brandon Avena⁴, M. Ete Chan^{1,2}

¹Bioengineering Education, Application and Research (BEAR) Vertically Integrated Project Team, ²Department of Biomedical Engineering, Stony Brook University, ³Department of Electrical Engineering, Stony Brook University, ⁴Department of Mechanical Engineering, Stony Brook University

Background

Clinical Unmet Needs:

- Chimeric Antigen Receptor (CAR) T-cell therapy is a promising immunotherapy for treating several types of cancer.¹
- Successful therapy depends on expanding large numbers of functional T-cells outside the body before reinfusion into patients.
- Current T-cell expansion methods require long culture times, significant resources, and high manufacturing costs.
- Improving T-cell proliferation is a key challenge for making CAR T-cell therapies more scalable and accessible.

LIV as a Potential Solution:

- Low-intensity vibration (LIV) is a non-drug mechanical stimulation approach that delivers controlled sinusoidal signals to biological systems.
- LIV signals are defined by parameters such as frequency, acceleration, displacement, treatment duration, and refractory period.
- By mimicking mechanical stimulation naturally produced during exercise, LIV may influence cellular activity and promote T-cell proliferation.
- Previous lab studies using voice-coil actuator LIV devices showed up to a 29% increase in T-cell proliferation compared to non-LIV controls.²

Needs for Optimization of LIV Signals:

- T-cell responses to LIV may vary based on cell type, cell size, patient age, immune status, health condition, and disease state, suggesting LIV is not a “one-size-fits-all” treatment.
- Identifying optimal LIV conditions requires testing multiple vibration parameters in parallel.
- Current voice-coil actuator systems are too large for high-throughput optimization and require large numbers of primary T-cells, which are difficult and expensive to obtain.




Fig 1. Typical LIV experimental setup and signal parameters.⁴

Design Objectives

- Develop PDMS-based microfluidic devices with multiple culture chambers to support parallel T-cell proliferation experiments using small cell populations.
- Adapt a proven primary T-cell microfluidic device that retains T-cells inside of the device during media renewal while limiting washout and shear stress.
- Integrate a piezoelectric actuator system to deliver controlled LIV across tunable parameters, including frequency, acceleration, displacement, treatment duration, and refractory period.
- Establish a PDMS casting workflow that allows microchannel and culture chamber geometry to be replicated and bonded to glass for live cell imaging.
- Establish a compact, high-throughput platform for optimizing LIV signal parameters while reducing device size and primary T-cell requirements.

Microfluidics Device Design




Fig. 2. Microfluidic device design and trapping mechanism. (A) Primary T-cell microfluidic culture device with parallel channels and repeated microchambers/wells for cell retention, media renewal, and long-term imaging.⁵ **(B)** Lab-in-a-Trench concept showing gravity-based cell settling into trench regions during media flow.³

- Device design is adapted from a multilevel microfluidic platform shown to successfully culture and monitor primary human T-cells over multiple days.⁵
- The chip uses parallel microchannels with small microchambers/wells that physically trap suspension T-cells and prevent them from being washed away during media renewal.
- Cells settle into the deeper center of each well, where they are protected from flow-induced shear stress while still receiving fresh culture media.
- This design enables long-term imaging of the same T-cell clusters, allowing cell growth and proliferation to be tracked over time.

Microfluidics Device Fabrication




Fig. 3. Proposed PDMS casting workflow.⁶ PDMS is poured over a master mold, cured to replicate the microchannel features, released from the mold, punched to create inlet/outlet ports, plasma-treated, and bonded to glass to form a sealed microfluidic device.

- Microfluidic device fabrication will follow the PDMS casting workflow.
- A master mold will be used to replicate the microchannel and culture chamber geometry.
- The master mold fabrication method is still being evaluated, with options including SU-8 lithography, 3D printing, Two-Photon Polymerization, or epoxy replica molding.
- The final PDMS device will be plasma-bonded to glass and integrated with the piezoelectric LIV platform.

Piezoelectric Actuator




Fig 4. 3D model of the proposed piezoelectric actuator platform and the selected actuator for LIV stimulation.

- Custom 3D-printed platform designed to support the microfluidic device during LIV stimulation.
- Four corner lead screws with hex nut inserts provide alignment, stability, and adjustable positioning.
- A springboard supports the device and allows controlled displacement.
- A piezoelectric actuator makes direct contact with the microfluidic device to deliver vibration.

Future Work/Long-term Goals

- Select and finalize the master mold fabrication method for PDMS device casting.
- Fabricate and test PDMS microfluidic devices for channel integrity, bonding quality, and media flow.
- Calibrate the piezoelectric platform to verify LIV parameters, including frequency, acceleration, and displacement.
- Optimize primary T-cell culture conditions based on the original microfluidic T-cell culture protocol and make necessary modifications for LIV stimulation experiments.
- Conduct control and LIV-treated T-cell culture experiments to evaluate growth, proliferation, and viability.
- Assess cell response using live-cell microscopy, CFSE proliferation assay, and Propidium Iodide (PI) viability assay.

Acknowledgements

We would like to acknowledge Dr. Chan and Andreas Kaimis for their mentorship, technical guidance, and support during this project.

References

- Ceja, Melanie Ayala, Mobina Khericha, Caitlin M. Harris, Cristina Puig-Saus, and Yvonne Y. Chen. 2024. “CAR-T Cell Manufacturing: Major Process Parameters and Next-generation Strategies.” *The Journal of Experimental Medicine* 221 (2). <https://doi.org/10.1084/jem.20230903>.
- Chan, M. Ete, Lia Strait, Christopher Ashdown, Sishir Pasumath, Abdullah Hassan, Steven Crimmarco, Chanpreet Singh, et al. 2023. “Low Intensity Mechanical Signals Promote Proliferation in a Cell-specific Manner: Tailoring a Non-drug Strategy to Enhance Biomanufacturing Yields.” *bioRxiv (Cold Spring Harbor Laboratory)*, July. <https://doi.org/10.1101/2023.07.05.547864>.
- O’Connell, Triona M., Damien King, Chandra K. Dixit, Brendan O’Connor, Dermot Walls, and Jens Duerée. 2014. “Sequential Glycan Profiling at Single Cell Level With the Microfluidic Lab-in-a-trench Platform: A New Era in Experimental Cell Biology.” *Lab on a Chip* 14 (18): 3629–39. <https://doi.org/10.1039/c4lc00618f>.
- Pasumath, S, S Crimmarco, C Singh, CT Rubin, and Department of Biomedical Engineering, Stony Brook University. n.d. “Low Intensity Vibration May Accelerate CAR T Cell Biomanufacturing for Cancer Immunotherapy.” *CAR-T Cell Therapy: Brief Introduction*. <https://bpb-us-e1.wpmucdn.com/you.stonybrook.edu/dist/c/4606/files/2020/11/BMES-2020-Presentation.pdf>.
- Peñaherrera-Pazmiño, Ana Belén, Gustavo Rosero, Dario Duarte, Julia Pinter, Karla Vizcete, Maximiliano Perez, Marie Follo, Betiana Lerner, and Roland Mertelsmann. 2025. “Activation and Expansion of Human T-Cells Using Microfluidic Devices.” *Biosensors* 15 (5): 270. <https://doi.org/10.3390/bios15050270>.
- Team, Elveflow. 2025. “PDMS: The Versatile Polymer for Microfluidic Chips | Elveflow.” Elveflow. April 29, 2025. <https://elveflow.com/microfluidic-reviews/the-polydimethylsiloxane-pdms-and-microfluidics/>.