InBody

BWA 2.0S

3MHz Technology

Remarkable precision with world's first 3 MHz frequency technology

Ergonomic Clamp Electrodes

Enhanced user experience with painless, reusable clamp electrodes

190+ Health Data

190+ health data in 30 seconds for thorough analysis

BWA2.0S_catalog_eng_B_250306.indd 1 2025-03-10 오전 9:12:15

A New Standard for Body Water Analysis

In the last 20 years, body composition analysis has established itself as a standard practice in various fields, and InBody has continuously strived to further expand its application to specialized areas, such as dialysis, rehabilitation, nutrition, and etc.

With the need for the precise measurement of body water, InBody introduces a new standard for body water analysis, BWA 2.0S.

The BWA 2.0S is equipped with state-of-the-art 3MHz technology and provides extensive research parameters for professionals to better suit diverse patients with different conditions and medical specialties than ever before.

- Cole-Cole Plot Graph for Monitoring Changes in Body Water and Cellular Integrity
- Statistical Analysis by Age, Based on InBody Big Data
- Clamp Electrode for High Reproducibility
- Covering Wide Range of Subjects / Patients and Conditions
- Extensive Research Parameters for Professionals

InBody Technology

InBody uses Bioelectrical Impedance Analysis (BIA) technology to measure human body composition. Impedance is the resistance of the human body generated when a micro alternating current flows through the human body. The human body is made of water that conducts electricity well, and the resistance varies depending on the amount of water. BIA is a technology that quantitatively measures body water through impedance that occurs when an electric current flows through the human body.
InBody provides diverse information on body composition based on the measured body water.

Direct Segmental Measurement-BIA

InBody measures impedance independently in the arms, legs, and trunk. Unlike traditional BIA, which models the human body as a single cylinder, DSM-BIA accounts for variations in shape, density, and cross-sectional area across body segments. This allows InBody devices to provide more accurate and segment-specific body composition analysis than conventional methods.



High Reproducibility and Accuracy Assured by 16-Point Clamp Electrodes

The 16-Point Clamp Electrodes were developed in a way so that the electrodes can be positioned on the wrist and ankle bone. It allows the instructor to place the electrode in the proper position and secures the reproducibility my minimizing the measurement errors.

Accurate Fluid Assessment Using Cole-Cole Plot, Z_0, and $Z\infty$

BWA2.0S utilizes frequencies up to 3 MHz to more precisely measure total body water, considering the frequency-dependent permeability of cell membranes. Segmental phase angle measurements at 5, 50, and 250 kHz enable accurate detection and monitoring of edema often associated with various clinical conditions. The zero frequency impedance (Z_0) better reflects extracellular water, while the infinite frequency impedance ($Z\infty$) offers a more accurate representation of both intracellular and extracellular water.



No Estimations or Empirical Estimations

InBody does not rely on empirical estimations based on age, gender, and more to ensure the accuracy of the measured data. By eliminating generalized assumptions and focusing solely on measurable data, InBody accurately detects even subtle changes in body composition. This approach has been widely adopted in clinical treatments and research that require continuous and precise monitoring of body composition changes.



Body Composition Evaluation by Age Based on InBody Big Data

InBody provides age-specific graphs for each body composition analysis parameter based on globally accumulated InBody Data. With this, a comprehensive analysis is provided so that you can compare your data of the young age group (T-score) and the same age group (Z-score).



* *InBody Big Data is used for the evaluation by age which is shown as T-Score and Z-score that indicate the relative position of subject. It does not affect the subjects' body composition analysis result.*

* *Depending on the country, the graph will be set differently.*

With Over 5,500 Research Studies and Counting

Study 1 ## HIGH ACCURACY AND REPRODUCIBILITY OF FAT FREE MASS & PERCENT BODY FAT MEASUREMENTS COMPARED WITH DEXA

The measurement (mean ± SD) for FFM with DXA was 52.8 ± 11.0, and BIA was 53.6 ± 11.0., Delta (S-MFBIA vs DXA) was 0.8 ± 2.2 (5 % limits of agreement −3.5 to +5.2), and concordance correlation coefficient (CCC) was 0.98 (95 % CI, 0.97–0.98).
The measurements (mean ± SD) for PBF with DXA was 37.5 ± 10.6 % and S-MFBIA was 36.6 ± 11.3 %. Delta (S-MFBIA vs DXA) was −0.9 ± 2.6 (5 % limits of agreement 6.0 to +4.2), and CCC was 0.97 (95 % CI, 0.96–0.98).

Hurt, Ryan T., et al., "The Comparison of Segmental Multifrequency Bioelectrical Impedance Analysis and Dual-Energy X-ray Absorptiometry for Estimating Fat Free Mass and Percentage Body Fat in an Ambulatory Population.,"
Journal of Parenteral and Enteral Nutrition (2020).

Study 2 ## HIGH CORRELATION WITH D2O DILUTION METHOD FOR TOTAL BODY WATER

The study concluded that the BIA device InBodyS10 showed good test-retest precision (%CV = 5.2 raw; 1.1 after outlier removal) and high accuracy to D_2O for Total Body Water[$TBWD_2O$ = 0.956 TBWBIA, R^2= 0.92, root mean squared error (RMSE) = 2.2 kg]. %Fat estimates from DXA, ADP, D_2O, and BIA all showed high correlation with the Lohman model.

Ng, Bennett K., etal., "Validation of rapid 4-component body composition assessment with the use of dual-energy X-ray absorptiometry and bioelectrical impedance analysis.,"
The American journal of clinical nutrition 108.4 (2018):708-715.

Study 3 ## HIGH ACCURACY WITH COMPUTED TOMOGRAPHY FOR MUSCLE MASS

It was suggested that estimating muscle mass using DXA and BIA (InBody720) is a preferred method for diagnosis of sarcopenia in kidney transplant recipients.
Both DXA and InBody showed high correlation with CT.

Yanishi, M.,etal., "Dual energy X-ray absorptiometry and bioimpedance analysis are clinically useful for measuring muscle mass in kidney transplant recipients with sarcopenia.,"
Transplantation proceedings.Vol.50.No.1.Elsevier, 2018.

Study 4 ## High Correlation of Fat Free Mass Between DEXA and BWA2.0S

Total of 150 results were analyzed, excluding duplicate data from the same subject.
Fat Free Mass measured by BWA2.0S had a very high correlation with DEXA of R^2=0.977 or higher. (P value < 0.05)

∗ *Internal Validation by InBody*



∗ *Total: 150 Male: 74, Female: 76*

FFM(kg)	Total	Male	Female
	Mean ± SD (range)	Mean ± SD (range)	Mean ± SD (range)
DEXA	49.09 ± 12.95 (27.2-80.8)	59.49 ± 9.19 (37.6-80.8)	38.97 ± 6.42 (27.2-57.6)
BWA2.0S	50.88 ± 13.61 (25.4-86.0)	61.82 ± 10.00 (38.6-86.0)	40.23 ± 6.17 (25.4-58.1)

Clinical Research & Application

Nutrition

Monitor body composition change for nutritional evaluation.

Kim, H.S., Lee, E.S., Lee, Y.J., Jae Ho Lee, C., T.L., & Cho, Y.J (2015) Clinical Application of Bioelectrical Impedance Analysis and its Phase Angle For Nutritional Assessment of Critically Ill Patients. Journal of the Korean Society for Parenteral and Enteral Nutrition, 7(2), 54-61

Nephrology

Gain valuable insights into the hydration and nutrition status of dialysis patients.

Ando, M., Suminaka, T., Shimada, N., Asano, K., Ono, J., I., Jikuya, K., & Mochizuki, S., (2018). Body water balance in hemodialysis patients reflects nutritional, circulatory, and body fluid status. Journal of Biorheology, 32(2), 46-55.

Geriatric

Monitor muscle mass to screen sarcopenia with SMI, which are related to risks of fall and frailty.

Yoshimura, Y., Wakabayashi, H., Bise, T., & Tanoue, M., (2018). Prevalence of sarcopenia and its association with activities of daily living and dysphagia in convalescent rehabilitation ward inpatients. Clinical Nutrition, 37(6), 2022-2028.



Rehabilitation

Monitor injury and post-surgical recovery.

Yoshimura, Y., Bise, T., Nagano, F., Shimazu, S., Shiraishi, A., Yamaga, M., & Koga, H., (2018). Systemic inflammation in the recovery stage of stroke: its association with sarcopenia and poor functional rehabilitation outcomes. Progress in Rehabilitation Medicine, 3, 20180011.

Cardiology

Pre-screen the risk factors of cardiovascular disease.

Thomas, E., Gupta, P., P., Fonarow, G., C., & Horwich, T., B., (2019). Bioelectrical impedance analysis of body composition and survival in patients with heart failure. Clinical cardiology, 42(1), 129-135.

Professional Sports

Manage body composition to enhance performance and minimize injury risk.

Almăjan-Guță, B., Rusu, A., M., Nagel, A., & Avram, C., (2015). Injury frequency and body composition of elite Romanian rugby players. Timisoara Physical Education and Rehabilitation Journal, 8(15), 17-21.

BWA2.0S Highlights

InBody's Accurate 3MHz Measurement Technology

The 3 MHz frequency penetrates cell membranes more effectively, providing a clearer reflection of Total Body Water. This technology allows for a more accurate distinction between Intracellular and Extracellular Water, particularly benefiting patients with unstable body water balance. It also enables precise measurements across a wide range of individuals, including athletes and those with extreme conditions, ensuring reliable results.

Clamp Electrode for High Reproducibility

The Clamp Electrode is a combination of two ergonomics electrodes, which acts as an indicator attached to the wrist and ankle for high reproducibility. The flexible design of the clamp ensures the electrodes to closely adhere to wrist and ankle even during the articular movements.

Covering Wide Range of Subjects / Patients and Conditions

More precise results can be obtained and utilized by entering the patient status information such as amputation, paralysis, lymphedema, and vascular access region.

Extensive Research Parameters for Professionals

Select from a range of distinct optional parameters for clinical and research purposes.

- Water Control Calculator: to set target ECW Ratio
- Age-specific graph: to evaluate and compare the body composition result by age
- BIVA (Bioelectrical Impedance Vector Analysis): Used to evaluate the hydration and nutritional status in comparison to their demographic group.



Comprehensive Parameters for Professionals

Body Water Balance

ECW Ratio Analysis

Whole Body ECW (Extracellular Water) Ratio and Segmental ECW Ratio offer a precise assessment of health status regarding the body water balance. This ratio is calculated by dividing Total Body Water (TBW) into Extracellular Water (ECW). And only in a healthy population, a balanced ratio between ECW and Intracellular Water (ICW) is maintained.
When health issues arise, this ratio can become imbalanced, indicating potential health concerns.

Cell Health Marker

Phase Angle

The human body comprises 36 trillion cells, and understanding cell health is crucial for overall well-being. The Phase Angle is a key parameter in assessing cell health and overall physiological status. It reflects the relationship between resistance in Total Body Water and reactance in cell membrane. A higher Phase Angle indicates better cell membrane integrity, and well-balanced fluid, suggesting healthier cells. Last but not least, with the addition of Whole Body Phase Angle History, users can intuitively track and monitor their health trends over time.

Sarcopenia Assessment

SMI(Skeletal Muscle Mass Index)

Sarcopenia, assigned the diagnosis code M62.84 by WHO, is acknowledged as a disease rather than just a natural phenomenon. It can be easily assessed and evaluated using the Skeletal Muscle Mass Index (SMI)* and Hand Grip Strength**, allowing for comprehensive evaluation and personalized consultations.

* Skeletal Muscle Mass Index (SMI) is calculated by taking the sum of the Appendicular Muscle Mass (in kilograms) and dividing it by the square of the person's height (in meters).

** Hand Grip Strength is available with connections to the InBody Handgrip Dynamometer (IB-HGS, optional).

InBody Big Data Solution

Evaluation Result Sheet

The InBody Big Data is consisted of over 130 million* body composition measurements collected globally. This extensive InBody's Big Data provides deeper insights for health solutions by comparing individual measurements with many other measurement variables.

* Over 130 million Data as of August 2024.

Extensive Research Parameters for Professionals

Select from a range of optional parameters for clinical and research purposes

Water Control Calculator

Set the Target ECW Ratio depending on the hydration status of dialysis and heart failure patients.



Up to 20 Optional Parameters

Provides up to 20 optional parameters for a customized experience. Select parameters to suit department needs.



Whole Body Phase Angle_50kHz

Skeletal Muscle mass Index

Bioelectrical Impedance Vector Analysis



Skeletal Muscle Mass Index and ECW Ratio



Outer Circumference



Device Overview

Various Features and Optional Components of BWA2.0S

LCD
Sharp 10.1" touch screen

Battery
BWA2.0S battery for mobile use

Test Posture
Measurable in a lying, seated or standing position

InBody USB
Easy data back up with InBody USB

Thermal Printer (Optional)
Easy printout BWA2.0S results

Clamp Electrode
Patented dual-structured Clamp Electrodes ensure high reproducibility

BWA2.0S Cart
Customized BWA2.0S Cart to easily arrange the Clamp Electrodes

InBody Tissue **Contains 300 packets/box*
Electrolyte tissues help hydrate the skin to ensure stable measurements



Adhesive Electrodes and Tape (Optional)
BWA2.0S Electrode Tapes for patients with difficulty in using Clamp Electrode



Body Water Result Sheet

Present the patients body water measurement in a clear, easy-to-read format, making it simple to understand.

BWA Body Water

[BWA2.0S]

ID	Height	Age	Gender	Test Date / Time
John Doe	173cm	41	Male	10.25.2025 15 : 44

6 **Customized Logo**

www.customized.com

1 Body Water Composition

	Under	Normal	Over
TBW Total Body Water (L)	31.4		
ICW Intracellular Water (L)	18.4		
ECW Extracellular Water (L)	13.0		

2 ECW Ratio Analysis

	Under	Normal	Over
ECW Ratio			0.415

3 Segmental Body Water Analysis

	Under	Normal	Over
Right Arm (L)	2.00		
Left Arm (L)	2.07		
Trunk (L)	17.2		
Right Leg (L)	5.04		
Left Leg (L)	5.01		

4 Segmental ECW Ratio Analysis



5 Body Water Composition History

	07.21.24 15:11	08.27.24 14:58	09.20.24 15:02	11.23.24 15:23	12.21.24 15:00	02.19.25 14:52	10.20.25 15:12	10.25.25 15:44
Weight (kg)	65.3	63.9	62.4	61.8	62.3	60.9	67.9	64.0
TBW Total Body Water (L)	28.3	28.0	28.0	27.9	27.9	27.6	34.5	31.4
ICW Intracellular Water (L)	17.0	16.9	16.9	16.8	16.8	16.7	20.0	18.4
ECW Extracellular Water (L)	11.3	11.1	11.1	11.0	11.1	10.9	14.5	13.0
ECW Ratio	0.399	0.398	0.396	0.396	0.397	0.396	0.421	0.415

☑ Recent ☐ Total

7 Body Composition Analysis

Protein	7.9 kg	(9.9 ~ 12.1)
Minerals	2.91 kg	(3.43 ~ 4.19)
Body Fat Mass	21.8 kg	(7.9 ~ 15.8)
Fat Free Mass	42.2 kg	(50.4 ~ 61.6)
Bone Mineral Content	2.37 kg	(2.82 ~ 3.44)

8 Muscle-Fat Analysis

Weight	64.0 kg	(55.9 ~ 75.7)
Skeletal Muscle Mass	22.0 kg	(28.2 ~ 34.4)
Soft Lean Mass	39.8 kg	(47.5 ~ 58.1)
Body Fat Mass	21.8 kg	(7.9 ~ 15.8)

9 Water Control

ECW Ratio	0.385	-1.5L	/ 62.5kg
ECW Ratio	0.395	-1.0L	/ 63.0kg
ECW Ratio	0.405	-0.5L	/ 63.5kg

*The water control item shows the water level to be controlled based on the extracellular water ratio. This item shows the water level which varies as the extracellular water ratio is set differently according to the presence or absence of complications as described in a paper published in the 2008 Journal of the Japan Society for Dialysis Therapy(JSDT).

10 Research Parameters

TBW/FFM	74.4 %
FFMI	14.1 kg/m2
FMI	7.3 kg/m2

11 Whole Body Phase Angle

Proximal

ϕ(°) 50 kHz | 3.8°

12 Segmental Phase Angle

Proximal	RA	LA	TR	RL	LL
ϕ(°) 5 kHz	2.2	2.0	2.2	1.6	1.5
50 kHz	4.9	4.8	5.0	2.8	2.6
250 kHz	4.8	4.7	5.9	3.1	2.8

13 Impedance (Z0, Z∞)

Z(Ω)	RA	LA	TR	RL	LL
0 Hz	327.4	279.6	38.9	242.7	228.5
∞ Hz	216.6	205.7	23.3	171.2	159.0

14 Impedance

Proximal



[Clamp Type, Lying Posture]
[000/000/000]

Copyright ©1996~by InBody Co.,Ltd. All rights reserved.

Result Sheet Interpretation

1 Body Water Composition

50-70 % of our body is composed of water. Body Water is distributed between all the cells and fluids in our body. Most of it is present in the cells while the rest is in the form of blood and interstitial fluid. The water inside the cell membrane is called Intracellular Water, and the water outside the cell membrane is called Extracellular Water.

2 ECW Ratio Analysis

The Extracellular Water Ratio shows the balance status of body water. The ratio between Intra and Extracellular Water remains constant at about 3:2 ratio in healthy individuals, and when this balance is broken down edema may occur.

3 Segmental Body Water Analysis

Segmental Body Water Analysis helps to understand the water balance by analyzing the Total Body Water in each segment of the body. Changes in Body Water corresponds to the changes in muscle mass. However, in the case of a subject who has health issue, the amount of Body Water may increase even if there is no increase in muscle mass. Therefore, it is necessary to check whether Extracellular Water Ratio is normal in segments.

4 Segmental ECW Ratio Analysis

Segmental ECW Ratio is displayed in a graph so you can easily determine if the ICW and ECW are balanced.
By analyzing the ECW Ratio, you can assess if there is a problem with Body Water circulation. This can help monitor the recovery of post-surgery or hemodialysis patients.

5 Body Water Composition History

Body Water History provides the changes in Weight, Skeletal Muscle Mass, Intracellular Water, Extracellular Water, Extracellular Water Ratio. Take the BWA2.0S Test periodically to monitor your progress.

6 Logo Customization

The Customized Logo can be applied on the Result Sheet. URL can also be applied at the bottom of the Result Sheet as well.

7 Body Composition Analysis

Body composition is a method of describing what the body is made of. BWA2.0S offers quantitative values and normal ranges for four core body components: Body Water, Protein, Minerals, and Fat.

8 Muscle-Fat Analysis

The balance between Skeletal Muscle Mass and Body Fat Mass is a key health indicator. Muscle-Fat Analysis shows this balance by comparing the length of the bars for Weight, Skeletal Muscle Mass, and Body Fat Mass

9 Water Control

The Water Control item provides a target for fluid adjustment based on the extracellular water (ECW) ratio. This reflects the patient's fluid status and may vary with conditions such as diabetes, hypoalbuminemia, or other factors affecting fluid retention.
By adjusting the ECW ratio to the patient's condition, it helps estimate dry weight and supports clinical decision-making in fluid management.

10 Research Parameters

Various nutritional outputs are provided such as Fat Free Mass, Basal Metabolic Rate, Visceral Fat Level, Recommended Calorie Intake per day and more.

11 Whole Body Phase Angle

Phase Angle is related to the health status of the cell membrane. Strengthening of the cellular membrane and structural function will increase the Phase Angle, while damage or a decrease in function will result in a decrease in the Phase Angle.

12 Segmental Phase Angle

Segmental Phase Angle indicates the Phase Angle of each part of the body, representing the level of structural integrity and function of the cell membrane.

13 Z0, Z∞ Frequency

At zero frequency, current does not pass through the cell membrane, so the impedance at zero frequency can be considered to reflect extracellular water, and at infinite frequency, the current can be seen to reflect both intracellular and extracellular water.

14 Impedance

Impedance is the resistance that occurs when weak alternating current is applied to the human body. BWA2.0S visualizes the impedance with the graph, so you can easily detect if there is reversed impedance error by checking crossed lines in the impedance graph. Below the impedance graph, you can also check the error codes.

Body Composition Result Sheet

Provides reference parameters to thoroughly evaluate patients' conditions across various medical practices.

BWA [BWA2.0S]

InBody
inbody.com

ID	Height	Age	Gender	Test Date / Time
John Doe	173cm	41	Male	10.25.2025 15 : 44

Body Composition Analysis

	Values	Total Body Water	Soft Lean Mass	Fat Free Mass	Weight
Total Body Water (L)	31.4 (37.0 ~ 45.2)	31.4	39.8 (47.5 ~ 58.1)	42.2 (50.4 ~ 61.6)	64.0 (55.9 ~ 75.7)
Protein (kg)	7.9 (9.9 ~ 12.1)				
Minerals (kg)	2.91 (3.43 ~ 4.19)	non-osseous			
Body Fat Mass (kg)	21.8 (7.9 ~ 15.8)				

Muscle-Fat Analysis

	Under	Normal	Over
Weight (kg)		64.0	
SMM Skeletal Muscle Mass (kg)		22.0	
Body Fat Mass (kg)			21.8

Obesity Analysis

	Under	Normal	Over
BMI Body Mass Index (kg/m²)		21.4	
PBF Percent Body Fat (%)			34.0

Segmental Lean Analysis

Based on ideal weight ■ Based on current weight ■

	Under	Normal	Over	ECW Ratio
Right Arm (kg) (%)		2.57 84.3		0.386
Left Arm (kg) (%)		2.65 87.1		0.387
Trunk (kg) (%)		21.9 89.9		0.415
Right Leg (kg) (%)		6.38 75.3		0.426
Left Leg (kg) (%)		6.33 74.8		0.428

ECW Ratio Analysis

	Under	Normal	Over
ECW Ratio		0.415	

Body Composition History

	07.20.24 15:12	08.27.24 15:44	09.20.24 15:02	11.23.24 15:23	12.21.24 15:00	02.19.25 14:52	10.20.25 15:12	10.25.25 15:44
Weight (kg)	64.5	64.3	64.1	64.4	64.7	66.8	67.9	64.0
SMM Skeletal Muscle Mass (kg)	21.5	21.6	21.5	21.6	21.7	23.0	24.1	22.0
PBF Percent Body Fat (%)	35.0	34.8	34.8	34.9	35.0	33.0	31.6	34.0
ECW Ratio	0.411	0.410	0.410	0.409	0.410	0.416	0.421	0.415

☑ Recent ☐ Total

InBody Score

69 / 100 Points

* Total score that reflects the evaluation of body composition. A muscular person may score over 100 points.

Visceral Fat Area

VFA(cm²)



Weight Control

Target Weight	65.8 kg
Weight Control	+ 1.8 kg
Fat Control	– 11.9 kg
Muscle Control	+ 13.7 kg

Research Parameters

Intracellular Water	18.4 L	(23.0~28.0)
Extracellular Water	13.0 L	(14.0~17.2)
Basal Metabolic Rate	1282 kcal	(1428~1663)
Waist-Hip Ratio	1.15	(0.80~0.90)
Body Cell Mass	26.3 kg	(32.8~40.2)

Whole Body Phase Angle

Proximal
ϕ(°) 50 kHz | 3.8°

Segmental Phase Angle

Proximal	RA	LA	TR	RL	LL
ϕ(°) 5 kHz	2.2	2.0	2.2	1.6	1.5
50 kHz	4.9	4.8	5.0	2.8	2.6
250 kHz	4.8	4.7	5.9	3.1	2.8

Sarcopenia Parameters

SMI	6.0 kg/m²	(< 7.0)
HGS	23.7 kg	(< 28.0)

Impedance

Proximal



[Clamp Type, Lying Posture]
[000/000/000]

Copyright ©1996~by InBody Co.,Ltd. All rights reserved.

Evaluation Result Sheet

Provides age-specific graphs based on InBody Big Data, enabling in-depth consultation and research application

BWA Evaluation

[BWA2.0S]

InBody
inbody.com

ID	Height	Age	Gender	Test Date / Time
John Doe	173cm	41	Male	03.20.2025 15 : 44

Research Parameters

Whole Body Phase Angle_50kHz



Ref. values from InBody

PhA (°)	Young adults (T-score)	Age-matched (Z-score)
3.8	- 6.1	- 6.1

Body Water Evaluation

Whole Body ECW Ratio



Ref. values from InBody

ECW/TBW	Young adults (T-score)	Age-matched (Z-score)
0.415	8.7	7.8

Muscle · Nutrition Evaluation

Skeletal Muscle mass Index



Cruz-Jentoft et al. Age ageing 2019 Jan 1:48(1):16-31
Ref. values from InBody

SMI (kg/m²)	Young adults (T-score)	Age-matched (Z-score)
6.0	- 3.4	- 3.7

Research Parameters

Body Cell Mass



Ref. values from InBody

BCM (kg)	Young adults (T-score)	Age-matched (Z-score)
26.3	- 2.9	- 3.0

Copyright ©1996~by InBody Co.,Ltd. All rights reserved.

Research Result Sheet

Provides comprehensive segmental body composition analysis through diverse research-based parameters

BWA Research

[BWA2.0S]

InBody
inbody.com

ID	Height	Age	Gender	Test Date / Time
John Doe	173cm	41	Male	03.20.2025 15 : 44

Body Composition Summary

	FFM·Lean Mass	FM	ICW	ECW	TBW	ECW/TBW
Right Arm	2.57 kg	1.6 kg	1.23 L	0.77 L	2.00 L	0.386
Left Arm	2.65 kg	1.5 kg	1.27 L	0.80 L	2.07 L	0.387
Trunk	21.9 kg	12.5kg	10.1 L	7.1 L	17.2 L	0.415
Right Leg	6.38 kg	2.5 kg	2.89 L	2.15 L	5.04 L	0.426
Left Leg	6.33 kg	2.5 kg	2.87 L	2.14 L	5.01 L	0.428
Whole Body	42.2 kg	21.8 kg	18.4 L	13.0 L	31.4 L	0.415
Weight	64.0 kg		* The difference between the whole body values and sum of segmental values are from the craniocervical region.			

Body Composition Analysis

Lean Mass ICW ECW
Fat Mass ECW/TBW



Research Parameters

Body Mass Index	21.4 kg/m²	(18.5~25.0)
Percent Body Fat	34.0 %	(10.0~20.0)
Skeletal Muscle Mass	22.0 kg	(28.2~34.4)
Soft Lean Mass	39.8 kg	(47.5~58.1)
Protein	7.9 kg	(9.9~12.1)
Minerals	2.91 kg	(3.43~4.19)
Bone Mineral Content	2.37 kg	(2.82~3.44)
Basal Metabolic Rate	1282 kcal	(1428~1663)
Waist Hip Ratio	1.15	(0.80~0.90)
Waist Circumference	100.1 cm	
Visceral Fat Area	144.2 cm²	
Obesity Degree	97 %	(90~110)
Body Cell Mass	26.3 kg	(32.8~40.2)
Arm Circumference	30.3 cm	
Arm Muscle Circumference	27.2 cm	
TBW/FFM	74.4 %	
Fat Free Mass Index	14.1 kg/m²	
Fat Mass Index	7.3 kg/m²	
Skeletal Muscle mass Index	6.0 kg/m²	

Whole Body Phase Angle

Proximal
ϕ(°) 50 kHz | 3.8°

Segmental Phase Angle

Proximal	RA	LA	TR	RL	LL
ϕ(°) 5 kHz	2.2	2.0	2.2	1.6	1.5
50 kHz	4.9	4.8	5.0	2.8	2.6
250 kHz	4.8	4.7	5.9	3.1	2.8

Impedance

Proximal



[Clamp Type, Lying Posture]
[000/000/000]

Copyright ©1996~by InBody Co.,Ltd. All rights reserved.

Comparison Result Sheet

Provides key segmental parameters and Cole-Cole plot graphs for clear before-and-after comparison

BWA Comparison

[BWA2.0S]

InBody
inbody.com

ID	Height	Age	Gender	Test Date / Time
John Doe	173cm	41	Male	10.25.2025 15:44

Standard median curve — Today's Results — Recent Results (10.25.2025 15:44)

Whole Body		Today	Recent	Difference
Weight	(kg)	64.0	67.9	-3.9
SMM	(kg)	22.0	24.1	-2.1
Body Fat Mass	(kg)	21.8	21.5	+0.3
ECW Ratio		0.415	0.421	-0.006
Phase Angle	(°)	3.8	3.9	-0.1



Right Arm		Today	Recent	Difference
Lean Mass	(kg)	2.57	2.82	-0.25
ECW Ratio		0.386	0.389	-0.003
Phase Angle	(°)	4.9	4.8	+0.1
Left Arm		**Today**	**Recent**	**Difference**
Lean Mass	(kg)	2.65	2.95	-0.30
ECW Ratio		0.387	0.388	-0.001
Phase Angle	(°)	4.8	4.7	+0.1



Right Leg		Today	Recent	Difference
Lean Mass	(kg)	6.38	6.86	-0.48
ECW Ratio		0.426	0.432	-0.006
Phase Angle	(°)	2.8	2.9	-0.1
Left Leg		**Today**	**Recent**	**Difference**
Lean Mass	(kg)	6.33	6.73	-0.40
ECW Ratio		0.428	0.433	-0.005
Phase Angle	(°)	2.6	2.6	0.0



Trunk		Today	Recent	Difference
Lean Mass	(kg)	21.9	23.2	-1.3
ECW Ratio		0.415	0.422	-0.007
Phase Angle	(°)	5.0	6.0	-1.0



Copyright ©1996~by InBody Co.,Ltd. All rights reserved.

Body Composition Result Sheet for Children

With the InBody Result Sheet for Children, you can assess and track a child's growth progress.

BWA [BWA2.0S]

InBody
inbody.com

ID	Height	Age	Gender	Test Date / Time
John Doe Jr.	139.4cm	10	Male	03.31.2025 16 : 40

Body Composition Analysis

Total amount of water in my body	Total Body Water	(L)	21.5 (18.0 ~ 22.0)
What I need to build muscles	Protein	(kg)	5.5 (4.9 ~ 5.9)
What I need for strong bones	Minerals	(kg)	1.55 (1.66 ~ 2.04)
Where my excess energy is stored	Body Fat Mass	(kg)	6.4 (3.8 ~ 7.7)
Sum of the above	Weight	(kg)	35.0 (27.3 ~ 36.9)

Muscle-Fat Analysis

	Under	Normal	Over
Weight (kg)	55 70 85 100 115 130 145 160 175 190 205 %	35.0	
SMM Skeletal Muscle Mass (kg)	70 80 90 100 110 120 130 140 150 160 170 %	14.9	
Body Fat mass (kg)	40 60 80 100 160 220 280 340 400 460 520 %	6.4	

Obesity Analysis

	Under	Normal	Over
BMI Body Mass Index (kg/m^2)	7.9 10.9 13.9 16.4 18.6 20.2 22.2 24.2 26.2 28.2 30.2	18.0	
PBF Percent Body Fat (%)	0.0 5.0 10.0 15.0 20.0 25.0 30.0 35.0 40.0 45.0 50.0	18.2	

Growth Graph

Height : 50 ~ 75%



BMI : 50 ~ 75%



* 7 growth charts of weights for ages were truncated at 10 years of age.

Body Composition History

	07.15.24 14:22	11.19.24 09:30	01.29.25 15:18	03.15.25 11:00	06.21.25 15:00	09.19.25 14:52	12.20.25 15:12	03.31.25 16:40
Height (cm)	134.5	135.2	136.4	137.2	137.9	138.5	139.0	139.4
Weight (kg)	30.8	31.3	32.0	32.8	33.5	34.0	34.4	35.0
SMM Skeletal Muscle Mass (kg)	12.5	12.7	12.8	13.0	13.1	13.1	13.2	13.3
PBF Percent Body Fat (%)	20.4	20.7	21.6	22.3	23.1	24.3	25.1	25.6

☑ Recent ☐ Total

Growth Score

97 / 100 Points

* If tall and within great body comparison standards, the growth score may surpass 100 points.

Nutrition Evaluation

Protein	☑ Normal	☐ Deficient	
Minerals	☐ Normal	☑ Deficient	
Body Fat	☑ Normal	☐ Deficient	☐ Excessive

Obesity Evaluation

BMI	☑ Normal	☐ Under	☐ Slightly Over	☐ Over
PBF	☑ Normal	☐ Slightly Over	☐ Over	

Body Balance Evaluation

Upper	☑ Balanced	☐ Slightly Unbalanced	☐ Extremely Unbalanced
Lower	☑ Balanced	☐ Slightly Unbalanced	☐ Extremely Unbalanced
Upper-Lower	☑ Balanced	☐ Slightly Unbalanced	☐ Extremely Unbalanced

Segmental Lean Analysis

Right Arm	1.53 kg
Left Arm	1.50 kg
Trunk	15.0 kg
Right Leg	5.18 kg
Left Leg	5.16 kg

Research Parameters

Child Obesity Degree 109 % (90 ~ 110)

Whole Body Phase Angle

Proximal

ϕ(°) 50 kHz | 4.4°

Segmental Phase Angle

Proximal	RA	LA	TR	RL	LL
ϕ(°) 5 kHz	1.8	1.7	4.7	1.7	1.6
50 kHz	4.5	4.1	5.7	4.0	3.8
250 kHz	4.3	3.8	5.6	2.9	2.9

Impedance

Proximal



[Clamp Type, Lying Posture]
[000/000/000]

Copyright ©1996-by InBody Co.,Ltd. All rights reserved.

Thermal Result Sheet

Thermal printout for quick access to essential results during mobile check-ups

BWA 10/25/2015 15:44

ID : John Doe
Height : 173cm **Age** : 41
Gender: Male **Weight** : 64.0kg

[Clamp Type, Lying Posture]

Muscle-Fat Analysis

Weight	64.0 kg	
Normal Range	(55.9~75.7)	
Skeletal Muscle Mass	22.0 kg	
Normal Range	(28.2~34.4)	
Soft Lean Mass	39.8 kg	
Normal Range	(47.5~58.1)	
Body Fat Mass	21.8 kg	
Normal Range	(7.9~15.8)	

Obesity Analysis

BMI	21.4 kg/m²	
Normal Range	(18.5~25.0)	
Percent Body Fat	34.0 %	
Normal Range	(10.0~20.0)	

Segmental ECW Ratio Analysis

Right Arm	0.386	
Normal Range	(0.360~0.390)	
Left Arm	0.387	
Normal Range	(0.360~0.390)	
Trunk	0.415	
Normal Range	(0.360~0.390)	
Right Leg	0.426	
Normal Range	(0.360~0.390)	
Left Leg	0.428	
Normal Range	(0.360~0.390)	

Body Water Analysis

Intracellular Water	18.4 L	
Normal Range	(23.0~28.0)	
Extracellular Water	13.0 L	
Normal Range	(14.0~17.2)	
Total Body Water	31.4 L	
Normal Range	(37.0~45.2)	

Proximal

Impedance



000, 000, 000

InBody
inbody.com

BWA 2021/03/31 15:44

ID : John Doe
Height : 173cm **Age** : 41
Gender: Male **Weight** : 64.0kg

Water Control

ECW Ratio	0.415
Target ECW Ratio	0.385
Over Hydration	-1.5 L
	(-1.65~-1.35)
Target Weight	65.5 kg

Data Management Program

LookinBody WEB (Cloud)

A cloud-based client and data management solution designed to optimize performance and deliver a better user experience. Try a free 1-month demonstration by contacting regional managers.



LookinBody120 (PC Software)

LookinBody120 allows you to view and manage all BWA2.0S data generated from your BWA2.0S device.



InBody Integration Solution



Specifications

BWA 2.0S Body Water Analyzer



Bioelectrical Impedance Analysis (BIA) Measurement Outputs	**Impedance (Z)**	30 Impedance Measurements by Using 6 Different Frequencies (5 kHz, 50 kHz, 250 kHz, 500 kHz, 1 MHz, 3 MHz) at Each of 5 Segments (Right Arm, Left Arm, Trunk, Right Leg, and Left Leg)
	Phase Angle (Ø)	15 Phase Angle Measurements by Using 3 Different Frequencies (5 kHz, 50 kHz, 250 kHz) at Each of the 5 Segments (Right Arm, Left Arm, Trunk, Right Leg, and Left Leg)
	Z0, Z∞	At zero frequency, current does not pass through the cell membrane, so the impedance at this frequency reflects Extracellular Water. At infinite frequency, the current reflects both Intracellular and Extracellular Water.
Measurement Method	· Direct Segmental Multi-Frequency Bioelectrical Impedance Analysis (DSM-BIA) · Simultaneous Multi-Frequency Bioelectrical Impedance Analysis (SMF-BIA)	
Electrode Method	16-Point Clamp Electrodes	
Body Composition Calculation Method	No use of Empirical Estimation	
Types of Result Sheet	Body Water Result Sheet, Body Composition Result Sheet, Evaluation Result Sheet, Research Result Sheet, Comparison Result Sheet, Result Sheet for Children, and Thermal Result Sheet	
Digital Results	LCD Screen, LookinBody Web, LookinBody120	
Data Storage	Test results can be saved using the member ID. The InBody can save up to 100,000 results.	
Test Mode	Medical Mode, Research Mode	
Test Duration	About 30 Seconds for Medical Mode, about 60 Seconds for Research Mode * Test duration may vary depending on the measurement posture or external environment.	
Weight Range	2 - 250 kg (4.4 - 551.2 lb)	
Height Range	95 - 220 cm (3 ft 1.4 in - 7 ft 2.6 in)	
Age Range	3+ years	
Administrator Menu	· Setup: Settings Configuration and Data Management · FAQ: Additional Guidance for Using the InBody	
USB Thumb Drive	Copy, Back Up, or Restore the InBody Test Data (which can be viewed in Excel or with LookinBody data management software).	
Backup Data	Backup data from the device using an InBody USB or a USB thumb drive, and restore results as needed.	
Dimensions	322 (W) × 282 (L) × 81.5 (H): mm 12.7 (W) × 11.1 (L) × 3.2 (H): in	
Device Weight	2.8 kg (6.17 lb, BWA2.0S only)	
Applied Rating Current	300 μA (± 30 μA)	
Operation Environment	10 - 40 °C (50 - 104 °F), 30 - 75 % RH, 70 - 106 kPa	
Storage Environment	−10 - 70 °C (14 - 158 °F),10 - 80 % RH, 50 - 106 kPa (No Condensaiotn)	
Display Type	1280 x 800 10.1inch Color TFT LCD	
Internal Interface	Touchscreen, Keypad	
External Interface	RS-232C 4 EA, USB Host 2 EA, USB Slave 1 EA, LAN(10/100 T) 1 EA, Bluetooth 1 EA, Wi-Fi (2.4 G / 5 G) 1 EA	
Adapter	**DELTA** — Power Input: AC 100 - 240 V, 50 - 60 Hz, 1.5 A - 0.75 A; Power Output: DC 12 V ⎓ , 5.0 A	
	Mean Well (GSM 40A12) — Power Input: AC 100 - 240 V, 50 / 60 Hz, 1.0 A - 0.5 A; Power Output: DC 12 V ⎓ , 3.34 A	
Wireless Connection	Bluetooth, Wi-Fi	
Compatible Items	Stadiometer, Blood pressure monitor, Thermal Printer (TP100), Serial Distributor (SD400), InGrip, BWA2.0S Portable Case, BWA2.0S Adhesive Electrodes and Tape, BWA2.0S Battery Pack	
Compatible Printer	Laser/Inkjet PCL 3 or above and SPL	
Notification Sounds and Voice Guidance	Notification sounds (test in progress, saving settings, personal information, etc.) and voice guidance during the test	
Logo Display	Name, address, and contact information can be shown on the InBody Result Sheet.	
QR Code	By scanning QR Code, you can send and verify the InBody results	
Language Support	InBody supports over 30 languages.	

Outputs (InBody Result Sheet)	Results and Interpretations · Body Composition Analysis (Total Body Water, Protein, Mineral, Body Fat Mass, Fat Free Mass, Soft Lean Mass, Weight) · Muscle-Fat Analysis (Weight, Skeletal Muscle Mass, Soft Lean Mass, Body Fat Mass) · Obesity Analysis (Body Mass Index, Percent Body Fat) · Segmental Lean Analysis · ECW Ratio Analysis · Body Composition History (Weight, Skeletal Muscle Mass, Percent Body Fat, ECW Ratio) · InBody Score · Whole Body Phase Angle (History) · SMI (History) · Visceral Fat Area (Graph) · Body Type · Weight Control (Target Weight, Weight Control, Fat Control, Muscle Control) · Nutrition Evaluation (Protein, Minerals, Body Fat) · Obesity Evaluation (BMI, Percent Body Fat) · Body Balance Evaluation (Upper, Lower, Upper-Lower) · Segmental Fat Analysis (Graph) · Segmental Fat Analysis · Body Water Composition (Total Body Water, Intracellular Water, Extracellular Water) · Segmental Body Water Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Segmental ICW Analysis · Segmental ECW Analysis	· Body Composition Analysis (Total Body Water, Protein, Mineral, Body Fat Mass, Fat Free Mass, Soft Lean Mass, Weight) · Muscle-Fat Analysis (Weight, Skeletal Muscle Mass, Body Fat Mass) · Obesity Analysis (Body Mass Index, Percent Body Fat) · Segmental Circumference · Waist-Hip Ratio (Graph) · Visceral Fat Level (Graph) · Water Contol · Research Parameters (Intracellular Water, Extracellular Water, Skeletal Muscle Mass, Fat Free Mass, Basal Metabolic Rate, Waist-Hip Ratio, Waist Circumference, Visceral Fat Level, Visceral Fat Area, Obesity Degree, Bone Mineral Content, Body Cell Mass, Arm Circumference, Arm Muscle Circumference, TBW/FFM, FFMI, FMI, SMI, SMM/WT, ECM/BCM, TBW/WT, Adjusted FFM, Adjusted SMI,Recommended Calorie Intake per day) · Calorie Expenditure of Exercise · Sarcopenia Parameters (SMI, HGS) · Blood Pressure (Systolic, Diastolic, Pulse, Mean Artery Pressure, Pulse, Rate Pressure Device) · QR code · Results Interpretation QR code · Whole Body ECW Ratio · Segmental ECW Ratio · Whole Body Phase Angle (50 kHz) · Segmental Body Phase Angle · Bioelectrical Impedance Vector Analysis · Impedance(Z0, Z∞) · Impedance (Each segment and each frequency)
Outputs (InBody Result Sheet for Children)	Results and Interpretations · Body Composition Analysis (Total Body Water, Protein, Mineral, Body Fat Mass, Weight) · Muscle-Fat Analysis (Weight, Skeletal Muscle Mass, Body Fat Mass) · Obesity Analysis (Body Mass Index, Percent Body Fat) · Growth Graph (Height, Weight, BMI) · Body Composition History (Height, Weight, Skeletal Muscle Mass, Percent Body Fat) · Whole Body Phase Angle (History) · SMI (History) · Growth Score · Weight Control (Target Weight, Weight Control, Fat Control, Muscle Control) · Nutrition Evaluation (Protein, Minerals, Body Fat) · Obesity Evaluation (BMI, Percent Body Fat) · Body Balance Evaluation (Upper, Lower, Upper-Lower)	· Segmental Lean Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Segmental Body Water Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Research Parameters (Intracellular Water, Extracellular Water, Skeletal Muscle Mass, Fat Free Mass, Basal Metabolic Rate, Child Obesity Degree, Bone Mineral Content, Body Cell Mass, FFMI, FMI, SMI, SMM/WT, ECM/BCM, TBW/WT) · Sarcopenia Parameters (SMI, HGS) · Blood Pressure (Systolic, Diastolic, Pulse, Mean Artery Pressure, Pulse, Rate Pressure Device) · QR code · Results Interpretation QR code · Whole Body Phase Angle (50 kHz) · Segmental Phase Angle (5 kHz, 50 kHz, 250 kHz: Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Impedance Graph (Each segment and each frequency)
Outputs (Body Water Result Sheet)	Results and Interpretations · Body Water Composition (Total Body Water, Intracellular Water, Extracellular Water) · ECW Ratio Analysis (ECW Ratio) · Segmental Body Water (Graph, Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Segmental ECW Ratio Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Body Water Composition History (Weight, Total Body Water, Intracellular Water, Extracellular Water, ECW Ratio) · InBody Score · Whole Body Phase Angle (History) · SMI (History) · Visceral Fat Area (Graph) · Body Type (Graph) · Weight Control · Nutrition Evaluation · Obesity Evaluation (BMI, Percent Body Fat) · Body Balance Evaluation · Segmental Fat Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Body Water Composition (Total Body Water, Intracellular Water, Extracellular Water) · Segmental Body Water Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Segmental ICW Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Segmental ECW Analysis (Right Arm, Left Arm, Trunk, Right Leg, Left Leg)	· Body Composition Analysis (Protein, Minerals, Body Fat Mass, Fat Free Mass, Bone Mineral Content) · Muscle-Fat Analysis (Weight, Skeletal Muscle Mass, Soft Lean Mass, Body Fat Mass) · Obesity Analysis (Body Mass Index, Percent Body Fat) · Segmental Circumference (Neck, Chest, Abdomen, Hip, Right Arm, Left Arm, Right Thigh, Left Thigh) · Waist-Hip Ratio (Graph) · Visceral Fat Level (Graph) · Water Contol · Research Parameters (Intracellular Water, Extracellular Water, Skeletal Muscle Mass, Fat Free Mass, Basal Metabolic Rate, Waist-Hip Ratio, Waist Circumference, Visceral Fat Level, Visceral Fat Area, Obesity Degree, Bone Mineral Content, Body Cell Mass, Arm Circumference, Arm Muscle Circumference, TBW/FFM, FFMI, FMI, SMI, SMM/WT, ECM/BCM, TBW/WT, Adjusted FFM, Adjusted SMI, Recommended calorie intake per day) · Calorie Expenditure of Exercise · Sarcopenia Parameters (SMI, HGS) · Blood Pressure (Systolic, Diastolic, Pulse, Mean Artery Pressure, Pulse, Rate Pressure Device) · QR code · Results Interpretation QR code · Whole Body Phase Angle (50 kHz) · Segmental Phase Angle (5 kHz, 50 kHz, 250 kHz: Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · BIVA (Bioelectrical Impedance Vector Analysis) · Impedance (Z0, Z∞) · Impedance (Each segment and each frequency)
Outputs (Evaluation Result Sheet)	· Bioelectrical Impedance Vector Analysis (BIVA) · Whole Body Phase Angle_50 kHz (PhA, Ø) (M ± SD, Percentile Graph) · Segmental Phase Angle_50 kHz (PhA, Ø) Balance · Whole Body ECW Ratio (ECW/TBW) (M ± SD, Percentile Graph) · ECW Ratio Balance (ECW/TBW) · TBW/WT (%) (M ± SD, Percentile Graph) · Percent Body Fat (PBF, %) (M ± SD, Percentile Graph) · Skeletal Muscle Mass and ECW Ratio (SMM, % & ECW/TBW) · Skeletal Muscle mass Index and ECW Ratio (SMI, kg/m² & ECW/TBW) · Skeletal Muscle mass Index (SMI, kg/m²) (M ± SD, Percentile Graph)	· Fat Free Mass Index (FFMI, kg/m²) (M ± SD, Percentile Graph) · Lean Mass (LM) Balance · Fat Mass Index (FMI, kg/m²) (M ± SD, Percentile Graph) · Skeletal Muscle Mass divided by WT (SMM/WT, %) (M ± SD, Percentile Graph) · Visceral Fat Area (VFA, cm²) (M ± SD, Percentile Graph) · Waist Hip Ratio (WHR) (M ± SD, Percentile Graph) · Weight (kg) (M ± SD, Percentile Graph) · Body Mass Index (BMI, kg/m²) (M ± SD, Percentile Graph) · Body Cell Mass (BCM, kg) (M ± SD, Percentile Graph) · ECM/BCM (M ± SD, Percentile Graph) · Outer Circumference (cm)
Outputs (Research Result Sheet)	· Body Composition Summary (Fat Free Mass, Fat Mass, Intracellular Water, Extracellular Water, Total Body Water, ECW/TBW: Whole Body, Right Arm, Left Arm, Trunk, Right Leg, Left Leg, Whole Body Weight) · Body Composition Analysis (Lean Mass, ICW, ECW, Fat Mass, ECW/TBW): Whole Body, Right Arm, Left Arm, Trunk, Right Leg, Left Leg	· Research Parameters (BMI, Percent Body Fat, Skeletal Muscle Mass, Soft Lean Mass, Protein, Minerals, Bone Mineral Content, Basal Metabolic Rate, Waist-Hip Ratio, Waist Circumference, Visceral Fat Area, Obesity Degree, Body Cell Mass, Arm Circumference, Arm Muscle Circumference, TBW/FFM, FFMI, FMI, SMI) · Whole Body Phase Angle (50 kHz) · Segmental Phase Angle (5 kHz, 50 kHz, 250 kHz: Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Impedance Graph (Each segment and each frequency)
Outputs (Comparison Result Sheet)	· Weight, Skeletal Muscle Mass, Body Fat Mass, ECW Ratio, Phase Angle: Whole Body (Today, Recent, Difference) · Lean Mass, ECW Ratio, Phase Angle: Right Arm, Left Arm, Trunk, Right Leg, Left Leg (Today, Recent, Difference)	· Cole-Cole Plot (Standard median curve, Today's Results, Previous Results)
Outputs (InBody Thermal Result Sheet)	· Personal Information · Muscle, Fat Analysis (Weight, Skeletal Muscle Mass, Soft Lean Mass, Body Fat Mass) · Obesity Analysis (BMI, Percent Body Fat) · Segmental Lean Analysis · Segmental ECW Ratio Analysis · Body Water Composition (Total Body Water, Intracellular Water, Extracellular Water) · Body Composition Analysis (Protein, Minerals, Body Fat Mass, Fat Free Mass, Bone Mineral Content) · Segmental Body Water Analysis · Segmental Fat Analysis · Segmental Lean & ECW Ratio Analysis (Assessment)	· Water Control · Research Parameters (Intracellular Water, Extracellular Water, Wholebody ECW Ratio, Skeletal Muscle Mass, Protein, Mineral, Bone Mineral Content, Body Cell Mass, Waist-Hip Ratio, Waist Circumference, Visceral Fat Area, Obesity Degree, Basal Metabolic Rate, Arm Circumference, Arm Muscle Circumference, FFMI, FMI, SMI, SMM/WT, TBW/FFM, ECM/BCM, TBW/WT, Adjusted FFM, Adjusted SMI) · Whole Body ECW Ratio · Segmental ECW Ratio · Whole Body Phase Angle (50 kHz: Right side of the body) · Segmental Body Phase Angle (5 kHz, 50 kHz, 250 kHz: Right Arm, Left Arm, Trunk, Right Leg, Left Leg) · Impedance (Each segment and each frequency)

* The above content is subject to change without prior notice for the purpose of improving device appearance and performance.
* Note that this is a medical device, and use it with proper care and knowledge of its precautions and instructions.
* The results about Blood Pressure or Hand Grip Stength are only available when integrated with InBody Blood Pressure Monitor. (BPBIO Series) or InBody Handgrip Dynamometer (InGrip).
* "QR Code" is registered trademark of DENSO WAVE INCORPORATED.

InBody

See what you're made of

The power of InBody

InBody maintains a high brand position with the highest level of technology.

Certifications obtained by InBody

InBody complies with the quality management system according to international standards. We satisfy country-specific regulatory requirements that apply to product safety and performance, and provide related services.

CE 1639 NAWI ISO9001 ISO13485 MDSAP GMP

InBody's Intellectual Property Rights

InBody owns patents and intellectual property rights around the world (Korea, U.S, China, Japan) and provides products with high accurancy and reproducibillity based on this technology.

InBody

InBody India [INDIA]
InBody India Pvt.Ltd.
Unit No. SF-B-03, 2nd Floor, B Wing, Art Guild House, Phoenix Market City, L.B.S Marg, Kurla West, Mumbai - 400070.
TEL: +91-22-6223-1911
Website: inbody.in E-mail: india@inbody.com

InBody UK [UNITED KINGDOM]
Orega, Belmont Road, Uxbridge, UB8 1HE, United Kingdom
TEL: +44-1530-569620
Website: uk.inbody.com
E-mail: uk@inbody.com

InBody Oceania [AUSTRALIA]
InBody Oceania Pty Ltd.
U2/82-86 Minnie Street, Southport, Queensland
TEL: +61-7-5681-1900
Website: au.inbody.com
Email: oceania@inbody.com

InBody HQ [KOREA]
InBody Co., Ltd.
InBody Bldg., 625, Eonju-ro, Gangnam-gu, Seoul 06106 Republic of Korea
TEL: +82-2-501-3939 FAX: +82-2-6919-2417
Website: inbody.com
E-mail: info@inbody.com

InBody Europe B.V. [NETHERLANDS]
InBody Europe B.V.
Gyroscoopweg 122, 1042 AZ, Amsterdam, The Netherlands
TEL: +31-20-238-6080 FAX: +31-6-5734-1858
Website: nl.inbody.com
E-mail: info.eu@inbody.com

Biospace Latin America [MÉXICO]
Insurgentes Sur 1457, Piso 15 Int.2. Col. Insurgentes Mixcoac, Alcaldia Benito Juarez, C.P. 03920, Ciudad de México, México
TEL: +52-55-5025-0147
Website: inbodymexico.com
E-mail: info.mx@inbody.com

InBody USA [USA]
Biospace Inc. dba InBody
13850 Cerritos Corporate Dr. Unit C Cerritos, CA 90703 USA
TEL: +1-323-932-6503 FAX: +1-323-952-5009
Website: inbodyusa.com
E-mail: info.us@inbody.com

InBody Europe B.V. Niederlassung Deutschland [GERMANY]
InBody Europe B.V.
Mergenthalerallee 15-21, 65760 Eschborn, Germany
TEL: +49-6196-76-916-62 FAX: +49-6196-76-916-11
Webseite: de.inbody.com
E-mail: erfolg@inbody.com

InBody Asia [MALAYSIA & SINGAPORE]
InBody Asia Sdn. Bhd.
Unit 3A-11, Oval Damansara, 685 Jalan Damansara Kuala Lumpur, WP KL 60000 Malaysia
TEL: +60-3-7732-0790 FAX: +60-3-7733-0790
Website: inbodyasia.com
E-mail: info@inbodyasia.com

InBody Japan [JAPAN]
InBody Japan Inc.
Tani Bldg., 1-28-6, Kameido, Koto-ku, Tokyo 136-0071 Japan
TEL: +81-3-5875-5780 FAX: +81-3-5875-5781
E-mail: inbody@inbody.co.jp

InBody BWA Inc. [USA]
InBody BWA Inc.
2550 Eisenhower Avenue, Suite C 209, Audubon, PA 19403
TEL: +1-610-348-7745
Website: inbodybwa.com
E-mail: bwainquiries@inbody.com

InBody China [CHINA]
Biospace China Co., Ltd.
Room 306b、307a、307b、308, MT2 Office Building, No. 3999 Hongxin Road, Minhang District, Shanghai
TEL: +86-21-6443-9705 FAX: +86-21-6443-9706
Website: inbodychina.com
E-mail: info@inbodychina.com

ⓒ2025 InBody Co., Ltd., All rights reserved. IC-ENG-U9-B-250306