# III. CT vs InBody



## Visceral fat area

| | Country | Research | Publication Year | Journal | Device | Subjects | Correlation Coefficient |
|---|---|---|---|---|---|---|---|
| 1 | China | Comparison of visceral fat area measured by CT and bioelectrical impedance analysis in Chinese patients with gastric cancer: a cross-sectional study | 2020 | BMJ Open | InBody 720 | 157 gastric cancer patients | Visceral fat<br>R=0.650 |
| 2 | Japan | InBody720 as a new method of evaluation visceral obesity | 2011 | Hepato-Gastroentertology | CT vs InBody720 | 50 Stomach Cancer surgery Patients, 37-85 years old | r= 0.759<br>CT: 128.9cm$^2$<br>InBody : 103.0cm$^2$ |

## Skeletal muscle mass

| | Country | Research | Publication Year | Journal | Device | Subjects | Correlation Coefficient |
|---|---|---|---|---|---|---|---|
| 1 | Japan | Feasibility of computed tomography-based assessment of skeletal muscle mass in hemodialysis patients | 2020 | Journal of Nephrology | unknown | 138 HD patients | SMM vs Psoas<br>R = 0.770 |
| 2 | Korea | Multifrequency Bioelectrical Impedance Analysis Compared With Computed Tomography for Assessment of Skeletal Muscle Mass in Primary Colorectal Malignancy: A Predictor of Short-Term Outcome After Surgery | 2019 | Nutrition in Clinical Practice | InBody 770 | 50 Colorectal Cancer Patients | SMM vs Psoas<br>R = 0.705 |
| 3 | Japan | Dual Energy X-ray Absorptiometry and Bioimpedance Analysis are Clinically Useful for Measuring Muscle Mass in Kidney Transplant Recipients With Sarcopenia | 2018 | Transplantation Proceedings | InBody 720 | 58 Kidney Transplant Recipients | SMM vs Psoas<br>R = 0.761 |

# III. CT vs InBody



CT vs InBody720

## InBody 720 as a New Method of Evaluating Visceral Obesity

*Hisataka Ogawa[1], Kazumasa Fujitani[1], Toshimasa Tsujinaka[1], Kenji Imanishi[2], Hatsumi Shirakata[2], Aiko Kantani[2], Motohiro Hirao[1], Yukinori Kurokawa[1], Shigetoshi Utsumi[2]*

[1]Department of Surgery and [2]Nutrition Support Service,
National Osaka Medical Center, Osaka, Japan.
Corresponding Author: Kazumasa Fujitani, MD,
Zip Code: 540-0006 2-1-14 Hoenzaka, Chuo-ku, Osaka, Japan
Tel: +81669421331, Fax: +81669436467, E-mail: fujitani@onh.go.jp





**FIGURE 3** A scatter plot of VFA.

# III. CT vs InBody



**CT vs InBody 770**

## Multifrequency Bioelectrical Impedance Analysis Compared With Computed Tomography for Assessment of Skeletal Muscle Mass in Primary Colorectal Malignancy: A Predictor of Short-Term Outcome After Surgery


Eun Young Kim, MD, PhD[1]; Seong Ryong Kim, MD[2]; Daeyoun David Won, MD[2]; Moon Hyung Choi, MD[3]; and In Kyu Lee, MD, PhD,[2]

[1]Division of Trauma and Surgical Critical Care, Department of Surgery, Seoul St. Mary's Hospital, Seoul, South Korea

[2]Division of Colorectal Surgery, Department of Surgery, Seoul St. Mary's Hospital, Seoul, South Korea

[3]Department of Radiology, Seoul St. Mary's Hospital, Seoul, South Korea






**Figure 2.** Bland-Altman plot for assessing agreement between SMI on CT scan and total body skeletal muscle mass determined by BIA. BIA, bioelectrical impedance analysis; CT, computed tomography; SMI, skeletal muscle index.

# III. CT vs InBody




**CT vs DEXA vs InBody 720**

**Dual Energy X-ray Absorptiometry and Bioimpedance Analysis are Clinically Useful for Measuring Muscle Mass in Kidney Transplant Recipients With Sarcopenia**

M. Yanishi[a,*], H. Kinoshita[a], H. Tsukaguchi[b], Y. Kimura[c], Y. Koito[a], E. Jino[a], M. Watanabe[a], M. Sugi[a], and T. Matsuda[a]

[a]Department of Urology and Andrology, Kansai Medical University, Osaka, Japan; [b]2nd Department of Internal Medicine, Division of Nephrology, Kansai Medical University, Osaka, Japan; and [c]Health Science Center, Kansai Medical University, Osaka, Japan




*Transplantation Proceedings*, 50, 150-154 (2018)