	InBody S10	BCM
Product		
Absolute Measurement	Impedance Reactance Segmental phase angle	Impedance Reactance Whole phase angle
Hydration	Body water Intracellular water Extracellular water ECW/TBW Dry weight (3criteria) Segmental ECW ratio	Body water Intracellular water Extracellular water ECW/ICW Dry weight

	InBody S10	BCM
Fat	Body fat mass Percent body fat BMI Visceral fat area Waist Hip ratio Arm circumference	Fat mass Adipose tissue mass Fat tissue index
Muscle	Soft lean mass Fat free mass Skeletal muscle mass Segmental lean analysis SMI	Lean tissue mass Lean tissue index
Other Nutritional Factors	Body cell mass Basal metabolic rate Protein Mineral Bone mineral content Weight control	Body cell mass -

	InBody S10	BCM
	Technical Specifications	
Software	LB120 Data management software	Fluid management tool
Data storage	100,000 data in device	Only in patient cards
Data output	USB storage device	Smart card writer
Result sheet	Body composition/ Body water I, II	Result available on software
Print	Print out through computer/USB cable	Print out through computer
Display	Touch screen Key pad	LC-Display buttons
Calibration	Pre calibration	Pre calibration
Time	1 min 50 sec	2 min
Weight	2kg	2kg
Dimension (mm)	220 x 322 x 53	272 x 302 x 130
Age range	3~99 years	2~95 years
Posture	Lying, seated, standing	lying

	InBody S10	BCM
	Technology	
Principal	BIA	BIS
Frequency	6 frequencies (1,5,50,250,500,1000)	50 discrete frequencies (5-1000)
Electrode	8-point thumb electrode	4-point electrode
	Touch or adhesive type (reusable)	Adhesive type (For one time)
Empirical estimations	X Shows the patient result as it is.	O
Measurement	Direct segmental	Whole body (Right half of the body)

InBody

FRESENIUS MEDICAL CARE

InBody S10

BCM

InBody Body Water

ID	Height	Age	Gender	Test Date / Time
27687	159cm	53	Male	2017.09.30. 13:58

Body Water Composition

	Under	Normal	Over	
TBW		32.2		(L)
ICW		18.5		(L)
ECW		13.7		(L)

ECW Ratio Analysis

	Under	Normal	Over	
ECW Ratio		0.424		

Segmental Body Water Analysis

	Under	Normal	Over	
Right Arm		1.53		(L)
Left Arm		1.78		(L)
Trunk		14.0		(L)
Right Leg		5.32		(L)
Left Leg		5.12		(L)

Segmental ECW Ratio Analysis

	Right Arm	Left Arm	Trunk	Right Leg	Left Leg
Over			0.423	0.430	0.432
Slightly Over		0.404			
Normal	0.398				

Body Water Composition History

Weight	(kg)	53.9	53.6
TBW	(L)	30.4	32.2
ICW	(L)	18.3	18.5
ECW	(L)	12.1	13.7
ECW Ratio		0.399	0.424

Recent / Total

Body Composition Analysis

Protein	8.0 kg	(8.4~10.2)
Minerals	3.02 kg	(2.89~3.53)
Body Fat Mass	10.3 kg	(6.7~13.4)
Fat Free Mass	43.2 kg	(42.5~52.0)
Bone Mineral Content	2.55 kg	(2.38~2.92)

Muscle-Fat Analysis

Weight	53.5 kg	(47.3~63.9)
Skeletal Muscle Mass	22.2 kg	(23.5~28.7)
Soft Lean Mass	40.7 kg	(40.1~49.1)
Body Fat Mass	10.3 kg	(6.7~13.4)

Obesity Analysis

BMI	21.2 kg/m²	(18.5~23.0)
PBF	19.2 %	(10.0~20.0)

Segmental Lean Analysis

Right Arm	1.95 kg
Left Arm	2.27 kg
Trunk	17.7 kg
Right Leg	6.71 kg
Left Leg	6.46 kg

Research Parameters

Basal Metabolic Rate	1304 kcal	
Visceral Fat Area	65.0 cm²	
Body Cell Mass	26.6 kg	(27.8~34.0)
Arm Circumference	25.9 cm	
Arm Muscle Circumference	24.1 cm	
TBW/FFM	74.4 %	
BMI	6.9 kg/m²	

Reactance

	RA	LA	TR	RL	LL
Xc 5kHz	5.9	4.6	0.6	2.8	2.2
50kHz	15.8	12.3	0.9	6.8	6.5
250kHz	22.0	16.6	0.8	3.7	4.4

Whole Body Phase Angle

2.8 °


	RA	LA	TR	RL	LL
Φ 50kHz	3.1	2.8	3.3	2.3	2.1

Impedance


	RA	LA	TR	RL	LL
Zan 1kHz	310.0	266.1	17.9	179.7	185.0
5kHz	307.3	264.0	17.6	179.2	183.9
50kHz	292.6	252.7	16.5	170.5	177.0
250kHz	274.3	237.3	15.3	161.8	168.1
500kHz	264.4	229.2	14.7	159.0	165.1
1000kHz	249.0	217.7	14.0	155.6	161.7

[Touch Type , Lying Posture]


Body Composition Plot




BP Pre Plot




Overhydration Plot



Nutrition Reference Plot



Hydration Reference Plot



Cole-Cole Plot

