


Sample Name: B03

=====

Acq. Operator	: Federico	Seq. Line	: 15
Acq. Instrument	: Q6120	Location	: Vial 15
Injection Date	: 1/17/2023 4:33:30 PM	Inj	: 1
		Inj Volume	: 1.000 µl
Sequence File	: D:\CHEM32\1\DATA\Sarah-Be\2023_01_17-BES-BD-006 2023-01-17 16-12-22\2023_01_17-BES-BD-006.S		
Acq. Method	: D:\CHEM32\1\DATA\SARAH-BE\2023_01_17-BES-BD-006 2023-01-17 16-12-22\ISO_A-B_FIA_05ML_1M		
Last changed	: 12/6/2021 6:41:11 PM by StefanP		
Analysis Method	: D:\CHEM32\1\DATA\SARAH-BE\2023_01_17-BES-BD-006 2023-01-17 16-12-22\ISO_A-B_FIA_05ML_1MIN_TARGET.M (Sequence Method)		
Last changed	: 12/6/2021 6:41:11 PM by StefanP		
Method Info	: Method for flow-injection analysis in positive ESI mode using eluents A and B		



=====
Area Percent Report
=====

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MSD1 TIC, MS File

Peak #	RetTime [min]	Type	Width [min]	Area	Height	Area %
1	0.293	BBA	0.3732	6.16293e6	2.25667e5	100.0000

Totals : 6.16293e6 2.25667e5

Signal 2: MSD2 TIC, MS File

Peak #	RetTime [min]	Type	Width [min]	Area	Height	Area %
1	0.172	BV	0.0609	3.43056e5	9.08288e4	15.0456
2	0.281	VBA	0.2449	1.93705e6	1.11371e5	84.9544

Totals : 2.28011e6 2.02199e5

Signal 3: MSD3 TIC, MS File

Peak #	RetTime [min]	Type	Width [min]	Area	Height	Area %
1	0.211	BBA	0.2095	9.43802e5	5.97351e4	100.0000

Totals : 9.43802e5 5.97351e4

Signal 4: VWD1 A, Wavelength=271 nm

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	0.153	BBA	0.0955	111.79093	16.00056	100.0000

Totals : 111.79093 16.00056

=====
*** End of Report ***