

M.J. Biopharm Private Limited Plot No. 18, International Biotech Park, Hinjawadi, Phase II, Pune 411057.	
Preliminary Investigation	

Deviation Number: DEV/QC/25/010

Where the deviation was observed (Department/Area)? Quality Control - Instrument Lab - II (FF-16)
When the deviation was observed? Date/Time of Occurrence: 21/03/2026 & 12:50 Date/Time of Identification: 21/03/2026 & 13:00 Date/Time of Logging: 21/03/2026 & 13:47
Product Name: NA
Batch Number: NA
Stage (If Applicable): Not Applicable
Equipment/System Detail (If any): TOC Analyzer (I/QC/018)
Brief Description of Deviation: While performing the system suitability test (SST) of TOC Analyzer, Instrument ID. I/QC/018, on 21/03/2026 at 13:00, in Instrument Lab - II (FF-16), analyst (Emp.ID. 615) observed that, obtained SST results (Recovery of Benzoquinone) is not within the acceptance criteria i.e. 115.1 % as per SOP/DP/QC/045. (Limit: 85 % to 115 %). This non-conformance is not in compliance with SOP No. SOP/DP/QC/045. As per SOP No.: SOP/DP/QC/045, Revision No: R02, Title: Operation, Calibration, Maintenance of Shimadzu TOC Analyzer Model No. TOC-L_{CPH} with LabSolutions Software, Section No.: 7.4.3.26 mentioned as 'The results of system suitability test must not be less than 85 % & must not more than 115 %'.
Possible Impact of the Deviation: The SST failure (Recovery of Benzoquinone 115.1%, above the limit of 85%–115% as per SOP/DP/QC/045) indicates that, TOC Analyzer, Instrument ID. I/QC/018, delivered the elevated results than the specified acceptance limit and it was identified during the calibration, therefore stopped the further calibration activity and as an immediate action, it was informed to perform the alternative test i.e. 'Oxidizable Substance' as per purified water standard test procedure i.e. DP/STP/UT/0003 for the scheduled water sample analysis. Therefore, impact on product quality and analyzed samples not anticipated at this stage..

M.J. Biopharm Private Limited
Plot No. 18, International Biotech Park, Hinjawadi, Phase II, Pune 411057.



Preliminary Investigation

CORRECTION (Immediate action taken):

1. The ongoing calibration activity was immediately stopped after observing that the SST result (% Recovery of Benzoquinone – 115.1%) was outside the acceptance criteria (85.0%–115.0%).
2. All glassware's which were used for blank and preparation of standard solutions were rechecked and quarantined the analysis tray.
3. The "Out of Calibration" Label was affixed on instrument as per SOP/DP/QA/020.
4. Verified the 2-point calibration result and noted that result is well within acceptance criteria.
5. As the System Suitability Test (SST) results did not meet the acceptance criteria, therefore as an immediate action, it was informed to perform the alternative test i.e. 'Oxidizable Substance' as per purified water standard test procedure i.e. DP/STP/UT/0003 for the scheduled water sample analysis.
6. The observed non-conformance immediately informed to Team lead and Head QC.

Root Cause or Most Probable Root Cause:

Based on the outcome of preliminary investigation and analyst interview, assignable cause could not be identified, however suspected cause for the occurred non-conformance could be dilution error which may have occurred inadvertently while preparing the standard solution, leading to a slightly higher recovery of the 1, 4-Benzoquinone, therefore to find out the root cause additional investigation followed by hypothesis testing needs to be performed.

CAPA Identified if any:

A detailed investigation shall be carried out, and based on the outcome of the investigation, appropriate CAPA shall be proposed and implemented to prevent recurrence.

Done By:	Name: <i>Nilesh Dushing</i>	Sign & Date: <i>[Signature]</i> 21/03/26
Reviewed By: (User <u>Head</u> /Designee)	Name: <i>Sagar Pawar</i>	Sign & Date: <i>[Signature]</i> 21/03/26
Reviewed By: (User <u>Head</u> /Designee)	Name: <i>Sagar K. Pali</i>	Sign & Date: <i>[Signature]</i> 21/03/26
Reviewed By: (User <u>Head</u> /Designee)	Name: <i>Ganpat Kodes</i>	Sign & Date: <i>[Signature]</i> 21/03/26
Reviewed By: (QA Representative)	Name: <i>Charanya Chougale</i>	Sign & Date: <i>[Signature]</i> 21/03/26
Approved By: (QA <u>Head</u> /Designee)	Name: <i>Pishan Jod</i>	Sign & Date: <i>[Signature]</i> 21/03/26