

Reference SOP No.: SOP/DP/QC/045

Unit: Drug Product

M.J. Biopharm Private Limited

Plot No. 18, International Biotech Park, Hinjawadi, Phase II, Pune 411057.

**CALIBRATION DATA SHEET FOR SYSTEM SUITABILITY TEST****1.0 Analysis details**

Analysis started on:	21/03/26	Analysis completed on:	21/03/26
----------------------	----------	------------------------	----------

2.0 Instrument details

Instrument Name	Instrument ID	Calibration Due On:
TOC Analyser	I1001018	21/03/26
Weighing Balance	I1001013	22/03/26
Micropipette	NA	NA

3.0 Standard Details

Standard Used :	1,4 Benzoquinone	Make :	USP
Batch/ Lot No.:	PS100210 PS1002123 ver	Expiry Date:	Current lot

4.0 Preparation of 1,4 Benzoquinone Standard Solution

Primary Standard Used:	1,4 Benzoquinone	Batch No.:	PS1002103 -002	Valid Up to:	Current lot
Benzoquinone (weight) 7.499 mg dissolved in 100 ml purified water. (50000 ppb)					
Pipetted out 1 ml of 50000 ppb, Benzoquinone solution & diluted to 100 ml with purified water (500 ppb)					
Prepared By:	A 21/03/24		Reviewed By:	B 21/03/26	
Sign & Date:			Sign & Date:		

5.0 Calculation

Area of Standard Benzoquinone solution (500 ppb) : 48.19			
Standard Deviation (SD)	1.227	Coefficient of Variation (CV)	2.55 %
Area of Standard sucrose solution (from 2 point calibration report) (500 ppb) : 42.58			
Standard Deviation (SD)	0.4564	Coefficient of Variation (CV)	1.07 %
Area of Blank (from 2 point calibration report) : 5.432			
Standard Deviation (SD)	0.1581	Coefficient of Variation (CV)	2.91 %

M.J. Biopharm Private Limited
Plot No. 18, International Biotech Park, Hinjawadi, Phase II, Pune 411057.



CALIBRATION DATA SHEET FOR SYSTEM SUITABILITY TEST

$$\text{Recovery of Benzoquinone} = \frac{\text{Area of Benzoquinone (Rss)} - \text{Area of Blank (Rw)} \times 100}{\text{Area of Sucrose (Rs)} - \text{Area of Blank (Rw)}}$$

$$= \frac{48.19 - 5.432 \times 100}{42.58 - 5.432} = 115.1017$$

= 115.1

Acceptance criteria:

Recovery of Benzoquinone: Between 85 -115 %

Standard Deviation: Not More than 0.50% or Coefficient Variation: Not more than 5.0%.


6.0 Result:

1. Calculated Benzoquinone recovery = 115 %

2. For 1,4 Benzoquinone solution (500 ppb)

Standard Deviation (SD)	1.227	Coefficient of Variation (CV)	% 2.55
-------------------------	-------	-------------------------------	--------

7.0 Conclusion: System suitability **Complies** / does not comply.

Analyzed by (Sign/date):	Reviewed by (Sign/date):
 21/03/26	 21/03/26