

Reference SOP No.: SOP/DP/QC/045

Unit: Drug Product

M.J. Biopharm Private Limited Plot No. 18, International Biotech Park, Hinjawadi, Phase II, Pune 411057.		
CALIBRATION DATA SHEET FOR TWO POINT CALIBRATION TEST.		

Instrument Name	TOC Analyzer		
Instrument ID	I1001018		
Start Date of Calibration	21/03/26	Calibration done on	21/03/26
Analytical Reference No.	212510438	Next calibration Due Date	21/04/26

1.0 Analysis details

Analysis started on:	21/03/26	Analysis completed on:	21/03/26
----------------------	----------	------------------------	----------

2.0 Instrument details

Instrument Name	Instrument ID	Calibration Due On:
Weighing Balance	I1001013	22/03/26
Micropipette	NA	NA

3.0 Standard Details

Standard Used :	Sucrose	Make :	SRL
Batch/ Lot No.:	9510091001	Expiry Date:	10/2029

4.0 Preparation of Sucrose Standard Solution

Primary Standard Used :	Sucrose	Batch No.:	9510091001	Valid Up to:	17/02/27
Sucrose (weight) <u>11.883</u> mg dissolved in <u>100</u> ml purified water.					
Pipetted <u>1</u> ml of 50000 ppb sucrose stock solution & diluted to <u>100</u> ml with purified water (500 ppb)					
Prepared By: <u>AD</u> Sign & Date: <u>21/03/26</u>			Reviewed By: <u>AS</u> Sign & Date: <u>21/03/26</u>		

5.0 Calculation

Area of Standard sucrose solution (500 ppb) : <u>42.58</u>			
Standard Deviation (SD)	<u>0.4564</u>	Coefficient of Variation(CV)	<u>1.07</u> %
Area of Blank : <u>5.432</u>			
Standard Deviation (SD)	<u>0.1581</u>	Coefficient of Variation(CV)	<u>2.91</u> %

M.J. Biopharm Private Limited
Plot No. 18, International Biotech Park, Hinjawadi, Phase II, Pune 411057.



CALIBRATION DATA SHEET FOR TWO POINT CALIBRATION TEST.

$$\text{TOC of Blank water in ppb} = \frac{500 \times \text{Area of Blank (rw)}}{\text{Area of Sucrose (rs)} - \text{Area of Blank (rw)}}$$

$$= \frac{500 \times 5.432}{42.58 - 5.432} = 73.1129 \approx 73 \text{ ppb}$$

$$= 73 \text{ ppb}$$

Acceptance criteria:


TOC of Blank water: Not More than 100 ppb

Standard Deviation: Not More than 0.50% or Coefficient Variation: Not more than 5.0%.

6.0 Result:

1. TOC of Blank Water: <u>73</u> ppb			
2. For 0 ppb			
Standard Deviation (SD)	0.1581	Coefficient of Variation (CV)	% 2.91
3. For 500 ppb			
Standard Deviation (SD)	0.4564	Coefficient of Variation (CV)	% 1.07

7.0 Conclusion: 2 point calibration Complies/ does not comply.

Analyzed by (Sign/date)	Reviewed by (Sign/date)
 21/03/26	 21/03/26