Equipment Care, Maintenance & Fault Log

Document No. AUR-SOP-09 · Version 0.1 (Draft) · Owner: General Manager · Effective: [Date]

> Elevé Health operating manuals, protocols and cleaning product lists have **not yet been received**. Every setting, check and product in this SOP is marked [per Elevé protocol — pending] and must be updated when the manuals arrive. Where Elevé guidance differs from this SOP, Elevé guidance applies.

1. Purpose and scope

Every machine at The Aura Reset Center is supplied, installed, calibrated and serviced by **Elevé Health**. We are Elevé Health's first partner clinic in the UAE. We hold **one unit of each machine**. If a unit develops a fault, that modality is **unavailable** until Elevé clears it. Careful daily checks, quick fault reporting and calm rebooking protect our guests and our schedule.

This SOP covers user checks by Hosts, the fault procedure, rebooking affected guests and the maintenance and calibration log. Treatment-by-treatment operating steps are in the separate modality SOPs [references]. Cleaning between guests is in AUR-SAF-02.

2. Our equipment

Unit	Elevé model / spec	Serial No.	Installed	Service interval
Whole-Body Cryotherapy	E-CRYO PRO electric chamber, to –110 °C, thermal-vision monitoring, no nitrogen	[___]	[Date]	[per Elevé protocol — pending]
Red-Light Bed	Full-body light bed, 660 / 850 nm	[___]	[Date]	[per Elevé protocol — pending]
Infrared Sauna	Full-spectrum infrared sauna room, 45–60 °C	[___]	[Date]	[per Elevé protocol — pending]
Hyperbaric Oxygen chamber	1.3–2.0 ATA	[___]	[Date]	[per Elevé protocol — pending]
Dry Float Bed	Warm membrane over heated water	[___]	[Date]	[per Elevé protocol — pending]
Salt Cabin	E-Salt Cabin: dry salt air, red light, oxygen, aromatherapy	[___]	[Date]	[per Elevé protocol — pending]
Cold Plunge	Elevé · Polar Recovery, filtered water at 5 °C	[___]	[Date]	[per Elevé protocol — pending]
Compression Recovery	Sequential pneumatic	[___]	[Date]	[per Elevé protocol —

Unit	Elevé model / spec	Serial No.	Installed	Service interval
	boots, 7 zones			pending]

> Hyperbaric Oxygen is offered as a non-medical wellness session [confirm licensing requirements for hyperbaric with DHA / licensing authority].

Only Hosts trained by Elevé may operate a unit. Only Elevé technicians may open panels, change settings, repair or calibrate a unit.

3. User checks

3.1 DAILY CHECKS (HOSTS)

Recorded on the opening checklist (AUR-SOP-01) and repeated at handover where shown.

Unit	Daily check	At handover
Whole-Body Cryotherapy	Reaches target temperature in normal time; touchscreen and thermal camera work; door opens freely from inside; floor dry	Temperature reached ☐
Red-Light Bed	All panels light evenly; no flicker; lid moves smoothly; eyewear clean	—
Infrared Sauna	Heats to set temperature; control panel and timer work; door, glass and bench undamaged	Temperature ☐
Hyperbaric Oxygen	Self-test passes; seals and door/zip intact; intercom and emergency release checked [per Elevé protocol — pending]	—
Dry Float Bed	Water heats to set temperature; membrane intact, no leaks	—
Salt Cabin	Salt, light, oxygen and aroma functions run on test; door seal intact	—
Cold Plunge	Water at 5 °C (logged, AUR-SOP-01); water clear; chiller and filter running; no leaks	Temperature logged ☐
Compression Recovery	Charged; all 7 zones inflate; hoses and zips undamaged	Charge ☐

3.2 WEEKLY CHECKS (SENIOR HOST ON [DAY])

Unit	Weekly check	Done (date / initials)
All units	Power cables, plugs and sockets	_____

Unit	Weekly check	Done (date / initials)
	undamaged; area around unit clear for ventilation	
Whole-Body Cryotherapy	[Filter / defrost check per Elevé protocol — pending]	_____
Infrared Sauna	Wood and bench inspected for cracks or splinters	_____
Hyperbaric Oxygen	[Weekly user check per Elevé protocol — pending]	_____
Dry Float Bed	Membrane inspected closely for wear	_____
Salt Cabin	[Salt top-up / aroma refill per Elevé protocol — pending]	_____
Cold Plunge	Filter checked; [water test / sanitiser level per Elevé protocol — pending]	_____
Compression Recovery	Garments inspected for wear; battery life normal	_____
Laundry machine(s), refreshment fridge	Seals, filters, temperature normal (AUR-SOP-05, AUR-SOP-06)	_____

4. Fault procedure: stop, tag out, notify

Use this procedure for any fault, error message, unusual noise, smell, leak, temperature outside range or guest report of something "not right".

1. **Stop.** If a guest is in the unit, end the session safely using the normal or emergency stop [per Elevé protocol — pending]. Help the guest out. Guest safety comes first; if the guest is unwell, follow the Emergency & Incident Response Procedure (AUR-SAF-04).
2. **Make safe.** Switch off and isolate the unit only if Elevé guidance allows [per Elevé protocol — pending]. Do not try to repair it.
3. **Tag out.** Hang the "OUT OF SERVICE — DO NOT USE" tag on the unit with the date, time, fault and your initials. Block the unit in [Booking system — to be chosen].
4. **Notify the General Manager** immediately by phone. If not reachable, follow the escalation chain in AUR-SOP-10.
5. **Contact Elevé Health** technical support: [Elevé support contact — pending]. Give the unit, serial number, error code, what happened and photos. Faults are reported to Elevé the same day (AUR-GOV-02).
6. **Log** the fault in the fault log (section 6).
7. **Rebook affected guests** (section 5).

8. **Return to service** only when an Elevé technician (or Elevé remote support, in writing) confirms the unit is safe. The General Manager removes the tag and signs the log.

> Only the General Manager may decide to take a unit out of service for longer than the day, or return it to service (AUR-GOV-02). Any Host may tag out a unit on safety grounds at any time.

5. Rebooking affected guests

Because we have one unit of each machine, a fault affects every guest booked on that unit until it is fixed.

1. Print or list every booking on the unit for the rest of the day and the next [2 days].
2. Contact each guest by WhatsApp or phone using the "unit unavailable" template (AUR-SOP-03), in time order.
3. Offer, in this order (AUR-CLI-03): a suitable alternative modality at the same time, a new time for the same modality, or a full refund of any prepaid amount.
4. Check the alternative against the guest's screening (AUR-FRM-03) and the Healer's plan.
5. For a session stopped part-way through because of a fault, follow AUR-CLI-03: replacement session plus [goodwill gesture], or full refund.
6. Members: the session is not counted against their allowance.
7. Record every guest contacted and the outcome in the fault log notes.

> **Script:** "I'm so sorry, [Name]. Our [modality] needs some attention from our technical team today. I'd love to offer you [alternative] at the same time, or I can move you to [time]. Which would you prefer?"

6. Fault log

Date / time	Unit	Fault description (error code)	Action taken (tag out, Elevé ref.)	Guests affected / rebooked	Returned to service (date, by whom)
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

7. Maintenance and calibration log

Elevé installs, calibrates and services every unit on its clinical service schedule. Record every Elevé visit and remote session.

Date	Unit	Type (service, calibration, repair, protocol update)	Elevé technician / ref.	Next due	Checked by (General Manager)
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

- Keep every Elevé service report and calibration certificate in the equipment file [location], with a scan in [shared drive].
- The General Manager checks the service schedule monthly and books visits before they fall due.
- Book planned maintenance outside peak times where possible, and block the unit in the booking system at least [7 days] ahead.

8. Protocol updates

When Elevé sends a protocol update:

1. The General Manager records it in the maintenance log and the version in the equipment file.
2. The relevant modality SOP and this SOP are updated (AUR-GOV-02: equipment and protocol changes).
3. All Hosts are briefed at the next shift huddle and sign that they have read it.

Update date	Unit	Summary	Briefed at huddle (date)	All Hosts signed (☐)
_____	_____	_____	_____	☐

Update date	Unit	Summary	Briefed at huddle (date)	All Hosts signed (☐)
_____	_____	_____	_____	☐

9. Power failure

If the power fails during a session, follow AUR-SAF-04 (power failure). Do not restart any unit until power is stable and the unit's start-up check passes [per Elevé protocol — pending]. Log the event in the fault log.