

Dated: -11.11.2022

CENTRAL MEDICAL SERVICES SOCIETY

Ministry of Health and Family Welfare (Government of India)
2nd floor, Vishwa Yuvak Kendra, Pt. Uma Shankar Dikshit Marg, Teen Murti Marg,
Opposite Police Station Chankaya Puri, New Delhi-110021

Telephones: 011-21410905, 21410906

Email: gmproc.cmss@gmail.com

agminderjeetyadav.cmss@gmail.com

Tender No: CMSS/PROC/2022-23/NTTLs/021

For

Procurement of Semi Major Equipments

Under

NTTL Programme

Minutes of Pre-bid meeting held on 20th October 2022 at 11.00 AM

1. Following officials were present during the Pre-bid meeting:-

- (i) Ms. Anjana, GM (Procurement), CMSS
- (ii) Mr. D. Mohapatra, GM (Finance), CMSS
- (iii) Dr. B.S Charan, Addl. DDG, DGHS MoHFW
- (iv) Ms. Seema Arora, Consultant NTTLs
- (v) Ms. Akansha Jain, AGM (QA), CMSS
- (vi) Mr. Inderjeet Yadav, AGM (Proc), CMSS

2. Following representative from prospective bidder was present during the Pre-bid meeting:-

- (i) Mr. Varun Jugran, Toshvin Shimadzu
- (ii) Mr. Nikhil Bansal, Metrohm
- (iii) Mr. Deepak, Metrohm
- (iv) Mr. Sidarth Kondal, PerkinElmer

अंजना
Anjana
General Manager (Procurement)
Central Medical Services Society
Vishwa Yuvak Kendra, Chanakyapuri
New Delhi - 110021

Minutes of Pre-bid meeting

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
A	M/s Precision Components & Engineers		
1	Query Please modify the Serial 1.4 Optics/Light Source parameter from Dual Source – Deuterium / Tungsten, Double Beam , Monochromatic system to Dual Source – Deuterium / Tungsten , Double Beam , DOUBLE MONOCHROMATOR system otherwise you may not be able to achieve the Stray light parameter against your Serial Number 1.5; Stray Light at 220nm: <0.005%T because this level of higher Stray light can only be achieved by implementing a spectrophotometer having High Quality Optical system with DOUBLE MONOCHROMATOR Design.	Technical Specification, Annex 1B Serial Number 1.4 Dual source- Deuterium /tungsten, Double beam monochromatic system.	Agreed. Should be read as: “Dual Source - Deuterium/ Tungsten, Double beam Double Monochromator system.”
2	Query Please also change the Serial Number 1.9, Scan Rate: Please enhance the scan rate at least to 8000 nm/minutes as Scan rate: 2000 nm/min is too slow to scan the spectra in desired wave length region i.e.190-1100nm.	Technical Specification, Annex 1B Serial Number 1.9 Scanning speed rate : 2000nm/min or better	Agreed. Should be read as: “ Scanning Speed rate: 8000 nm/min or better.”
B	M/s SHIVA GLOBAL TEKNO SYS		
1.	Query: Optics/ Light source: Dual source Deuterium/ tungsten Double beam, Double Monochromator System Please insert DOUBLE in front of Monochromator in	Technical Specification, Annex 1B Serial Number 1.4 Dual source- Deuterium /tungsten, Double beam monochromatic system.	Agreed. Should be read as: “Dual Source - Deuterium/ Tungsten, Double beam Double Monochromator system.”


General Manager (Procurement)
Central Medical Services Society
Vishwa Yuvak Kendra, Chhankaspuri
New Delhi - 110021

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	order to get the stray light at 220 nm as <0.005% T because Higher stray light is the feature of Monochromator design. Normally this level of Stray light can only be achieved with high quality & sensitive type of Optics having Double Monochromator Design spectrometer. Therefor the revised specification should be as <u>Optics/ Light source: Dual source Deuterium/ tungsten Double beam, Double Monochromator System.</u>		
2.	Query: Photometric Range (Abs) :-4.0to + 4.0 Abs or better Please make it as 4.0to + 4.0 Abs or better as this level of absorbance is required when highly concentrated solutions are used. Or colored samples are analyzed, directly.	Technical Specification, Annex 1B Serial Number 1.8 Photometric Range (Abs): ± 3.5 Abs or better	Not agreed. Already covered.
3.	Query: Scan rate: 12000 nm/min or better Higher the scan speed better the scanning capability or through put and scan rate 2000 nm nm/min is too slow to scan.	Technical Specification, Annex 1B Serial Number 1.9 Scanning Speed rate: 2000nm/min or better	Should be read as: "Scanning Speed rate: 8000 nm/min or better."
reason to remove the line: 4.	Query: Minimum 75mm or better Integrating sphere or above for transmission & diffuse reflectance measurements on liquid and solid samples along with white standards. Please remove the necessity of black and color standards along with integrating sphere for transmission & diffuse	Technical Specification, Annex 1B Serial Number 4.2 4. Desirable Accessories Minimum of 75mm or better Integrating sphere or above for transmission and diffuse reflectance measurements on liquid and solid samples along with white, black and colour standards.	Agreed. From the specification point 4.2, remove the following "along with white, black and colour standards" (as integrated sphere itself comes with white standard plate for calibrations). So, no

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	reflectance measurements in case you are not doing color comparison or color measurement as per the US Chromaticity chart 1931 for the determination of Hunter LAB value, LUV value, LAB value etc.		additional standard required, initially.
C	M/s TOSHVIN Analytical Pvt. Ltd		
1.	Spectral Resolution Request to change to 0.9 cm⁻¹ to 16 cm⁻¹ To Read As: Variable from 0.9 cm⁻¹ to 16 cm⁻¹ Reason for Changes In case of FTIR the method are performed at spectral resolution of 4cm⁻¹ or 8 cm⁻¹ for all the application. 0.5 cm⁻¹ is specific range and does not allow other vendors to participate.	Technical Specification, Annex 1A Spectral Resolution ± 0.5 cm⁻¹ or better for full range.	Agreed. "Spectral Resolution 4 cm⁻¹ or better for full range."
2.	Detector Request to delete option of LiTaO₃ detector. To be read as DTGS/DLATGS detector with temperature control mechanism or better detectors Reasons for changes Below two detectors are equally comparable in terms of holding higher curie temperature approximate 30 0C more during the application is run on the FTIR system and are superior detectors for sustaining IR heat released during the analysis	Description and Specification Annex 1A Detectors DTGS/LiTaO₃ Detector or better detectors.	Not agreed. No change required.

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	is run on FTIR. 1). "DTGS with temperature control" and 2). Detector "DLATGS with temperature control". However, specific vendor has both DTGS and LiTaO₃ detector which can be verified on their website. LiTaO₃ detector they are pushing to be Lowest bidder in the tender without considering the quality of detector performance. We request committee to mention detector "DTGS with LiTaO₃ detector which can be verified on their website. LiTaO₃ detector they are pushing to be Lowest bidder in the tender without considering the quality of detector performance. We request committee to mention detector "DTGS with temperature control" / "DLATGS with temperature control". As the DTGS with Temperature Stabilization or DLATGS are considered the superior detectors as they can hold curie temperature higher as compared to other detectors. It is useful for sustain IR heat released during the analysis. This is the third time tender is floated but due to specification favouring to single vendor and does not encourage other vendors to participate. LiTaO₃ shall be deleted from the tender specification. Whereas LiTaO₃ detector are less sensitivity and S/N		

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	ratio performance will definitely be compromised in that case. Hence requested to delete the LiTaO3 from the tender specification.		
3.	Measurable Concentration Range Request to delete the point as it has no meaning in the tender specification Reason for Changes There shall be clarity to the vendors that which standard solution has to be run to prove "50 micromol or lower sensitivity" of the FTIR along with the method. The technical committee to provide standard solution and method to perform this during installation. The method shall be well proven and published in the journal. Otherwise this specification has not mean to be part of the tender specification. There shall be some base to prove this tender specification by the bidder.	Description and Specification Annex 1A Measurable Concentration Range: 50 micromolar lower	Agreed. From the specification remove the following: Measurable concentration range: 50 micromolar lower
4.	Additional Accessories:- To be deleted Reason for Changes Tender has asked for ATR and software to be part of the main instruments. Moreover only ATR and software is required for recording and performing Spectra. These two items "ATR and software" cannot be quoted as optional. It is typographic error.	Additional Accessories:- Accessories for recording spectra of liquids, pastes, films and solids should be quoted both in the main tender document and as Separate entity.	Agreed. From the specification remove the following "and as separate entity"

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
4.	Comprehensive Warranty As the equipment is covered for the warranty of 5 years and additional CAMC is so applicable in the tender. Asking 10 year for interferometer laser is beyond the vendor understanding. Either the entire equipment shall be covered under 10 year warranty or else 5 year of warranty shall be applicable. We have raised this point and would like to document it that why only a particular part/component has to be claimed for 10 years. There shall be complete comprehensive warranty of 10 years on complete instrument. Otherwise it is favoring to specific vendors to only participate in the tender. This again discourage other vendors to participate in the tender as it is extremely favoring to certain vendor only. Hence request to delete 10 year for interferometer laser. To be read as Comprehensive warranty of 5 years comprehensive maintenance. Reason for Change: We have raised this point and would like to document it that why only a particular part is required to be claim for 10 years. This discourage the vendors to participate in the tender as it is extremely favouring to certain vendor only. The complete equipment shall comes under the warranty and not	Description and Specification Annex 1A, Standards, safety and training Point 4 Comprehensive warranty for 5 years comprehensive maintenance and 10 year for interferometer laser.	Agreed. From the specification remove the following “and 10 years for interferometer laser.” It is suggested to reduce warranty and CAMC period to 2 years and 3 years respectively.

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	only a particular part or component shall be covered for 10 years. We want to raise a strong protest in this paragraph. Hence request to delete 10 year for interferometer laser.		
5.	Query for UV-VIS Spectrophotometer Request to delete Photo diode detector. To read as: PMT or a better detection System Reason for Change: PMT detector is most advance more superior detector as compared to Photo diode detector. In case of PMT detector the stray light sensitivity i.e < 0.005% or better can be achieved. This sensitivity is not possible by Photo diode detector. All reputed vendors has the model with PMT detector and can achieve required tender specification stray sensitivity. We request committee not to dilute the stray light sensitivity of < 0.005% or better. It will defeat the purpose of tender.	Description and Specification Annex 1B, Detector:- Photo diode/ PMT or a better detection system covering the whole range 190-1100nm.	Not Agreed. Already both options are given in tender specifications.
5.	Query for UV-VIS Spectrophotometer Wavelength scanning range The specification requested in the tender like stray light at 220 nm < 0.005% T or better is achievable with highly sensitive PMT detector which is the most advance detector for high sensitive application. However PMT detector range is from 185 nm to 900 nm. This is applicable to all the vendors.	Description and Specification Annex 1B, Wavelength scanning range : 190nm-1100nm or better	Not Agreed.

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	To be read as 190 – 900 nm or better Reason for Changes The routine application can be covered by Photo diode detector but when we consider the high sensitive application where the stray light play extremely important parameter in this tender specification case i.e. < 0.005% or better can only be possible with PMT detector. We also discourage the committee for diluting the stray light sensitivity i.e. < 0.005% or better otherwise it will defeat the purpose of actual use of the equipment during real time analysis in respective laboratories. Where high sensitivity stray light is required. Hence we request committee to consider range from 190 nm to 900 nm.		
6.	Query for UV-VIS Spectrophotometer We request the committee to consider the range from 60 mm or better. As all the vendors has the range from 60 mm or better. 60 mm diameter size is sufficient for the application. To be read as Integrated Sphere (with a diameter of 60 mm or better) Reason for Changes All the application related to reflectance or diffuse reflectance is possible with 60 mm diameter of integrating sphere.	Description and Specification Annex 1B, Point 1.11 Integrating sphere (with a minimum of 75 mm or better) for analysis of solid reflectance and transmission modes – please quote separately under desirable accessories as well as in the main tender documents.	Not Agreed. It will compromise the sensitivity of equipment, it is available with all vendors.



S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	Hence request the technical committee to consider it to 60 mm diameter.		
7.	Desirable Accessories To delete white, black and colour standards. To consider 60 mm integrating sphere. To be read as Minimum of 60 mm or better Integrating sphere or above for transmission & diffuse reflectance measurements on liquid and solid samples along with the colour measurement software. Reason for Changes white, black and colour standards is that there is no such standard available for UV-VIS Spectrophotometer application. Or otherwise, we request customer to provide the white, black and color standards name which are available for UV-VIS Spectrophotometer in the market. Also, rather than going for colour standard it is better to have colour measurement software which can take care of all colour measurement application. All the vendors has colour measurement software.	Description and Specification Annex 1B, 4. Desirable Accessories Point 4.2 Minimum of 75 mm or better Integrating sphere or above for transmission and diffuse reflectance measurements on liquid and solid samples along with white, black and color standards	Agreed. From the specification remove the following “along with white, black and color standards.”
8.	Desirable Accessories Request committee to delete this complete paragraph. Reason for Changes We request the technical	Description and Specification Annex 1B, 4. Desirable Accessories Point 4.3 Variable angle reflectance accessory for examination of solid	Agreed. Please remove this point from the specifications.

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	committee to delete complete parameters as Integrated sphere accessory is already asked with the main instrument and Integrating sphere has variable angles to perform reflectance studies. This paragraph has no role in performing the application on UV-VIS Spectrophotometer standalone or with Integrated sphere. Hence request to delete the same.	surface and their interfaces at various angles of reflection.	
9.	Delivery of instruments to site in 60 days from LOA and 30 days for instrument depending of site readiness.	For Delivery- 45 days from LOA and 15 days for installation & commissioning the equipment from the date of handing over the site.	For Delivery- 60 days from LOA and 30 days for installation & commissioning the equipment from the date of handing over the site.
D.	M/S PERKIN ELMER (INDIA) PVT		
1.	Query for FTIR with ATR Request to amend Wavelength accuracy to 0.1 cm-1 or better. Remarks To get wide participation, wavelength accuracy to be amended to 0.1cm-1.	Description and Specification Annex 1A, Wavelength Accuracy: 0.01 cm-1 or better.	Agreed. Please change it to: Wavelength Accuracy: 0.1 cm-1 or better
2.	Query for FTIR with ATR Request to amend Signal to Noise ratio as 30000:1 or better Remarks Higher signal to Noise ratio represents better performance of detector (Instrument)	Description and Specification Annex 1A, Signal to noise ratio : 20000:1 or better)	Not Agreed. All vendors are already covered.
3.	Query for FTIR with ATR Requested to specify Library name Or list of compounds to be included	Description and Specification Annex 1A, IR Spectral Library with required software to facilitate search important compounds (please quote separately)	Agreed. Library for compounds related to tobacco constituent.
4.	Query for UV-VIS Spectrophotometer Stray light at 220nm Requested to amend it to Stray	Description and Specification Annex 1B, Point 1.5 Stray light at 220nm:	Not Agreed.

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	Light at 220nm :<0.02%T or better Remarks 0.02%T straylight criteria passes the validation or calibration mentioned in different Pharmacopeia or different Regulatory guidelines.	<0.005%T or better.	
5.	Query for UV-VIS Spectrophotometer Requested to amend Integrating Sphere to 50nm or better. Remarks 50mm Integrating sphere having low noise and high energy throughput . Also, it will allow more vendor participation	Description and Specification Annex 1B, Point 1.11 Integrated sphere (with a minimum size of 75nm or better) for analysis of solid reflectance and transmission modes-please quote separately under desirable accessories as well as in the main tender document.	Not Agreed. Sensitivity compromised. All vendors have this feature.
6.	Query for Spectrofluorophotometer Requested to amend lamp source to Pulse Xenon Lamp Remarks Pulse xenon lamp prevent photobleaching of sample material. And it is preferred for all kind of sample analysis.	Description and Specification Annex 1C, Lamp Source: 150W Xenon Lamp	Not Agreed. It is vendor specific.
7.	Query for Spectrofluorophotometer Requested to amend Signal-to-noise level to 750:1 or better. Remarks This will allow wider vendor participation.	Description and Specification Annex 1C, Signal-to-noise level: 10000:1 RMS or better	Not Agreed. Sensitivity will reduce.
8.	Query for Spectrofluorophotometer Reference detector Please clarify the phenomena need to monitor during reaction progress. 1. Single emission & excitation 2. Single excitation & emission range recording	Description and Specification Annex 1C, Reference detector: Standard photo diode/PMT	Not Agreed. Not required.


 General Manager (Procurement)
 Central Medical Services Society
 Vishwa Yuva Kendra, Chanakyapuri
 New Delhi - 110021

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	3. Duration of time		
E	M/S ELRON INSTRUMENT CO. PVT. LTD		
1.	Query for UV VIS Spectrophotometer Please change the photometric range from -3.0 Abs nm to +3.0 Abs	Description and Specification Annex 1B, Point 1.8 Photometric Range (Abs): ± 3.5 Abs or better	Not Agreed. Not required.
2.	Query for UV VIS Spectrophotometer Optics/ Light Source: Please change it to Dual Source- Deuterium/Tungsten, Double beam Double Monochromator System .[Here the word "MONOCHROMATIC" should be replaced by the word "MONOCHROMATOR" . Again the Word DOUBLE has to be prefixed before MOMOCHROMATOR, Otherwise the word "MONOCHROMATIC" does not carry any significance or meaning & completely incorrect.	Description and Specification Annex 1B, Point 1.4 Optics/ Light Source: Dual source- Deuterium /tungsten, Double beam monochromatic system.	Already agreed.
3.	Query for UV VIS Spectrophotometer Color Software: The diffuse reflectance/specular reflectance measurement are often performed in visible region i.e color zone. It is also advisable to opt for a color measurement software for quantitative measurement of colors using various International system and Index viz: CIE system, GAB system, XYZ,XY, LUV, LAB values, Hunter lab values and all possible and commercial available color measurement index and values to measure absorption, reflectance (diffused and specular),	Description and Specification Annex 1B, Point 1.10 The system should be deployable for analysis of a broad range of solid, filmsand paste through reflectance and transmission spectral measurements.	Not Agreed. Color standard are already deleted.


Anjana
General Manager (Procurement)
Central Medical Services Society
Vishwa Yuvaat Kendra, Chanakyapuri
New Delhi - 110021

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	transmission and Fluorescence measurement on 1931 & 1960 UCS Chromaticity Charts for the evaluation of color difference.		
F	M/s METROHM INDIA PVT LTD		
1.	Query for Ion Chromatograph Flow Rate Requested to amend Flow rate from 0.001ml/min – 20ml/min or more Reason-Upper value of flow rate is missing in the specification. All brands have upper flow rate 20ml/min or more	Description and Specification Annex 1D, Flow Rate: 0.001ml/min or more	Not Agreed.
2.	Query for Ion Chromatography Photo Diode array detector Channels: 8 measuring channels (upto eight wavelength can be selected). Reasons Gives user more flexibility and extends application expandability for the user	Description and Specification Annex 1D Five measuring channel and one reference channel	Not Agreed.
3.	Description and Specification Annex 1D No of Diodes minimum 1024 Reasons:-1024diodes are basic number of diodes. 512 diodes have less sensitivity and old technology not available	Description and Specification Annex 1D No of Diodes minimum 512	No of Diodes minimum 1024
4	Delivery of instruments to site in 60 days from LOA and 30 days for instrument depending of site readiness. Reasons: As the equipment is to be installed at different location and delivery term is DDP please provide sufficient time to complete delivery and installation.	For Delivery- 45 days from LOA and 15 days for installation & commissioning the equipment from the date of handing over the site.	For Delivery- 60 days from LOA and 30 days for installation & commissioning the equipment from the date of handing over the site.
5	As the tender is invited under international competitive bidding it should be allowed to quote directly	17.1 Payment Terms Payment shall be made in Indian Rupees as specified.....	No Change


Anjana
 General Manager (Procurement)
 Central Medical Services Society
 Vishwa Yuvak Kendra, Chanakyaपुरी
 New Delhi - 110021

S. No.	Query Received from Prospective Bidders	Standard Tender Clause	Response
	in internationally accepted foreign currency like EURO, USD, CHF (Swiss franc), JY etc Reasons:- Quoting in foreign currency will enable the principal company to quote directly and both purchase order and payment can be made directly to them saving 18% GST for the buyer (NTTL/CMSS)		
G	M/S Eclectic Scientific Instruments Pvt Ltd		
1.	Query for UV-VIS Spectrophotometer Scanning Speed Rate Please upgrade the scan rate from 2,000 nm/minutes to 10000 nm/ Minutes . The scan rate 2000 nm/ minutes is very slow in order to scan the spectra in whole UV-Visible region and become inclined toward a particular make and defeat the whole purpose of your applications.	Description and Specification Annex 1B, Point 1.9, Scanning Speed Rate: 2000nm/min or better	Agreed. Should be read as: "Scanning Speed rate: 8000 nm/min or better."
2.	Query for UV-VIS Spectrophotometer Spectral Band Width Please modify the band as from "0.2 to 4nm " instead of from "0.5 to 4nm" or higher in steps or continuous mode .	Description and Specification Annex 1B, Page 50 Point 1.6, Spectral Bandwidth: 0.5nm to 4nm variable or higher	Not Agreed.
3.	Query for UV-VIS Spectrophotometer Optics/Light Source Kindly ask for Dual Source; Deuterium & Tungsten, Double Beam, Double Monochromator System. It requires Double Beam, Double Monochromator type Optics for the Stray Light parameter less than 0.005%T @ 220 nm as demanded by you otherwise please dilute it to 0.05%T @220nm.	Description and Specification Annex 1B, Page 50 Point 1.4, Optics/Light Source Dual source- Deuterium/tungsten, Double beam monochromatic system	Agreed.

S. No	Existing Terms & Conditions	Revised Terms & Conditions																								
1	Tender Clause no 4.(b) Average Annual turnover for Tenderers in the last three years i.e. 2019-20, 2020-21 and 2021-2022 shall not be less than the following: <table><tr><th>Schedule</th><th>Amount (in Rs.)</th></tr><tr><td>I</td><td>50 Lakh</td></tr><tr><td>II</td><td>50 Lakh</td></tr><tr><td>III</td><td>70 Lakh</td></tr><tr><td>IV</td><td>80 Lakh</td></tr><tr><td>V</td><td>60 Lakh</td></tr></table> The turnover benchmark will not apply to Micro and Small Enterprises (MSE). Note : MSE Traders will not get the benefits of MSE Firm Note- For Authorized Agencies (non-manufacturers), the bidders can utilize the financial and past supply credentials of the principal	Schedule	Amount (in Rs.)	I	50 Lakh	II	50 Lakh	III	70 Lakh	IV	80 Lakh	V	60 Lakh	Tender Clause no 4.(b) Average Annual turnover for Tenderers in the last three years i.e. 2019-20, 2020-21 and 2021-2022 shall not be less than the following: <table><tr><th>Schedule</th><th>Amount (in Rs.)</th></tr><tr><td>I</td><td>40 Lakh</td></tr><tr><td>II</td><td>46 Lakh</td></tr><tr><td>III</td><td>60 Lakh</td></tr><tr><td>IV</td><td>72 Lakh</td></tr><tr><td>V</td><td>47 Lakh</td></tr></table> The turnover benchmark will not apply to Micro and Small Enterprises (MSE). Note : MSE Traders will not get the benefits of MSE Firm Note- For Authorized Agencies (non-manufacturers), the bidders can utilize the financial and past supply credentials of the principal	Schedule	Amount (in Rs.)	I	40 Lakh	II	46 Lakh	III	60 Lakh	IV	72 Lakh	V	47 Lakh
Schedule	Amount (in Rs.)																									
I	50 Lakh																									
II	50 Lakh																									
III	70 Lakh																									
IV	80 Lakh																									
V	60 Lakh																									
Schedule	Amount (in Rs.)																									
I	40 Lakh																									
II	46 Lakh																									
III	60 Lakh																									
IV	72 Lakh																									
V	47 Lakh																									
2	Tender Clause no 19 WARRANTY - Warranty period- 3 years from the date of installation and commissioning. - ACMC (Annual Comprehensive Maintenance Contract) period- 5 Years from the date of expiry of warranty period.	Tender Clause no 19 WARRANTY - Warranty period- 2 years from the date of installation and commissioning. - ACMC (Annual Comprehensive Maintenance Contract) period- 3 Years from the date of expiry of the warranty period.																								
3	Total Tender Value including 3 yrs warranty + 5 yrs CMC- Rs. 7.59 Cr.	Total Tender Value including 2 yrs warranty + 3yrs CMC- Rs. 6.52 Cr.																								
4	Price Bid BOQ- The bidders are requested to quote the following:- Unit price Ex. Works including 3 years	The bidders are requested to quote the following:- Unit price Ex. Works including 2 years warranty at column no 7 of BOQ,																								


 Anjan
 Central Manager (Procurement)
 Central Medical Services Society
 Yashwantrao Chavan Vardha, Chhatrapati
 Shivaji Maharaj - 411002

	warranty at column no 7 of BOQ, GST (in Rs.) in line with column no 7 at Column no 8, and first 3yrs for CMC. The bidders are requested not to quote ACMC charges of 4 th year after completion of warranty at Column no 13 & 5th Year after completion of warranty at column no 14. The BOQ as available at CPPP cannot be edited at this juncture, therefore, bidders are advised to quote prices as described above.	GST (in Rs.) in line with column no 7 at Column no 8, and first 3yrs for CMC. The bidders are requested not to quote ACMC charges of 4 th year after completion of warranty at Column no 13 & 5th Year after completion of warranty at column no 14. The BOQ as available at CPPP cannot be edited at this juncture, therefore, bidders are advised to quote prices as described above.
5	Revised Annexure III (Details of EMD) attached	

Rest all Terms and Conditions remains same.
Enclosure- Annexure III- Revised Details of EMD


 General Manager (Procurement)
 Anjana
 General Manager (Procurement)
 Central Medical Services Society
 Vishwa Yuvak Kendra, Chanakyapuri
 New Delhi - 110021

Annexure-III

DETAILS OF E.M.D. SUBMITTED

We herewith submit the E.M.D. of Rs. _____ in the form of **Account Payee Demand Draft/FD Receipt/ Banker's Cheque/RTGS/ NEFT and Bank Guarantee** vide document Ref. No. _____ Dated: _____ Bank: _____ in favour of **Central Medical Services Society** for the following items:-

Sch No.	Name of the Product/ Item	UOM	Quantity in Bid	Amount of EMD Payable (in INR) for 100% quantity	Quantity Quoted	Amount of EMD
1	3	4	5	6	8	9
I.	FTIR with ATR Arrangement	Nos.	3	1,92,000		
II.	UV- vis Spectrophotometer	Nos.	3	2,28,000		
III.	Spectrofluorometer	Nos.	3	2,96,000		
IV.	Ion Chromatograph	Nos.	3	3,56,400		
V.	Texture Analyzer	Nos.	3	2,35,200		



अंजना
Anjana
General Manager (Procurement)
Central Medical Services Society
Vishwa Yuvak Kendra, Chanakyaपुरी
New Delhi - 110021