

CERTIFICATE OF ANALYSIS

Catalog Number	P1013-15BF 471310885092	Mfg. Date	2018.06
Lot Number	1806050	Sample Id	N/A
Product Description	15ML Conical Centrifuge Tube, Pre-sterilized (25tubes/pk,20pks/cs)		

Lot Number Designation: 7 Digit Lot Number: First 2 digits -Year of manufacture; Next 2 digits –Month of manufacture, Last 3 Run number

Expiration Statement: This product has an expiration date of 5 years from the date of manufacture for the attributes contained within this certificate.

OD X H (mm): 17 x 120 mm

Maximum RCF: 12000 g

Component Materials:

Flat Cap Polypropylene, meets the respective FDA regulations and 21 CFR 177.1520(b) certified in compliance with US Pharmacopoeia Plastic Class VI.

Sterilized: Gamma- Radiation sterilized

Autoclavable: The product can be autoclavable under 121°C, 15 psi, 20min

Temperature Range: -20°C to 121°C

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Result
Integrity Test	Pass
Visual Attributes	Pass
Tightness Test	Pass
Packaging	Pass
DNase / RNase Test	Pass

Endotoxin contaminated diagnosis:

Check to LAL Regulatory Guideline .Results: less than 0.06 EU/mL

Ivy Wu

ExtraGene Inc.
Quality Assurance Department