



Certified ISO9001 / 13485 / CE

ExtraGene Inc.

USA Add:
1424 Yukon St. Davis
California 95616
USA
Tel:530-753-2363
Fax:530-792-8352

ExtraGene Inc.

Taiwan Add:
60, Renhuagone 16th Rd.
Dali Dist. Taichung City
41278 Taiwan
Te:886-424930999
Fax: 886-424922777

CERTIFICATE OF ANALYSIS

Catalog Number	P1013-15BF 471310885092	Mfg. Date	2018.06
Lot Number	1806050	Sample Id	N/A
Product Description	15ML Conical Centrifuge Tube, Pre-sterilized (25tubes/pk,20pks/cs)		

Lot Number Designation: 7 Digit Lot Number: First 2 digits -Year of manufacture; Next 2 digits –Month of manufacture,
Last 3 Run number

Expiration Statement: This product has an expiration date of 5 years from the date of manufacture for the attributes
contained

within this certificate.

OD X H (mm) 17 x 120 mm

Maximum RCF 12000 g

Component Materials:

Flat Cap Polypropylene, meets the respective FDA regulations and 21 CFR 177.1520(b) certified in compliance with US
Pharmacopoeia Plastic Class VI.

Sterilized: Gamma- Radiation sterilized

Autoclavable: The product can be autoclavable under 121°C, 15 psi, 20minumus

Temperature Range: -20? to 121?

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has
been released by Quality Assurance for the following characteristics:

Test/Procedure:	Result
Integrity Test	Pass
Visual Attributes	Pass
Tightness Test	Pass
Packaging	Pass
DNase / RNase Test	Pass

Endotoxin contaminated diagnosis:

Check to LAL Regulatory Guideline .Results: less than 0.06 EU/mL

Ivy Wu

ExtraGene Inc.
Quality Assurance Department