

AED Inspection Report

Demo AED Org · Inspection Period: 2026-05

Device Information

Manufacturer	Philips
Model	HeartStart FRx
Serial Number	AED-DEMO-001
Installation Location	본관 1층 로비
Device Expiry	2035-05-01
Pad Replacement Due	2028-05-01

Inspection Items

Code	Item	Result	Remark
OP_POWER	Main unit power indicator	OK	
OP_PAD	Patient electrode pads	OK	
OP_BATTERY	Battery charge status	OK	
BX_ALARM	Anti-theft alarm	OK	
BX_GUIDE	Cabinet usage notice	OK	
BX_EMG	Emergency contact list	OK	
BX_CPR	CPR instructions	OK	
BX_EXP	Expiration date label	NG	
LOC_ENT	Entrance signage	OK	
LOC_DIR	Direction/location signage	OK	
DOC_FILE	Management documents	OK	
TIME_24	24-hour accessibility	OK	

Summary

Total Items	12

OK	11
NG	1

Inspector

Name	Inspector One
Email	inspector1+e8both@gmail.com
Phone	-

Signature

[Signature Missing]

Notes

본체 외관 상단 측면 미세 흠집 확인 — 작동에는 영향 없으나 모니터링 필요.

Inspected At: 2026-05-01 11:09