

# AED Inspection Report

Organization : Demo AED Org    Inspection Period : 2026-05

## Device Information

| Manufacturer | Product / Model Name | Serial Number | Manufactured Date |
|--------------|----------------------|---------------|-------------------|
| Philips      | HeartStart FRx       | AED-DEMO-001  | 2025-05-01        |

## Maintenance Period

| Purchase Date | Device Expiry | Pad Replacement Due |
|---------------|---------------|---------------------|
| 2025-05-01    | 2035-05-01    | 2028-05-01          |

|                       |          |
|-----------------------|----------|
| Installation Location | 본관 1층 로비 |
|-----------------------|----------|

| No. | Inspection Item                                         | Result   |        |
|-----|---------------------------------------------------------|----------|--------|
|     |                                                         | Abnormal | Normal |
| 1   | Main Unit Operation                                     |          |        |
|     | ㉠ Main unit operating status (power indicator blinking) |          | ✓      |
|     | ㉡ Patient pads present                                  |          | ✓      |
|     | •Pad Replacement Due : 2028-05-01                       |          |        |
|     | ㉢ Battery charge state                                  |          | ✓      |
|     | •Battery Replacement Due : 2030-05-01                   |          |        |
| 2   | Cabinet Status                                          |          |        |
|     | ㉠ Anti-theft alarm working                              |          | ✓      |
|     | ㉡ Cabinet guidance labels state                         |          | ✓      |
|     | ㉢ Emergency contact list displayed                      |          | ✓      |
|     | ㉣ CPR guidance booklet present                          |          | ✓      |
|     | ㉤ Pad/battery expiry date labeled                       | ✓        |        |
| 3   | AED Location Signage                                    |          |        |
|     | ㉠ Building entrance signage                             |          | ✓      |
|     | ㉡ Installation location and direction signage           |          | ✓      |
| 4   | Management Documentation                                |          |        |
|     | Management documentation prepared and present           |          | ✓      |
| 5   | Manager Change                                          |          |        |
|     |                                                         | Yes      | No     |
|     | ㉠ Manager Name :      Manager Phone (Emergency) :       |          | ✓      |
| 6   | Operating Hours                                         |          |        |
|     | ㉠ Available 24 hours                                    | ✓        |        |
|     | ㉡ Not available 24 hours                                |          |        |

\* Notes : 본체 외관 상단 측면 미세 흠집 확인 — 작동에는 영향 없으나 모니터링 필요.

Inspection Date : 2026-05-01

Inspector : Inspector One (Signature)