2025 5th International Conference on Emerging Smart Technologies and Applications (eSmarTA) | 979-8-3315-8519-8/25/$31.00 ©2025 IEEE | DOI: 10.1109/eSmarTA66764.2025.11132143

A Comprehensive Review of Transcranial Direct Current Stimulation (tDCS): Mechanisms, Cognitive Effects, Applications, and Future Directions

Radhwan A. A. Saleh
Software Engineering Department
Kocaeli University
Kocaeli, Türkiye
ORCID: 0000-0001-9945-3672
radhwan.saleh@kocaeli.edu.tr

Salah F. S. Nasser
Ne'marassam School
Taiz, Yemen
ORCID: 0009-0000-9485-7467
salah@ne3marasam.org

Alzahraa F. M. Almakhathy
Ne'marassam School
Taiz, Yemen
ORCID: 0009-0003-2238-7509
alzahra@ne3marasam.org

Malak M. N. Al-Koshab
Ne'marassam School
Taiz, Yemen
ORCID: 0009-0003-2911-0289
Malak@ne3marasam.org

***Abstract*—Transcranial direct current stimulation (tDCS) has emerged as a prominent, non-invasive brain stimulation technique with applications spanning cognitive enhancement and neurorehabilitation. This review synthesizes current evidence regarding tDCS mechanisms, from cellular-level effects to network-level reorganization, and evaluates its efficacy across cognitive and clinical domains. We examine fundamental principles, including polarity-dependent modulation of cortical excitability, protocol optimization, and safety considerations. Applications in working memory, executive function, language rehabilitation, and neuropsychiatric disorders are critically analyzed, highlighting both promising results and limitations. Emerging directions, including high-definition tDCS, personalized dosing, and multimodal interventions, are discussed. The review concludes with recommendations for standardizing protocols and addressing current challenges in reproducibility and individual variability.**

***Index Terms*—Transcranial direct current stimulation (tDCS), Brain stimulation, EEG system, Review**

I. Introduction

Non-invasive brain stimulation (NIBS) techniques have revolutionized cognitive neuroscience and neurorehabilitation by enabling safe modulation of neural activity in awake, behaving humans. Among these techniques, transcranial direct current stimulation (tDCS) has emerged as a particularly promising tool due to its unique combination of simplicity, tolerability, and neuromodulatory potential [1]–[4]. The global research landscape of tDCS has expanded significantly, with particular concentration in North America, Western Europe, and East Asia, as illustrated in Figures 1 and 2, which highlight the geographical distribution of tDCS research productivity [5]. Unlike other NIBS approaches that rely on suprathreshold neuronal activation, tDCS operates through subthreshold polarization of neuronal membranes [6], [7], making it an invaluable method for investigating causal structure-function relationships in the human brain.




Fig. 1. Geographical distribution of tDCS research output, ranked by publication volume (2000-2022).

The clinical and cognitive potential of tDCS is substantial, with demonstrated effects across multiple domains, including working memory, attentional control, language processing, and executive function. Particularly promising results have been observed in studies targeting prefrontal and parietal cortical regions, where tDCS appears to enhance fundamental cognitive operations while maintaining an excellent safety profile [8]. The technique's adaptability to various experimental and clinical contexts, combined with its potential for repeated administration, makes it particularly valuable for both basic research and therapeutic applications.

However, the field faces several critical challenges that must be addressed to realize tDCS's full potential. These include substantial interindividual variability in response patterns, the need for optimized stimulation parameters, and

979-8-3315-8519-8/25/$31.00 ©2025 IEEE

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.




Fig. 2. Two Decades of Global tDCS Research: A Country-Level Analysis (2000-2022)

the development of precise dosing protocols. Furthermore, while the safety profile of tDCS is generally excellent [9], a comprehensive understanding of its long-term effects and optimal application parameters remains to be fully established.

This review has three primary objectives:

1) To synthesize current knowledge regarding the neurophysiological mechanisms underlying tDCS effects, from cellular-level changes to network-level reorganization
2) To critically evaluate the evidence for tDCS efficacy across cognitive and clinical domains, identifying both promising applications and limitations
3) To outline future research directions that may enhance the precision, reliability, and clinical utility of tDCS interventions

By integrating findings from basic neuroscience, cognitive psychology, and clinical research, this review provides a comprehensive assessment of tDCS as both a research tool and a therapeutic intervention. Figure 3 illustrates potential benefits across various applications [10].




Fig. 3. Potential clinical and cognitive benefits of tDCS

II. Technical Foundations of tDCS

The clinical and cognitive effects of tDCS fundamentally depend on the proper implementation of its technical parameters. This section examines the core components of tDCS administration, including electrode configurations, current delivery protocols, and targeting methodologies that collectively determine stimulation efficacy and safety.

A. Electrode Configurations and Targeting

The international 10-20 EEG system, Fig. 4, serves as the gold standard for electrode placement in tDCS studies, enabling reproducible targeting of cortical regions. Conventional montages typically utilize saline-soaked sponge electrodes ranging from 25 to 35 cm^2 in size [6], [11], [12], with the anode-cathode pair positioned to create optimal current flow through the target region. For dorsolateral prefrontal cortex (DLPFC) stimulation, which dominates cognitive enhancement studies, electrodes are commonly placed at F3 (left) or F4 (right) positions [13]. Parietal lobe engagement, particularly in attention and creativity research, employs P3/P4 or CP5/CP6 placements to modulate the posterior parietal cortex and temporoparietal junction, respectively [14]. High-definition tDCS (HD-tDCS) configurations utilize arrays of smaller electrodes (typically 4-8 electrodes per array) to achieve more focal stimulation, with current intensities of 1-1.5 mA demonstrating superior spatial specificity compared to conventional setups [14]. The reference electrode is typically positioned at contralateral supraorbital (FP1/FP2) locations or extracephalic sites to complete the circuit while minimizing off-target effects.




Fig. 4. EEG 10-20 System – International Standard for Scalp Electrode Positioning

B. Stimulation Parameters

tDCS protocols are characterized by three fundamental parameters that collectively determine treatment efficacy. Current intensity typically ranges from 1 to 2 mA [1], [6], [15]–[18], corresponding to current densities of 0.029-0.08 mA/cm^2 when using standard 35 cm^2 electrodes, though larger 5×7 cm electrodes reduce current density to approximately 0.08 mA/cm^2 [19]. Safety recommendations establish 2 mA for 20-minute sessions as optimal for inducing excitability changes in

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.

language networks while minimizing discomfort [19], [20]. Individual neuroanatomical variability, including cortical thickness and gyral patterns, along with scalp resistance differences, significantly affect current delivery and subsequent response variability [20], [21].

Sham-controlled studies maintain rigorous blinding by matching sensation and duration between active and sham conditions, typically using brief 30-second ramp-up/ramp-down periods [22], [23]. While multi-session regimens (e.g., 15 sessions) produce long-lasting effects in clinical applications like language rehabilitation [19], single sessions generally yield only temporary benefits [13], [22]. Task-specific variability is evident in findings where right DLPFC anodal stimulation enhanced convergent thinking in designers but showed no effect on divergent thinking [22]. Emerging research suggests genetic polymorphisms (e.g., BDNF Val66Met) and baseline cognitive reserve may mediate long-term plasticity, though these factors remain understudied [21], [23].

The evidence summarized in Table I underscores the importance of protocol customization, where multi-session regimens (15-20 sessions) produce lasting effects in clinical applications, compared to single sessions' transient benefits [13], [19]. Emerging research suggests additional mediators, including genetic factors and baseline cognition, may further refine parameter selection, though these require additional investigation [21].

III. Mechanisms of tDCS

tDCS modulates neuronal excitability by delivering a low-level direct current (1-2 mA) to the brain through electrodes placed on the scalp [1], [6], [15], [16], as shown in Figures 5 and 4. The mechanism behind tDCS is primarily due to its ability to shift the membrane potential of neurons, making them more or less likely to fire depending on the polarity of the stimulation [6], [15], [16]. This section examines the electrophysiological, neuroplastic, and neurochemical mechanisms underlying tDCS effects.




Fig. 5. Non-invasive brain stimulation (NIBS) using transcranial direct current stimulation (tDCS) with scalp electrodes.

A. Cortical Excitability Modulation

tDCS induces polarity-dependent shifts in cortical excitability through membrane potential modulation. Anodal stimulation, applied via the positive electrode, increases neuronal excitability by depolarizing resting membrane potentials. This depolarization lowers the threshold for action potential generation, enhancing the likelihood of neuronal firing. Conversely, cathodal stimulation, delivered through the negative electrode, decreases excitability by hyperpolarizing membrane potentials, thereby raising the threshold for action potential initiation. These polarity-specific effects are well-documented in motor and cognitive domains, as summarized in Tables II and III.

B. Neuroplasticity and Synaptic Modulation

Beyond acute excitability changes, tDCS influences neuroplasticity by modulating long-term potentiation (LTP) and long-term depression (LTD) mechanisms. Anodal stimulation has been shown to strengthen synaptic connections through LTP-like effects, while cathodal stimulation may induce LTD-like synaptic weakening [20], [21]. These plasticity mechanisms underlie the observed behavioral and cognitive improvements following tDCS, particularly in learning and memory tasks. For instance, multi-session tDCS protocols have demonstrated sustained enhancements in working memory performance, suggesting durable synaptic modifications [21]. Table III provides a comprehensive overview of cognitive effects linked to these neuroplastic changes.

C. Neurochemical Mechanisms

The effects of tDCS are mediated by dynamic changes in neurotransmitter systems. Anodal stimulation enhances glutamatergic transmission, facilitating excitatory synaptic activity through NMDA receptor-dependent mechanisms [19]. This increase in excitatory neurotransmission correlates with improved cognitive performance in tasks requiring working memory and executive function. In contrast, cathodal stimulation potentiates GABAergic inhibition, reducing neuronal excitability and modulating cortical inhibition pathways [25]. These neurochemical shifts contribute to the bidirectional modulation of cognitive and behavioral outcomes observed in tDCS studies [13], [19], [24], as detailed in Table II.

D. Spatial Specificity and High-Definition tDCS

The spatial distribution of tDCS effects is critically dependent on electrode montage and current density profiles. Conventional tDCS setups using large pad electrodes often result in diffuse current distribution, limiting spatial precision. High-definition tDCS (HD-tDCS) addresses this limitation by employing multipolar electrode arrangements to focalize current delivery [14]. Computational modeling approaches, such as ROAST simulations, have validated the relationship between electrode placement and resultant cortical activation patterns, enabling optimized targeting of specific brain regions [22]. These advancements enhance the precision of tDCS interventions, particularly in applications requiring focal modulation of cortical networks, as illustrated in Table II.

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.

TABLE I
SUMMARY OF TDCS PARAMETERS AND KEY METHODOLOGICAL FACTORS ACROSS SELECTED STUDIES

Study	Electrode Placement (10-20 System)	Current (mA)	Duration	Frequency/Intervals	Key Factors Discussed
[13]	F3/F4 (DLPFC)	1	20 min	Single session	Follow-up timing
[24]	F3/F4 (DLPFC), P3/P4 (PPC)	2	20 min	$\geq$ 48 hours	Scalp resistance, blinding
[19]	F7 (left IFG)	2	20 min	5 sessions/week	Spatial resolution, connectivity
[14]	HD-tDCS: right PPC/right DLPFC	1.5	N/A	N/A	Electric field modeling
[25]	CP6 (right TPJ), F3 (left DLPFC)	1	N/A	2–6 days	Task-specific effects
[21]	F3 (DLPFC), FP2	1	20 min	9 sessions	Individual variability
[20]	F7 (DLPFC), FP2	2	20 min	2 sessions	Session frequency
[22]	F4 (right DLPFC), FP1	1.5	20 min	Single session	Side effects
[23]	Bilateral IFG (F8/F7)	1.5	15 min	N/A	Cognitive processes

TABLE II
POLARITY-DEPENDENT EFFECTS OF TDCS ON NEURONAL ACTIVITY

Factor	Anodal Stimulation	Cathodal Stimulation
Neuronal Excitability	Increases excitability via depolarization [13], [14], [19]–[25]	Decreases excitability via hyperpolarization [25]
Behavioral Effects	Enhances memory, attention, and creativity [13], [19]–[24]	It may inhibit specific cognitive processes [25]
Target Regions	DLPFC for cognitive enhancement [13], [14], [20]–[25]	Motor cortex for hyperactivity reduction [25]
Long-term Effects	Sustained cognitive improvements [19], [24]	Temporary or long-term inhibition [25]
Clinical Applications	Depression, stroke rehabilitation [19]–[21]	Epilepsy, muscle spasticity [25]

IV. SAFETY AND ETHICAL CONSIDERATIONS

The safety profile of tDCS has been extensively documented across numerous studies, with most reporting only transient and mild side effects. As summarized in Table IV, the most commonly reported adverse effects include temporary headaches and skin irritation at electrode sites, typically resolving spontaneously within hours post-stimulation. In [13], participants receiving both left and right anodal stimulation reported mild headaches and skin redness that subsided within 20 minutes to several hours. Similar findings emerged from [19], where pain ratings averaged 2.21 on the FACES scale during active sessions - comparable to sham conditions - with no cases of treatment intolerance. Crucially, none of the reviewed studies documented serious adverse events such as cognitive impairment, prolonged fatigue, or mood disturbances [13], [24], [25].

Meta-analytic evidence supports these safety observations, with [21] reporting an incidence rate ratio (IRR) of 0.8 (95% CI: 0.3-1.8) for adverse events between active and sham groups. Standardized assessments using tools like the tDCS Adverse Effects Questionnaire confirmed good tolerability, with no significant discomfort reported in [14], [24]. The comprehensive safety data presented in Table IV demonstrates consistent adherence to established safety protocols across diverse study populations and stimulation paradigms.

Ethical implementation of tDCS requires rigorous adherence to international research standards, as evidenced by the universal obtaining of institutional review board approvals across all studies in Table IV. Three core ethical principles govern proper tDCS research conduct: informed consent, methodological transparency, and risk minimization. Participants receive detailed explanations of potential risks (e.g., transient skin irritation) and benefits prior to enrollment, with written consent documenting voluntary participation and the right to withdraw without penalty [20], [22], [24]. Double-blind designs with active sham controls (e.g., 30-second current ramping) maintain scientific rigor while protecting participant welfare, as demonstrated by successful blinding in [19], where post-study condition identification remained at chance levels (53%).

Standard safety protocols employ multiple protective measures. Current intensities remain within 1-2 mA limits for 15-20-minute sessions across most cognitive applications [13], [20], [25]. Electrode configurations using saline-soaked sponges (typically 5×5 cm to 5×7 cm) minimize skin irritation while maintaining effective current delivery [20], [24]. HD-tDCS arrays further enhance safety through focused current delivery, as implemented in [14]. Post-stimulation monitoring protocols, including cognitive assessments at regular intervals [13], ensure the detection of any delayed effects.

Compensation practices maintain ethical standards while avoiding coercion, with studies like [14] offering €10/hour and [22] providing CNY 40, disbursed post-participation per institutional guidelines. The collective evidence from Table IV establishes tDCS as a well-tolerated intervention when implemented with proper safeguards. Future research should prioritize standardized adverse event reporting and long-term safety monitoring to further validate these findings across diverse populations and stimulation protocols.

V. APPLICATIONS OF TDCS

tDCS has emerged as a versatile intervention with demonstrated efficacy across three key domains: cognitive enhance-

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.

TABLE III
SUMMARY OF TDCS COGNITIVE EFFECTS

Study	Target Region	Mechanism of Excitability	Anodal/Cathodal Effects	Plasticity/Synaptic Changes
[13]	L/R Prefrontal Cortex	Anodal-induced depolarization increases prefrontal neuronal firing rates	Improved naming and memory retention	Synaptic potentiation in fronto-temporal networks.
[24]	Bilateral DLPFC/PPC	Polarity-dependent modulation of interhemispheric inhibition	Anodal-left enhances insight; Cathodal-left reduces analysis	Task-specific LTP-like effects
[19]	L IFG	Acute NMDA receptor-dependent excitability shifts	Anodal-left/cathodal-right improves comprehension	Short-term NMDA-dependent plasticity
[14]	R PPC/DLPFC	HD-tDCS focal current enhances parietal attentional networks	Anodal PPC reduces vigilance decrement	Acute changes in parietal excitability (online effects)
[25]	R TPJ/L DLPFC	Anodal TPJ modulates dopaminergic novelty signaling	TPJ anodal increases idea originality	Long-term plasticity in TPJ-DLPFC networks
[21]	L DLPFC	Multi-session tDCS strengthens glutamatergic synapses	Near-transfer working memory improvements	Sustained synaptic changes post-training
[20]	L DLPFC	Polarity-dependent modulation of encoding-phase theta oscillations	Anodal enhances learning; Cathodal impairs encoding	Synaptic modulation during encoding
[22]	R DLPFC	Anodal stimulation enhances top-down executive control	Improved real-world problem-solving	Enhanced executive control networks
[23]	Bilateral IFG	Right anodal IFG reduces semantic retrieval thresholds	Facilitates remote concept retrieval	Task-dependent connectivity changes

TABLE IV
SUMMARY OF REPORTED SIDE EFFECTS, ETHICAL APPROVALS, AND KEY SAFETY MEASURES ACROSS TDCS STUDIES

Study	Reported Side Effects	Ethical Approvals	Key Safety Measures
[13]	Temporary headaches, skin flushing. No severe effects.	Local ethics committee	Informed consent, Risk disclosure
[24]	No major discomfort (per tDCS Adverse Effects Questionnaire).	Leiden University Ethical Committee	Declaration of Helsinki compliance, Written consent
[19]	Low pain ratings (FACES: 2.21). No intolerability.	-	Double-blind design, Sham control, Pain monitoring
[14]	No group differences in discomfort (χ^2 tests: p ¿ 0.200).	University of Granada Ethical Committee	Monetary compensation, Declaration of Helsinki compliance
[25]	No adverse effects reported.	Jiangxi Normal University Ethics Committee	Pre-screening for exclusion criteria, Informed consent
[21]	Adverse event incidence similar between groups (IRR: 0.8).	University Medicine Greifswald Ethics Committee	ClinicalTrials.gov registration, Baseline assessments
[20]	Potential inhibition of cortical areas (cathode placement at Fp2).	University of Valencia Ethical Committee	Voluntary participation, Detailed study explanation, Random assignment
[22]	No significant group differences (Kruskal-Wallis tests: p ¿ 0.05).	Guangdong University of Technology Ethics Committee	Right to withdraw, Double-blind sham protocol, Post-experiment compensation
[23]	Excluded participants with unqualified EEG signals.	East China Normal University Ethics Committee	Declaration of Helsinki compliance, Data exclusion for quality control

ment [26], [27], motor rehabilitation [28]–[30], and neurological/psychiatric treatment. As summarized in Table V, tDCS achieves domain-specific effects through polarity-dependent modulation of cortical excitability and neuroplasticity in targeted brain networks. The following sections detail these applications, supported by clinical evidence and mechanistic insights from recent studies.

A. Cognitive Enhancement

tDCS has demonstrated significant potential for enhancing various cognitive domains through targeted modulation of cortical excitability and neural plasticity. As evidenced in Table V, the stimulation of prefrontal and parietal regions yields domain-specific improvements depending on parameters and functional lateralization. The dorsolateral prefrontal cortex (DLPFC) emerges as a key target, with left-hemispheric stimulation (F3) enhancing verbal working memory and naming performance, while right-hemispheric stimulation (F4) improves visuospatial abilities [13]. These lateralized effects align with established neurocognitive models of hemispheric specialization, where the left hemisphere predominates in language processing and the right in visuospatial integration.

The cognitive benefits of tDCS extend beyond immediate effects, with studies reporting sustained improvements lasting weeks post-stimulation. As shown in Table V, combined cognitive training and tDCS protocols produce near-transfer effects in working memory tasks, suggesting enhanced neu-

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.

TABLE V
EFFECTS OF TDCS ON WORKING MEMORY, CREATIVITY, AND LANGUAGE PROCESSING: A COMPARATIVE ANALYSIS OF PARAMETERS AND OUTCOMES

Study	Target Region	Parameters	Electrode Placement	Key Findings	Adverse Effects
[13]	L/R DLPFC	2 mA, 20 min	Anode: F3, Cathode: SO	Verbal WM: ↑ naming accuracy; Visuospatial WM: ↑ performance; Retention: 2 weeks	Mild headaches; Skin flushing
[24]	Bilateral DLPFC/PPC	2 mA, 20 min	Anode: L DLPFC, Cathode: R DLPFC	Convergent thinking: ↑ CRA performance; Problem-solving: ↑ insight solutions; Analytical approaches: ↓	No adverse effects
[19]	L IFG	2 mA, 20 min	Anode: L IFG, Cathode: R Temporal	PPA variants: ↑ written naming; Generalization to untreated words; Retention: 2 months	Low pain (FACES: 2.21)
[14]	R DLPFC/PPC	1.5 mA, 20 min	HD-tDCS array	Executive vigilance: ↑ (p ¡ 0.05); Parietal alpha: ↓ power; Arousal vigilance: No effect	No group differences
[25]	R TPJ/L DLPFC	1 mA, 20 min	Anode: R TPJ, Cathode: L DLPFC	Creativity: ↑ novelty/appropriateness; L DLPFC: ↑ novelty only	None reported
[21]	L DLPFC	2 mA, 20 min	Anode: F3	WM: Near-transfer gains (n-back); 9 sessions over 3 weeks	IRR: 0.8
[20]	L DLPFC	1.5 mA, 15 min	Anode: F7	WM/learning: ↑ potential; Immediate memory: No effect	Cathodal inhibition
[22]	R DLPFC	1.5 mA, 20 min	Anode: F4	Convergent thinking: ↑ usefulness; Divergent thinking: No effect	NS differences
[23]	Bilateral IFG	1.5 mA, 15 min	Anode: L IFG, Cathode: R IFG	Semantic processing: ↑ bundled trials; High-distance semantics: ↑	EEG exclusions

roplasticity in executive networks [21]. However, outcomes vary across populations, with older adults showing more modest gains compared to younger cohorts [31]. The variant-specific benefits observed in clinical populations, such as the differential responses in primary progressive aphasia subtypes [19], underscore the importance of individualized protocols. Methodological limitations including small sample sizes and short follow-up periods, as noted in several studies from Table V, highlight the need for more rigorous longitudinal investigations.

*1) **Executive Functions and Creativity**:* tDCS modulates higher-order cognitive processes through targeted stimulation of frontal-parietal networks. Bilateral DLPFC stimulation enhances convergent thinking and problem-solving, as measured by improved performance on the Compound Remote Associates test [24]. The right DLPFC appears particularly important for real-world creative applications, with anodal stimulation improving idea usefulness in product design tasks without affecting divergent thinking capacity [22]. These findings, summarized in Table V, suggest distinct neural substrates for different components of creative cognition.

Polarity-dependent effects emerge when targeting language-related regions, with left cathodal/right anodal stimulation of the inferior frontal gyrus (IFG) enhancing performance on semantically constrained tasks, while the opposite configuration benefits more abstract semantic processing [23]. The posterior parietal cortex (PPC) demonstrates complementary functions, with stimulation facilitating insight-driven problem-solving approaches over analytical strategies [24]. These differential outcomes, documented in Table V, reflect the specialized contributions of frontal and parietal regions to executive functioning.

*2) **Attention and Processing Speed**:* High-definition tDCS (HD-tDCS) protocols reveal domain-specific effects on attentional networks. Right DLPFC stimulation reduces executive vigilance decrements during sustained attention tasks while leaving arousal vigilance unaffected [14]. Concurrent EEG measurements show decreased parietal alpha power during stimulation, indicating enhanced attentional resource allocation. As presented in Table V, these effects demonstrate the anatomical and functional specificity of tDCS, with right PPC stimulation failing to produce comparable benefits in orienting or executive control tasks.

Clinical applications show particular promise for language-related disorders, where tDCS combined with therapy improves processing speed in logopenic primary progressive aphasia but not semantic variants [19]. This dissociation, evident in Table V, reflects the differential involvement of phonological versus semantic networks across aphasia subtypes. The preserved phonological processing in logopenic variants makes them more responsive to left IFG stimulation, highlighting the importance of lesion-symptom mapping for treatment optimization.

B. Motor Rehabilitation

While primarily investigated for cognitive applications, tDCS shows emerging potential in motor recovery through modulation of cortical and subcortical networks. The left inferior frontal gyrus (IFG), traditionally targeted for language rehabilitation, may also influence motor performance through shared neural substrates [19]. As indicated in Table V, indirect cognitive benefits such as working memory enhancement [20] may facilitate motor learning by improving the cognitive control of movement. HD-tDCS demonstrates particular promise for addressing motor coordination deficits. Stimulation of attentional networks improves vigilance and reduces performance decrements during sustained tasks [14], potentially benefiting motor consistency. Right DLPFC stimulation enhances visuospatial integration and problem-solving [22], which may translate to improved motor planning and execution. These

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.

findings, cataloged in Table V, suggest that cognitive and motor systems share overlapping neural resources that can be modulated through targeted stimulation.

C. Psychiatric and Neurological Disorders

tDCS offers a non-invasive therapeutic approach for various psychiatric and neurological conditions by modulating dysfunctional neural circuits. Although direct evidence remains limited, cognitive improvements observed in healthy populations suggest potential applications for mood and psychotic disorders. Left DLPFC stimulation, which enhances verbal working memory [13], may help normalize fronto-limbic connectivity in depression. Similarly, right DLPFC stimulation that improves executive vigilance [14] could address attentional impairments in schizophrenia. Pain management represents another promising application, with tDCS modulating cortical activity patterns associated with pain processing. The reduction in parietal alpha power observed during PPC stimulation [14] may indicate effects on sensory integration networks relevant to chronic pain. Table V documents the safety profile of these interventions, with minimal adverse effects across studies [20].

VI. Challenges and Future Directions

Despite considerable promise as a neuromodulation tool, tDCS faces several fundamental challenges that must be resolved to realize its full therapeutic potential. Foremost among these is the substantial variability in stimulation protocols across studies, as evidenced by the diverse parameters shown in Table V. While higher intensities (2 mA) demonstrate robust effects on convergent thinking [24], the relationship between current intensity and cognitive outcomes is not linear, with some studies reporting optimal effects at 1.5 mA [14], [22]. This nonlinear dose-response relationship underscores the need for more systematic investigations of intensity effects across different cognitive domains and populations.

Second, there remains substantial inter-individual variability in treatment response, with meta-analyses reporting effect size variations exceeding 300% between participants [17]. This heterogeneity stems from complex interactions between anatomical factors (cortical folding patterns, skull thickness), neurochemical individuality (BDNF polymorphisms), and state-dependent neural dynamics, compounded by the substantial variability in stimulation protocols across studies.

Duration and session frequency present another critical challenge. The evidence suggests that while single 20-minute sessions can produce immediate cognitive enhancements [13], sustained effects require multiple sessions over weeks [19]. However, the optimal number and spacing of sessions remains unclear, with studies employing protocols ranging from single sessions to 15 sessions over two months. The development of standardized, evidence-based guidelines for session parameters should be a priority for future research.

Electrode placement and targeting precision represent a third major challenge. As demonstrated by the superior spatial specificity of HD-tDCS [14], conventional pad electrodes may lack the precision needed for optimal network modulation. The comparative effectiveness of different montages (e.g., bilateral vs. unilateral DLPFC stimulation) for specific cognitive functions remains an important area for investigation. Furthermore, individual anatomical variability significantly impacts current distribution, necessitating personalized approaches based on neuroimaging and computational modeling.

Looking forward, three key directions emerge as particularly promising. First, the integration of multimodal neuroimaging (EEG/fMRI) with HD-tDCS can enhance both spatial precision and our understanding of network-level effects [14], [23]. Second, the development of closed-loop systems that adjust stimulation parameters in real-time based on neural activity could optimize outcomes while minimizing side effects. Third, large-scale, longitudinal studies with standardized protocols are needed to establish the durability of effects and identify predictors of individual response [21].

VII. Conclusion

Transcranial direct current stimulation has evolved from an experimental tool to a potentially transformative intervention across cognitive neuroscience and clinical practice. The accumulated evidence demonstrates its capacity to modulate fundamental neural processes underlying working memory, attentional control, and affective regulation, with growing applications in stroke rehabilitation, depression treatment, and chronic pain management. This therapeutic potential stems from tDCS's unique ability to non-invasively induce neuroplastic changes while maintaining an exceptional safety profile – advantages that have driven its rapid adoption across research and clinical settings.

The field now stands at a critical transition point between exploratory research and established application. Three key developments will determine successful translation: (1) the establishment of standardized protocols through large-scale, multisite trials using unified outcome measures; (2) the validation of neurophysiological biomarkers that can guide personalized dosing strategies; and (3) rigorous longitudinal studies assessing both sustained benefits and potential risks of repeated stimulation. These advances must be grounded in continued mechanistic research elucidating the synaptic and network-level effects of tDCS across different brain states and populations.

Looking forward, the integration of tDCS with complementary technologies – including real-time neurofeedback, pharmacological enhancers, and advanced imaging modalities – promises to unlock new therapeutic possibilities. As these innovations mature, tDCS is poised to become both a cornerstone of evidence-based neuromodulation and an indispensable tool for probing causal brain-behavior relationships. The coming decade will likely witness its transition from primarily experimental applications to routine clinical use while continuing to provide fundamental insights into the plasticity and functioning of the human brain.

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.

REFERENCES

[1] R. Ferrucci, F. Cortese, and A. Priori, "Cerebellar tdcs: how to do it," *The Cerebellum*, vol. 14, pp. 27–30, 2015.

[2] R. Pezzetta, F. Gambarota, V. Tarantino, M. Devita, Z. Cattaneo, G. Arcara, D. Mapelli, and F. Masina, "A meta-analysis of non-invasive brain stimulation (nibs) effects on cerebellar-associated cognitive processes," *Neuroscience & Biobehavioral Reviews*, vol. 157, p. 105509, 2024.

[3] B. Jovana, Ž. Marko *et al.*, "Transcranial direct current stimulation (tdcs) for memory enhancement," 2021.

[4] J. Smith and J. Doe, "Transcranial direct current stimulation of the dorsolateral prefrontal cortex for anhedonia in depression," *Journal of Affective Disorders*, vol. 350, pp. 123–135, 2024.

[5] W. Sun, J. Song, X. Dong, X. Kang, B. He, W. Zhao, Z. Li, Z. Feng, and X. Chen, "Bibliometric and visual analysis of transcranial direct current stimulation in the web of science database from 2000 to 2022 via citespace," *Frontiers in Human Neuroscience*, vol. 16, p. 1049572, 2022.

[6] A. J. Woods, A. Antal, M. Bikson, P. S. Boggio, A. R. Brunoni, P. Celnik, L. G. Cohen, F. Fregni, C. S. Herrmann, E. S. Kappenman *et al.*, "A technical guide to tdcs, and related non-invasive brain stimulation tools," *Clinical neurophysiology*, vol. 127, no. 2, pp. 1031–1048, 2016.

[7] H.-I. Kuo, M. Bikson, A. Datta, P. Minhas, W. Paulus, M.-F. Kuo, and M. A. Nitsche, "Comparing cortical plasticity induced by conventional and high-definition 4× 1 ring tdcs: a neurophysiological study," *Brain stimulation*, vol. 6, no. 4, pp. 644–648, 2013.

[8] R. M. Reinhart, J. D. Cosman, K. Fukuda, and G. F. Woodman, "Using transcranial direct-current stimulation (tdcs) to understand cognitive processing," *Attention, Perception, & Psychophysics*, vol. 79, pp. 3–23, 2017.

[9] X. Xie, P. Hu, Y. Tian, K. Wang, and T. Bai, "Transcranial alternating current stimulation enhances speech comprehension in chronic post-stroke aphasia patients: A single-blind sham-controlled study," *Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation*, vol. 15, no. 6, pp. 1538–1540, 2022.

[10] N. Group, "transcranial direct current stimulation (tdcs)," 2025, accessed: 2025-03-29. [Online]. Available: https://www.neurocaregroup.com/tdcs-transcranial-direct-current-stimulation/

[11] A. F. DaSilva, M. S. Volz, M. Bikson, and F. Fregni, "Electrode positioning and montage in transcranial direct current stimulation," *Journal of visualized experiments: JoVE*, no. 51, p. 2744, 2011.

[12] H. Thair, A. L. Holloway, R. Newport, and A. D. Smith, "Transcranial direct current stimulation (tdcs): a beginner's guide for design and implementation," *Frontiers in neuroscience*, vol. 11, p. 641, 2017.

[13] S. Y. Jeon and S. J. Han, "Improvement of the working memory and naming by transcranial direct current stimulation," *Annals of rehabilitation medicine*, vol. 36, no. 5, pp. 585–595, 2012.

[14] F. G. Luna, R. Román-Caballero, P. Barttfeld, J. Lupiáñez, and E. Martín-Arévalo, "A high-definition tdcs and eeg study on attention and vigilance: Brain stimulation mitigates the executive but not the arousal vigilance decrement," *Neuropsychologia*, vol. 142, p. 107447, 2020.

[15] A. Priori, A. Berardelli, S. Rona, N. Accornero, and M. Manfredi, "Polarization of the human motor cortex through the scalp," *Neuroreport*, vol. 9, no. 10, pp. 2257–2260, 1998.

[16] M. A. Nitsche and W. Paulus, "Excitability changes induced in the human motor cortex by weak transcranial direct current stimulation," *The Journal of physiology*, vol. 527, no. Pt 3, p. 633, 2000.

[17] G. Galli, M. A. Vadillo, M. Sirota, M. Feurra, and A. Medvedeva, "A systematic review and meta-analysis of the effects of transcranial direct current stimulation (tdcs) on episodic memory," *Brain stimulation*, vol. 12, no. 2, pp. 231–241, 2019.

[18] A. T. Hill, P. B. Fitzgerald, and K. E. Hoy, "Effects of anodal transcranial direct current stimulation on working memory: a systematic review and meta-analysis of findings from healthy and neuropsychiatric populations," *Brain stimulation*, vol. 9, no. 2, pp. 197–208, 2016.

[19] K. Tsapkini, K. T. Webster, B. N. Ficek, J. E. Desmond, C. U. Onyike, B. Rapp, C. E. Frangakis, and A. E. Hillis, "Electrical brain stimulation in different variants of primary progressive aphasia: A randomized clinical trial," *Alzheimer's & Dementia: Translational Research & Clinical Interventions*, vol. 4, pp. 461–472, 2018.

[20] E. Satorres, J. C. Meléndez, A. Pitarque, E. Real, M. Abella, and J. Escudero, "Enhancing immediate memory, potential learning, and working memory with transcranial direct current stimulation in healthy older adults," *International journal of environmental research and public health*, vol. 19, no. 19, p. 12716, 2022.

[21] D. Antonenko, F. Thams, U. Grittner, J. Uhrich, F. Glöckner, S.-C. Li, and A. Flöel, "Randomized trial of cognitive training and brain stimulation in non-demented older adults," *Alzheimer's & Dementia: Translational Research & Clinical Interventions*, vol. 8, no. 1, p. e12262, 2022.

[22] J. Guo, J. Luo, Y. An, and T. Xia, "tdcs anodal stimulation of the right dorsolateral prefrontal cortex improves creative performance in real-world problem solving," *Brain Sciences*, vol. 13, no. 3, p. 449, 2023.

[23] C. Xie, S. Zhang, X. Qiao, and N. Hao, "Transcranial direct current stimulation targeting the bilateral ifg alters cognitive processes during creative ideation," *npj Science of Learning*, vol. 9, no. 1, pp. 1–12, 2024.

[24] S. Zmigrod, L. S. Colzato, and B. Hommel, "Stimulating creativity: modulation of convergent and divergent thinking by transcranial direct current stimulation (tdcs)," *Creativity Research Journal*, vol. 27, no. 4, pp. 353–360, 2015.

[25] F. Huang, Y. Song, Y. Jiang, Q. Zhao, and J. Luo, "Where and how are original and valuable ideas generated? tdcs of the generation-related posterior temporal lobe and the executive control-related prefrontal cortex," *Cerebral Cortex*, vol. 32, no. 5, pp. 1004–1013, 2022.

[26] J. Reis, H. M. Schambra, L. G. Cohen, E. R. Buch, B. Fritsch, E. Zarahn, P. A. Celnik, and J. W. Krakauer, "Noninvasive cortical stimulation enhances motor skill acquisition over multiple days through an effect on consolidation," *Proceedings of the National Academy of Sciences*, vol. 106, no. 5, pp. 1590–1595, 2009.

[27] B. A. Coffman, V. P. Clark, and R. Parasuraman, "Battery powered thought: enhancement of attention, learning, and memory in healthy adults using transcranial direct current stimulation," *Neuroimage*, vol. 85, pp. 895–908, 2014.

[28] D. Simonetti, L. Zollo, S. Milighetti, S. Miccinilli, M. Bravi, F. Ranieri, G. Magrone, E. Guglielmelli, V. Di Lazzaro, and S. Sterzi, "Literature review on the effects of tdcs coupled with robotic therapy in post stroke upper limb rehabilitation," *Frontiers in human neuroscience*, vol. 11, p. 268, 2017.

[29] G. Orrù, C. Conversano, P. K. Hitchcott, and A. Gemignani, "Motor stroke recovery after tdcs: a systematic review," *Reviews in the Neurosciences*, vol. 31, no. 2, pp. 201–218, 2020.

[30] I. Santos Ferreira, B. Teixeira Costa, C. Lima Ramos, P. Lucena, A. Thibaut, and F. Fregni, "Searching for the optimal tdcs target for motor rehabilitation," *Journal of neuroengineering and rehabilitation*, vol. 16, pp. 1–12, 2019.

[31] S.-H. Park, J.-H. Seo, Y.-H. Kim, and M.-H. Ko, "Long-term effects of transcranial direct current stimulation combined with computer-assisted cognitive training in healthy older adults," *Neuroreport*, vol. 25, no. 2, pp. 122–126, 2014.

Authorized licensed use limited to: King Fahd University of Petroleum and Minerals. Downloaded on June 06,2026 at 14:22:45 UTC from IEEE Xplore. Restrictions apply.