

PulseSync Medical - Cardiac Device Services

Patient: Mark Smith	Patient ID: PSM491002	Date: 18-07-2025
Document: Interrogation Report (Post-Implant Follow-up)		

Device & Lead Overview

Manufacturer / Model	PulseSync DuoRate DR-220
Implant Date	21-06-2025
Mode	DDD 60–130 bpm
Battery Status	Good; Est. longevity 9.8 years
Atrial Lead	P-wave 3.4 mV; Threshold 0.5 V @ 0.4 ms; Impedance 535 Ω
Ventricular Lead	R-wave 10.5 mV; Threshold 0.7 V @ 0.4 ms; Impedance 600 Ω

Event Summary

Atrial High Rate Episodes	0
Ventricular High Rate Episodes	0
Mode Switches	0
Pacing Percent (A / V)	A: 20% V: 92%

Programming Changes

No programming changes. Outputs and sensing within optimal range.

Rhythm Observations

Appropriate atrial sensing and ventricular capture. Underlying rhythm remains high-grade AV block; device providing reliable AV synchrony.

Plan

Continue remote monitoring. Routine follow-up in 3 months.