

GOVERNMENT OF INDIA

LEGAL METROLOGY DEPARTMENT

NAWI Compliance Verification Report

Report ID: 9020b854-e2cf-45f5-8360-d36fc94904d7
Date: 2026-09-27T13:55:30.556904

Instrument Information:

Manufacturer	Essae Teraoka Pvt. Ltd.	Model	DS-852
Serial Number	ET-DS852-2026-0471	Class	III
Max Capacity	15.0 kg	Min Capacity	0.04 kg
Verification Scale Int (e)	0.005 kg		

Test Results Summary:

Overall Status	FAIL
Compliance Score	15.36%
Risk Level	CRITICAL
Combined Uncertainty	0.000550

Detailed Load Point Errors:

Load (kg)	Reading (kg)	Direction	Error (kg)	MPE (kg)	Result
0.000	0.000	increasing	0.0000	0.0025	PASS
1.000	1.002	increasing	0.0020	0.0025	PASS
3.000	3.003	increasing	0.0030	0.0050	PASS
5.000	5.004	increasing	0.0040	0.0050	PASS
7.500	7.503	increasing	0.0030	0.0050	PASS
10.000	10.005	increasing	0.0050	0.0050	FAIL
12.000	12.004	increasing	0.0040	0.0075	PASS
15.000	15.006	increasing	0.0060	0.0075	PASS
0.000	0.000	increasing	0.0000	0.0025	PASS
1.000	1.001	increasing	0.0010	0.0025	PASS
5.000	5.003	increasing	0.0030	0.0050	PASS
10.000	10.004	increasing	0.0040	0.0050	PASS
15.000	15.007	increasing	0.0070	0.0075	PASS
0.000	0.000	increasing	0.0000	0.0025	PASS

5.000	5.005	increasing	0.0050	0.0050	PASS
10.000	10.006	increasing	0.0060	0.0050	FAIL
15.000	15.005	increasing	0.0050	0.0075	PASS
15.000	15.007	decreasing	0.0070	0.0075	PASS
10.000	10.006	decreasing	0.0060	0.0050	FAIL
5.000	5.005	decreasing	0.0050	0.0050	PASS
1.000	1.002	decreasing	0.0020	0.0025	PASS
0.000	0.001	decreasing	0.0010	0.0025	PASS
5.000	5.004	increasing	0.0040	0.0050	PASS
5.000	5.003	increasing	0.0030	0.0050	PASS
5.000	5.006	increasing	0.0060	0.0050	FAIL
5.000	5.002	increasing	0.0020	0.0050	PASS



Error Analysis Breakdown:

Metric	Value	Description
Repeatability	0.000090	Max stdev of repeated readings / capacity
Linearity	0.000467	Max deviation from ideal / capacity
Hysteresis	0.000076	Max inc/dec difference / capacity
Eccentricity	0.000267	Max positional difference / capacity
Combined	0.000550	Root sum of squares

Legal Framework and Citations:

Reference	Applicable Section
OIML R 76-1 (2006)	Sections 3, 5, 6 - Metrological requirements and MPE
The Legal Metrology Act, 2009	Section 12 - Approval of models
LM (Approval of Models) Rules, 2011	Rule 7 - Tests for approval
LM (General) Rules, 2011	Rule 2 - Classification
IS 9281:2002 / ISO 76:2017	Indian Standard for electronic weighing instruments

Validity Period: 12 months | **Next Calibration Due:** 2027-09-27

Inspector: Rajesh Kumar Sharma (ID: INS-DL-2024-0087)

SHA-256 Hash: f42680b0e0f452db21ec3ac2eeaf37b2baeebf0d03b8ed54c582285d69ac161c

Prepared in accordance with Section 12 of the Legal Metrology Act, 2009.