A company in the service of biodiagnostics since 1976

ARK DiagnoSys



Biochemistry, Turbilatex & Immuno Turbimetry Reagents



QCA QUÍMICA CLÍNICA APLICADA SPAIN

Química Clínica Aplicada, Spain
ISO 9001 - ISO 13485 Certified Company

PRODUCT	PACK SIZE		METHOD	WORKING REAGENT STABILITY	MODE	WAVELENGTH	REAG	SAMPLE	LINEARITY
ACP (Acid phosphatase)	18 x 2 ml 10 x 10 ml		Hillmann	10 days at 2-8°C	Kinetic	405 nm	1 ml	100 ul	75 U/L
ADA (Adenosine deaminase)	1 x 50 ml	R1 : 40 ml + R2 : 10 ml	GLDH UV	2 weeks at 2-8°C	Kinetic	340 nm	1 ml	50 ul	200 IU/L
Alkaline Phosphatase Liq	1 x 50 ml 1 x 125 ml	R1 : 40 ml + R2 : 10 ml R1 : 100 ml + R2 : 25 ml	IFCC	5 weeks at 2-8°C	Kinetic	405 nm	1 ml	20 ul	1200 U/L
Alk.Phos pNPP	1 x 125 ml	R1 : 100 ml + R2 : 25 ml	DGKC	4 weeks at 2-8°C	Kinetic	405 nm	1 ml	20 ul	1000 U/L
Amylase	1 x 60 ml 1 x 120 ml	R1 : 50 ml + R2 : 10 ml R1 : 100 ml + R2 : 20 ml	IFCC	1 weeks at 2-8°C	Kinetic	405 nm	1 ml	33 ul	1500 U/L
Cholinesterase	1 x 50 ml 5 x 25 ml	R1 : 50 ml + R2 : 2 x 1 ml R1 :5 x 25 ml + R2 : 5 x 1 ml	Butyryl sub.	till expiry	Kinetic	405 nm	1.55 ml	10 ul	22500 U/L
CK-MB	20 x 2.5 ml	R1 : 20 x 2.5 ml R2 : 1 x 50 ml Control : 1 x 2 ml (Lyo.)	Immunological	12 days at 2-8°C	Kinetic	340 nm	1 ml	40ul	600 U/L
CK-MB Liq	1 x 50 ml	R1 : 40 ml + R2 : 10 ml Control : 1 x 2 ml (Lyo.)	Immunological	3 weeks at 2-8°C	Kinetic	340 nm	1 ml	40ul	945 U/L
CK Nac Liq	1 x 50 ml 1 x 125 ml	R1 : 40 ml + R2 : 10 ml R1 : 100 ml + R2 : 25 ml	IFCC	2 weeks at 2-8°C	Kinetic	340 nm	1 ml	40ul	1500 U/L
GGT Liq	1 x 50 ml 1 x 125 ml	R1 : 40 ml + R2 : 10 ml R1 : 100 ml + R2 : 25 ml	SZASZ Modf.	5 weeks at 2-8°C	Kinetic	405 nm	1 ml	100ul	600 U/L
SGOT UV Lyophilised	12 x 16 ml	R1 : 12 vials substrate R2 : 2 x 100 ml buffer	IFCC	4 weeks at 2-8°C	Kinetic	340 nm	1 ml	100ul	430 U/L
SGOT UV Liq	1 x 50 ml 1 x 250 ml 1 x 940 ml	R1 : 40 ml + R2 : 10 ml R1 : 2 x 100 ml + R2 : 1 x 50 ml R1 : 3 x 250 ml + R2 : 1 x 190 ml	IFCC	4 weeks at 2-8°C	Kinetic	340 nm	1 ml	100ul	680 U/L
SGPT UV Lyophilised	12 x 16 ml	R1 : 12 vials substrate R2 : 2 x 100 ml buffer	IFCC	4 weeks at 2-8°C	Kinetic	340 nm	1 ml	100ul	350 U/L
GPT UV Liq	1 x 50 ml 1 x 250 ml 1 x 940 ml	R1 : 40 ml + R2 : 10 ml R1 : 2 x 100 ml + R2 : 1 x 50 ml R1 : 3 x 250 ml + R2 : 1 x 190 ml	IFCC	4 weeks at 2-8°C	Kinetic	340 nm	1 ml	100ul	550 U/L
LDH Liq	1 x 50 ml 1 x 125 ml	R1 : 40 ml + R2 : 10 ml R1 : 100 ml + R2 : 25 ml	SFBC	9 weeks at 2-8°C	Kinetic	340 nm	1 ml	20ul	1800 U/L
Lipase	1 x 80 ml	R1 : 50 ml + R2 : 30 ml con.: 1 x 1ml	Colorimetric	15 days at 2-8°C	Kinetic	578 nm	1.6 ML	10ul	300 U/L
Albumin	3 x 100 ml 2 x 250 ml	3 x 100 ml 2 x 250 ml	BCG	till expiry	End point	630 nm	2.5 ML	10ul	6 g/dL
Bilirubin (T+D)	1 x 245 ml	R1 : 1 x 40 ml R2 : 1 x 100 ml R3 : 1 x 100 ml R4 : 1 x 5 ml	Jendrassik Grof	till expiry	End point	546 nm	2.2 ml	200 ul	25 mg/dL
Bilirubin Direct	1 x 250 ml	R1 : 1 x 50 ml R2 : 2 x 100 ml R3 : 1 x 2ml	Jendrassik Grof	Diazo reagent 2 days at RT	End point	546 nm	1.1 ml	100ul	15 mg/dL
Bilirubin Total	1 x 250 ml	R1 : 1 x 50 mlR2 : 2 x 100 ml R3 : 1 x 2ml	Jendrassik Grof	Diazo reagent 10 days at RT	End point	546 nm	1.1 ml	100ul	25 mg/dL
Bilirubin Direct DPD	1 x 300 ml	R1 : 1 x 240 ml + R2 : 1 x 60 ml	DPD	till expiry	End point	546 nm	1 ml	80 ul	10 mg/dL
Bilirubin Total DPD	1 x 300 ml	R1 : 1 x 240 ml + R2 : 1 x 60 ml	DPD	till expiry	End point	546 nm	1 ml	80 ul	30 mg/dL
Cholesterol Liq	1 x 100 ml 3 x 100 ml 2 x 250 ml	1 x 100 ml 3 x 100 ml 2 x 250 ml	CHOD-POD	till expiry	End point	505 nm	1 ml	10 ul	700 mg/dL
Creatinine Jaffe	2 x 100 ml 4 x 250 ml	R1 : 1 x 100 ml + R2 : 1 x 100 ml R1 : 2 x 250 ml + R2 : 2 x 250 ml	Modf. JAFFE	15 days at RT	Fixed time kinetic	510 nm	1 ml	100 ul	15 mg/dL
Creatinine DMSO	2 x 100 ml	R1 : 1 x 100 ml + R2 : 1 x 100 ml	Modf. JAFFE	21 days at RT	Fixed time kinetic	510 nm	1 ml	100 ul	15 mg/dL
HDL Chol Direct	1 x 40 ml 1 x 80 ml 1 x 400 ml	R1 : 1 x 30 ml + R2 : 1 x 10 ml R1 : 1 x 60 ml + R2 : 1 x 20 ml R1 : 3 x 100 ml + R2 : 1 x 100 ml	Colorimetric	till expiry	Fixed time kinetic	600 nm	400 ul	4 ul	200 mg/dL
LDL-Cholesterol (ppt)	1 x 10 ml	1 x 10 ml	Polyvinyl sulphate	till expiry		505 nm	1 ml	10 ul	-
LDL-Chol Direct	1 x 40 ml 1 x 80 ml 1 x 400 ml	R1 : 1 x 30 ml + R2 : 1 x 10 ml R1 : 1 x 60 ml + R2 : 1 x 20 ml R1 : 3 x 100 ml + R2 : 1 x 100 ml	Colorimetric	till expiry	Fixed time kinetic	600 nm	400 UL	5 UL	500 mg/dL
Glucose Liq	4 x 250 ml	4 x 250 ml	GOD POD	till expiry	End point	505 nm	1 ml	10 ul	500 mg/dL
Total Protein	3 x 100 ml	3 x 100 ml	Biuret	till expiry	End point	540 nm	1 ml	20 ul	12 g/dL

PRODUCT	PACK SIZE		METHOD	WORKING REAGENT STABILITY	MODE	WAVELENGTH	REAG	SAMPLE	LINEARITY
Triglycerides Liq	1 x 100 ml 3 x 100 ml 2 x 250 ml	1 x 100 ml 3 x 100 ml 2 x 250 ml	GPO	till expiry	End point	505 nm	1 ml	10 ul	1000 mg/dL
Urea Enzymatic	2 x 100 ml 5 x 100 ml	R1 : 2 x 50 ml + R2 : 1 x 3 ml R1 : 5 x 100 ml + R2 : 1 x 15 ml	Modif. Berthelot	3 weeks at 2-8°C	End point	578 nm	2 ml	10 ul	400 mg/dL
Urea UV Lyophilised	12 x 16 ml	R1 : 12 vials of substrate R2 : 2 x 100 ml buffer	Urease-GLDH	6 weeks at 2-8°C	Fixed time kinetic	340 nm	1 ml	10 ul	300 mg/dL
Urea UV Liq	1 x 80 ml 4 x 50 ml 4 x 100 ml	R1 : 1 x 60 ml + R2 : 1 x 20 ml R1 : 3 x 50 ml + R2 : 1 x 50 ml R1 : 3 x 100 ml + R2 : 1 x 100 ml	Urease-GLDH	5 weeks at 2-8°C	Fixed time kinetic	340 nm	1 ml	10 ul	300 mg/dL
Uric Acid Liq	1 x 100 ml 3 x 100 ml 2 x 250 ml	1 x 100 ml 3 x 100 ml 2 x 250 ml	Urease-GLDH	till expiry	End point	505 nm	1 ml	20 ul	25 mg/dL
Urinary CSF Protein	2 x 100 ml	2 x 100 ml	Pyrogallol red	-	End point	600 nm	1 ml	10 ul	200 mg/dl
Calcium OCPC	2 x 100 ml	R1 : 1 x 100 ml + R2 : 1 x 100 ml	O-Cresolphthalein	3 days at RT	End point	565 nm	1 ml	10 ul	20 mg/dL
Calcium Arsenazo III	2 x 100 ml	2 x 100 ml	Colorimetric	till expiry	End point	650 nm	1 ml	10 ul	15 mg/dL
Chlorides	2 x 100 ml	R1 : 2 x 100 ml + R2 : 1 x 5 ml	Thiocyanate	till expiry	End point	450 nm	1 ml	10 ul	140 mEq/L
Copper	2 x 50 ml	2 x 50 ml	Colorimetric	till expiry	End point	582 nm	1 ml	50 ul	500 ug/dL
Ing. Phosphorus UV	3 x 100 ml	3 x 100 ml	Fiske – Subbarow	till expiry	End point	340 nm	1 ml	10 ul	15 mg/dL
Iron - CAB	2 x 100 ml	2 x 100 ml	Chromazurol B	6 weeks at RT	End point	620 nm	1 ml	50 ul	500 ug/dL
Iron Ferrozine	2 x 100 ml	R1 : 2 x 100 ml R2 : 1 x 15 ml	Colorimetric	3 months at RT	End point	578 nm	1 ml	200 ul	1000 ug/dL
TIBC	100 tests	R1 :2 x 100 ml + R2 : 1 x 20 gm	FeCl3 + MgCO3	till expiry	-	-	2 ml	1 ml	-
Magnesium	2 x 100 ml	2 x 100 ml	Xylidyl Blue	till expiry	End point	520 nm	1 ml	10 ul	5 mg/dL
Zinc	2 x 50 ml	2 x 50 ml	Colorimetric	till expiry	End point	560 nm	1 ml	50 ul	400 ug/dL
ASO Turbidimetric	50 ml 100 ml 300ml	R1 :1 x 10 ml + R2 : 1 x 40 ml R1 :1 x 20 ml + R2 : 1 x 80 ml R1 :1 x 60 ml + R2 : 3 x 80 ml	Turbilatex	7 days at 2-8°C	Fixed time kinetic	540 nm	1 ml	10 ul	800 UI/mL
CRP turbidimetric	50 ml 100 ml 300ml	R1 :1 x 10 ml + R2 : 1 x 40 ml R1 :1 x 20 ml + R2 : 1 x 80 ml R1 :1 x 60 ml + R2 : 3 x 80 ml	Turbilatex	4 days at 2-8°C	Fixed time kinetic	540 nm	1 ml	8 ul	150 mg/L
RF turbidimetric	50 ml 100 ml 300ml	R1 :1 x 10 ml + R2 : 1 x 40 ml R1 :1 x 20 ml + R2 : 1 x 80 ml R1 :1 x 60 ml + R2 : 3 x 80 ml	Turbilatex	10 days at 2-8°C	Fixed time kinetic	630 nm	1 ml	33 ul	100UI/mL
Ferritin	30 ml 90 ml	R1 :1 x 20 ml + R2 : 1 x 10 ml R1 :1 x 60 ml + R2 : 1 x 30 ml	Turbilatex	till expiry	Fixed time kinetic	630/700 nm	900 ul	100 ul	300 ng/mL
Microalbumin	24 + 4 ml	R1 :1 x 24 ml R2 : 1 x 4 ml	Turbilatex	till expiry	Fixed time kinetic	340 nm	350 ul	10 ul	50 mg/L
Automatic Glycohemoglobin HbA1c	1 x 40 ml	R1 : 1 x 30 ml R2 : 1 x 10 ml R3 : 2 x 70 ml	Immuno-turbidimetric	till expiry	End point	660 nm	240 ul	5 ul	16%
Calibrator for Fully Automated	1 x 7 ml								
Seriscann Abnormal Control	10 x 5 ml								
Sersicann Normal Control	10 x 5 ml								

***NOTE : Wherever required Standards / Calibrators are provided in the kits**
Immuno Turbimetry Reagents also available

BIOCHEMISTRY

ENZYMES

Product
ACID PHOSPHATASE
ADENOSINE DEAMINASE (ADA)
ADENOSINE DEAMINASE, CONTROL KIT
ALKALINE PHOSPHATASE
ALKALINE PHOSPHATASE p-NPP
ALKALINE PHOSPHATASE LIQUID
AMYLASE BPS
AMYLASE LIQUID
CHOLINESTERASE
CK-NAC ACTIVATED
CK-NAC LIQUID
CK-MB
CK-MB LIQUID
gamma-GT
gamma-GT LIQUID
GLDH ACTIVATED
GOT/AST UV
GOT/AST UV LIQUID
GOT/AST COLORIMETRIC
GPT/ALT UV
GPT/ALT UV LIQUID
GPT/ALT COLORIMETRIC
LDH
LDH LIQUID
LIPASE

SUBSTRATES

Product
ALBUMIN
BILIRUBIN
BILIRUBIN TOTAL
BILIRUBIN DIRECT
CHOLESTEROL LIQUID
HDL-CHOLESTEROL
HDL-CHOLESTEROL DIRECT
LDL-CHOLESTEROL
LDL-CHOLESTEROL DIRECT
CREATININE
CREATININE DMSO
GLUCOSE LIQUID
GLYCOHEMOGLOBIN (HbA1c) AUTOMATIC
GLYCOHEMOGLOBIN (HbA1c) AUTOMATIC, CALIBRATOR SET
GLYCOHEMOGLOBIN (HbA1c) AUTOMATIC, CONTROL SET
GLYCOHEMOGLOBIN (HbA1)
GLYCOHEMOGLOBIN HbA1, CONTROL SET
HEMOGLOBIN
TOTAL PROTEIN
TRIGLYCERIDES LIQUID
UREA ENZYMATIC
UREA UV
UREA UV LIQUID
URIC ACID LIQUID

Product
URINE / C.S.F. PROTEINS
URINE / C.S.F. PROTEINS

ELECTROLYTES

Product
AMMONIA
AMMONIA CONTROLS
CALCIUM-ARSENAZO III
CALCIUM
CHLORIDE
COPPER
INORGANIC PHOSPHORUS VISIBLE
INORGANIC PHOSPHORUS UV
IRON - FERROZINE
IRON - CAB
TIBC
MAGNESIUM
MAGNESIUM - CALMAGITE
POTASSIUM
ZINC

CONTROLS AND CALIBRATORS

Product
CALIBRATOR FOR AUTOANALYZERS
SERISCANN® ABNORMAL
SERISCANN® NORMAL

IMMUNOLOGY

AUTO-IMMUNE DISEASES

Product
ANTI-N-DNA LATEX

RHEUMA LINE SLIDE TEST

Product
AS DIRECT LATEX
CRP DIRECT LATEX
RF DIRECT LATEX

TURBIDIMETRIC RHEUMA LINE

Product
AS TURBIDIMETRIC LATEX (1:5), KIT
CRP TURBIDIMETRIC LATEX (1:5), KIT
RF TURBIDIMETRIC LATEX (1:5), KIT
RHEUMA LINE CONTROL POLYVALENT HIGH
RHEUMA LINE CONTROL POLYVALENT LOW

SLO HEMOLYTIC TEST

Product
STREPTOLYSIN O FREEZE DRIED
SLO PHOSPHATE BUFFER CONC.1:10

INFECTIOUS SEROLOGY

Product
FEBRILE ANTIGEN: Brucella abortus
FEBRILE ANTIGEN: Brucella melitensis
FEBRILE ANTIGEN: Proteus OX 19

Product
FEBRILE ANTIGEN: S. paratyphi A, H
FEBRILE ANTIGEN: S. paratyphi B, H
FEBRILE ANTIGEN: S. paratyphi C, H
FEBRILE ANTIGEN: S. typhi H
FEBRILE ANTIGEN: S. typhi O
FEBRILE ANTIGEN: S. paratyphi A, O
FEBRILE ANTIGEN: S. paratyphi B, O
FEBRILE ANTIGEN: S. paratyphi C, O
FEBRILE ANTIGENS SLIDE KIT
FEBRILE ANTIGENS: NEGATIVE CONTROL
FEBRILE ANTIGENS: Brucella, POSITIVE CONTROL
FEBRILE ANTIGENS: Proteus, POSITIVE CONTROL
FEBRILE ANTIGENS: Salmonella, POSITIVE CONTROL
BENGAL ROSE
INFECTIOUS MONONUCLEOSIS
RPR CARBON

PLASMA PROTEINS

Product
alpha-1 ACID GLYCOPROTEIN
alpha-1 ANTITRYPSIN
APOLIPOPROTEIN A1
APOLIPOPROTEIN B
APOLIPOPROTEIN CALIBRATOR
APOLIPOPROTEIN CONTROL
C3 COMPLEMENT
C4 COMPLEMENT
FERRITIN
FERRITIN, CALIBRATOR
FERRITIN, CONTROL
IMMUNOGLOBULIN A (Ig A)
IMMUNOGLOBULIN G (Ig G)
IMMUNOGLOBULIN M (Ig M)
MICROALBUMIN
TRANSFERRIN
PROTEINS CALIBRATOR
PROTEINS CONTROL

HAEMOSTASIS

Product
ANTITHROMBIN III (AT III)
D-DIMER
D-DIMER CALIBRATOR
D-DIMER CONTROL
FIBRINOGEN
HEMOSCANN®
PLASMASCANN®
PLASMASCANN® HS
PLASMASCANN® LIQUID
CALCIUM CHLORIDE (0.025 M)
EDTA K3
SODIUM CITRATE, 3.2 % (0,129M)
COAGULATION CONTROLS

Samarth Agency

'Moreshwar', Plot No.10,
Abhiyantanagar, Islampur
Pin -415 409 Tal.Walwa,
Dist. Sangli, Maharastra

ph-
+918329160577 +919699637355
+918412821206 +919699627366

email-
sharadmore7517@gmail.com

QUÍMICA CLÍNICA APLICADA SPAIN