BLOOD/HEMO-DIALYSIS SOLUTION MEASURING SYSTEM
ELECTROLYTE ANALYZER


EX-Ds



- **Measurement Parameter:** Na⁺, K⁺, Cl⁻
- **Measurement Speed:** 100 Tests/Hour
- **Measurement Sample:** Whole Blood, Plasma, Serum
- Newly developed high Sensitive electode allows only one time calibration in a ay to obtain very accurate result of C.V.≤0.5%

everlife
CPCDiagnostics
Enabling Better Outcomes

ONLY ONE TIME CALIBRATION IN A DAY, VERY STABLE DATA CAN BE OBTAINED.
IT IS DESIGNED FOR EASY OPERATION AND ECONOMICAL PERFORMANCE.



Aluminum Pack



WITH ALUMINUM REAGENT PACKS COMPLETELY SEALED UP, REAGENTS HAVE ALMOST NO CONCENTRATION OR NO DILUTION.
THEREFORE, YOU CAN GET DATA MORE ACCURATELY & RELIABLY

Specification of EX-Ds

Measurement method	Fully automated system with ion selective electrode		
Sample volume	95µ(blood mode), 300µ(hemo - dialysis mode)		
Measurement speed	35 seconds /sample with start mode		
Test parameters		Blood, Serum, Plasma	Urine (Diluted)
	Na^+	70.0-200.0 (mmol/l)	10.0-500.0 (mmol/L)
	K^+	1.00-20.00 (mmol/l)	1.00-100.0 (mmol/L)
	Cl^-	70.0-200.0 (mmol/l)	10.0-550.0 (mmol/L)
Function	Electrical Potential of each electrodes can be checked		
Calibration	Two - Point automatic calibration (interval is adjustable)		
Test Functions	Calculation of the X and C.V% Setting up normal range		
Data Memory	500 patient data 100 QC data		
On-Line	RS - 232C Built in		
Display	LCD Display with back light (20x4 Lines)		
Printer	Thermal Printer		
Barcode reader	Option		
Weight	Approx. 13 kg		
Dimensions	240 (W) x 410 (D) x 350 (H) mm		
Power	AC 110/220V+ 10% 50/60 Hz		

everlife
CPCDiagnostics
Enabling Better Outcomes

Samarth Agency
'Moreshwar', Plot No.10,
Abhiyantanagar, Islampur
Pin -415 409 Tal.Walwa,
Dist. Sangli, Maharastra

ph-
+918329160577 +919699637355
+918412821206 +919699627366

email-
sharadmore7517@gmail.com